



Designation: D7457 – 12

Standard Test Method for Determining Chloride in Aromatic Hydrocarbons and Related Chemicals by Microcoulometry¹

This standard is issued under the fixed designation D7457; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This test method covers the measurement of chloride in aromatic hydrocarbons, their derivatives, and related chemicals as defined by the method. Aromatics typically do not have any source of chlorides other than organic. The presence of metallic or inorganic chlorides is theoretically possible but not likely and this method does not purport to address recovery of those compounds.

1.2 This test method is applicable to samples with chloride concentrations from 0.24 to 5.0 mg/kg. The test method limit of detection (LOD) is 0.07 mg/kg.

1.3 This test method is preferred over Test Method D5194 for products, such as styrene, that are polymerized by the sodium biphenyl reagent.

1.4 In determining the conformance of the test results using this method to applicable specifications, results shall be rounded off in accordance with the rounding-off method of Practice E29.

1.5 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.6 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.* For specific hazard statements, see Section 9.

¹ This test method is under the jurisdiction of ASTM Committee D16 on Aromatic Hydrocarbons and Related Chemicals and is the direct responsibility of Subcommittee D16.04 on Instrumental Analysis.

Current edition approved Jan. 1, 2012. Published February 2012. Originally approved in 2011. Last previous edition approved in 2011 as D7457 - 11 ^{e1}. DOI: 10.1520/D7457-12.

2. Referenced Documents

2.1 *ASTM Standards:*²

D1193 Specification for Reagent Water

D1555M Test Method for Calculation of Volume and Weight of Industrial Aromatic Hydrocarbons and Cyclohexane [Metric]

D3437 Practice for Sampling and Handling Liquid Cyclic Products

D3505 Test Method for Density or Relative Density of Pure Liquid Chemicals

D4052 Test Method for Density, Relative Density, and API Gravity of Liquids by Digital Density Meter

D5194 Test Method for Trace Chloride in Liquid Aromatic Hydrocarbons

D6809 Guide for Quality Control and Quality Assurance Procedures for Aromatic Hydrocarbons and Related Materials

E29 Practice for Using Significant Digits in Test Data to Determine Conformance with Specifications

E691 Practice for Conducting an Interlaboratory Study to Determine the Precision of a Test Method

2.2 *Other Document:*

OSHA Regulations, 29 CFR paragraphs 1910.1000 and 1910.1200³

3. Terminology

3.1 *Definitions of Terms Specific to This Standard:*

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ Available from U.S. Government Printing Office Superintendent of Documents, 732 N. Capitol St., NW, Mail Stop: SDE, Washington, DC 20401, http://www.access.gpo.gov.

3.1.1 *calibration factor/recovery factor, n*—an indication of the efficiency of the measurement computed by dividing the measured value of a standard by its theoretical value.

3.1.2 *dehydrating tube or drying vessel, n*—a chamber containing concentrated sulfuric acid that scrubs the combustion effluent gases to remove water vapor prior to entering the titration cell.

3.1.3 *endpoint routine/test titration, n*—processes which allow the coulometer to set the endpoint and gain values to be used for sample analysis.

3.1.4 *oxidative pyrolysis, n*—a process in which a sample is evaporated in an inert gas atmosphere and combusted afterwards completely in an oxygen-rich atmosphere at high temperature to break down the components of the sample into elemental oxides.

3.1.5 *reference electrode, n*—used in conjunction with sensor electrode to measure the potential of the titration cell.

3.1.6 *sensor electrode, n*—detects changes in silver ion concentration.

3.1.7 *titration cell, n*—vessel that contains the sensor-reference electrode pair and generator electrode pair along with the cell electrolyte.

3.1.8 *titration parameters, n*—various instrumental conditions that can be changed for different types of analysis and analyzers.

3.1.9 *working electrode pair (generator electrode), n*—an electrode pair consisting of an anode and a cathode.

4. Summary of Test Method

4.1 An aliquot of sample is introduced into a combustion tube maintained at 900 to 1100°C having a flowing stream of oxygen combustion gas and argon inert gas. Oxidative pyrolysis converts the organic and inorganic halides to hydrogen halides that then flow into a titration cell where it reacts with silver ions present in the electrolyte. The silver ion thus consumed is coulometrically replaced and the total electrical work to replace it is a measure of the amount of organic halides in the specimen, which was introduced (see [Annex A1](#)).

5. Significance and Use

5.1 This test method is useful for determining organic as well as inorganic chloride compounds that can prove harmful to equipment and reactions in processes involving hydrocarbons. The combination of both the organic and inorganic chloride is commonly termed “total chloride” and since the inorganic chlorides are partially recovered, the result for total chloride will be biased low.

5.2 Maximum chloride levels are often specified for process streams and for hydrocarbon products.

5.3 Organic chloride species are potentially damaging to refinery processes. Hydrochloric acid can be produced in hydro treating or reforming reactors and this acid accumulates in condensing regions of the refinery.

6. Interferences

6.1 Both nitrogen and sulfur interfere at concentrations greater than approximately 0.1 %.

6.2 Bromides and iodides, if present, will be calculated as chlorides. However, fluorides are not detected by this test method.

6.3 Both organic and inorganic chloride in the sample will be measured due to conversion of both species to HCL during the oxidative pyrolysis process which is commonly termed as total chloride. The results for total chloride will be biased low since the measurement of total chloride is the sum of both organic and inorganic chloride compounds in the sample and inorganic compounds (if present) are only partially converted. The measurement of only the “organic chloride” in the sample can be measured by water washing the sample prior to analysis to remove the inorganic chloride in the sample. Partial loss of organic halogens which are partially soluble into the water wash, such as chloroacetic acid, may occur. The water washing procedure can not be applied to water soluble samples.

7. Apparatus

7.1 *Pyrolysis Furnace*, which can maintain at 900 to 1100°C and sufficient to pyrolyze the organic matrix and convert all chlorine present in the sample to hydrogen chloride. Furnace systems with the furnace orientated horizontally or vertically can be used.

7.2 *Pyrolysis Tube*, made of quartz and constructed in a way that the sample can be evaporated in an inert gas stream, and be pyrolyzed afterwards in the presence of oxygen. The inlet end of the tube must have a sample inlet port with a septum through which the sample can be injected by syringe. The inlet end must also have side arms for the introduction of oxygen and inert carrier gas. The pyrolysis tube must be of ample volume, so that complete pyrolysis of the sample is ensured.

7.3 *Titration Cell*, containing sensor electrode, reference electrode and working electrode pair. An inlet from the pyrolysis tube and magnetic stirring is also required. (**Warning**—Excessive stirring speed will decouple the stirring bar and cause it to bounce in the titration cell, possibly damaging the electrodes. A slight vortex in the cell is adequate.)

7.4 *Microcoulometer*, with connections for the electrodes, capable of measuring the potential of the sensor electrode, and comparing this potential with a bias potential, and amplifying the difference to the working electrode pair to generate a current. The microcoulometer output voltage signal should be proportional to the generating current.

7.5 *Controlling Unit*, for programming and setting the operating parameters of the instrument and integration of data.

7.6 *Flow Control*—The apparatus must be quipped with flow controllers capable of maintaining a constant supply and flow of oxygen and argon gas.

7.7 *Sample Inlet (Boat Inlet or Direct Injection)*.

7.7.1 *Boat Inlet Systems (Horizontal Furnace)*.

7.7.1.1 *Automated Boat Drive (Required)*, having variable stops, such that the sample boat may be driven into the furnace and stopped at various points as it enters the furnace thereby ensuring the controlled combustion of the sample and preventing the formation of soot or coke, both of which indicate

incomplete sample combustion. If direct injection is used with a horizontal system, the automated boat drive is not required.

7.7.1.2 *Boat Inlet Cooler (Required)*—Sample volatility and injection volume require an apparatus capable of cooling the sample boat prior to sample introduction. Thermoelectric coolers (peltier) or re-circulating refrigerated liquid devices have both been found effective. Other approaches can be used as long as the performance criteria of the method are not affected. If direct injection is used with a horizontal system, the boat inlet cooler is not required.

7.7.1.3 *Quartz or Ceramic Sample Boats*, of sufficient size to hold 90 to 250 μL or mg of sample.

7.7.2 *Direct Injection (Both Horizontal or Vertical Furnace)*.

7.7.2.1 *Direct Injection (Required)*—A sample introduction system capable of directly injecting sample into the furnace (either vertical or horizontal) at a controlled rate thereby ensuring the controlled combustion of the sample and preventing the formation of soot or coke, both of which indicate incomplete sample combustion.

7.8 *Autosampler (Required)*, capable of accurately delivering 90 to 240 μL of sample into the pyrolysis tube or sample boat if a boat inlet with automated boat drive and inlet system is used. An autosampler is required to ensure the accuracy and performance of the method is maintained.

NOTE 1—Multiple rinsing with clean solvent and/or sample between sample injections and/or between sampling from different sample vials is recommended to minimize carryover contamination. An independent solvent flush from a separate vial and different from the solvent wash thereby providing a clean, uncontaminated solvent may also be used. These features may be used as long as they do not degrade the performance and accuracy of the method.

7.9 *Dehydrating Tube or Drying Vessel*, positioned at the outlet of the pyrolysis tube so that effluent gases are bubbled through concentrated sulfuric acid. The water vapor formed in the combustion gases are trapped while all other gases including hydrogen halides are allowed to flow through into the titration cell.

7.10 *Gas-Tight Sampling Syringe*, having a capacity up to 250 μL and capable of accurately delivering 90 to 240 μL of sample. Syringes of differing sizes (100 or 250 μL) are permissible to be used as long as the precision and accuracy of the method are not degraded.

7.11 *Balance*, analytical, with sensitivity to 0.0001 g.

7.12 *Gas Regulators*—Two-stage gas regulators must be used for the reactant and carrier gas.

8. Reagents and Materials

8.1 *Purity of Reagents*—Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society,

where such specifications are available.⁴ Other grades may be used, provided that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

8.2 *Purity of Water*—Unless otherwise indicated, references to water shall be understood to mean reagent water conforming to Specification D1193, Type I. Type II can be used as long as the precision and accuracy of the method is not degraded.

8.3 *Solvent*—The solvent of choice must be capable of dissolving the chloride sample. The solvent of choice should have a boiling point similar to the sample being analyzed. Suggested possibilities include, but are not limited to isooctane, toluene, *p*-xylene and methanol.

8.4 *Purity of Solvent*—The blank value of the solvent used must be smaller than 0.1 $\mu\text{g}/\text{Kg}$ chlorides. Before using, the blank value of the solvent must be checked by carrying out a chlorine determination with the device being used for this analysis.

8.5 *Acetic Acid*—Glacial acetic acid (CH_3COOH), concentration: 97 %.

8.6 *Cell Electrolyte Solution*—Several electrolyte solutions based on acetic acid solutions. Please follow instrument manufacturer's recommendation.

NOTE 2—Bulk quantities of the electrolyte should be stored in a dark place and dark bottle and is recommended to be prepared fresh at least every two weeks.

8.7 *Inner Electrolyte Solution*—Please follow the instrument manufacturer's recommendation.

8.8 *Outer Electrolyte Solutions*—Please follow the instrument manufacturer's recommendation.

8.9 *Gases*.

8.9.1 *Argon*, 99.996 % minimum purity required as inert and carrier gas.

8.9.2 *Oxygen*, 99.995 % minimum purity is required as the oxidation gas.

8.10 *Sodium Acetate*, anhydrous, (NaCH_3CO_2), fine granular.

8.11 *Sodium Sulfate*, (Na_2SO_4), crystalline and anhydrous.

8.12 *Sulfuric Acid*, (95 to 98 %), (H_2SO_4) concentrated.

8.13 *2,4,5-Trichlorophenol or 2,4,6-Trichlorophenol (TCP)*, ($\text{C}_6\text{H}_3\text{OCl}_3$) fine granular.

8.14 *100 $\mu\text{g}/\text{mL}$ —Chloride Standard Stock Solution*—Weigh either 2,4,5-trichlorophenol or 2,4,6-trichlorophenol in the range of 0.0850 to 0.0950 g (target weight of 0.0929 g) to the

⁴ *Reagent Chemicals, American Chemical Society Specifications*, American Chemical Society, Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.