



SLOVENSKI STANDARD
oSIST prEN IEC 61010-2-020:2024
01-november-2024

Varnostne zahteve za električno opremo za meritve, nadzor in laboratorijsko uporabo - 2-020. del: Posebne zahteve za laboratorijske centrifuge

Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 2-020: Particular requirements for laboratory centrifuges

Sicherheitsbestimmungen für elektrische Mess-, Steuer-, Regel- und Laborgeräte - Teil 2-020: Besondere Anforderungen an Laborzentrifugen

Règles de sécurité pour appareils électriques de mesure, de régulation et de laboratoire - Partie 2-020: Exigences particulières pour centrifugeuses de laboratoire

Ta slovenski standard je istoveten z: prEN IEC 61010-2-020:2024

[oSIST prEN IEC 61010-2-020:2024](http://standards.researchgate.net/publication/352549744/figure/fig/1/figure-pdf/352549744/352549744.pdf)

<http://standards.researchgate.net/publication/352549744/figure/fig/1/figure-pdf/352549744/352549744.pdf>

ICS:

19.080	Električno in elektronsko preskušanje	Electrical and electronic testing
71.040.10	Kemijski laboratoriji. Laboratorijska oprema	Chemical laboratories. Laboratory equipment

oSIST prEN IEC 61010-2-020:2024

en,fr,de



66/820/CDV

COMMITTEE DRAFT FOR VOTE (CDV)

PROJECT NUMBER:

IEC 61010-2-020 ED4

DATE OF CIRCULATION:

2024-09-06

CLOSING DATE FOR VOTING:

2024-11-29

SUPERSEDES DOCUMENTS:

66/816/RR

IEC TC 66 : SAFETY OF MEASURING, CONTROL AND LABORATORY EQUIPMENT	
SECRETARIAT: United Kingdom	SECRETARY: Ms Stephanie Lavy
OF INTEREST TO THE FOLLOWING COMMITTEES:	HORIZONTAL FUNCTION(S): TC 66 Horizontal Group Safety
ASPECTS CONCERNED: Safety	
<input checked="" type="checkbox"/> SUBMITTED FOR CENELEC PARALLEL VOTING Attention IEC-CENELEC parallel voting The attention of IEC National Committees, members of CENELEC, is drawn to the fact that this Committee Draft for Vote (CDV) is submitted for parallel voting. The CENELEC members are invited to vote through the CENELEC online voting system.	<input type="checkbox"/> NOT SUBMITTED FOR CENELEC PARALLEL VOTING

<https://standards.iteh.ai/standards.iteh.ai/Document-Preview>

<https://standards.iteh.ai/standards.iteh.ai/Document-Preview>

This document is still under study and subject to change. It should not be used for reference purposes.

Recipients of this document are invited to submit, with their comments, notification of any relevant patent rights of which they are aware and to provide supporting documentation.

Recipients of this document are invited to submit, with their comments, notification of any relevant "In Some Countries" clauses to be included should this proposal proceed. Recipients are reminded that the CDV stage is the final stage for submitting ISC clauses. (SEE AC/22/2007 OR NEW GUIDANCE DOC).

TITLE:

Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 2-020: Particular requirements for laboratory centrifuges

PROPOSED STABILITY DATE: 2025

NOTE FROM TC/SC OFFICERS:

Copyright © 2024 International Electrotechnical Commission, IEC. All rights reserved. It is permitted to download this electronic file, to make a copy and to print out the content for the sole purpose of preparing National Committee positions. You may not copy or "mirror" the file or printed version of the document, or any part of it, for any other purpose without permission in writing from IEC.

CONTENTS

1		
2		
3	FOREWORD	3
4	1 Scope and object	5
5	2 Normative references	6
6	3 Terms and definitions	6
7	4 Tests	8
8	5 Marking and documentation	8
9	6 Protection against electric shock	11
10	7 Protection against mechanical HAZARDS	11
11	8 Resistance to mechanical stresses	18
12	9 Protection against the spread of fire	18
13	10 Equipment temperature limits and resistance to heat	18
14	11 Protection against HAZARDS from fluids and solid foreign objects	18
15	12 Protection against radiation, including laser sources, and against sonic and 16 ultrasonic pressure	19
17	13 Protection against liberated gases and substances, explosion and implosion	19
18	14 Components and subassemblies	20
19	15 Protection by interlocks	20
20	16 Hazards resulting from application	20
21	17 Risk assessment	20
22	Annexes	21
23	Annex L Index of defined terms	21
24	Annex AA (normative) Dynamic microbiological test method for BIOSEALS	22
25	Annex BB (informative) General guidance and rationale for particular subclauses	25
26	Annex CC (informative) General guidance for an empirical method to determine the 27 kinetic energy of a ROTOR	29
28	Bibliography	32
29		
30	Figure 101 – Rotor test setup	30
31		
32	Table 101 – Time-temperature conditions	18
33		
34		
35		

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR
MEASUREMENT, CONTROL, AND LABORATORY USE –****Part 2-020: Particular requirements for LABORATORY CENTRIFUGES**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 61010-2-020 has been prepared by IEC technical committee 66: Safety of measuring, control and laboratory equipment.

This fourth edition cancels and replaces the third edition published in 2016. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) alignment with changes introduced by Amendment 1 of IEC 61010-1:2010.

It has the status of a product safety publication in accordance with IEC Guide 104.

The text of this International Standard is based on the following documents:

CDV	Report on voting
66/xxx/CDV	66/xxx/RVC

88
89 Full information on the voting for the approval of this International Standard can be found in
90 the report on voting indicated in the above table.

91 This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

92 A list of all parts of the IEC 61010 series, published under the general title *Safety*
93 *requirements for electrical equipment for measurement, control, and laboratory use*, can be
94 found on the IEC website.

95 This Part 2-020 is intended to be used in conjunction with the latest edition of IEC 61010-1. It
96 was established on the basis of the third edition (2010) and its Amendment 1 (2016),
97 hereinafter referred to as Part 1.

98 This Part 2-020 supplements or modifies the corresponding clauses in IEC 61010-1 so as to
99 convert that publication into the IEC standard: *Particular requirements for LABORATORY*
100 *CENTRIFUGES*.

101 Where a particular subclause of Part 1 is not mentioned in this Part 2-020, that subclause
102 applies as far as is reasonable. Where this Part 2-020 states "addition", "modification" or
103 "replacement", the relevant requirement, test specification or note in Part 1 shall be adapted
104 accordingly.

105 In this standard:

106 1) the following print types are used:

107 – requirements: in roman type;

108 – NOTES: in small roman type; <https://standards.iteh.ai/>

109 – *conformity and tests*: in italic type; <https://standards.iteh.ai/>

110 – terms used throughout this standard which have been defined in Clause 3: SMALL
111 ROMAN CAPITALS.

112 2) subclauses, tables or figures which are additional to those in Part 1 are numbered starting
113 from 101. Additional annexes are lettered starting from AA.

114 The committee has decided that the contents of this document will remain unchanged until the
115 stability date indicated on the IEC website under "http://webstore.iec.ch" in the data related to
116 the specific document. At this date, the document will be

- 117 • reconfirmed,
- 118 • withdrawn,
- 119 • replaced by a revised edition, or
- 120 • amended.

121

122

SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE –

Part 2-020: Particular requirements for LABORATORY CENTRIFUGES

1 Scope and object

This clause of Part 1 is applicable except as follows:

1.1 Scope

1.1.1 Equipment included in scope

Replacement:

This part of IEC 61010 is applicable to electrically powered LABORATORY CENTRIFUGES.

It is possible that all or part of the equipment falls within the scope of one or more other Part 2 standards of IEC 61010 as well as within the scope of this standard. In that case, the requirements of those other Part 2 standards will also apply.

1.1.2 Equipment excluded from scope

Addition:

Add the following new item:

aa) IEC 60034 (Rotating electrical machinery).

1.2 Object

1.2.1 Aspects included in scope

Addition:

Add the following new items:

- aa) contact with moving parts (see 7.3);
- bb) LABORATORY CENTRIFUGE movement during any DISRUPTION (see 7.4.101);
- cc) high energy chemical reaction after ROTOR DISRUPTION (see 7.7.2.2 l));
- dd) ineffectiveness of BIOSEALS (see 13.101).

1.2.2 Aspects excluded from scope

Addition:

Add the following new items:

- aa) additional precautions which may need to be observed when centrifuging materials which are flammable or explosive (see 5.4.101);
- bb) additional precautions which may need to be observed when centrifuging materials that could react chemically with sufficient vigour to cause a HAZARD (see 5.4.101).

158 **1.4 Environmental conditions**

159 **1.4.1 Normal environmental conditions**

160 *Replacement:*

161 *Replace item c) by the following:*

162 c) temperature 2 °C to 40 °C;

163 **1.4.2 Extended environmental conditions**

164 *Replacement:*

165 *Replace item c) by the following:*

166 c) ambient temperatures below 2 °C or above 40 °C;

167 **2 Normative references**

168 This clause of Part 1 is applicable except as follows:

169 Addition:

170 ISO 3864 (all parts), *Graphical symbols – Safety colours and safety signs*

171 **3 Terms and definitions**

172 This clause of Part 1 is applicable except as follows:

173 **3.1 Equipment and states of equipment**

174 *Addition:*

175 *Add the following new terms and definitions:*

176 **3.1.101**

177 **LABORATORY CENTRIFUGE**

178 apparatus intended for laboratory use that applies a centrifuging effect to sample materials

179 **3.1.102**

180 **CENTRIFUGE-ROTOR COMBINATION**

181 LABORATORY CENTRIFUGE and ROTOR ASSEMBLY that are intended to operate together and which
182 have to be evaluated together

183 **3.1.103**

184 **DISRUPTION**

185 event in which the ROTOR ASSEMBLY, or part of it, fails or becomes detached during rotation

186 **3.2 Parts and accessories**

187 *Addition:*

188 *Add the following new terms and definitions:*

189 **3.2.101**

190 **CHAMBER**

191 enclosed space within a LABORATORY CENTRIFUGE in which the ROTOR ASSEMBLY rotates

192 **3.2.102**

193 **ROTOR**

194 primary component of a LABORATORY CENTRIFUGE which holds the material to be subjected to
195 centrifugal force and which is rotated by the DRIVE SYSTEM

196 **3.2.103**

197 **BUCKET**

198 sub-assembly of a ROTOR designed to support one or more containers

199 **3.2.104**

200 **PROTECTIVE CASING**

201 casing which completely surrounds the ROTOR ASSEMBLY and which includes the LID and its
202 securing devices

203 **3.2.105**

204 **LID**

205 access cover of the CHAMBER

206 **3.2.106**

207 **ROTOR ASSEMBLY**

208 ROTOR carrying a combination of ROTOR accessories specified by the manufacturer

209 *Note 1 to entry:* In the context of a ROTOR ASSEMBLY, ROTOR accessories include all components used with or in
210 the centrifuge ROTOR for the purpose of holding samples, including adaptors, buckets, tubes and bottles.

211 **3.2.107**

212 **DRIVE SYSTEM**

213 all components of the LABORATORY CENTRIFUGE associated with the provision of torque to, or
214 the rotational support of, the ROTOR ASSEMBLY

215 **3.2.108**

216 **BIOSEAL**

217 device or mechanism additional to, or integral with, a ROTOR or BUCKET and a closure
218 assembly, and which is designed to prevent the escape of contents, for example micro-
219 biological material, during centrifuging

220 **3.5 Safety terms**

221 *Addition:*

222 *Add the following new terms and definitions:*

223 **3.5.101**

224 **CLEARANCE ENVELOPE**

225 space around a LABORATORY CENTRIFUGE which is needed for safety

226 **3.5.102**227 **MCA**228 **MAXIMUM CREDIBLE ACCIDENT**

229 event chosen to represent worst-case conditions for a test that will evaluate the inherent
230 mechanical safety of a CENTRIFUGE-ROTOR COMBINATION (see 7.7 and Annex BB)

231 **4 Tests**

232 This clause of Part 1 is applicable.

233 **5 Marking and documentation**

234 This clause of Part 1 is applicable except as follows.

235 **5.1.2 Identification**

236 *Addition:*

237 *Add the following new list items c):*

238 aa) serial number or other means to identify the production batch of the equipment.

239

240 **5.1.3 MAINS supply**

241 *Addition:*

242 *Add the following note after the compliance statement.*

243 NOTE 101 The maximum power or input current considered is usually during the acceleration phase of the ROTOR,
244 with any options such as cooling or heating energized.

245 *Add the following new subclause:*

246 **5.1.101 ROTORS and accessories**

247 All OPERATOR-replaceable ROTORS and ROTOR ASSEMBLIES, including ROTOR ACCESSORIES, shall
248 be marked with the manufacturer's or supplier's name or registered trade mark, and
249 identification code (such as ID code, serial number or batch number).

250 If components are too small, or are not suitable for such marking, the required information
251 shall be marked on the original packaging, as well as being stated in the documentation.

252 NOTE Packaging can be the outer box, an insert, etc.

253 If the manufacturer specifies that an individual part, for example a BUCKET, is to be fitted only
254 to a specific ROTOR or in specific ROTOR positions for balance or some other reason, each
255 BUCKET and ROTOR position shall be identified by marking with corresponding numbers or
256 letters.

257 *Conformity is checked by inspection.*

258 **5.4.2 Equipment RATINGS**

259 *Addition:*

260 *Add the following new list items:*

- 261 aa) a list of all ROTORS and ROTOR accessories specified for use with a LABORATORY
262 CENTRIFUGE, together with their RATED rotational frequencies;
- 263 bb) any restrictions by the manufacturer warning against the use of particular materials to
264 be centrifuged;
- 265 cc) density and volume limits for ROTOR ASSEMBLY loading and, if applicable, derating
266 instructions.

267 **5.4.3 Equipment installation**

268 *Addition:*

269 *Add, after list item a), the following sublist items:*

- 270 i) floor or bench area required for the CLEARANCE ENVELOPE for the intended use (see
271 7.4.101);
- 272 ii) total weight of the LABORATORY CENTRIFUGE;
- 273 iii) instructions for site preparation;
- 274 iv) methods for levelling of the LABORATORY CENTRIFUGE;
- 275 v) means for securing to the mounting surface;

276 **5.4.4 Equipment operation**

277 *Addition:*

278 *Add the following new items:*

- 279 aa) loading and balancing procedures;
- 280 bb) ROTOR changing procedure;
- 281 cc) any specific requirement for an OPERATOR to be present at stated phases of the
282 centrifuging procedure;
- 283 dd) necessary safeguards for personnel. Instructions shall include at least the following:
- 284 – not to lean on a LABORATORY CENTRIFUGE;
- 285 – not to stay within the CLEARANCE ENVELOPE longer than necessary for operational
286 reasons;
- 287 – not to deposit any potentially hazardous materials within the CLEARANCE ENVELOPE;
- 288 – methods for safe operation during open LID procedures (see 7.3.102.2);
- 289 ee) instructions for use of BIOSEALS and other biocontainment components, including the
290 proper closure techniques. These instructions shall indicate that BIOSEALS and related
291 components are intended to be part of biocontainment systems, as specified in
292 international and national biosafety guidelines. They are not to be relied on as the only
293 means of safeguarding workers and the environment when handling pathogenic micro-
294 organisms.

295 **5.4.5 Equipment maintenance and service**

296 *Addition:*

297 *Add the following new text at the end of the subclause (before the compliance statement):*

298 Where applicable, the instructions shall specify: