

INTERNATIONAL STANDARD

NORME INTERNATIONALE



**Ultrasonics – Pulse-echo scanners –
Part 1: Techniques for calibrating spatial measurement systems and
measurement of system point-spread function response**

**Ultrasons – Scanners à impulsion et écho –
Partie 1: Techniques pour l'étalonnage des systèmes de mesure spatiaux et
des mesures de la réponse de la fonction de dispersion ponctuelle du système**

<https://standards.iteh.ai/catalog/standards/iec/2bb4f020-0145-4ffc-8892-42fc1225a313/iec-61391-1-2006>





THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2017 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

IEC Catalogue - webstore.iec.ch/catalogue

The stand-alone application for consulting the entire bibliographical information on IEC International Standards, Technical Specifications, Technical Reports and other documents. Available for PC, Mac OS, Android Tablets and iPad.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and also once a month by email.

Electropedia - www.electropedia.org

The world's leading online dictionary of electronic and electrical terms containing 21 000 terms and definitions in English and French, with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

IEC Glossary - std.iec.ch/glossary

67 000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Catalogue IEC - webstore.iec.ch/catalogue

Application autonome pour consulter tous les renseignements bibliographiques sur les Normes internationales, Spécifications techniques, Rapports techniques et autres documents de l'IEC. Disponible pour PC, Mac OS, tablettes Android et iPad.

Recherche de publications IEC - webstore.iec.ch/advsearchform

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études,...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

IEC Just Published - webstore.iec.ch/justpublished

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et aussi une fois par mois par email.

Electropedia - www.electropedia.org

Le premier dictionnaire en ligne de termes électroniques et électriques. Il contient 21 000 termes et définitions en anglais et en français, ainsi que les termes équivalents dans 16 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

Glossaire IEC - std.iec.ch/glossary

67 000 entrées terminologiques électrotechniques, en anglais et en français, extraites des articles Termes et Définitions des publications IEC parues depuis 2002. Plus certaines entrées antérieures extraites des publications des CE 37, 77, 86 et CISPR de l'IEC.

Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: sales@iec.ch.



IEC 61391-1

Edition 1.1 2017-07
CONSOLIDATED VERSION

INTERNATIONAL STANDARD

NORME INTERNATIONALE



**Ultrasonics – Pulse-echo scanners –
Part 1: Techniques for calibrating spatial measurement systems and
measurement of system point-spread function response**

**Ultrasons – Scanners à impulsion et écho –
Partie 1: Techniques pour l'étalonnage des systèmes de mesure spatiaux et
des mesures de la réponse de la fonction de dispersion ponctuelle du système**

<https://standards.iteh.ai/catalog/standards/iec/2bb4f020-0145-4ffc-8892-42fc1225a313/iec-61391-1-2006>

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 17.140.50

ISBN 978-2-8322-4589-7

**Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

REDLINE VERSION

VERSION REDLINE



**Ultrasonics – Pulse-echo scanners –
Part 1: Techniques for calibrating spatial measurement systems and
measurement of system point-spread function response**

**Ultrasons – Scanners à impulsion et écho –
Partie 1: Techniques pour l'étalonnage des systèmes de mesure spatiaux et
des mesures de la réponse de la fonction de dispersion ponctuelle du système**

<https://standards.iteh.ai/catalog/standards/iec/2bb4f020-0145-4ffc-8892-42fc1225a313/iec-61391-1-2006>

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	6
1 Scope.....	7
2 Normative references.....	7
3 Terms and definitions	7
4 Symbols and abbreviated terms	15
5 General conditions.....	16
6 Techniques for calibrating 2D-measurement systems.....	17
6.1 Test methods	17
6.2 Instruments	17
6.3 Test settings.....	18
6.4 Test parameters	19
7 Methods for calibrating 3D-measurement systems.....	21
7.1 General	21
7.2 Types of 3D-reconstruction methods.....	21
7.3 Test parameters associated with reconstruction problems.....	22
7.4 Test methods for measurement of 3D-reconstruction accuracy.....	23
8 Measurement of point-spread and line-spread functions (high-contrast spot size)	27
8.1 General	27
8.2 Test methods	28
8.3 Instruments	28
8.4 Test settings.....	29
8.5 Test parameters	31
Annex A (normative) Test objects – Calibration of 2D-spatial measurement systems.....	36
Annex B (normative) Test objects – Measurement and calibration of 3D-image reconstruction accuracy	39
Annex C (normative) Test objects – Measurement of point-spread function response.....	43
Annex D (informative) Quality parameters derived by PSF-mapping analysis.....	48
Bibliography	64
Figure A.1 – Concentric circular arrays of nylon filaments	37
Figure A.2 – Regular 2D-array of nylon filaments	38
Figure B.1 – Tissue mimicking ovoid target phantom	40
Figure B.2 – Composite of two cross-sectional views of test object shown in Figure B.1	40
Figure B.3 – Projection view from top of test object shown in Figure B.1	41
Figure B.4 – Projection view from end window of test object shown in Figure B.1	42
Figure C.1 – Filament test object for measuring the LSF.....	44
Figure C.2 – Axial resolution test object.....	45
Figure C.3 – Movable single filament in water.....	46
Figure C.5 – Slice thickness measurement and calculation	47
Figure D.1 – Principal schematic of the PSF-analyser function.....	51
Figure D.2 – Principle of elimination of internal multi-reflections in the spherical target using filtration in time domain	54

Figure D.3 – A pixel maximum level and PSF-trace estimation in ROI stored digital data.....	56
Figure D.4 – The derivation of FWHM-value from the lateral-amplitude profile of PSF received-echo signal ($a_{r,max}(x_i,z_k) = 221$) with reflector positioned on axis of the ultrasound beam in point (x_i,z_k) of the measuring grid	57
Figure D.5 – The FWHM-derivation from the elevational (transversal) profile of MER in one point of the measuring grid	58
Figure D.6 – The derivation of the reflected-signal axial-profile from the ROI	59
Figure D.7 – The enumeration of the $W_{H,HM}$ parameter from the axial-profile line: $a_{r,max}(x,z) = 243$	59
Figure D.8 – The distribution of FWHM over a scan area of width 20 mm to depth 80 mm, made with a monofocal scan using a linear 5 MHz transducer	60
Figure D.9 – The distribution of FWHM over a scan area of width 20 mm to depth 80 mm, made with the same system as for Figure D.7 but using three focal points (F1, F2 and F3) for the scan	60
Figure D.10 – Plot of the distribution of MER-intensity over the scanning area 30 mm wide and 40 mm deep	61
Figure D.11 – The elevational profile recorded from $a_{r,max}(x_i,y,z_k)$ values for a spherical target passing perpendicularly to the scanning plane in one point (x_i,z_k) of the measuring grid	62
Table 1 – Expected values for the two ellipsoidal objects in Figure B.3	27
Table 2 – Suggested table of reported values	27

itehStandards
(<https://standards.iteh.ai>)
Document Preview

[IEC 61391-1:2006](#)

<https://standards.iteh.ai/catalog/standards/iec/2bb4f020-0145-4ffc-8892-42fc1225a313/iec-61391-1-2006>

INTERNATIONAL ELECTROTECHNICAL COMMISSION

ULTRASONICS – PULSE-ECHO SCANNERS –**Part 1: Techniques for calibrating spatial measurement systems
and measurement of system point-spread function response**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

This consolidated version of the official IEC Standard and its amendment has been prepared for user convenience.

IEC 61391-1 edition 1.1 contains the first edition (2006-07) [documents 87/336/FDIS and 87/343/RVD] and its amendment 1 (2017-07) [documents 87/650/FDIS and 87/653/RVD].

In this Redline version, a vertical line in the margin shows where the technical content is modified by amendment 1. Additions are in green text, deletions are in strikethrough red text. A separate Final version with all changes accepted is available in this publication.

International Standard IEC 61391-1 has been prepared by IEC technical committee 87: Ultrasonics.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

Terms in **bold** in the text are defined in clause 3.

This standard is intended to be published in two or more parts:

- Part 1 deals with techniques for calibrating spatial measurement systems and measurement of system point-spread function response;
- Part 2 will deal with measurement of system sensitivity, dynamic range, and low-contrast resolution.

The committee has decided that the contents of the base publication and its amendment will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

iTeh Standards

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

[IEC 61391-1:2006](https://standards.iteh.ai/catalog/standards/iec/2bb4f020-0145-4ffc-8892-42fc1225a313/iec-61391-1-2006)

<https://standards.iteh.ai/catalog/standards/iec/2bb4f020-0145-4ffc-8892-42fc1225a313/iec-61391-1-2006>

INTRODUCTION

An ultrasonic pulse-echo scanner produces images of tissue in an ultrasonic **scan plane** by sweeping a narrow pulsed beam of **ultrasound** through the section of interest and detecting the echoes generated at tissue boundaries. A variety of **ultrasonic transducer** types are employed to operate in a transmit/receive mode for the ultrasonic signals. Ultrasonic scanners are widely used in medical practice to produce images of many soft-tissue organs throughout the human body.

This standard describes test procedures that should be widely acceptable and valid for a wide range of types of equipment. Manufacturers should use the standard to prepare their specifications; the users should employ the standard to check specifications. The measurements can be carried out without interfering with the normal working conditions of the machine. Typical **test objects** are described in the annexes. The structures of the **test objects** have not been specified in detail, rather suitable types of overall and internal structures are described. The specific structure of a **test object** should be reported with the results obtained using it. Similar commercial versions of these **test objects** are available.

The performance parameters specified and the corresponding methods of measurement have been chosen to provide a basis for comparison with the manufacturer's specification and between similar types of apparatus of different makes, intended for the same kind of diagnostic application. The manufacturer's specification should allow comparison with the results obtained from the tests in this standard. Furthermore, it is intended that the sets of results and values obtained from the use of the recommended methods will provide useful criteria for predicting the performance of equipment in appropriate diagnostic applications. This standard concentrates on measurements of images by digital techniques. Methods suitable for inspection by eye are covered here as well. Discussion of other visual techniques can be found in IEC 61390 [1] ¹⁾.

Where a diagnostic system accommodates more than one option in respect of a particular system component, for example the **ultrasonic transducer**, it is intended that each option be regarded as a separate system. However, it is considered that the performance of a machine is adequately specified, if measurements are undertaken for the most significant combinations of machine control settings and accessories. Further evaluation of equipment is obviously possible but this should be considered as a special case rather than a routine requirement.

1) Figures in square brackets refer to the Bibliography.

ULTRASONICS – PULSE-ECHO SCANNERS –

Part 1: Techniques for calibrating spatial measurement systems and measurement of system point-spread function response

1 Scope

This International Standard describes methods of calibrating the spatial measurement facilities and **point-spread function** of ultrasonic imaging equipment in the ultrasonic frequency range 0,5 MHz to 15 MHz. This standard is relevant for ultrasonic scanners based on the pulse-echo principle of the types listed below:

- mechanical sector scanners;
- electronic phased-array sector scanners;
- electronic linear-array scanners;
- electronic curved-array sector scanners;
- water-bath scanners based on any of the above four scanning mechanisms;
- 3D-volume reconstruction systems.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60050-801:1994, *International Electrotechnical Vocabulary – Chapter 801: Acoustics and electroacoustics*

IEC 60050-802:2011, *International Electrotechnical Vocabulary – Part 802: Ultrasonics*

~~IEC 61102:1991, *Measurement and characterisation of ultrasonic fields using hydrophones in the frequency range 0,5 MHz to 15 MHz*~~

IEC 61685:2001, *Ultrasonics – Flow measurement systems – Flow test object*

IEC 62127-1:2007, *Ultrasonics – Hydrophones – Part 1: Measurement and characterization of medical ultrasonic fields up to 40 MHz*

3 Terms and definitions

For the purposes of this document, the ~~following~~ terms and definitions given in IEC 60050-801:1994, IEC 60050-802:2011, IEC 62127-1:2007 and the following apply. See also related International Standards, Technical Specifications and Technical Reports for definitions and explanations [1-5] [1] [2] [3] [4] [34] [35] [36] [37] [38] [39].

3.1

A-scan

class of data acquisition geometry in one dimension, in which echo strength information is acquired from points lying along a single **beam axis** and displayed as amplitude versus time of flight or distance

3.2

acoustic coupling agent (also, coupling agent)

a material, usually a gel or other fluid, that is used to ensure acoustic contact between the transducer and the patient's skin, or between the transducer and the surface of a sealed test object

3.3

acoustic working frequency

arithmetic mean of the frequencies f_1 and f_2 at which the amplitude of the acoustic pressure spectrum is 3 dB below the peak amplitude

(See 3.4.2 of IEC 61102)

3.4

automatic time-gain compensation

ATGC

automatic working time gain control based on the observed decrease in echo amplitudes due to the attenuation in ultrasonic pulse amplitude with depth

3.5

axial resolution

minimum separation along the beam axis of two equally scattering volumes or **targets** at a specified depth for which two distinct echo signals can be displayed

3.6

backscatter coefficient

mean acoustic power scattered in the 180° direction by a specified object with respect to the direction of the incident beam, per unit solid angle per unit volume, divided by the incident beam intensity. For a volume filled with many scatterers, the scatterers are considered to be randomly distributed. The mean power is obtained from different spatial realisations of the scattering volume

NOTE **Backscatter coefficient** is commonly referred to as the differential scattering cross-section per unit volume in the 180° direction

3.7

backscatter contrast (normalized)

difference between the **backscatter coefficients** from two defined regions divided by the square root of the product of the two **backscatter coefficients**

3.8

beam axis

the longitudinal axis of the pulse-echo response pattern of a given **B-mode scan line**, a pulse-echo equivalent to the transmitted beam axis of IEC 61828 [2]

3.9

B-scan

class of data acquisition geometry in which echo information is acquired from points lying in an ultrasonic **scan plane** containing interrogating ultrasonic beams. See **B-mode** below.

NOTE B-scan is a colloquial term for **B-mode** scan or image. (See 3.10)

3.10

Brightness-modulated display

B-mode

method of presentation of **B-scan** information in which a particular section through an imaged object is represented in a conformal way by the scan plane of the display and echo amplitude is represented by local brightness or optical density of the display

[IEC 60854: definition 3.18, modified]

3.11**displayed dynamic range**

ratio, expressed in decibels, of the amplitude of the maximum echo that does not saturate the display to the minimum echo that can be distinguished in the display under the scanner test settings

3.12**elevational resolution**

minimum separation perpendicular to the ultrasonic **scan plane** of two equally scattering **targets** at a specified depth for which two distinct echo signals can be displayed. Often used here informally for slice thickness for purposes of 3D-scanning

3.13**field-of-view**

area in the ultrasonic **scan plane** which is insonated by the **ultrasound beam** during the acquisition of echo data to produce one image frame

3.14**frame rate**

number of sweeps comprising the full-frame refresh rate that the ultrasonic beam makes per second through the **field-of-view**

3.15**gain**

ratio of the output to the input of a system, generally an amplifying system, usually expressed in decibels

3.16**grey scale**

range of values of image brightness, being either continuous between two extreme values or, if discontinuous, including at least three discrete values

[IEC 60854: definition 3.14]

3.17**lateral resolution**

minimum separation of two **line targets** at a specified depth in a **test object** made of **tissue-mimicking material** for which two distinct echo signals can be displayed. The **line targets** should be perpendicular to the scanned plane; the separation between the **targets** should be perpendicular to the beam-alignment axis

3.18**line-spread function****LSF**

characteristic response in three dimensions of an imaging system to a high-contrast line target

3.19**line target**

cylindrical reflector whose diameter is so small that the reflector cannot be distinguished by the imaging system from a cylindrical reflector with diameter an order of magnitude smaller, except by signal amplitude. The backscatter from a standard **line target** should be a simple function of frequency over the range of frequencies studied

3.20**M-mode
time-motion mode**

method of presentation of **M-scan** information in which the motion of structures along a fixed **beam axis** is depicted by presenting their positions on a line which moves across a display to show the variation with time of the echo

3.21**M-scan
time-motion scan**

class of acquisition geometry in which echo information from moving structures is acquired from points lying along a single beam axis. The echo strength information is presented using an **M-mode** display

3.22**nominal frequency (of a transducer)**

intended **acoustic working frequency** of a transducer as quoted by the designer or manufacturer

[adapted from definition 3.7 of IEC 60854]

3.23**pixel**

picture element

smallest spatial unit or cell size of a digitized 2-dimensional array representation of an image. Each **pixel** has an address (x-and y-coordinates corresponding to its position in the array) and a specific brightness level

NOTE **Pixel** is a contraction of 'picture element'.

3.24**point target**

reflector whose scattering surface dimensions are so small that it cannot be distinguished (except by signal amplitude) by the imaging system from a similar **target** whose scattering surface is an order of magnitude smaller. The backscatter cross section of a standard point **target** should be a simple function of frequency over the range of frequencies studied.

3.25**point-spread function****PSF**

characteristic response in three dimensions of an imaging system to a high-contrast **point target**.

NOTE For most ultrasound systems, an individual ultrasound **PSF** cannot be used as the overall system impulse response, due to changes in the **PSF** with depth, with other positions in the region of use and with system focal and frequency settings. The problem is solved by PSF mapping – see Annex D.

3.26**scan line**

one of the component lines which form a **B-mode** image on an ultrasound monitor. Each line is the envelope-detected **A-scan** line in which the echo amplitudes are converted to brightness values

3.27**scan plane**

a plane containing the **ultrasonic scan lines**

[IEC 61102: definition 3.38, modified]