

INTERNATIONAL      **ISO/ASTM**  
STANDARD                      **51631**

Fourth edition  
2020-02

---

---

**Practice for use of calorimetric  
dosimetry systems for dose  
measurements and dosimetry system  
calibration in electron beams**

*Pratique de l'utilisation des systèmes dosimétriques calorimétriques  
pour pour des mesures de dose délivrée par un faisceau d'électrons et  
pour l'étalonnage de dosimètres*

*iteh standards*  
**(<https://standards.iteh.ai>)**  
**Document Preview**

[ISO/ASTM 51631:2020](https://standards.iteh.ai/catalog/standards/iso/39dac662-4823-4e35-afb2-0cb73dd4feff/iso-astm-51631-2020)

<https://standards.iteh.ai/catalog/standards/iso/39dac662-4823-4e35-afb2-0cb73dd4feff/iso-astm-51631-2020>



Reference number  
ISO/ASTM 51631:2020(E)

© ISO/ASTM International 2020

iTeh Standards  
(<https://standards.iteh.ai>)  
Document Preview

[ISO/ASTM 51631:2020](https://standards.iteh.ai/catalog/standards/iso/39dac662-4823-4e35-afb2-0cb73dd4feff/iso-astm-51631-2020)

<https://standards.iteh.ai/catalog/standards/iso/39dac662-4823-4e35-afb2-0cb73dd4feff/iso-astm-51631-2020>



**COPYRIGHT PROTECTED DOCUMENT**

© ISO/ASTM International 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester. In the United States, such requests should be sent to ASTM International.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Fax: +41 22 749 09 47  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

ASTM International  
100 Barr Harbor Drive, PO Box C700  
West Conshohocken, PA 19428-2959, USA  
Phone: +610 832 9634  
Fax: +610 832 9635  
Email: [khooper@astm.org](mailto:khooper@astm.org)  
Website: [www.astm.org](http://www.astm.org)

Published in Switzerland

## Contents

	Page
<b>1</b> Scope .....	<b>1</b>
<b>2</b> Referenced Documents .....	<b>1</b>
<b>3</b> Terminology .....	<b>2</b>
<b>4</b> Significance and use .....	<b>2</b>
<b>5</b> Interferences .....	<b>3</b>
<b>6</b> Apparatus .....	<b>3</b>
<b>7</b> Calibration procedures .....	<b>4</b>
<b>8</b> Dose measurement procedures .....	<b>5</b>
<b>9</b> Calibration of other dosimetry systems .....	<b>6</b>
<b>10</b> Documentation .....	<b>7</b>
<b>11</b> Measurement uncertainty .....	<b>7</b>
<b>12</b> Keywords .....	<b>7</b>
<b>Annexes</b> .....	<b>7</b>
<b>Figure 1</b> Example of a polystyrene calorimeter used for routine measurements at a 10-MeV industrial electron accelerator .....	<b>3</b>
<b>Figure 2</b> Absorber (phantom) for irradiation at 10 MeV electron irradiation facility of routine and transfer-standard dosimeters <b>(10)</b> . Material: Polystyrene .....	<b>4</b>
<b>Figure 3</b> Example of measurements of temperature of a graphite calorimeter before and after irradiation <b>(7)</b> .....	<b>6</b>
<b>Figure 4</b> Example of on-line measurements of a graphite calorimeter <b>(5)</b> .....	<b>6</b>
<b>Table 1</b> Measurement uncertainties of routine polystyrene calorimetric dosimetry systems from Risø high dose reference laboratory (in percent, at $k = 2$ ) <b>(9)</b> .....	<b>7</b>
<b>Table A1.1</b> Results for alanine and calorimeter dose measurements .....	<b>8</b>
<b>Table A2.1</b> Thickness and size of several graphite calorimetric bodies designed at NIST for use at specific electron energies .....	<b>8</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted (see [www.iso.org/directives](http://www.iso.org/directives)).

ASTM International is one of the world's largest voluntary standards development organizations with global participation from affected stakeholders. ASTM technical committees follow rigorous due process balloting procedures.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO and ASTM International shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement. (<https://standards.iteh.ai>)

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

<https://standards.iteh.ai/catalog/standards/iso/39dac662-4823-4e35-afb2-0cb73dd4feff/iso-astm-51631-2020>  
This document was prepared by ASTM Committee E61, *Radiation processing* (as ASTM E1631-94), and drafted in accordance with its editorial rules. It was assigned to Technical Committee ISO/TC 85, *Nuclear energy, nuclear technologies and radiation protection*, and adopted under the “fast-track procedure”.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

## ISO/ASTM 51631:2020(E)



# Standard Practice for Use of Calorimetric Dosimetry Systems for Dose Measurements and Routine Dosimetry System Calibration in Electron Beams<sup>1</sup>

This standard is issued under the fixed designation ISO/ASTM 51631; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision.

## 1. Scope

1.1 This practice covers the preparation and use of semi-adiabatic calorimetric dosimetry systems for measurement of absorbed dose and for calibration of routine dosimetry systems when irradiated with electrons for radiation processing applications. The calorimeters are either transported by a conveyor past a scanned electron beam or are stationary in a broadened beam.

1.2 This document is one of a set of standards that provides recommendations for properly implementing dosimetry in radiation processing, and describes a means of achieving compliance with the requirements of ISO/ASTM Practice 52628 for a calorimetric dosimetry system. It is intended to be read in conjunction with ISO/ASTM Practice 52628.

1.3 The calorimeters described in this practice are classified as Type II dosimeters on the basis of the complex effect of influence quantities. See ISO/ASTM Practice 52628.

1.4 This practice applies to electron beams in the energy range from 1.5 to 12 MeV.

1.5 The absorbed dose range depends on the calorimetric absorbing material and the irradiation and measurement conditions. Minimum dose is approximately 100 Gy and maximum dose is approximately 50 kGy.

1.6 The average absorbed-dose rate range shall generally be greater than 10 Gy·s<sup>-1</sup>.

1.7 The temperature range for use of these calorimetric dosimetry systems depends on the thermal resistance of the calorimetric materials, on the calibration range of the temperature sensor, and on the sensitivity of the measurement device.

1.8 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate*

*safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.*

1.9 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

## 2. Referenced Documents

### 2.1 ASTM Standards:<sup>2</sup>

E666 Practice for Calculating Absorbed Dose From Gamma or X Radiation

E668 Practice for Application of Thermoluminescence-Dosimetry (TLD) Systems for Determining Absorbed Dose in Radiation-Hardness Testing of Electronic Devices

E3083 Terminology Relating to Radiation Processing: Dosimetry and Applications

### 2.2 ISO/ASTM Standards:<sup>2</sup>

ISO/ASTM 51631-1 Practice for Calibration of Routine Dosimetry Systems for Radiation Processing

ISO/ASTM 51649 Practice for Dosimetry in an Electron Beam Facility for Radiation Processing at Energies Between 300 keV and 25 MeV

ISO/ASTM 51707 Guide for Estimating Uncertainties in Dosimetry for Radiation Processing

ISO/ASTM 52628 Practice for Dosimetry in Radiation Processing

### 2.3 International Commission on Radiation Units and Measurements (ICRU) Reports:<sup>3</sup>

ICRU Report 34 The Dosimetry of Pulsed Radiation

ICRU Report 35 Radiation Dosimetry: Electron Beams with Energies Between 1 and 50 MeV

ICRU Report 80 Dosimetry Systems for use in Radiation Processing

ICRU Report 85a Fundamental Quantities and Units for Ionizing Radiation

<sup>1</sup> This practice is under the jurisdiction of ASTM Committee E61 on Radiation Processing and is the direct responsibility of Subcommittee E61.02 on Dosimetry Systems, and is also under the jurisdiction of ISO/TC 85/WG 3.

Current edition approved by ASTM May 15, 2019. Published February 2020. Originally published as E1631-94. The present Fourth Edition of International Standard ISO/ASTM 51631:2020(E) is a minor revision of the Third Edition of ISO/ASTM 51631-2013(E).

<sup>2</sup> For referenced ASTM and ISO/ASTM standards, visit the ASTM website, [www.astm.org](http://www.astm.org), or contact ASTM Customer Service at [service@astm.org](mailto:service@astm.org). For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

<sup>3</sup> Available from the Commission on Radiation Units and Measurements, 7910 Woodmont Ave., Suite 800, Bethesda, MD 20814, U.S.A.



2.4 *Joint Committee for Guides in Metrology (JCGM) Reports:*

JCGM 100:2008, GUM 1995, with minor corrections, Evaluation of measurement data – Guide to the Expression of Uncertainty in Measurement<sup>4</sup>

JCGM 200:2012, VIM International vocabulary of metrology – Basic general concepts and general terms<sup>5</sup>

### 3. Terminology

#### 3.1 Definitions:

3.1.1 *primary-standard dosimetry system*—dosimetry system that is designated or widely acknowledged as having the highest metrological qualities and whose value is accepted without reference to other standards of the same quantity.

3.1.2 *reference standard dosimetry system*—dosimetry system, generally having the highest metrological quality available at a given location or in a given organization, from which measurements made there are derived.

3.1.3 *transfer standard dosimetry system*—dosimetry system used as an intermediary to calibrate other dosimetry systems.

3.1.4 *type II dosimeter*—dosimeter, the response of which is affected by influence quantities in a complex way that cannot practically be expressed in terms of independent correction factors.

#### 3.2 Definitions of Terms Specific to This Standard:

3.2.1 *adiabatic*—no heat exchange with the surroundings.

3.2.2 *calorimeter*—assembly consisting of calorimetric body (absorber), thermal insulation, and temperature sensor with wiring that, when irradiated, exhibits increase in the absorber temperature that can be related to absorbed dose. This language parallels that of dosimeter.

3.2.3 *calorimetric body*—mass of material absorbing radiation energy and whose temperature is measured.

3.2.4 *calorimetric dosimetry system*—dosimetry system consisting of calorimeter, measurement instruments and their associated reference standards, and procedures for the system's use.

3.2.5 *endothermic reaction*—chemical reaction that consumes energy.

3.2.6 *exothermic reaction*—chemical reaction that releases energy.

3.2.7 *heat defect (thermal defect)*—amount of energy released or consumed by chemical reactions caused by the absorption of radiation energy.

3.2.8 *specific heat capacity*—amount of energy required to raise 1 kg of material by the temperature of 1 K.

<sup>4</sup> Document produced by Working Group 1 of the Joint Committee for Guides in Metrology (JCGM/WG 1). Available free of charge at the BIPM website (<http://www.bipm.org>).

<sup>5</sup> Document produced by Working Group 2 of the Joint Committee for Guides in Metrology (JCGM/WG 2). Available free of charge at the BIPM website (<http://www.bipm.org>).

3.2.9 *thermistor*—electrical resistor with a well-defined relationship between resistance and temperature of the thermistor.

3.2.10 *thermocouple*—junction of two metals producing an electrical voltage with a well-defined relationship to junction temperature.

3.3 Definitions of other terms used in this standard that pertain to radiation measurement and dosimetry may be found in ASTM Terminology E3083. Definitions in E3083 are compatible with ICRU Report 85a; that document, therefore, may be used as an alternative reference.

### 4. Significance and use

4.1 This practice is applicable to the use of calorimetric dosimetry systems for the measurement of absorbed dose in electron beams, the qualification of electron irradiation facilities, periodic checks of operating parameters of electron irradiation facilities, and calibration of other dosimetry systems in electron beams. Calorimetric dosimetry systems are most suitable for dose measurement at electron irradiation facilities utilizing conveyor systems for transport of product during irradiation.

NOTE 1—For additional information on calorimetric dosimetry system operation and use, see ICRU Report 80. For additional information on the use of dosimetry in electron accelerator facilities, see ISO/ASTM 51649, and ICRU Reports 34 and 35, and Refs (1-3).<sup>6</sup>

4.2 The calorimetric dosimetry systems described in this practice are not primary standard dosimetry systems. The calorimeters are classified as Type II dosimeters (ISO/ASTM 52628). They might be used as internal standards at an electron beam irradiation facility, including being used as transfer standard dosimetry systems for calibration of other dosimetry systems, or they might be used as routine dosimeters. The calorimetric dosimetry systems are calibrated by comparison with transfer standard dosimeters.

4.3 The dose measurement is based on the measurement of the temperature rise (dosimeter response) in an absorber (calorimetric body) irradiated by an electron beam. Different absorbing materials are used, but the response is usually defined in terms of dose to water.

NOTE 2—The calorimetric bodies of the calorimeters described in this practice are made from low atomic number materials. The electron fluences within these calorimetric bodies are almost independent of energy when irradiated with electron beams of 1.5 MeV or higher, and the mass collision stopping powers are approximately the same for these materials.

4.4 The absorbed dose in other materials irradiated under equivalent conditions can be calculated. Procedures for making such calculations are given in ASTM Practices E666 and E668, and Ref (1).

4.4.1 Calorimeters for use at industrial electron accelerators have been constructed using graphite, polystyrene or a Petri dish filled with water as the calorimetric body (4-10). The thickness of the calorimetric body should be less than the range of the incident electrons.

4.4.2 Polymeric materials other than polystyrene might also be used for calorimetric measurements. Polystyrene is used

<sup>6</sup> The boldface numbers in parentheses refer to the bibliography at the end of this practice.