

~~ISO/TC 79/SC 5~~

~~Date: 2023-05~~

ISO/DTS 4181:20XX(E)

ISO/TC 79/SC 5

Secretariat: SAC

Date: 2023-07-10

Magnesium and magnesium alloys — Determination of strontium — Inductively coupled plasma optical emission spectrometric method

~~Magnésium et alliages de magnésium — Détermination du strontium — Méthode par spectrométrie d'émission atomique avec source à plasma induit~~

(standards.iteh.ai)

ISO/DTS 4181

<https://standards.iteh.ai/catalog/standards/sist/e94fbcde-7508-4768-ad7a-af589ea8634f/iso-dts-4181>

FDIS stage

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO/DTS 4181

<https://standards.iteh.ai/catalog/standards/sist/e94fbcde-7508-4768-ad7a-af589ea8634f/iso-dts-4181>

© ISO 2023

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office

CP 401 • Ch. de Blandonnet 8

CH-1214 Vernier, Geneva

Phone: + 41 22 749 01 11

E-mail: copyright@iso.org

Website: www.iso.org

Published in Switzerland

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO/DTS 4181

<https://standards.iteh.ai/catalog/standards/sist/e94fbcde-7508-4768-ad7a-af589ea8634f/iso-dts-4181>

Contents

Foreword	v
Introduction	viii
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Principle	1
5 Reagents	1
6 Apparatus	2
6.1 Inductively coupled plasma optical emission spectrometer	2
6.1.1 General	2
6.1.2 Wavelengths	2
6.1.3 Limit of detection	2
6.1.4 Linearity of the calibration curve	3
7 Sampling and sample preparation	3
8 Procedure	3
8.1 General	3
8.2 Test sample	3
8.3 Determination	3
8.3.1 Preparation of the test solution for strontium contents between 0,1 % (mass fraction) and 1,0 % (mass fraction)	3
8.3.2 Preparation of the calibration solutions for strontium contents between 0,1 % (mass fraction) and 1,0 % (mass fraction)	3
8.3.3 Preparation of the test solution for strontium contents between 1,0 % (mass fraction) and 3,5 % (mass fraction)	4
8.3.4 Preparation of the calibration solutions for strontium contents between 1,0 % (mass fraction) and 3,5 % (mass fraction)	4
8.4 Adjustment of the apparatus	5
8.5 Measurement of the calibration solutions	5
8.6 Calibration curve	5
8.7 Measurements of the test solution	5
9 Expression of results	5
9.1 Method of calculation	5
9.2 Precision	5

10	Test report	6
	Annex A (normative) Limit of detection	7
	Annex B (informative) Information on the precision test	8
	Annex C (informative) Graphical representation of precision data	10
	Bibliography	12

Foreword — iv

Introduction — v

1 Scope — 1

2 Normative references — 1

3 Terms and definitions — 1

4 Principle — 1

5 Reagents — 1

6 Apparatus — 2

7 Sampling and sample preparation — 3

8 Procedure — 3

8.1 General — 3

8.2 Test sample — 3

8.3 Determination — 3

8.3.1 Preparation of the test solution for strontium contents between 0,1 % (mass fraction) and 1,0 % (mass fraction) — 3

8.3.2 Preparation of the calibration solutions for strontium contents between 0,1 % (mass fraction) and 1,0 % (mass fraction) — 4

8.3.3 Preparation of the test solution for strontium contents between 1,0 % (mass fraction) and 3,5 % (mass fraction) — 4

8.3.4 Preparation of the calibration solutions for strontium contents between 1,0 % (mass fraction) and 3,5 % (mass fraction) — 4

8.4 Adjustment of the apparatus — 5

8.5 Measurement of the calibration solutions — 5

8.6 Calibration curve — 5

8.7 Measurements of the test solution — 5

9 Expression of results — 6

9.1 Method of calculation	6
9.2 Precision	6
10 Test report	6
Annex A (normative) Limit of detection	8
Annex B (informative) Information on the precision test	9
Annex C	10
Bibliography	11

iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO/DTS 4181

<https://standards.iteh.ai/catalog/standards/sist/e94fbcde-7508-4768-ad7a-af589ea8634f/iso-dts-4181>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO ~~documents~~document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Field Code Changed

~~Attention is drawn~~ISO draws attention to the possibility that ~~some of the elements~~implementation of this document may ~~be involve~~ the ~~subject~~use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights. ~~Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).~~

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) ~~see www.iso.org/iso/foreword.html~~, see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 79, *Light metals and their alloys*, Subcommittee SC 5, *Magnesium and alloys of cast or wrought magnesium*.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Magnesium and magnesium alloys are one kind of light metallic materials and show several advantageous properties, such as low density, high specific stiffness and strength, good damping capacity, castability, weldability and machinability, etc. Adding a proper amount of strontium to a magnesium alloy can reduce supercooling and grain growth rate. ~~So~~**Thus**, the grain size is refined and the compactness of magnesium alloys is increased.

ISO 16220 and ISO 3116 give the chemical composition of several magnesium alloy products containing strontium. The contents of strontium are "0,20-% to 0,30-%", "0,9-% to 1,5-%", "1,8-% to 2,3-%", and "2,1-% to 2,8-%". In order to cover this range, ~~the present~~**this** document specifies the determination of strontium from "0,02-% to 3,5-%".

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO/DTS 4181

<https://standards.iteh.ai/catalog/standards/sist/e94fbcde-7508-4768-ad7a-af589ea8634f/iso-dts-4181>

Magnesium and magnesium alloys — Determination of strontium — Inductively coupled plasma optical emission spectrometric method

1 Scope

This document specifies an inductively coupled plasma optical emission spectrometric method for the determination of strontium contents between 0,1-% (mass fraction) and 3,5-% (mass fraction) in magnesium and magnesium alloys.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 648, *Laboratory glassware* — Single-volume pipettes

ISO 1042, *Laboratory glassware* — One-mark volumetric flasks

ISO 3696, *Water for analytical laboratory use* — Specification and test methods

3 Terms and definitions

No terms and definitions are listed in this document.

ISO and IEC maintain ~~terminological~~terminology databases for use in standardization at the following addresses:

— ISO Online browsing platform: available at <https://www.iso.org/obp>

— IEC Electropedia: available at <http://www.electropedia.org/>

4 Principle

After dissolution of a test sample with nitric acid and hydrochloric acid, the solution is nebulized into an inductively coupled plasma optical emission spectrometer and the intensity of the emitted light from strontium is measured. The concentrations of strontium in the test solutions are derived from magnesium-based calibration curves.

5 Reagents

During the analysis, use only reagents of recognized analytical grade and only grade 2 water as specified in ISO 3696, or water of equivalent purity.

5.1 Pure magnesium, purity $\geq 99,99\%$ (mass fraction), free from strontium.

5.2 Strontium carbonate, purity $\geq 99,99\%$ (mass fraction).

5.3 Hydrochloric acid, ρ about 1,19 g/ml.

5.4 Nitric acid, ρ about 1,42 g/ml.

5.5 Hydrochloric acid solution 1 + 1.

ISO/DTS 4181:(E)

Add 500 ml of hydrochloric acid (5.3)(5.3) to 500 ml of water and mix.

5.6—Nitric acid solution 1 + 1.

Add 500 ml of nitric acid (5.4)(5.4) to 500 ml of water and mix.

5.7 Strontium standard solution, 1-g/l.

Dry several grams of strontium carbonate (5.2)(5.2) in an oven at $110\text{ }^{\circ}\text{C} \pm 5\text{ }^{\circ}\text{C}$ for at least 2 h and cool to room temperature in a desiccator. Weigh 1,684-9 g of the dried strontium carbonate, transfer into a 500 ml glass beaker, and add 50-ml of nitric acid solution (5.6)(5.6) and 15-ml of hydrochloric acid solution (5.5)(5.5). Cool and transfer quantitatively into a 1000-ml one-mark volumetric flask. Dilute to the mark with water and mix.

1 ml of this solution contains 1,0 mg of strontium.

5.8 Strontium standard solution, 0,1-g/l.

Transfer 10,00 ml of the strontium standard solution (5.7)(5.7) to a one-mark 100 ml volumetric flask. Add 10 ml of nitric acid solution (5.6)(5.6). Dilute to the mark with water and mix.

1 ml of this solution contains 0,1 mg of strontium.

6 Apparatus

All volumetric glassware shall be class A and calibrated in accordance with ISO 648 or ISO 1042, as appropriate.

6.1 Inductively coupled plasma optical emission spectrometer

6.1.1 General

The instrument used first shall be optimized in accordance with the manufacturer's instructions and then shall meet the performance criteria given in 6.1.3 to 6.1.4.

6.1.2 Wavelengths

This method does not specify any particular wavelength. It is mandatory that each laboratory shall carefully investigate the wavelengths available on its own equipment to find the most suitable ones regarding the sensitivity and the absence of interferences.

In Table 1, however, a suggestion is given in Table 1 together with possible interferences. This wavelength has been carefully investigated. It is recommended to use Sr 407,771 nm / Sr 421,552 nm because of their high sensitivity.

Table 1 — Examples of a wavelength for strontium determination

Element	Wavelength nm	Possible interferences
-		-
Sr	407,771	Fe
	421,552	Cr, Cu

6.1.3 Limit of detection

Calculate the limit of detection of strontium, according to Annex A in accordance with Annex A. The limit of detection should be equal to or less than 0,01-mg/l.

6.1.4 ~~6.1.4~~—Linearity of the calibration curve

The linearity of the calibration curve is checked by calculating the corresponding correlation coefficient. This coefficient shall be higher than 0,999.

7 Sampling and sample preparation

Sampling shall be carried out in accordance with an appropriate national standard for magnesium. If it is suspected that the laboratory sample is contaminated with oil or grease from the milling or drilling process, it shall be cleaned with ethanol or acetone and then dried in air. The sample shall be in the form of fine drillings, chips or millings with a maximum thickness of 1-mm. Sampling position shall be selected so as to be representative of the sample. In order to avoid oxidation of the surface, the chips shall be taken from an inner part of a bulk sample just before starting ~~an~~the analytical procedure.

8 Procedure

8.1 General

The same reagents should be used for the preparation of calibration solutions and sample solutions.

8.2 Test sample

Mix the sample well so that any portion weighed represents the average composition.

8.3 Determination

8.3.1 Preparation of the test solution for strontium contents between 0,1 % (mass fraction) and 1,0 % (mass fraction)

Weigh, to the nearest 0,1-mg, 0,25-g of the test sample (~~8.2~~see 8.2) and transfer- into a 300 ml glass beaker.

Add about 10 ml of water, and, in small portions, -10-ml of nitric acid solution (~~5.6~~)(5.6) and 3-ml of hydrochloric acid solution (~~5.5~~)(5.5). Cover with a watch-glass and, if necessary, heat gently to complete the dissolution.

Allow to cool at room temperature. Transfer the solution quantitatively into a 100 ml one-mark volumetric flask. Dilute to the mark with water and mix.

8.3.2 Preparation of the calibration solutions for strontium contents between 0,1-% (mass fraction) and 1,0 % (mass fraction)

Weigh, to the nearest ~~of~~1-mg, 0,25-g of pure magnesium (~~5.1~~)(5.1) and transfer into a series of 300-ml glass beakers.

Add about 10 ml of water and, in small portions, 10-ml of nitric acid solution (~~5.6~~)(5.6) and 3-ml of hydrochloric acid solution (~~5.5~~)(5.5). Cover with a watch-glass and, if necessary, heat gently to complete the dissolution. -Cool and transfer quantitatively into a series of 100 ml one-mark volumetric flasks.

In each volumetric flask, add the volume of -strontium standard solution (~~5.8~~)(5.8) shown in ~~Table 2~~ Table 2.

Dilute to the mark -with water and mix.

Table 2 — Calibration solution for -strontium contents between 0,1-% and 1,0-%