

---

---

**Cryolite primarily used for the  
production of aluminium —  
Determination of elements —  
Wavelength-dispersive X-ray  
fluorescence spectrometric method  
using pressed powder tablets**

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[ISO/PRF 4443](https://standards.iteh.ai/catalog/standards/sist/787a29d7-dd81-4342-98a5-87e23536dd5b/iso-prf-4443)

<https://standards.iteh.ai/catalog/standards/sist/787a29d7-dd81-4342-98a5-87e23536dd5b/iso-prf-4443>

**PROOF/ÉPREUVE**

---

---



Reference number  
ISO 4443:2021(E)

© ISO 2021

## iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO/PRF 4443

<https://standards.iteh.ai/catalog/standards/sist/787a29d7-dd81-4342-98a5-87e23536dd5b/iso-prf-4443>



### **COPYRIGHT PROTECTED DOCUMENT**

© ISO 2021

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

Page

Foreword.....	iv
Introduction.....	v
1 Scope.....	1
2 Normative references.....	1
3 Terms and definitions.....	1
4 Principle.....	1
5 Reagents and materials.....	1
6 Apparatus.....	2
7 Test procedure.....	2
7.1 Test specimen preparation.....	2
7.2 Test tablet preparation.....	2
7.3 X-ray fluorescence spectrometer application.....	2
7.3.1 Instrumental conditions.....	2
7.3.2 Calibration and calibration reference materials.....	3
7.3.3 Verification of the calibration.....	3
7.4 Monitoring of the sample for correction of instrumental drift.....	3
7.5 Analyses, calculation and expression of results.....	4
8 Precision.....	4
8.1 General.....	4
8.2 Repeatability.....	4
8.3 Reproducibility.....	4
9 Test report.....	5
Annex A (informative) Conversion table.....	6
Annex B (informative) Example of how to use Table 2 using sulfur (SO <sub>4</sub> ).....	7
Bibliography.....	8

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 226, *Materials for the production of primary aluminium*.

<https://standards.iteh.ai/catalog/standards/sist/787a29d7-dd81-4342-98a5-85b231d54455/iso-4443>

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

## Introduction

In the production of aluminium, cryolite is used as a flux in order to lower the melting points of aluminium smelting baths and increase the conductivity of the electrolyte. This use is critical during the start-up and normal operation stages of electrolysis cells.

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

ISO/PRF 4443

<https://standards.iteh.ai/catalog/standards/sist/787a29d7-dd81-4342-98a5-87e23536dd5b/iso-prf-4443>

## **iTeh STANDARD PREVIEW** **(standards.iteh.ai)**

ISO/PRF 4443

<https://standards.iteh.ai/catalog/standards/sist/787a29d7-dd81-4342-98a5-87e23536dd5b/iso-prf-4443>

# Cryolite primarily used for the production of aluminium — Determination of elements — Wavelength-dispersive X-ray fluorescence spectrometric method using pressed powder tablets

## 1 Scope

This document specifies a wavelength-dispersive X-ray fluorescence spectrometric (XRF) method for the determination of cryolite ( $\text{Na}_3\text{AlF}_6$ ) from the content of fluorine to the content of trace elements. The calibration reference materials are not specified in this method.

The method is applicable to cryolite, which is primarily used for the production of aluminium. [Annex A](#) provides conversion factors for converting elements to compounds. The validity and precision of test results for concentrations outside these ranges has not been determined.

The concentration range of fluorine (given as F) is from 510 g/kg to 560 g/kg. The concentration range of aluminium (given as Al) is from 120 g/kg to 150 g/kg. The concentration range of sodium (given as Na) is from 270 g/kg to 330 g/kg. The concentration range of silicon (given as  $\text{SiO}_2$ ) is from the lowest limit of detection (LLD) to 4,0 g/kg. The concentration range of iron (given as  $\text{Fe}_2\text{O}_3$ ) is from LLD to 0,37 g/kg. The concentration range of sulfur (given as  $\text{SO}_4$ ) is from LLD to 5,0 g/kg. The concentration range of phosphorus (given as  $\text{P}_2\text{O}_5$ ) is from LLD to 0,40 g/kg. The concentration range of calcium (given as CaO) is from LLD to 0,80 g/kg.

## 2 Normative references

[ISO/PRF 4443](#)

There are no normative references in this document.

## 3 Terms and definitions

No terms and definitions are listed in this document.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

## 4 Principle

A representative sample of cryolite is milled. A test portion is packed and pressed on a powder tablet press to make the test tablets.

The test tablets are analysed on an X-ray fluorescence spectrometer instrument that has been calibrated using a series of cryolite reference materials covering the required concentration range of the elements to be determined.

## 5 Reagents and materials

**5.1 Acetone or ethyl alcohol**, analytical pure, used as dispersing agent.

**5.2 Boric acid**, analytical pure.

## 6 Apparatus

**6.1 Wavelength-dispersive X-ray fluorescence spectrometer (XRF)**, with vacuum path and equipped with crystals required as shown in [Table 1](#).

**6.2 Vibratory disc mill**, with tray, ring and puck made of non-contaminating material. Tungsten-carbide and chrome steel have been found to be satisfactory.

**6.3 Tablet press**, capable of providing 340 kN for 20 s.

**6.4 Balance**, with precision  $\pm 0,01$  g.

**6.5 Flat spatula**.

## 7 Test procedure

### 7.1 Test specimen preparation

See ISO 1619 for guidance.

### 7.2 Test tablet preparation

**7.2.1** Dry the sample to constant mass at  $(110 \pm 2)$  °C. A typical drying time is 2 h.

**7.2.2** Weigh 10,0 g of sample and put it into the mill. Then add 10 drops of acetone or ethyl alcohol and vibrate for 60 s to make the particle size below 45 µm fineness.

**7.2.3** Transfer 2,0 g of milled sample to the tablet press, use boric acid backing to prepare the pressed tablet with a pressure of 340 kN and hold for 20 s. Tablet thickness should be a minimum of 4 mm, with a diameter of 40 mm.

**7.2.4** Remove the tablets from the tablet press. Touch the edge of the tablets only during measurement. Trim the edges. Avoid polluting the surface during X-ray measurement.

Always use the same mass and proportion of sample and reagents as for the calibration reference materials.

Results are particle-size dependent. Short milling times give larger spread in intralaboratory precision. The particle size should be determined by using a suitable sieving technique. Erroneous data will be collected if analytical samples contain particles varying significantly in size that cause different beam penetration depth and a different surface roughness from those in the calibration reference samples.

The surface of the pressed powder tablet should be smooth and firm and should not drop out power after trimming. The measuring surface should not be mixed with boric acid.

### 7.3 X-ray fluorescence spectrometer application

#### 7.3.1 Instrumental conditions

Follow the control setting and operation instructions of the spectrometer manufacturer, including monitoring and calibrating angles and monitoring and correcting for X-ray tube intensity.



Suggested conditions of measurement are given in Table 1. All measurements shall be made under vacuum. The  $K\alpha$  analytical lines are preferred. Use spectral lines overlap correction if required. Correct for background, using the lines recommended in Table 1.

**Table 1 — Measurement lines and suggested conditions of measurement**

Element	Crystal	Angle °2 $\theta$	Counter	Time for peak s	Time for background s	Collimator °	Voltage kV	Current mA
Fluorine	XS-55	38,694	Flow	60	30 % of time for peak	0,46	27	111
Aluminium	PET	144,584	Flow	30	30 % of time for peak	0,46	27	60
Sodium	XS-55	24,962	Flow	60	30 % of time for peak	0,46	27	80
Silicon	PET	108,985	Flow	10	30 % of time for peak	0,46	27	111
Iron	LiF200	57,556	Scintillation	2	30 % of time for peak	0,46	60	50
Sulfur	PET	75,73	Flow	10	30 % of time for peak	0,46	27	111
Phosphorus	Ge	141,009	Flow	10	30 % of time for peak	0,46	27	111
Calcium	LiF200	113,135	Flow	2	30 % of time for peak	0,46	50	60

### 7.3.2 Calibration and calibration reference materials

The calibration reference materials are not specified in this method. The user of this document shall obtain suitable calibration reference materials, such as the China National Set GSB 04-2016-2006 series GFC01 to GFC06 for F, Al, Na, SiO<sub>2</sub>, Fe<sub>2</sub>O<sub>3</sub>, SO<sub>4</sub>, P<sub>2</sub>O<sub>5</sub>, CaO.<sup>1)</sup>

The X-ray fluorescence spectrometer application software is used to make a cryolite application on the XRF instrument and the set of calibration reference materials are analysed. Elemental concentrations in the calibration samples shall bracket the values expected in the analytical samples. The number of calibration reference materials shall be six or more for each element. If six are not available, the number used shall be included in the test report.

Calibration is performed using the XRF instrument control software for application development and is not detailed in this document. If necessary, the instrument manufacturer should be consulted for recommendations on optimal tube anodes, crystal options and any concentration limit restrictions.

### 7.3.3 Verification of the calibration

Reference sample tablets can be used daily for checking the instrument condition. Limits for acceptable deviation shall be established by the laboratory; if the deviation is unacceptable, the instrument shall be reviewed and, if necessary, recalibrated.

## 7.4 Monitoring of the sample for correction of instrumental drift

Instrumental drift correction shall be performed regularly in accordance with the instrument manufacturer's instructions to ensure that measured intensities are corrected for any spectrometer

1) The China National Set GSB 04-2016-2006 series GFC01 to GFC06 for F, Al, Na, SiO<sub>2</sub>, Fe<sub>2</sub>O<sub>3</sub>, SO<sub>4</sub>, P<sub>2</sub>O<sub>5</sub>, CaO is an example of a suitable product available commercially. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO of this product.

drift. If monitor samples are measured infrequently, then each batch of analysed samples shall be drift corrected.

## 7.5 Analyses, calculation and expression of results

The analysis shall be run on three tablets made from the same test portion. The triplicate measurements are averaged to give the concentration to be used in the test report.

However, for any element, if any two results differ by more than twice the intralaboratory repeatability for that element as shown in [Table 2](#), the analysis should be repeated.

## 8 Precision

### 8.1 General

The precision given in [Table 2](#) was computed using ASTM E691. For details of the computation of the precision see References [3] and [4]. [Annex B](#) provides an example of how to use [Table 2](#).

**Table 2 — Precision for elements in cryolite based on average of triplicate measurements**

Element	Range	Repeatability $r$	Repeatability standard deviation $s_r$	Decimals <sup>a</sup> $D_r$	Reproducibility $R$	Reproducibility standard deviation $s_R$	Decimals <sup>a</sup> $D_R$
Fluorine	510 to 560	3	1,1	0	8	2,9	0
Aluminium	120 to 150	1	0,4	0	4	1,6	0
Sodium	270 to 330	3	0,9	0	12	4,4	0
Silicon	LLD to 4,0	0,1	0,03	1	0,3	0,09	1
Iron	LLD to 0,37	0,02	0,007	2	0,03	0,010	2
Sulfur	LLD to 5,0	0,2	0,21	1	0,4	0,14	1
Phosphorus	LLD to 0,40	0,02	0,007	2	0,03	0,010	2
Calcium	LLD to 0,80	0,1	0,02	1	0,3	0,09	1

<sup>a</sup>  $D_r$  and  $D_R$  are the recommended number of decimals.

### 8.2 Repeatability

The maximum permissible difference due to test error between two test results obtained by one operator on the same material using the same test method and the same test instrument is given by the repeatability limit ( $r$ ). The 95 % interval is given in [Table 2](#).

Two test results that do not differ by more than the repeatability limit ( $r$ ) will be considered to be from the same population; conversely, two test results that differ by more than the repeatability limit ( $r$ ) will be considered to be from different populations.

More than two test results for which the standard deviation does not differ by more than 1,96 ( $s_r$ ) (from [Table 2](#)) will be considered to be from the same population; conversely, more than two test results for which the standard deviation differs by more than 1,96 ( $s_r$ ) (from [Table 2](#)) will be considered to be from different populations.

### 8.3 Reproducibility

The maximum permissible difference due to test error between two test results obtained by two operators in different laboratories on the same material using the same test method is given by the reproducibility limit ( $R$ ). The 95 % interval is given in [Table 2](#).