

~~ISO/TC156-N7363~~

ISO/FDIS 4631

~~Date: 2023-05-09~~

~~ISO/TC_156/AVG 11~~

Secretariat: SAC

~~Date: 2023-07-17~~

Corrosion of metals and alloys — Measurement of the electrochemical critical localized corrosion potential (E-CLCP) for Ti alloys fabricated via additive manufacturing method in simulated biomedical solutions

(<https://standards.iteh.ai>)
Corrosion des métaux et alliages
Document Preview

ISO 4631

<https://standards.iteh.ai/catalog/standards/sist/f633af24-9493-4e7c-9c72-536e04bfb13f/iso-4631>

FDIS stage

ISO/FDIS 4631:2023(E)

© ISO 2023

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

iTeh Standards (<https://standards.itih.ai>) Document Preview

ISO 4631

<https://standards.itih.ai/catalog/standards/sist/f633af24-9493-4e7c-9c72-536e04bfb13f/iso-4631>

Contents — Page

Foreword.....	v
Introduction.....	vi
1 Scope.....	1
2 Normative references.....	1
3 Terms and definitions.....	1
4 Principles.....	2
5 Apparatus.....	3
5.1 Potentiostat.....	3
5.2 Heating bath with temperature controller.....	4
5.3 Specimen holder and connections.....	4
5.4 Test cell.....	5
5.5 Auxiliary electrode and reference electrode.....	7
6 Test solutions.....	7
7 Test specimen.....	7
8 Procedure.....	7
8.1 Preparation of reference electrodes.....	7
8.2 Preparation of biomedical AM Ti alloy specimen.....	7
8.3 Preparation of solution.....	7
8.4 Setting up the E-CLCP test.....	7
8.5 Ending test.....	9
9 Examination and evaluation.....	9
10 Test report.....	9
Annex A (informative) Validity of the E-CLCP by the correlation between E-CLCP and E-CLCT.....	11
Bibliography.....	13

Foreword —	2
Introduction —	3
1 — Scope —	4
2 — Normative Reference —	4
3 — Terms and definitions —	4
4 — Principles —	4
5 — Apparatus —	6
6 — Test Solutions —	7
7 — Test Specimen —	7
8 — Procedure —	8
9 — Examination and Evaluation —	9
10 — Report —	9
Annex A (informative) The validity of the E-CLCP by the correlation between E-CLCP and E-CLCT —	11

ISO/FDIS 4631:2023(E)

Bibliography — 12

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

ISO 4631

<https://standards.iteh.ai/catalog/standards/sist/f633af24-9493-4e7c-9c72-536e04bfb13f/iso-4631>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

~~Attention is drawn~~ **ISO draws attention** to the possibility that ~~some of the elements~~ **implementation** of this document may ~~be involve~~ **the subject of (a) patent(s)**. ~~ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).~~

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 156, *Corrosion of metals and alloys*.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Field Code Changed

Introduction

Ti alloys such as Ti-6Al-4V are considered to be the most promising biomedical materials. Due to a unique combination of high strength, low modulus, lower density, and outstanding corrosion resistance, their applications have become more widespread in a wide range of industries, e.g. aerospace, automobile, marine and biomedical fields. Especially the medical grade Ti alloys are implanted in patients worldwide every year and also have a significantly higher strength to weight ratio than competing stainless steels. It has well established that Ti alloys are excellent corrosion resistance by all body fluids and tissue and are thus completely biocompatible. These Ti alloys are conventionally produced by wrought or cast processes which are the subtractive manufacturing (SM) methods, but recently the new additive manufacturing (AM) method is emerging, called "3D printing". This new methodology has gained worldwide attention as a way to cut costs and improve efficiency. Comparing SM with AM, the ratio of the mass of the starting, raw material to the mass of the final, finished part can be as high as 20:1. In terms of mechanical view point, both strength and ductility of Ti alloys such as Ti-6Al-4V fabricated by AM are comparable to or above their properties made by conventional manufacturing methods, because of their unique microstructure. However, the resistance to corrosion of Ti alloys produced by AM is still unknown, whether it is comparable to those of the conventionally manufactured Ti alloys or not, because of their defects such as porosity, the formation of martensite phase resulting from the rapid solidification, and directional difference with the stacking. Therefore, the new standards should be made how to evaluate the resistance to localized corrosion on the stacked alloys produced by AM methods. For this use, the measurement method such as electrochemical critical localized corrosion temperature (E-CLCT) has been adopted as ISO 22910:2020. However, at the normal ranges of pH and temperatures corresponding to human body, the use of E-CLCT measurement method should be limited because of the temperature scan during testing. Therefore, electrochemical critical localized corrosion potential (E-CLCP) is newly introduced and proposed as a new criterion for the evaluation on the resistance to the localized corrosion on the biomedical AM Ti alloys in humane body environments. This new test method is controlled by ~~Potentiostat~~potentiostat through potentiodynamic –galvanostatic-potentiostatic polarization processes and electrochemical polarization cell is used in the artificial physiological fluids.

ISO 4631

<https://standards.iteh.ai/catalog/standards/sist/f633af24-9493-4e7c-9c72-536e04bfb13f/iso-4631>