



SLOVENSKI STANDARD

SIST ENV 725-2:2000

01-december-2000

Advanced technical ceramics - Methods of test for ceramic powders - Part 2: Determination of impurities in barium titanate

Advanced technical ceramics - Methods of test for ceramic powders - Part 2:
Determination of impurities in barium titanate

Hochleistungskeramik - Prüfverfahren für keramische Pulver - Teil 2: Bestimmung von
Verunreinigungen in Bariumtitanat

Céramiques techniques avancées - Méthodes d'essai pour poudres céramiques - Partie
2: Détermination des impuretés dans le titane de baryum

<https://standards.iteh.ai/catalog/standards/sist/8735ae93-d21a-4997-83ac-982b73c1ae24/sist-env-725-2-2000>

Ta slovenski standard je istoveten z: **ENV 725-2:1994**

ICS:

81.060.30 Sodobna keramika Advanced ceramics

SIST ENV 725-2:2000

en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST ENV 725-2:2000

<https://standards.iteh.ai/catalog/standards/sist/8735ae93-d21a-4997-83ac-982b73c1ae24/sist-env-725-2-2000>

EUROPEAN PRESTANDARD

ENV 725-2

PRÉNORME EUROPÉENNE

EUROPÄISCHE VORNORM

January 1994

UDC 666.3-121:620.1:543.422

Descriptors: Ceramics, powdery materials, tests, determination, impurities, barium, emission spectrophotometry

English version

**Advanced technical ceramics - Methods of test for
ceramic powders - Part 2: Determination of
impurities in barium titanate**

Céramiques techniques avancées - Méthodes
d'essais pour poudres céramiques - Partie 2:
Détermination des impuretés dans le titane de
baryum

Hochleistungskeramik - Prüfverfahren für
keramische Pulver - Teil 2: Bestimmung von
Verunreinigungen in Bariumtitanat

(standards.iteh.ai)

SIST ENV 725-2:2000

<https://standards.iteh.ai/catalog/standards/sist/8735ae93-d21a-4997-83ac-982b73c1ae24/sist-env-725-2-2000>

This European Prestandard (ENV) was approved by CEN on 1992-12-08 as a prospective standard for provisional application. The period of validity of this ENV is limited initially to three years. After two years the members of CEN will be requested to submit their comments, particularly on the question whether the ENV can be converted into an European Standard (EN).

CEN members are required to announce the existence of this ENV in the same way as for an EN and to make the ENV available promptly at national level in an appropriate form. It is permissible to keep conflicting national standards in force (in parallel to the ENV) until the final decision about the possible conversion of the ENV into an EN is reached.

CEN members are the national standards bodies of Austria, Belgium, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

CEN

European Committee for Standardization
Comité Européen de Normalisation
Europäisches Komitee für Normung

Central Secretariat: rue de Stassart, 36 B-1050 Brussels

Contents list

	Page
Foreword	3
1 Scope	4
2 Normative references	4
3 Principle	4
4 Reagents	4
5 Apparatus	5
6 Preparation of sample solution	5
7 Calibration	6
8 Procedure	6
9 Test report	7

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST ENV 725-2:2000

<https://standards.iteh.ai/catalog/standards/sist/8735ae93-d21a-4997-83ac-982b73c1ae24/sist-env-725-2-2000>

Foreword

This European prestandard has been prepared by Technical Committee CEN/TC 184 "Advanced technical ceramics", of which the secretariat is held by BSI.

EN 725 consists of 11 Parts:

- Part 1: Determination of impurities in alumina
- Part 2: Determination of impurities in barium titanate (ENV)
- Part 3: Determination of oxygen content of non-oxides by thermal extraction with a carrier gas
- Part 4: Determination of oxygen content in aluminium nitride by XRF analysis (ENV)
- Part 5: Determination of particle size distribution
- Part 6: Determination of the specific surface area
- Part 7: Determination of absolute density
- Part 8: Determination of tapped bulk density
- Part 9: Determination of untamped bulk density
- Part 10: Determination of compaction properties
- Part 11: Determination of densification on natural sintering (ENV)

CEN/TC 184 approved this European prestandard by resolution 1 during its seventh meeting held in Berlin, 1992-12-08/09.

In accordance with the CEN/CENELEC Internal Regulations, the following countries are bound to announce this European prestandard: Austria, Belgium, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

1 Scope

This Part of EN 725 describes a method for the determination of impurities in barium titanate powders, using inductively coupled plasma atomic emission (ICP-AES).

The method is applicable only to stoichiometric barium titanate. The maximum concentrations measured for each impurity are as follows:

- Sr: 4 000 ppm
- Ca: 500 ppm
- K: 200 ppm
- Na, Mg, Al, Fe, Nb: 100ppm

The minimum concentration or detection limits are from ppm to 5 ppm.

2 Normative references

This European pre-standard incorporates by dated or undated reference provisions from other publications. These normative references are cited at appropriate places in the text and in the publications listed hereafter. For dated references, subsequent amendments to or revisions of any of these publications apply to this European pre-standard only when incorporated in it by amendment or revision. For undated references the latest edition of the publication referred to applies:

- <https://standards.iteh.ai/catalog/standards/sist/8735ae93-d21a-4997-83ac-721777777777/standard/en-725-2-2000>
- EN 45001 General criteria for the operation of testing laboratories.
- ISO 13527 Chemical analysis of ferrous materials - Guidelines on the use of inductively coupled plasma atomic emission spectrometry.†

3 Principle

Calcined barium titanate powder is dissolved in a mixture of hydrochloric acid and hydrogen peroxide, and the impurities in the solution are determined by ICP-AES analysis, using an addition method of synthetic matrix matched solutions for the calibration.

4 Reagents

Spectroscopic grade reagents shall be used, including:

- 4.1 Quartz distilled water or water of equivalent purity;
- 4.2 Hydrochloric acid, density = 1.18 g/cm³ at 20°C;
- 4.3 Hydrogen peroxide, 110 volume;
- 4.4 Barium chloride solution;
- 4.5 Titanium chloride, TiCl₄, in solution.

† In course of preparation.

5 Apparatus

- 5.1 Spectrometer for inductively coupled plasma atomic emission, with a quartz torch hand pneumatic cross-flow nebuliser. See ISO 13527.
- 5.2 Muffle furnace for operation at $900\text{ }^{\circ}\text{C} \pm 50\text{ }^{\circ}\text{C}$.
- 5.3 Platinum crucible and lid.
- 5.4 Balance to weigh to $\pm 0,001\text{ g}$.
- 5.5 PTFE beakers, 100 ml, with paraffin film.
- 5.6 Polyethylene pipettes, micro-pipettes and volumetric flasks.
- 5.7 Magnetic stirrer with a small bar.
- 5.8 Desiccator.

6 Preparation of sample solution

6.1 Sample calcination

Weigh about 1,2 g of the sample in the platinum crucible. Calcine the sample for two hours in the muffle furnace at $900\text{ }^{\circ}\text{C} \pm 50\text{ }^{\circ}\text{C}$. Remove it from the hot furnace and allow it to cool in the desiccator.

6.2 Test sample

Weigh 1,000 g of the calcined powder from 6.1, to the nearest 0,001 g, and place this test sample in a 100 ml PTFE beaker.

[SIST ENV 725-2:2000](#)

6.3 Dissolution of the test sample

Place the magnetic bar (5.7) in the beaker containing the test sample and carefully add 25 ml of water (4.1), 25 ml of hydrochloric acid (4.2) and 8 ml of hydrogen peroxide (4.3). Cover the beaker with paraffin film. Heat the beaker at $60\text{ }^{\circ}\text{C}$, with magnetic stirring, until either the sample is dissolved or the amount of residue is constant.

NOTE : A 'double boiler' is recommended for heating.

If there is residue, remove this by sedimentation and repeat the dissolution procedure.

Transfer the cool solution to a 100 ml volumetric flask and make up to volume with the distilled water (4.1). Use the solution for the analysis (clause 8) within 48 h.

7 Calibration

7.1 Samples with unknown impurity levels

Samples with unknown impurity levels shall be calibrated by an addition method. Prepare the final test solutions by diluting the sample solution (6.3) so that:

- a) the final test solutions with and without added amounts of standard solutions (4.4 and 4.5) are in the concentration range given in table 1.
- b) the amount of standard solution added is similar to the amount of impurities in the final test solution.

7.2 Samples with known impurity levels

Samples with known impurity levels shall be calibrated by a synthetic method. This method uses solutions which are matrix-matched with BaCl_2 , TiCl_4 , HCl and H_2O_2 . The concentration of matrix components in these solutions shall be such that the concentration of the impurities is within the range given in table 1.

Table 1 : Concentration range for impurities

Element	Concentration range (10^{-3} g ML ⁻¹)
Na	0 - 10
K	0 - 5
Mg	0 - 0,3
Ca	0 - 3
Sr	0 - 5
Al	0 - 10
Fe	0 - 1
Nb	0 - 15

8 Procedure

Adjust the ICP spectrometer (see 5.1) to provide a stable plasma, a high signal-to-noise ratio and a separation of spectral overlaps for the impurities being determined. Follow the manufacturer's instructions for igniting the plasma and wear UV eye protectors as soon as the plasma can be seen.

Use the wavelengths given in table 2 for determination of each impurity.