



# FINAL DRAFT International Standard

## ISO/FDIS 4064-2

### Water meters for cold potable water and hot water —

#### Part 2: Test methods

*Compteurs d'eau potable froide et d'eau chaude —*

*Partie 2: Méthodes d'essai*

ISO/TC 30/SC 7

Secretariat: **BSI**

Voting begins on:  
**2024-06-27**

Voting terminates on:  
**2024-08-22**

iTeh Standards  
(<https://standards.iteh.ai>)  
Document Preview

ISO/FDIS 4064-2

<https://standards.iteh.ai/catalog/standards/iso/117ef2f8-b617-4573-9533-3398b6fea8bd/iso-fdis-4064-2>

**ISO/CEN PARALLEL PROCESSING**

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

iTeh Standards  
(<https://standards.iteh.ai>)  
Document Preview

ISO/FDIS 4064-2

<https://standards.iteh.ai/catalog/standards/iso/117ef2f8-b617-4573-9533-3398b6fea8bd/iso-fdis-4064-2>



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2024

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

## Contents

Page

<b>Foreword</b>	<b>vii</b>
<b>1 Scope</b>	<b>1</b>
<b>2 Normative references</b>	<b>1</b>
<b>3 Terms and definitions</b>	<b>2</b>
<b>4 Reference conditions</b>	<b>2</b>
<b>5 Symbols, units and equations</b>	<b>3</b>
<b>6 External examination</b>	<b>3</b>
6.1 General	3
6.2 Object of the examination	3
6.3 Preparation	3
6.4 Examination procedures	4
6.4.1 General	4
6.4.2 Marks and inscriptions ( ISO 4064-1:— OIML R 49-1:—, 6.6)	4
6.4.3 Indicating device ( ISO 4064-1:— OIML R 49-1:—, 6.7)	4
6.4.4 Protection devices ( ISO 4064-1:— OIML R 49-1:—, 6.8)	8
<b>7 Performance tests for all water meters</b>	<b>8</b>
7.1 General	8
7.2 Required conditions for all tests	8
7.2.1 Water quality	8
7.2.2 General rules concerning test installation and location	8
7.3 Static pressure test ( ISO 4064-1:— OIML R 49-1:—, 4.2.10)	9
7.3.1 Object of the test	9
7.3.2 Preparation	9
7.3.3 Test procedure	9
7.3.4 Acceptance criteria	9
7.4 Determination of intrinsic errors (of indication) (ISO 4064-1:— OIML R 49-1:—, 7.2.3)	10
7.4.1 Object of the test	10
7.4.2 Preparation	10
7.4.3 Combination meters	15
7.4.4 Test procedure	16
7.4.5 Acceptance criteria	16
7.4.6 Interchange test on all types of cartridge meters and meters with exchangeable metrological modules ( ISO 4064-1: — OIML R 49-1: —, 7.2.7)	17
7.5 Water temperature test ( ISO 4064-1: — OIML R 49-1: —, 4.2.8)	18
7.5.1 Object of the test	18
7.5.2 Preparation	18
7.5.3 Test procedure	18
7.5.4 Acceptance criteria	18
7.6 Overload water temperature test ( ISO 4064-1: —  OIML R 49-1: —, 7.2.5)	18
7.6.1 Object of the test	18
7.6.2 Preparation	18
7.6.3 Test procedure	18
7.6.4 Acceptance criteria	18
7.7 Water pressure test ( ISO 4064-1: —  OIML R 49-1: — 4.2.8)	19
7.7.1 Object of the test	19
7.7.2 Preparation	19
7.7.3 Test procedure	19
7.7.4 Acceptance criteria	19
7.8 Reverse flow test ( ISO 4064-1: —  OIML R 49-1: —, 4.2.7)	19
7.8.1 Object of the test	19
7.8.2 Preparation	19
7.8.3 Test procedure	19
7.8.4 Acceptance criteria	20

## ISO/FDIS 4064-2:2024(en)

7.9	Pressure loss test ( ISO 4064-1: — OIML R 49-1: —, 6.5)	20
7.9.1	Object of the test	20
7.9.2	Equipment for pressure loss test	21
7.9.3	Test procedure	22
7.9.4	Calculation of the actual $\Delta p$ of a water meter	24
7.9.5	Acceptability criteria	24
7.10	Flow disturbance tests ( ISO 4064-1: — OIML R 49-1: —, 6.3.4)	24
7.10.1	Object of the tests	24
7.10.2	Preparation	24
7.10.3	Test procedure	25
7.10.4	Acceptance criteria	25
7.11	Durability tests ( ISO 4064-1: —  OIML R 49-1: —, 7.2.6)	25
7.11.1	General	25
7.11.2	Discontinuous flow test	25
7.11.3	Continuous flow test	29
7.12	Magnetic field testing	31
7.13	Tests on ancillary devices of a water meter	31
7.13.1	Object of the test	31
7.13.2	Preparation	31
7.13.3	Test procedure	31
7.13.4	Acceptance criteria	32
7.14	Environmental testing	32
7.15	Software evaluation	32
<b>8</b>	<b>Performance tests related to influence factors and disturbances</b>	<b>32</b>
8.1	General requirements ( ISO 4064-1: — OIML R 49-1: —, A.1)	32
8.1.1	Overview	32
8.1.2	Environmental classification	32
8.1.3	Electromagnetic classification	32
8.1.4	Reference conditions ( ISO 4064-1: — OIML R 49-1: —, 7.1)	32
8.1.5	Test volumes for measuring error (of indication) of a water meter	33
8.1.6	Influence of the water temperature ( ISO 4064-1: — OIML R 49-1: —, A.5)	33
8.1.7	Requirements for environmental tests	33
8.1.8	Equipment under test ( ISO 4064-1: — OIML R 49-1: —, 7.2.12.3)	34
8.1.9	Determination of required tests	35
8.2	Dry heat (non-condensing) ( ISO 4064-1: — OIML R 49-1: —, A.5)	36
8.2.1	Object of the test	36
8.2.2	Preparation	36
8.2.3	Test procedure (in brief)	36
8.2.4	Acceptance criteria	37
8.3	Cold ( ISO 4064-1: — OIML R 49-1: —, A.5)	37
8.3.1	Object of the test	37
8.3.2	Preparation	37
8.3.3	Test procedure (in brief)	37
8.3.4	Acceptance criteria	38
8.4	Damp heat, cyclic (condensing) ( ISO 4064-1: — OIML R 49-1: —, A.5)	38
8.4.1	Object of the test	38
8.4.2	Preparation	38
8.4.3	Test procedure (in brief)	38
8.4.4	Acceptance criteria	39
8.5	Power supply variation ( ISO 4064-1: — OIML R 49-1: —, A.5)	39
8.5.1	Water meters powered by direct AC or by AC/DC converters ( ISO 4064-1: — OIML R 49-1: —, A.5)	39
8.5.2	Water meters powered by external DC voltage or by primary DC batteries ( ISO 4064-1: — OIML R 49-1: —, A.5)	40
8.5.3	Interruption in battery supply	41
8.6	Vibration (random) ( ISO 4064-1: — OIML R 49-1: —, A.5)	41
8.6.1	Object of the test	41
8.6.2	Preparation	41

8.6.3	Test procedure (in brief)	42
8.6.4	Acceptance criteria	42
8.7	Mechanical shock ( ISO 4064-1:— OIML R 49-1:—, A.5)	42
8.7.1	Object of the test	42
8.7.2	Preparation	42
8.7.3	Test procedure (in brief)	43
8.7.4	Acceptance criteria	43
8.8	AC mains voltage dips, short interruptions and voltage variations ( ISO 4064-1:— OIML R 49-1:—, A.5)	43
8.8.1	Object of the test	43
8.8.2	Preparation	43
8.8.3	Test procedure (in brief)	43
8.8.4	Acceptance criteria	45
8.9	Bursts on signal lines ( ISO 4064-1:— OIML R 49-1:—, A.5)	45
8.9.1	Object of the test	45
8.9.2	Preparation	46
8.9.3	Test procedure (in brief)	46
8.9.4	Acceptance criteria	46
8.10	Bursts (transients) on AC and DC mains (ISO 4064-1:— OIML R 49-1:—, A.5)	47
8.10.1	Object of the test	47
8.10.2	Preparation	47
8.10.3	Test procedure (in brief)	47
8.10.4	Acceptance criteria	47
8.11	Electrostatic discharge (ISO 4064-1:— OIML R 49-1:—, A.5)	48
8.11.1	Object of the test	48
8.11.2	Preparation	48
8.11.3	Test procedure (in brief)	48
8.11.4	Acceptance criteria	49
8.12	Radiated electromagnetic fields ( ISO 4064-1:— OIML R 49-1:—, A.5)	49
8.12.1	Object of the test	49
8.12.2	Preparation	49
8.12.3	Test procedure (in brief)	49
8.12.4	Acceptance criteria	50
8.13	Conducted electromagnetic fields ( ISO 4064-1:— OIML R 49-1:—, A.5)	51
8.13.1	Object of the test	51
8.13.2	Preparation	51
8.13.3	Test procedure (in brief)	51
8.13.4	Acceptance criteria	52
8.14	Surges on signal, data and control lines (ISO 4064-1 OIML R 49-1, A.5)	52
8.14.1	Object of the test	52
8.14.2	Preparation	52
8.14.3	Test procedure (in brief)	52
8.14.4	Acceptance criteria	53
8.15	Surges on AC and DC mains power lines ( ISO 4064-1: — OIML R 49-1: —, A.5)	53
8.15.1	Object of the test	53
8.15.2	Preparation	53
8.15.3	Test procedure (in brief)	53
8.15.4	Acceptance criteria	54
8.16	Static magnetic field (ISO 4064-1 OIML R 49-1, 7.2.8)	54
8.16.1	Test conditions	54
8.16.2	Object of the test	55
8.16.3	Preparation	55
8.16.4	Test procedure in brief	55
8.16.5	Acceptance criteria	55
8.17	Absence of flow test	56
8.17.1	Object of the test	56
8.17.2	Preparation	56
8.17.3	Test procedure	56
8.17.4	Acceptance criteria	56

8.18	Mains power frequency electromagnetic fields test.....	56
8.18.1	Object of the test.....	56
8.18.2	Preparation.....	56
8.18.3	Test procedure.....	56
8.18.4	Acceptance criteria.....	57
<b>9</b>	<b>Test program for type evaluation.....</b>	<b>57</b>
9.1	Number of samples required.....	57
9.2	Performance test applicable to all water meters.....	57
9.3	Performance tests applicable to electronic water meters, mechanical water meters fitted with electronic devices, and their separable parts.....	58
9.4	Type evaluation of separable parts of a water meter.....	59
9.5	Families of water meters.....	59
<b>10</b>	<b>Tests for initial verification.....</b>	<b>59</b>
10.1	Initial verification of complete and combined water meters.....	59
10.1.1	Object of the test.....	59
10.1.2	Preparation.....	60
10.1.3	Test procedure.....	60
10.1.4	Acceptance criteria.....	60
10.2	Initial verification of separable parts of a water meter.....	61
10.2.1	Object of the test.....	61
10.2.2	Preparation.....	61
10.2.3	Test procedure.....	61
10.2.4	Acceptance criteria.....	61
<b>11</b>	<b>Presentation of results.....</b>	<b>62</b>
11.1	Object of the reports.....	62
11.2	Identification and test data to be included in records.....	62
11.2.1	Type evaluation.....	62
11.2.2	Initial verification.....	62
<b>Annex A (normative)</b>	<b>Type examination and testing of checking facilities of electronic devices.....</b>	<b>63</b>
<b>Annex B (normative)</b>	<b>Calculating the relative error (of indication) of a water meter.....</b>	<b>69</b>
<b>Annex C (normative)</b>	<b>Installation requirements for flow disturbance tests.....</b>	<b>74</b>
<b>Annex D (normative)</b>	<b>Type evaluation of a family of water meters.....</b>	<b>75</b>
<b>Annex E (informative)</b>	<b>Examples of methods and components used for testing concentric water meters.....</b>	<b>77</b>
<b>Annex F (informative)</b>	<b>Maximum uncertainties in the measurement of influence factors and disturbances.....</b>	<b>80</b>
<b>Annex G (informative)</b>	<b>Pressure loss test pressure tappings, hole and slot details.....</b>	<b>83</b>
<b>Annex H (normative)</b>	<b>Flow disturbers.....</b>	<b>86</b>
<b>Annex I (normative)</b>	<b>Software evaluation for software-controlled water meters.....</b>	<b>96</b>
<b>Annex ZA (informative)</b>	<b>Relationship between this European Standard and the essential requirements of Directive 2014/32/EU aimed to be covered.....</b>	<b>97</b>
<b>Bibliography</b>	<b>.....</b>	<b>102</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at [www.iso.org/patents](http://www.iso.org/patents). ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 30, *Measurement of fluid flow in closed conduits*, Subcommittee SC 7, *Volume methods including water meters* and OIML Technical Subcommittee TC 8/SC 5 *Water meters*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 92, *Test methods and equipment for cold water meters*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This fifth edition of ISO 4064-1 cancels and replaces the fourth edition (ISO 4064-1:2014), which has been technically revised.

The main changes are as follows:

- a few editorial and technical changes were done throughout the document.

This edition of ISO 4064-1 is identical to the corresponding edition of OIML R 49-1, which has been issued concurrently. OIML R 49-1 was approved for final publication by the International Committee of Legal Metrology at its 59th meeting in October 2024. It will be submitted to the International Conference on Legal Metrology in XXX for formal sanction.

A list of all parts in the ISO 4064 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).





# Water meters for cold potable water and hot water —

## Part 2: Test methods

### 1 Scope

This document is applicable to the type evaluation and initial verification testing of water meters for cold potable water and hot water as defined in ISO 4064-1:—|OIML R 49-1:—. OIML Certificates of conformity can be issued for water meters under the scope of the OIML Certificate System, provided that this document, ISO 4064-1:—|OIML R 49-1:2024 and ISO 4064-3:—|OIML R 49-3:— are used in accordance with the rules of the system.

This document sets out details of the test programme, principles, equipment and procedures to be used for the type evaluation, and initial verification of a meter type.

The provisions of this document also apply to ancillary devices, if required by national regulations.

The provisions include requirements for testing the complete water meter and for testing the measurement transducer (including the flow or volume sensor) and the calculator (including the indicating device) of a water meter as separate units.

### 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60068-2-1, *Environmental testing — Part 2-1: Tests — Test A: Cold*

IEC 60068-2-2, *Environmental testing — Part 2-2: Tests — Test B: Dry heat*

IEC 60068-2-30, *Environmental testing — Part 2-30: Tests — Test Db: Damp heat, cyclic (12 h + 12 h cycle)*

IEC 60068-2-31, *Environmental testing — Part 2-31: Tests — Test Ec: Rough handling shocks, primarily for equipment-type specimens*

IEC 60068-2-47, *Environmental testing — Part 2-47: Tests — Mounting of specimens for vibration, impact and similar dynamic tests*

IEC 60068-2-64, *Environmental testing — Part 2-64: Tests — Test Fh: Vibration, broadband random and guidance*

IEC 60068-3-4, *Environmental testing — Part 3-4: Supporting documentation and guidance — Damp heat tests*

IEC 60654-2, *Operating conditions for industrial process measurement and control equipment — Part 2: Power*

IEC 61000-2-1, *Electromagnetic compatibility (EMC) — Part 2: Environment — Section 1: Description of the environment — Electromagnetic environment for low-frequency conducted disturbances and signaling in public power supply systems*

IEC 61000-2-2, *Electromagnetic compatibility (EMC) — Part 2-2: Environment — Compatibility levels for low-frequency conducted disturbances and signaling in public low-voltage power supply systems*

IEC 61000-4-1, *Electromagnetic compatibility (EMC) — Part 4-1: Testing and measurement techniques — Overview of IEC 61000-4 series*

IEC 61000-4-2, *Electromagnetic compatibility (EMC) — Part 4-2: Testing and measurement techniques — Electrostatic discharge immunity test*

IEC 61000-4-3, *Electromagnetic compatibility (EMC) — Part 4-3: Testing and measurement techniques — Radiated, radio frequency, electromagnetic field immunity test*

IEC 61000-4-4, *Electromagnetic compatibility (EMC) — Part 4-4: Testing and measurement techniques — Electrical fast transient/burst immunity test*

IEC 61000-4-5, *Electromagnetic compatibility (EMC) — Part 4-5: Testing and measurement techniques — Surge immunity test*

IEC 61000-4-6, *Electromagnetic compatibility (EMC) — Part 4-6: Testing and measurement techniques — Immunity to conducted disturbances, induced by radio-frequency fields*

IEC 61000-4-8, *Electromagnetic compatibility (EMC) - Part 4-8: Testing and measurement techniques - Power frequency magnetic field immunity test*

IEC 61000-4-11, *Electromagnetic compatibility (EMC) — Part 4-11: Testing and measurement techniques — Voltage dips, short interruptions and voltage variations immunity tests*

IEC 61000-4-20, *Electromagnetic compatibility (EMC) - Part 4-20: Testing and measurement techniques - Emission and immunity testing in transverse electromagnetic (TEM) waveguides*

IEC 61000-6-1, *Electromagnetic compatibility (EMC) — Part 6-1: Generic standards — Immunity for residential, commercial and light-industrial environments*

IEC 61000-6-2, *Electromagnetic compatibility (EMC) — Part 6-2: Generic standards — Immunity for industrial environments*

OIML D 11:2013, *General requirements for measuring instruments -Environmental conditions*

OIML G 13, *Planning of metrology and testing laboratories*

ISO 4064-1, *Water meters for cold potable water and hot water — Part 1: Metrological and technical requirements*

ISO 4064-3, *Water meters for cold potable water and hot water — Part 3: Test report format*

ISO/IEC Guide 98-3:2008, *Uncertainty of measurement — Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

### 3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 4064-1|OIML R 49-1 apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

### 4 Reference conditions

All applicable influence quantities, except for the influence quantity being tested, shall be held at the following values during type evaluation tests on a water meter. However, for influence factors and disturbances for electronic water meters, it is permissible to use the reference conditions defined in the applicable IEC standard:

Flow rate:	$0,7 \times (Q_2 + Q_3) \pm 0,03 \times (Q_2 + Q_3)$
Water temperature:	T30, T50 is $20\text{ °C} \pm 5\text{ °C}$ T70 to T180 is $20\text{ °C} \pm 5\text{ °C}$ and $50\text{ °C} \pm 5\text{ °C}$ T30/70 to T30/180 is $50\text{ °C} \pm 5\text{ °C}$
Water pressure:	Within rated operating conditions (see ISO 4064-1:— OIML R 49-1:—, 6.4)
Ambient temperature range:	$15\text{ °C}$ to $25\text{ °C}$
Ambient relative humidity range:	$25\%$ to $75\%$
Ambient atmospheric pressure range:	$86\text{ kPa}$ to $106\text{ kPa}$ [ $0,86\text{ bar}$ to $1,06\text{ bar}$ ]
Power supply voltage (mains AC):	Nominal voltage, $U_{\text{nom}} \pm 5\%$
Power supply frequency:	Nominal frequency, $f_{\text{nom}} \pm 2\%$
Power supply voltage (battery):	A voltage $V$ in the range $U_{\text{bmin}} \leq V \leq U_{\text{bmax}}$

During each test, the temperature and relative humidity shall not vary by more than  $5\text{ °C}$  or  $10\%$ , respectively, within the reference range. The reference conditions are permitted to deviate from the defined tolerance values during the performance tests if evidence can be given to the body responsible for type approval that the type of meter under consideration is not affected by the deviation of the condition in question. The actual values of the deviating condition, however, shall be measured and documented as part of the performance test documentation.

## 5 Symbols, units and equations

Equations, symbols and their units, concerning the calculation of the error (of indication) of a water meter used in this part of ISO 4064|OIML R 49, are given in [Annex B](#).

## 6 External examination

### 6.1 General

During the external examination, all relevant values, dimensions, and observations shall be recorded.

NOTE 1 For presentation of the results of type examinations, see [11](#).

NOTE 2 The relevant subclauses of ISO 4064-1:—|OIML R 49-1:— are shown in parentheses in the following.

### 6.2 Object of the examination

To verify that a water meter meets the requirements of ISO 4064-1:—|OIML R 49-1:— with respect to the design of the indicating device, the marking of the meter and the application of protection devices.

### 6.3 Preparation

Linear measurements that have to be taken of a meter shall be made using traceable, calibrated measuring devices.

The actual or apparent dimensions of the scales of the indicating device shall be taken without removing the meter lens or disassembling the meter.

NOTE A travelling microscope (cathetometer) can be used to measure the width, spacing and height of the scale divisions and the height of the numerals.

## 6.4 Examination procedures

### 6.4.1 General

The following aspects of a meter design shall be examined on at least one meter from the sample.

Either the same meter sample may be used for all the external examinations or different meters from the samples submitted may be used for some of the examinations.

### 6.4.2 Marks and inscriptions ( ISO 4064-1:—|OIML R 49-1:—, 6.6)

- a) Verify that a place is provided for affixing the verification mark which is visible without dismantling the water meter.
- b) Verify that the water meter is clearly and indelibly marked with the information presented in ISO 4064-1:—|OIML R 49-1:—, 6.6.2.
- c) Complete the section reference ISO 4064-1:—|OIML R 49-1:—, 6.6.1 and 6.6.2 (r) in ISO 4064-3:—|OIML R 49-3:—, 4.4.1.

### 6.4.3 Indicating device ( ISO 4064-1:—|OIML R 49-1:—, 6.7)

#### 6.4.3.1 Function ( ISO 4064-1:—|OIML R 49-1:—, 6.7.1.1)

- a) Verify that the indicating device provides an easily read, reliable and unambiguous visual indication of the indicated volume.
- b) Verify that the indicating device includes visual means for testing and calibration.
- c) If the indicating device includes additional elements for testing and calibration by other methods, e.g. for automatic testing and calibration, record the type(s) of device.
- d) If the meter is a combination meter with two indicating devices, [6.4.3](#) applies to both indicating devices.
- e) Complete the section reference ISO 4064-1:—|OIML R 49-1:—, 6.7.1.1 in ISO 4064-3:—|OIML R 49-3:—, 4.4.1.

#### 6.4.3.2 Unit of measurement, symbol, and its placement ( ISO 4064-1:—|OIML R 49-1:—, 6.7.1.2)

- a) Verify that the indicated volume of water is expressed in cubic metres.
- b) Verify that the symbol  $\text{m}^3$  appears on the dial or immediately adjacent to the numbered display.
- c) Complete the section reference ISO 4064-1:—|OIML R 49-1:—, 6.7.1.2 in ISO 4064-3:—|OIML R 49-3:—, 4.4.1.

#### 6.4.3.3 Indicating range ( ISO 4064-1:—|OIML R 49-1:—, 6.7.1.3)

- a) Verify that the indicating device is able to record the indicated volume in cubic metres given in ISO 4064-1:—|OIML R 49-1:—, Table 5 corresponding to the permanent flow rate  $Q_3$  without passing through zero.
- b) Complete the section reference ISO 4064-1:—|OIML R 49-1:—, 6.7.1.3, in ISO 4064-3:—|OIML R 49-3:—, 4.4.1.

#### 6.4.3.4 Colour coding for indicating devices (ISO 4064-1:—|OIML R 49-1:—, 6.7.1.4)

- a) Verify that either:
  - 1) the colour black is used to indicate the cubic metre and its multiples; and
  - 2) the colour red is used to indicate sub-multiples of a cubic metre; and
  - 3) the colours are applied either to the pointers, indexes, numbers, wheels discs, dials or aperture frames;
  - 4) or other means of indicating the cubic metre are used in which there is no ambiguity in distinguishing between the primary indication and alternative displays, e.g. sub-multiples for verification and testing.
- b) Complete the section reference ISO 4064-1:—|OIML R 49-1:—, 6.7.1.4, in ISO 4064-3:—|OIML R 49-3:—, 4.4.1.

#### 6.4.3.5 Types of indicating device (ISO 4064-1:—|OIML R 49-1:—, 6.7.2)

##### 6.4.3.5.1 Type 1 — Analogue device (ISO 4064-1:—|OIML R 49-1:—, 6.7.2.1)

- a) If a type 1 indicating device has been used, verify that volume is indicated by:
  - either continuous movement of one or more pointers moving relative to graduated scales;
  - or continuous movement of one or more circular scales or drums, each passing an index.
- b) Verify that the value expressed in cubic metres for each scale division is of the form  $10^n$ , where  $n$  is a positive or a negative whole number or zero, thereby establishing a system of consecutive decades.
- c) Verify that each scale is either graduated in values expressed in cubic metres or accompanied by a multiplication factor ( $\times 0,001$ ;  $\times 0,01$ ;  $\times 1$ ;  $\times 10$ ;  $\times 100$ ;  $\times 1\ 000$ , etc.).
- d) Verify that the rotational movements of the pointers or circular scales are clockwise.
- e) Verify that the linear movement of the pointers or scales is from left to right.
- f) Verify that the movement of the numbered roller indicators is upwards.
- g) Complete the section reference ISO 4064-1:—|OIML R 49-1:—, 6.7.2.1, in ISO 4064-3:—|OIML R 49-3:—, 4.4.1.

##### 6.4.3.5.2 Type 2 — Digital device (ISO 4064-1:—|OIML R 49-1:—, 6.7.2.2)

- a) Verify that the indicated volume is given by a line of digits, appearing in one or more apertures.
- b) Verify that the advance of one digit is completed while the digit of the next immediately lower decade changes from 9 to 0.
- c) Verify that the actual or apparent height of the digits is at least 4 mm.
- d) For non-electronic devices:
  - 1) verify that the movement of the numbered roller indicators (drums) is upwards;
  - 2) if the lowest value decade has a continuous movement, verify that the aperture is large enough to permit a digit to be read without ambiguity.

For electronic devices:

- 1) verify that for non-permanent displays the volume can be displayed at any time for at least 10 s;
- 2) check the entire display visually in the following sequence:
  - i) for seven segment type verify that all the elements can be correctly displayed (e.g. an “eights” test),

- ii) for seven segment type verify that all the elements can be blanked (a “blanks” test),
  - iii) for graphical displays use an equivalent test to verify that display faults cannot result in any digit being misinterpreted,
  - iv) verify that each step of the sequence lasts at least 1 s.
- e) Complete the section reference ISO 4064-1:—|OIML R 49-1:—, 6.7.2.2 in ISO 4064-3:—|OIML R 49-3:—, 4.4.1.

#### **6.4.3.5.3 Type 3 — Combination of analogue and digital devices ( ISO 4064-1:—|OIML R 49-1:—, 6.7.2.3)**

- a) If the indicating device is a combination of type 1 and 2 devices, verify that the respective requirements of each apply (see [6.4.3.5.1](#) and [6.4.3.5.2](#)).
- b) Complete the section reference ISO 4064-1:—|OIML R 49-1:—, 6.7.2.3 in ISO 4064-3:—|OIML R 49-3:—, 4.4.1.

#### **6.4.3.6 Verification devices — First element of an indicating device — Verification scale interval ( ISO 4064-1:—|OIML R 49-1:—, 6.7.3)**

##### **6.4.3.6.1 General requirements (ISO 4064-1:—|OIML R 49-1:—, 6.7.3.1)**

- a) Verify that the indicating device has the means for visual, non-ambiguous verification testing and calibration.
- b) Note whether the visual verification display has a continuous or a discontinuous movement.
- c) Note whether, in addition to the visual verification display, the indicating device includes provisions for rapid testing by the inclusion of complementary elements (e.g. star wheels or discs), providing signals through externally attached sensors. Note the relationship, stated by the manufacturer, between the visual indication of volume and the signals emitted by these complementary devices.
- d) Complete the section reference ISO 4064-1:—|OIML R 49-1:—, 6.7.3.1 in ISO 4064-3:—|OIML R 49-3:—, 4.4.1.

##### **6.4.3.6.2 Visual verification display ( ISO 4064-1:—|OIML R 49-1:—, 6.7.3.2)**

##### **6.4.3.6.2.1 Value of the verification scale interval (ISO 4064-1:—|OIML R 49-1:—, 6.7.3.2.1)**

- a) Verify that the value of the verification scale interval, expressed in cubic metres, is of the form  $1 \times 10^n$ , or  $2 \times 10^n$ , or  $5 \times 10^n$ , where  $n$  is a positive or negative whole number, or zero.
- b) For analogue and digital indicating devices with continuous movement of the first element, verify that the verification scale interval is formed from the division into 2, 5 or 10 equal parts of the interval between two consecutive digits of the first element.
- c) For analogue and digital indicating devices with continuous movement of the first element, verify that numbering is not applied to the divisions between consecutive digits of the first element.
- d) For digital indicating devices with discontinuous movement of the first element, the verification scale interval is the interval between two consecutive digits or incremental movements of the first element.
- e) Complete the section reference ISO 4064-1:—|OIML R 49-1:—, 6.7.3.2.1 in ISO 4064-3:—|OIML R 49-3:—, 4.4.1.

##### **6.4.3.6.2.2 Form of the verification scale (ISO 4064-1:—|OIML R 49-1:—, 6.7.3.2.2)**

- a) If the indicating device has continuous movement of the first element, check that the apparent scale spacing is not less than 1 mm and not more than 5 mm.