

ISO/FDIS 23402-3:2023(E)

ISO TC 106/SC 6/WG 8

Secretariat: DIN

Dentistry — Portable dental equipment ~~for use~~used in non-permanent healthcare ~~environment~~environments — Part 3: Portable suction equipment

First edition ~~Updated~~

2023-~~10-30~~12-12

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO/FDIS 23402-3](#)

<https://standards.iteh.ai/catalog/standards/sist/a7b667aa-8905-494e-b075-cbeb1036c1b5/iso-fdis-23402-3>

© ISO 2023

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ~~ISO's~~ISO's member body in the country of the requester.

ISO ~~copyright office~~ Copyright Office

CP 401 • CH ~~de Blandonnet 8~~

~~CH~~-1214 Vernier, Geneva

Phone: + 41 22 749 01 11

Email: copyright@iso.org ~~copyright@iso.org~~

Website: www.iso.org

Published in Switzerland.

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO/FDIS 23402-3](#)

<https://standards.iteh.ai/catalog/standards/sist/a7b667aa-8905-494c-b075-cbeb1036c1b5/iso-fdis-23402-3>

Contents

Foreword.....	5
Introduction.....	6
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Classification	3
4.1 Applicable classifications	3
4.2 Classification according to the degree of integration and location	3
4.2.1 General	3
4.2.2 Type A: Fully integrated	3
4.2.3 Type B: Remote	3
4.2.4 Type C: Standalone	3
4.3 Classification according to air flow rate	3
5 Requirements	4
5.1 General	4
5.2 Protection against electric shock	4
5.3 Safety requirements	4
5.4 Electromagnetic compatibility	5
5.5 Utility requirements	5
5.5.1 Compressed air supply	5
5.5.2 Suction waste container	5
5.6 Operational requirements	5
5.6.1 General	5
5.6.2 Cleaning	5
5.7 Characterization of portable suction equipment flow rate performance	6
5.8 Maximum suction pressure	6
5.9 Amalgam separator	6
5.10 Bacterial filter	6
6 Sampling	6
7 Measurement and test methods	6
7.1 General	6
7.1.1 General provisions for tests	6
7.1.2 Reference conditions (air flow rates)	7
7.1.3 Atmospheric conditions	7
7.1.4 Other conditions	7
7.2 Performance tests	7
7.2.1 Measurement of flow rate performance	7
7.2.2 Measurement of maximum suction pressure of portable suction equipment types A and C	8
7.2.3 Measurement of maximum suction pressure of type B portable suction equipment ..	8
7.2.4 Measures to prevent excessive suction pressure	8
7.3 Bacterial filters	8
7.4 Visual inspection	8
7.4.1 Visual inspection equipment	8
7.4.2 Visual inspection of documentation	9

7.5	Resistance to cleaning agents	9
7.5.1	Principle	9
7.5.2	Test specimens	9
7.5.3	Test conditions	9
7.5.4	Cleaning agent solutions	9
7.5.5	Immersion procedure	9
7.5.6	Inspection	10
8	Information supplied by the manufacturer	11
8.1	General	11
8.2	Instructions for use	11
8.3	Technical description	12
9	Marking	13
9.1	General	13
9.2	Marking on the outside of electrically operated portable suction equipment	13
9.3	Marking of packaging	13
10	Packaging	14
	Bibliography	15

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO/FDIS 23402-3](https://standards.iteh.ai/catalog/standards/sist/a7b667aa-8905-494e-b075-cbeb1036c1b5/iso-fdis-23402-3)

<https://standards.iteh.ai/catalog/standards/sist/a7b667aa-8905-494e-b075-cbeb1036c1b5/iso-fdis-23402-3>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO ~~documents~~ document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

~~Attention is drawn~~ ISO draws attention to the possibility that ~~some of the elements~~ implementation of this document may ~~be involve~~ the subject use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ~~ISO~~ ISO shall not be held responsible for identifying any or all such patent rights. ~~Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see).~~

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 6, *Dental equipment*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 55, *Dentistry*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

A list of all parts in the ISO 23402 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

The ISO 23402 series is a series of standards ~~(planned or published)~~, with the objective of standardizing requirements for portable dental equipment for use in non-permanent healthcare environments. ~~ISO 23402 consists of the following parts, under the general title, Dentistry — Portable dental equipment for use in non-permanent healthcare environments.~~

~~— Part 1: General requirements~~

~~The following parts are in preparation:~~

~~— Part 2: Portable dental units~~

~~— additional parts are planned to be added in the future~~

iTeh Standards (<https://standards.iteh.ai>) Document Preview

[ISO/FDIS 23402-3](https://standards.iteh.ai/catalog/standards/sist/a7b667aa-8905-494e-b075-cbeb1036c1b5/iso-fdis-23402-3)

<https://standards.iteh.ai/catalog/standards/sist/a7b667aa-8905-494e-b075-cbeb1036c1b5/iso-fdis-23402-3>

Dentistry — Portable dental equipment ~~for use~~used in non-permanent healthcare ~~environment~~environments — Part 3: Portable suction equipment

1 Scope

This document specifies terminology, classification, requirements and test methods for portable suction equipment primarily intended to be used by dental professionals in non-permanent healthcare environments.

This document applies to portable suction equipment incorporated in a portable dental unit and free-standing portable suction equipment.

The requirements in this document focus on portability ~~and certain technical aspects appropriate for international standardization~~.

This document specifies requirements for information to be supplied by the manufacturer on the performance, operation and maintenance of portable suction equipment designed and constructed to be transported for use in non-permanent healthcare environments. ~~This document~~ also specifies requirements for the instructions to be supplied by the manufacturer on assembling, disassembling, and packing for human transport between non-permanent healthcare environments.

This document does not apply to stationary dental equipment, wearable equipment (such as headlamps and loupes), mobile dental equipment or portable dental equipment that is not intended to be used in non-permanent healthcare environments or not designed to be disassembled, folded or packed for human transport between non-permanent healthcare environments. Also, requirements for stationary dental equipment that ~~may~~can be installed in a dental mobile medical facility (e.g., vehicular or containerized mobile dental clinic) are not considered in this document.

This document specifies requirements for portable suction equipment used to provide reduced pressure and flow at the cannula connector.

This document does not apply to portable suction equipment used for life support or for scavenging halogenated anaesthetic gases.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 1942, *Dentistry — Vocabulary*

ISO 5167-1:2022, *Measurement of fluid flow by means of pressure differential devices inserted in circular cross-section conduits running full — Part 1: General principles and requirements*

ISO 7494-2, *Dentistry — Stationary dental units and dental patient chairs — Part 2: Air, water, suction and wastewater systems*

ISO/FDIS 23402-3:2023(E)

ISO 9687, *Dentistry — Graphical symbols for dental equipment*

ISO 11143, *Dentistry — Amalgam separators*

ISO 23402-1:2020, *Dentistry — Portable dental equipment for use in non-permanent healthcare environment — Part 1: General requirements*

ISO 29463-1:2017, *High efficiency filters and filter media for removing particles from air — Part 1: Classification, performance, testing and marking*

IEC 60335-1, *Household and similar electrical appliances — Safety — Part 1: General requirements*

IEC 60601-1:2005, +AMD1:2012+ AMD2:2020, *Medical electrical equipment — Part 1: General requirements for basic safety and essential performance*

IEC 61000-6-2, *Electromagnetic compatibility (EMC) — Part 6-2: Generic standards — Immunity standard for industrial environments*

IEC 61000-6-3, *Electromagnetic compatibility (EMC) — Part 6-3: Generic standards — Emission standard for residential, commercial and light-industrial environments*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 1942, ISO 7494-2, ISO 23402-1:2020, IEC 60335-1, IEC 60601-1:2005, IEC 60601-1:2005+AMD1:2012+AMD2:2020 and the following apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

— ISO Online browsing platform: available at <https://www.iso.org/obp>

— IEC Electropedia: available at <https://www.electropedia.org/>

3.1

air separator

apparatus which separates liquids and solids from the suction system air flow

[SOURCE: ISO 10637:2018, 3.1]

3.2

cannula connector

component at the inlet end of the dental suction operating hose, which joins the cannula to the operating hose

[SOURCE: ISO 10637:2018, 3.2]

3.3

suction machine

suction equipment component that produces pressure that is lower than atmospheric pressure

EXAMPLE Pump, side channel blower.

3.4

patient environment

area contained within the walls of an operatory or in the absence of walls, within a 1,5 ~~metre m~~ radius of the patient

Note 1 to entry: The area within a 1,5 m radius of the patient's body has been ~~justified~~defined as the patient environment ~~in practice~~. See IEC 60601-1:2005, Figure A.9.

4 Classification**4.1 Applicable classifications**

Classifications are in accordance with ~~clause 4 of~~ ISO 23402-1:2020, Clause 4, except ~~as follows: those given in 4.2 and 4.3.~~

4.2 Classification according to the degree of integration and location**4.2.1 General**

Portable suction equipment is classified according to the degree of integration into the dental unit and the distance from the patient.

4.2.14.2.2 Type A: Fully integrated

The suction machine, air separator, solids filter, suction operating hose and cannula connector are contained in the portable dental unit, which is located within the patient environment.

4.2.24.2.3 Type B: Remote

The suction machine is not contained in the portable dental unit and ~~may~~can be located outside the patient environment. The air separator and solids filter are integrated in either the portable dental unit or portable suction equipment or both. ISO/FDIS 23402-3

4.2.34.2.4 Type C: ~~Stand alone: Standalone~~

The suction machine, air separator, solids filter, suction operating hose and cannula connector are contained in a self-contained portable suction unit, which is located within the patient environment and separate from the portable dental unit.

4.3 Classification according to air flow rate

Portable suction equipment is classified according to the minimum air flow rate intended to be produced by the portable suction equipment under normal use conditions.

NOTE 1 This classification is intended to facilitate communication between parties involved in the specification, design, procurement, operation and maintenance of portable suction equipment.

NOTE 2 Reference conditions for ~~NL~~ (normal litres (NL)) are defined in 7.1.2.

— Type 1: portable suction equipment is intended to supply a minimum air flow rate of 250 NL/min at one suction cannula connector on the dental unit.

NOTE 3 Type 1 portable suction equipment is often commercially referred to as “high-volume suction” in certain regions.

- Type 2: portable suction equipment is intended to supply a minimum air flow rate of 170 l/min at one suction cannula connector on the dental unit.
- Type 3: portable suction equipment is intended to supply a minimum air flow rate of 90 l/min at one suction cannula connector on the dental unit.

NOTE 4 Type 3 portable suction equipment is often commercially referred to as “medium-volume suction” in certain regions.

These specified classifications do not restrict other specifications for minimum air flow rate that may be deemed appropriate in certain applications, regions or markets. In instances where none of the specified classifications are applicable, an alternative value may be specified for the minimum air volume flow rate, to be supplied by the manufacturer.

NOTE 5 The target performance indicated for each of these types of portable suction equipment is intended to assist in communicating the performance requirements of complete portable suction equipment. See also ISO 10637:2018, 4.2.

5 Requirements

5.1 Applicable requirements

5.1 General

Requirements are ~~in accordance with clause 5 of~~ according to ISO 23402-1:2020, Clause 5 except where superseded by ~~the following: 5.2 to 5.10.~~

5.2 Protection against electric shock

If applicable, portable suction equipment classified as Class I in accordance with IEC 60601-1 shall have an external protective earth terminal or conductor, and the instructions for use shall include instructions for the connection to the earth.

Devices with a voltage of 24VAC or 34VDC or less, can be used without an external protective earth terminal or conductor, when specified in the instructions for use.

5.3 Safety requirements

For the safety requirements, the following requirements shall be applied according to the classification specified in 4.2.

- For portable suction equipment intended to be located within patient environment, the requirements of IEC 60601-1:~~2005+AMD1:2012+AMD2:2020~~ shall apply.

Testing shall be carried out in accordance with IEC 60601-1:~~2005+AMD1:2012+AMD2:2020.~~

- For type B portable suction equipment intended to be located outside of patient environment, the requirements of IEC 60335-1 shall apply.

Testing shall be carried out in accordance with ~~IEC 60335~~ IEC 60335-1.