



FINAL DRAFT International Standard

ISO/FDIS 15883-1

Washer-disinfectors —

Part 1: General requirements, terms and definitions and tests

Laveurs désinfecteurs —

Partie 1: Exigences générales, termes et définitions et essais

[ISO/FDIS 15883-1](https://standards.iteh.ai/catalog/standards/iso/20244107-f392-40b6-b1e3-cf9489759430/iso-fdis-15883-1)

<https://standards.iteh.ai/catalog/standards/iso/20244107-f392-40b6-b1e3-cf9489759430/iso-fdis-15883-1>

ISO/CEN PARALLEL PROCESSING

ISO/TC 198

Secretariat: **ANSI**

Voting begins on:
2024-03-19

Voting terminates on:
2024-06-11

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO/FDIS 15883-1](https://standards.iteh.ai/catalog/standards/iso/20244107-f392-40b6-b1e3-cf9489759430/iso-fdis-15883-1)

<https://standards.iteh.ai/catalog/standards/iso/20244107-f392-40b6-b1e3-cf9489759430/iso-fdis-15883-1>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2024

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword	vi
Introduction	vii
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Performance requirements	10
4.1 General.....	10
4.2 Cleaning.....	14
4.2.1 General.....	14
4.2.2 Pre-wash stage.....	14
4.2.3 Washing stage.....	14
4.2.4 Rinsing stage.....	15
4.3 Disinfection.....	15
4.3.1 Thermal disinfection.....	15
4.3.2 Chemical disinfection.....	15
4.4 Final rinsing.....	16
4.5 Drying.....	17
4.6 Process chemicals.....	17
4.7 Self-disinfection.....	17
5 Mechanical and process requirements	18
5.1 Materials, design and manufacture/construction.....	18
5.1.1 Materials.....	18
5.1.2 Design and manufacture/construction.....	19
5.1.3 Load carrier(s).....	20
5.2 Safety.....	20
5.3 Tanks.....	21
5.4 Loading and unloading doors and their controls.....	22
5.4.1 General.....	22
5.4.2 Control of manually operated doors.....	23
5.4.3 Control of doors of a double-ended WD.....	23
5.4.4 Internal doors and access ports.....	23
5.4.5 Continuous process machine without doors.....	23
5.5 Pipework and fittings.....	24
5.6 Spray systems.....	24
5.7 Dosing systems.....	24
5.8 Load temperature protection.....	25
5.9 Process temperature control limits.....	26
5.10 Switches, gauges and indicating devices.....	27
5.11 Process verification.....	27
5.12 Instrumentation and controls.....	28
5.13 Temperature indicating systems.....	29
5.14 Pressure indicating systems.....	30
5.15 Volume/flow indicating devices.....	30
5.16 Timing equipment.....	31
5.17 Operating cycle indicating equipment.....	31
5.18 Recording systems (if fitted).....	31
5.18.1 Cycle control recorders.....	31
5.18.2 Process verification system.....	32
5.19 Control systems.....	33
5.20 Override of automatic control.....	34
5.21 Microprocessor control systems.....	35
5.22 Fault indication systems.....	35
5.23 Water supply.....	36

ISO/FDIS 15883-1:2024(en)

5.24	Venting and drainage systems.....	36
5.25	Drainage.....	37
5.26	Air filters installed within the WD.....	37
5.27	Load handling and supports for use within the WD.....	37
5.28	Trolleys.....	38
5.29	Environment.....	38
6	Testing for conformity.....	38
6.1	General.....	38
6.1.1	Inter-relationship of tests.....	38
6.1.2	Conformity of WD, as supplied, with ISO 15883-1.....	38
6.1.3	Conformity of WD, as installed, with ISO 15883-1.....	39
6.1.4	Confirmation of validation.....	40
6.1.5	Requalification.....	40
6.1.6	Routine and periodic tests.....	41
6.2	Instrumentation for testing.....	41
6.2.1	Temperature sensors.....	41
6.2.2	Temperature recording instruments.....	41
6.2.3	Calibration.....	42
6.3	Tests on doors, interlocks and fault indications.....	42
6.3.1	Operating cycle start interlock.....	42
6.3.2	Door locking during operating cycle.....	42
6.3.3	Door interlocks on double-ended WD.....	42
6.3.4	Cycle complete door interlocks.....	43
6.3.5	Fault indication on sensor failure.....	43
6.3.6	Fault indication on service failure.....	44
6.3.7	Failed cycle interlock.....	44
6.3.8	Blocked drain protection.....	44
6.4	Tests on water quality and water volume.....	45
6.4.1	General.....	45
6.4.2	Quality of final rinse water.....	45
6.4.3	Volume of water used per stage.....	46
6.4.4	Quality of water used during testing.....	46
6.5	Tests on pipework.....	46
6.5.1	Estimation of dead volume pipework.....	46
6.5.2	Leakage.....	47
6.5.3	Free draining [chamber, load carrier(s), tanks].....	47
6.5.4	Pipework flow to discharge point.....	47
6.5.5	Venting.....	48
6.5.6	Load contamination from ductwork of the WD.....	48
6.6	Tests on instrumentation fitted to the WD.....	49
6.6.1	Verification of calibration.....	49
6.6.2	Legibility.....	49
6.7	Tests on load carrier(s) and trolleys.....	49
6.7.1	Load carrier(s) used within the chamber.....	49
6.7.2	Trolleys.....	50
6.8	Thermometric tests.....	50
6.8.1	General.....	50
6.8.2	Load and load carrier(s) temperature test during operating cycle.....	50
6.8.3	Chamber wall temperature test.....	51
6.8.4	Temperature tests on tanks.....	52
6.8.5	Load temperature protection.....	52
6.9	Chemical dosing tests.....	53
6.9.1	Dispensed volume.....	53
6.9.2	Indication of insufficient process chemical for a cycle.....	53
6.10	Tests of cleaning efficacy.....	54
6.10.1	General.....	54
6.10.2	Cleaning type test.....	54
6.10.3	Cleaning performance qualification test.....	54

ISO/FDIS 15883-1:2024(en)

6.10.4	Routine process monitoring.....	54
6.10.5	Tests for process residuals.....	55
6.11	Tests of air quality.....	55
6.11.1	General.....	55
6.11.2	Procedure.....	55
6.11.3	Results.....	55
6.12	Test of load dryness.....	55
6.12.1	General/principle.....	55
6.12.2	Materials.....	55
6.12.3	Exterior surface drying.....	56
6.12.4	Lumen load drying.....	56
6.13	Test of automatic control.....	56
6.13.1	General.....	56
6.13.2	Procedure.....	57
6.13.3	Results.....	57
7	Documentation.....	57
8	Information to be supplied.....	58
8.1	General.....	58
8.2	Before delivery of the WD and for installation.....	58
8.3	At delivery of the WD.....	59
9	Marking, labelling and packaging.....	61
9.1	Marking and labelling.....	61
9.2	Packaging.....	61
10	Information to be requested from the purchaser by the supplier of the WD.....	61
Annex A	(informative) Test programme.....	63
Annex B	(informative) A_0 concept — Comparative lethality of moist heat processes.....	67
Annex C	(normative) Microbiological recovery medium for estimation of bacterial contamination of water.....	71
Annex ZA	(informative) Relationship between this European Standard the General Safety and Performance Requirements of Regulation (EU) 2017/745 aimed to be covered.....	72
Bibliography	79

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 198, *Sterilization of health care products*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 102, *Sterilizers for medical purposes*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 15883-1:2006), which has been technically revised. It also incorporates ISO 15883-1:2006/Amd 1:2014.

The main changes are as follows:

- alignment of terms and definitions with ISO 11139:2018;
- addition of requirements for load carrier(s);
- clarification of water quality requirements;
- relocation of former [Annex C](#), *Test methods for the detection and assessment of residual proteinaceous contamination*, to ISO 15883-5:2021;
- redesignation of former Annex D, *Microbiological recovery medium for estimation of bacterial contamination of water*, as [Annex C](#);
- increase in the minimum temperature limit for thermal disinfection and calculation of A_0 values from 65 °C to 70 °C (see Reference [\[48\]](#));
- revision of the Bibliography.

A list of all parts in the ISO 15883 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

This document is the first part of the ISO 15883 series of standards specifying the performance of washer-disinfectors (WDs) and specifies the general requirements for performance applicable to all WDs. The requirements given in this document are applicable to all WDs specified in subsequent parts of the ISO 15883 series, except insofar as they are modified or added to by a subsequent part, in which case the requirements of that particular part apply.

Fields of application within the scope of ISO 15883 series can include laboratory, veterinary, dental and pharmaceutical applications and other specific applications, such as WDs for bedsteads and transport carts and the disinfection of crockery and cutlery intended for use with immunologically compromised patients.

WDs are used only for processing the type of loads specified by the manufacturer of the WD.

In selecting the appropriate WD, references are made both to this document and to the relevant parts of ISO 15883 series. It is the user's responsibility to ensure that the choice of type of WD, operating cycle or quality of services or process chemicals is appropriate for any particular load.

This document has been prepared on the basis that each individual WD is subject to validation tests (e.g. installation qualification, operational qualification, and performance qualification on first installation) and that in use continued conformance is established by periodic tests.

In respect of the potential adverse effects on the quality of water intended for human consumption or environmental impacts caused by the WD and its intended use, it is noteworthy that the ISO 15883 series of standards provides no information as to whether the WD may be used without restrictions in any of the ISO member states.

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO/FDIS 15883-1](https://standards.iteh.ai/catalog/standards/iso/20244107-f392-40b6-b1e3-cf9489759430/iso-fdis-15883-1)

<https://standards.iteh.ai/catalog/standards/iso/20244107-f392-40b6-b1e3-cf9489759430/iso-fdis-15883-1>

Washer-disinfectors —

Part 1: General requirements, terms and definitions and tests

1 Scope

This document specifies general performance requirements for washer-disinfectors (WD) and WD accessories that are intended to be used for cleaning and disinfection of reusable medical devices. It specifies performance requirements for cleaning and disinfection as well as for the accessories that can be required to achieve the necessary performance. The methods and instrumentation required for validation, routine control and monitoring and requalification, periodically and after essential repairs, are also specified.

NOTE 1 The requirements can be applied to WDs intended for use with other articles used in the context of medical, dental, pharmaceutical and veterinary practice.

The requirements for WDs intended to process specific loads are specified in ISO 15883-2, ISO 15883-3, ISO 15883-4, ISO 15883-6 and ISO 15883-7. For WDs intended to process loads of two or more different types, the requirements of ISO 15883-2, ISO 15883-3, ISO 15883-4, ISO 15883-6 and ISO 15883-7 apply.

This document does not specify requirements intended for machines for use for laundry or general catering purposes.

This document does not include requirements for machines which are intended to sterilize the load, or which are designated as “sterilizers” and addressed in other standards.

The specified performance requirements of this document do not ensure the inactivation or removal of the causative agent(s) (prion protein) of transmissible spongiform encephalopathies.

NOTE 2 Chemicals in some cleaning agents and disinfectants can react with prion protein in a manner that can inhibit its removal or inactivation. If the presence of prion protein is considered a possibility, then this can influence the choice of cleaning agent and disinfectant.

NOTE 3 This document can be used by prospective purchasers and manufacturers as the basis of agreement on the specification of a WD. The test methods for demonstration of conformity with the requirements of this document can also be employed by users to demonstrate continued conformity of the installed WD throughout its service life. Guidance on a routine test programme is given in [Annex A](#).

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 7000, *Graphical symbols for use on equipment — Registered symbols*

ISO 10012, *Measurement management systems — Requirements for measurement processes and measuring equipment*

ISO 14644-3, *Cleanrooms and associated controlled environments — Part 3: Test methods*

ISO 14971, *Medical devices — Application of risk management to medical devices*

ISO/FDIS 15883-1:2024(en)

ISO 15883-2:—¹⁾, *Washer-disinfectors — Part 2: Requirements and tests for washer-disinfectors employing thermal disinfection for critical and semi-critical medical devices*

ISO 15883-3, *Washer-disinfectors — Part 3: Requirements and tests for washer-disinfectors employing thermal disinfection for human waste containers*

ISO 15883-4, *Washer-disinfectors — Part 4: Requirements and tests for washer-disinfectors employing chemical disinfection for thermolabile endoscopes*

ISO 15883-5:2021, *Washer-disinfectors — Part 5: Performance requirements and test method criteria for demonstrating cleaning efficacy*

ISO 15883-6, *Washer-disinfectors — Part 6: Requirements and tests for washer-disinfectors employing thermal disinfection for non-invasive, non-critical medical devices and healthcare equipment*

ISO 15883-7:—²⁾, *Washer-disinfectors — Part 7: Requirements and tests for washer-disinfectors employing chemical disinfection for thermolabile non-critical medical devices and healthcare equipment*

IEC 60417-DB, *Graphical symbols for use on equipment*

IEC 60584-1:2013, *Thermocouples — Part 1: EMF specifications and tolerances*

IEC 60751:2008, *Industrial platinum resistance thermometer and platinum temperature sensors*

IEC 61010-2-040:2020, *Safety requirements for electrical equipment for measurement, control and laboratory use – Part 2-040: Particular requirements for sterilizers and washer-disinfectors used to treat medical materials*

IEC 61326-1:2021, *Electrical equipment for measurement, control and laboratory use. EMC requirements – Part 1: General requirements*

IEC 80416-1, *Basic principles for graphical symbols for use on equipment — Part 1: Creation of graphical symbols for registration*

European Pharmacopoeia Assays – 2.5.30 Oxidising substances; Biological tests - 2.6.14 Bacterial endotoxins

United States Pharmacopeia Chemical tests <541> Titrimetry, Oxidation-Reduction (Redox) titrations; Biological tests <85> Bacterial endotoxins test

<https://standards.iteh.ai/catalog/standards/iso/20244107-f392-40b6-b1e3-cf9489759430/iso-fdis-15883-1>

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

— ISO Online browsing platform: available at <https://www.iso.org/obp>

— IEC Electropedia: available at <https://www.electropedia.org/>

3.1

A_0
measure of microbiological lethality delivered by a moist heat *disinfection* (3.14) process expressed in terms of the equivalent time in seconds at 80 °C with reference to a *microorganism* (3.33) with a z value of 10 K

Note 1 to entry: See [Annex B](#).

[SOURCE: ISO 11139:2018, 3.1, modified — Note 1 to entry has been added.]

1) Under preparation. Stage at time of publication: ISO/DIS 15883-2:2023.

2) Under preparation. Stage at time of publication: ISO/DIS 15883-7:2023

3.2

automatic controller

device that directs the equipment sequentially through required stages of the cycle in response to programmed cycle parameters

[SOURCE: ISO 11139:2018, 3.18]

3.3

bedpan washer disinfecter

washer-disinfecter (3.58) for human waste containers that additionally empties and flushes

[SOURCE: ISO 11139:2018, 3.22]

3.4

bioburden

population of viable *microorganisms* (3.33) on or in product and/or sterile barrier system

[SOURCE: ISO 11139:2018, 3.23]

3.5

calibration

operation that, under specified conditions, in a first step, establishes a relation between the quantity values with measurement uncertainties provided by measurement standards and corresponding indications with associated measurement uncertainties and, in a second step, uses this information to establish a relation for obtaining a measurement result from an indication

[SOURCE: ISO 11139:2018, 3.31]

3.6

calorifier

closed vessel, at a pressure greater than atmospheric, in which water is indirectly heated by the flow of heated *fluid* (3.21) through a heat exchanger

[SOURCE: ISO 11139:2018, 3.32]

3.7

chamber

part of equipment in which a *load* (3.28) is processed

Note 1 to entry: The chamber does not include steam generators, pipework, e.g. drain and fittings from which it can be isolated.

[SOURCE: ISO 11139:2018, 3.36, modified — Note 1 to entry has been added.]

3.8

chemical disinfection

disinfection (3.14) achieved by the action of one or more chemicals

[SOURCE: ISO 11139:2018, 3.42]

3.9

cleaning

removal of contaminants to the extent necessary for further processing or for intended use

[SOURCE: ISO 11139:2018, 3.46]

3.10

continuous process machine

equipment that moves one work unit at a time between each step of the process with the product generally remaining in motion

Note 1 to entry: This is contrasted with batch process equipment, which would expose the entire batch to each step of the process, one step at a time.

[SOURCE: ISO 11139:2018, 3.62]

**3.11
cycle complete**

message from the automatic controller that the operating cycle has ended successfully

[SOURCE: ISO 11139:2018, 3.71]

**3.12
D value**

D_{10} value

time or dose required under stated conditions to achieve inactivation of 90% of a population of the test *microorganisms* (3.33)

[SOURCE: ISO 11139:2018, 3.75]

**3.13
dead volume**

<washer-disinfector> enclosed space of pipework which is not purged by the usual flow of liquids during the *operating cycle* (3.36)

[SOURCE: ISO 11139:2018, 3.318.2]

**3.14
disinfection**

process to inactivate viable *microorganisms* (3.33) to a level previously specified as being appropriate for a defined purpose

[SOURCE: ISO 11139:2018, 3.84]

**3.15
disinfection temperature**

minimum temperature on which the evaluation of the *disinfection* (3.14) efficacy is based

Note 1 to entry: Several *disinfection* (3.14) temperatures can prevail during the *disinfection* (3.14) stage.

[SOURCE: ISO 11139:2018, 3.85, modified — Note 1 to entry has been added.]

3.16

disinfection time

period for which the *process variable(s)* (3.42) is/are maintained at or above that/those specified

Note 1 to entry: Examples of *process variables* (3.42) include temperature of the *load* (3.28), disinfectant concentration in the chamber.

[SOURCE: ISO 11139:2018, 3.86]

**3.17
double-ended**

having separate doors for loading and unloading in separate areas

[SOURCE: ISO 11139:2018, 3.92]

**3.18
endoscope washer-disinfector**

washer-disinfector (3.58) intended to clean and disinfect *loads* (3.28) comprising flexible endoscopes

[SOURCE: ISO 11139:2018, 3.100]

3.19

fail safe

attribute of equipment, or its associated services, that ensures that a malfunction will not give rise to a hazardous situation

[SOURCE: ISO 11139:2018, 3.115]

3.20

fault

situation in which one or more of the process or cycle parameters is/are outside its/their specified tolerance(s)

[SOURCE: ISO 11139:2018, 3.116]

3.21

fluid

substance that continually deforms (flows) under applied shear stress

EXAMPLE Liquid, gas, vapour, plasma.

[SOURCE: ISO 11139:2018, 3.120]

3.22

flushing

purging

removing by displacement with a *fluid* (3.21)

[SOURCE: ISO 11139:2018, 3.121]

3.23

free draining

allowing the unimpeded flow of liquids towards the discharge point under the influence of gravity

[SOURCE: ISO 11139:2018, 3.124]

3.24

holding time

period during which process parameters are maintained, within their specified tolerances

[SOURCE: ISO 11139:2018, 3.133]

3.25

human waste

body *fluids* (3.21) and excretions

EXAMPLE Faeces, urine, blood, pus, vomit, mucus.

[SOURCE: ISO 11139:2018, 3.134]

3.26

human waste container

vessel for holding and transporting *human waste* (3.25)

[SOURCE: ISO 11139:2018, 3.135]

3.27

installation qualification

IQ

process of establishing by objective evidence that all key aspects of the process equipment and ancillary system installation comply with the approved specification

[SOURCE: ISO 11139:2018, 3.220.2]

3.28

load

product, equipment, or materials to be processed together within an *operating cycle* (3.36)

[SOURCE: ISO 11139:2018, 3.155]

3.29

loading door

means of access through which a *load* (3.28) is passed into the chamber before processing

[SOURCE: ISO 11139:2018, 3.157]

3.30

lumen device

item that consists of tube(s) or pipe(s)

[SOURCE: ISO 11139:2018, 3.158]

3.31

medical device

instrument, apparatus, implement, machine, appliance, implant, reagent for in vitro use, or software material, or other similar or related article, intended by the manufacturer to be used, alone or in combination, for human beings, for one or more of the specific medical purpose(s) of:

- diagnosis, prevention, monitoring, treatment, or alleviation of disease;
- diagnosis, monitoring, treatment, alleviation of or compensation for an injury;
- investigation, replacement, modification or support of the anatomy or of a physiological process;
- supporting or sustaining life;
- control of conception;
- disinfection of medical devices;
- providing information by means of in vitro examination of specimens derived from the human body, and does not achieve its primary intended action by pharmacological, immunological or metabolic means, but which may be assisted in its intended function by such means

Note 1 to entry: Products which may be considered to be medical devices in some jurisdictions but not in others include:

- items specifically intended for cleaning or sterilization of medical devices;
- pouches, reel goods, sterilization wrap, and reusable containers for packaging of medical devices for sterilization;
- disinfection substances;
- aids for persons with disabilities;
- devices incorporating animal and/or human tissues;
- devices for in vitro fertilization or assisted reproduction technologies.

[SOURCE: ISO 11139:2018, 3.166]

3.32

microbial reduction factor

extent to which the *bioburden* (3.4) is reduced in tenfold increments

Note 1 to entry: It is expressed as \log_{10} .

[SOURCE: ISO 11139:2018, 3.174]