

FINAL
DRAFT

INTERNATIONAL
STANDARD

ISO/FDIS
2612

ISO/TC 193/SC 1

Secretariat: NEN

Voting begins on:
2023-08-24

Voting terminates on:
2023-10-19

Analysis of natural gas — Biomethane — Determination of ammonia content by tuneable diode laser absorption spectroscopy

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO 2612](#)

<https://standards.iteh.ai/catalog/standards/sist/ed9590d4-f425-4872-bb4b-a225eac73165/iso-2612>

ISO/CEN PARALLEL PROCESSING

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.



Reference number
ISO/FDIS 2612:2023(E)

© ISO 2023

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

ISO 2612

<https://standards.iteh.ai/catalog/standards/sist/ed9590d4-f425-4872-bb4b-a225eac73165/iso-2612>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2023

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword.....	v
Introduction.....	vi
1 Scope.....	1
2 Normative references.....	1
3 Terms and definitions.....	2
4 Safety precautions.....	3
5 Principle.....	3
6 Apparatus.....	3
6.1 NH ₃ analyser.....	3
6.2 Gas delivery system.....	3
6.3 Temperature sensor.....	4
6.4 Pressure sensor.....	4
6.5 Calibration equipment.....	4
6.5.1 General.....	4
6.5.2 Pressure regulators for the NH ₃ cylinders.....	4
6.5.3 Flow controller.....	4
6.5.4 Flow meter.....	4
6.5.5 Dilution device.....	4
6.5.6 Output manifold.....	5
7 Reagents and materials.....	5
7.1 Methane.....	5
7.2 Calibration gases.....	5
7.3 Inert gas.....	5
8 Sampling.....	5
8.1 General.....	5
8.2 Construction materials.....	5
8.3 Cleanliness.....	6
8.4 Installation of the calibration gas cylinder.....	6
8.5 Pressure control.....	6
8.6 Purging of reduction valve and transfer lines.....	6
8.7 Flow control.....	6
8.8 Leak control.....	7
9 Calibration.....	7
9.1 Calibration procedures.....	7
9.2 Frequency of calibration.....	7
9.2.1 Multipoint calibration.....	7
9.2.2 Zero and span point calibration.....	7
10 Interferences.....	7
10.1 Interfering absorbers.....	7
10.2 Matrix gas.....	7
10.3 Secondary level spectroscopic effects: Gas temperature, gas pressure, spatial homogeneity.....	8
10.4 Humidity and carbon dioxide.....	8
11 Measurement procedure.....	8
12 Expression of results.....	8
12.1 Quantities and units.....	8
12.2 Uncertainty.....	8
13 Test report.....	9

Annex A (informative) Spectroscopic analyser performance characteristics for NH₃ analysis in biomethane	10
Bibliography	11

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO 2612](https://standards.iteh.ai/catalog/standards/sist/ed9590d4-f425-4872-bb4b-a225eac73165/iso-2612)

<https://standards.iteh.ai/catalog/standards/sist/ed9590d4-f425-4872-bb4b-a225eac73165/iso-2612>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 193, *Natural gas*, Subcommittee SC 1, *Analysis of natural gas*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 408, *Natural gas and biomethane for use in transport and biomethane for injection in the natural gas grid*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Ammonia is a common trace impurity found in biomethane. It is a product from the anaerobic digestion of biomass, formed from the breakdown of amino acids during the production of biogas. As an impurity in biogas and biomethane, ammonia is corrosive when it dissolves in the presence of water, damaging process equipment and leads to an increase in anti-knock processes in combustion engines when used as a fuel for vehicles. Ammonia is detrimental to the environment and as an air pollutant forms particulates which are damaging to public health. Additionally, when present in the combustion of biomethane, ammonia can lead to the formation of nitrogen oxides (NO_x), which are regulated pollutants as they are toxic and affect air quality. Therefore, the presence of ammonia in biogas and biomethane is undesirable to gas distributors and their customers.

Measuring ammonia content in mixtures of methane at the trace level (i.e. mg m^{-3}) is technically difficult due to the adsorptive nature (i.e. “stickiness”) of ammonia. Particularly spectral NH_3 measurements can be severely hampered by spectral interferences from the matrix gas components, which further increases the complexity of these measurements. Measurements in biogas or biomethane are also dangerous due to the potentially explosive nature of methane, when mixed with an oxidizer like ambient air.

This method supports the implementation of specifications for biomethane and biogas such as EN 16723-1^[8] and EN 16723-2^[9] when used in the natural gas grids and when using it as a transportation fuel. Implementation of these specifications require fit-for-purpose measurement methods with known performance and acceptable metrological traceability to support the trade of renewable gases as well as conformity assessment. Currently, methods are referenced in standards such as EN 16723-1 which have not been validated for use with biomethane and biogas. This document describes measurement methods that meet these requirements and can be implemented by laboratories and industry, also those seeking accreditation on the basis of, e.g. ISO/IEC 17025.

The methods described are based on commercially available spectroscopic analysers, specific to the measurement of ammonia. They have been shown to perform at an acceptable level when quantifying the ammonia content of biomethane at the 10 mg m^{-3} level, as specified in, e.g. EN 16723-1.

[ISO 2612](https://standards.iteh.ai/catalog/standards/sist/ed9590d4-f425-4872-bb4b-a225eac73165/iso-2612)

<https://standards.iteh.ai/catalog/standards/sist/ed9590d4-f425-4872-bb4b-a225eac73165/iso-2612>