

ISO/TC 193/SC 1

Date: 2023-07-18

**ISO/FDIS 2612:2023(E)**

ISO/TC 193/SC 1/WG 25

Secretariat: NEN

Date: 2023-08-10

## Analysis of natural gas — Biomethane — Determination of ammonia content by tuneable diode laser absorption spectroscopy

*Analyse du gaz naturel — Biométhane — Détermination de la teneur en ammoniac par spectroscopie d'absorption laser à diode accordable*

(<https://standards.iteh.ai>)  
Document Preview

ISO 2612

<https://standards.iteh.ai/catalog/standards/sist/ed9590d4-f425-4872-bb4b-a225eac73165/iso-2612>

**FDIS stage**

© ISO 2023

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office

CP 401 • Ch. de Blandonnet 8 • CP 401

CH-1214 Vernier, Geneva, Switzerland

Tel. Phone: + 41 22 749 01 11

Fax + 41 22 749 09 47

E-mail: [copyright@iso.org](mailto:copyright@iso.org)

Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

# iTeh Standards (<https://standards.iteh.ai>) Document Preview

ISO 2612

<https://standards.iteh.ai/catalog/standards/sist/ed9590d4-f425-4872-bb4b-a225eac73165/iso-2612>

**Contents—Page**

Foreword .....	vi
Introduction.....	vii
1 Scope .....	1
2 Normative references.....	1
3 Terms and definitions .....	2
4 Safety precautions .....	3
5 Principle.....	3
6 Apparatus.....	4
6.1 NH <sub>3</sub> analyser.....	4
6.2 Gas delivery system .....	4
6.3 Temperature sensor .....	4
6.4 Pressure sensor .....	4
6.5 Calibration equipment.....	4
6.5.1 General.....	4
6.5.2 Pressure regulators for the NH <sub>3</sub> cylinders .....	4
6.5.3 Flow controller.....	5
6.5.4 Flow meter.....	5
6.5.5 Dilution device.....	5
6.5.6 Output manifold.....	5
7 Reagents and materials .....	5
7.1 Methane .....	5
7.2 Calibration gases .....	5
7.3 Inert gas .....	6
8 Sampling.....	6
8.1 General .....	6
8.2 Construction materials .....	6
8.3 Cleanliness.....	6
8.4 Installation of the calibration gas cylinder.....	6
8.5 Pressure control .....	6
8.6 Purging of reduction valve and transfer lines.....	7
8.7 Flow control .....	7
8.8 Leak control.....	7
9 Calibration .....	7
9.1 Calibration procedures.....	7

9.2	Frequency of calibration .....	7
9.2.1	Multipoint calibration .....	7
9.2.2	Zero and span point calibration .....	8
10	Interferences .....	8
10.1	Interfering absorbers .....	8
10.2	Matrix gas .....	8
10.3	Secondary level spectroscopic effects: Gas temperature, gas pressure, spatial homogeneity .....	8
10.4	Humidity and carbon dioxide .....	8
11	Measurement procedure .....	9
12	Expression of results .....	9
12.1	Quantities and units .....	9
12.2	Uncertainty .....	9
13	Test report .....	9
Annex A (informative)	Spectroscopic analyser performance characteristics for NH <sub>3</sub> analysis in biomethane .....	11
Bibliography	.....	12

~~Foreword~~ — v

~~Introduction~~ — vi

~~1~~ — ~~Scope~~ — 1

~~2~~ — ~~Normative references~~ — 1

~~3~~ — ~~Terms and definitions~~ — 2

~~4~~ — ~~Safety Precautions~~ — 3

~~5~~ — ~~Principle~~ — 3

~~6~~ — ~~Apparatus~~ — 4

~~6.1~~ — ~~NH<sub>3</sub> analyser~~ — 4

~~6.2~~ — ~~Gas Delivery System~~ — 4

~~6.3~~ — ~~Temperature sensor~~ — 4

~~6.4~~ — ~~Pressure sensor~~ — 4

~~6.5~~ — ~~Calibration equipment~~ — 4

~~6.5.1~~ — ~~General~~ — 4

~~6.5.2~~ — ~~Pressure regulators for the NH<sub>3</sub> cylinders~~ — 4

~~6.5.3~~ — ~~Flow controller~~ — 5

~~6.5.4~~ — ~~Flow meter~~ — 5

~~6.5.5~~ — ~~Dilution device~~ — 5

~~6.5.6~~ — ~~Output manifold~~ — 5

iv

iv

iTeh Standards  
(<https://standards.itih.ai>)  
Document Preview

ISO 2612

<https://standards.itih.ai/standards/sist/ed9590d4-f425-4872-bb4b-a225eac73165/iso-2612>

<del>7</del>	<del>Reagents and materials</del>	<del>5</del>
<del>7.1</del>	<del>Methane</del>	<del>5</del>
<del>7.2</del>	<del>Calibration gases</del>	<del>5</del>
<del>7.3</del>	<del>Inert Gas</del>	<del>6</del>
<del>8</del>	<del>Sampling</del>	<del>6</del>
<del>8.1</del>	<del>General</del>	<del>6</del>
<del>8.2</del>	<del>Construction materials</del>	<del>6</del>
<del>8.3</del>	<del>Cleanliness</del>	<del>6</del>
<del>8.4</del>	<del>Installation of the calibration gas cylinder</del>	<del>6</del>
<del>8.5</del>	<del>Pressure control</del>	<del>7</del>
<del>8.6</del>	<del>Purging of reduction valve and transfer lines</del>	<del>7</del>
<del>8.7</del>	<del>Flow control</del>	<del>7</del>
<del>8.8</del>	<del>Leak control</del>	<del>7</del>
<del>9</del>	<del>Calibration</del>	<del>7</del>
<del>9.1</del>	<del>Calibration procedures</del>	<del>7</del>
<del>9.2</del>	<del>Frequency of calibration</del>	<del>8</del>
<del>9.2.1</del>	<del>Multipoint calibration</del>	<del>8</del>
<del>9.2.2</del>	<del>Zero and span point calibration</del>	<del>8</del>
<del>10</del>	<del>Interferences</del>	<del>8</del>
<del>10.1</del>	<del>Interfering Absorbers</del>	<del>8</del>
<del>10.2</del>	<del>Matrix Gas</del>	<del>8</del>
<del>10.3</del>	<del>Secondary level spectroscopic effects: Gas temperature, gas pressure, spatial homogeneity</del>	<del>8</del>
<del>10.4</del>	<del>Humidity and carbon dioxide</del>	<del>9</del>
<del>11</del>	<del>Measurement procedure</del>	<del>9</del>
<del>12</del>	<del>Expression of results</del>	<del>9</del>
<del>13</del>	<del>Test report</del>	<del>9</del>
<del>Annex A (informative)</del>	<del>Spectroscopic Analyser Performance Characteristics for NH<sub>3</sub> Analysis in Biomethane</del>	<del>10</del>
<del>Bibliography</del>	<del></del>	<del>11</del>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO ~~documents~~document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

~~Attention is drawn~~ISO draws attention to the possibility that ~~some of the~~elementsimplementation of this document may ~~be involve~~the subjectuse of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at [www.iso.org/patents](http://www.iso.org/patents). ISO shall not be held responsible for identifying any or all such patent rights. ~~Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see -).~~

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation ~~of~~ the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html)the following URL: <https://www.iso.org/iso/2612>

This document was prepared by Technical Committee ISO/TC 193, *Natural gas*, Subcommittee SC 1, *Analysis of natural gas*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 408, *Natural gas and biomethane for use in transport and biomethane for injection in the natural gas grid*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

## Introduction

Ammonia is a common trace impurity found in biomethane. It is a product from the anaerobic digestion of biomass, formed from the breakdown of amino acids during the production of biogas. As an impurity in biogas and biomethane, ammonia is corrosive when it dissolves in the presence of water, damaging process equipment and leads to an increase in anti-knock processes in combustion engines when used as a fuel for vehicles. Ammonia is detrimental to the environment and as an air pollutant forms particulates which are damaging to public health. Additionally, when present in the combustion of biomethane, ammonia can lead to the formation of nitrogen oxides (NO<sub>x</sub>), which are regulated pollutants as they are toxic and affect air quality. Therefore, the presence of ammonia in biogas and biomethane is undesirable to gas distributors and their customers.

Measuring ammonia content in mixtures of methane at the trace level (i.e. mg m<sup>-3</sup>) is technically difficult due to the adsorptive nature (i.e. “stickiness”) of ammonia. Particularly spectral NH<sub>3</sub> measurements can be severely hampered by spectral interferences from the matrix gas components, which further increases the complexity of these measurements. Measurements in biogas or biomethane are also dangerous due to the potentially explosive nature of methane, when mixed with an oxidizer like ambient air.

This method supports the implementation of specifications for biomethane and biogas such as EN 16723-1 [9][1] and EN 16723-2 [9][2] when used in the natural gas grids and when using it as a transportation fuel. Implementation of these specifications require fit-for-purpose measurement methods with known performance and acceptable metrological traceability to support the trade of renewable gases as well as conformity assessment. Currently, methods are referenced in standards such as EN 16723-1 which have not been validated for use with biomethane and biogas. This document describes measurement methods that meet these requirements and can be implemented by laboratories and industry, also those seeking accreditation on the basis of, e.g. ISO/IEC 17025.

The methods described are based on commercially available spectroscopic analysers, specific to the measurement of ammonia. They have been shown to perform at an acceptable level when quantifying the ammonia content of biomethane at the 10 mg m<sup>-3</sup> level, as specified in, e.g. EN 16723-1.

[ISO 2612](https://standards.iteh.ai/ISO/2612)

<https://standards.iteh.ai/catalog/standards/sist/ed9590d4-f425-4872-bb4b-a225eac73165/iso-2612>





# Analysis of natural gas — Biomethane — Determination of ammonia content by tuneable diode laser absorption spectroscopy

**WARNING** — The use of this document can involve hazardous materials, operations and equipment. This document does not aim to address all of the safety problems associated with the materials specified. It is the responsibility of the user of this document to establish appropriate safety and health practices and to determine the applicability of any other restrictions prior to use.

## 1 Scope

This document describes several test methods for measuring the ammonia amount fraction in natural gas and biomethane at the trace level ( $\mu\text{mol mol}^{-1}$ ). The suitable handling and sampling of pressurised mixtures of ammonia in methane that are applied to several different ammonia measurement systems are described. The measurement systems ~~comprise~~are comprised of readily available commercial spectroscopic analysers that are specific to ammonia. These  $\text{NH}_3$  analysers are considered as a black box in terms of their operation, which is dependent on the instructions of the manufacturer. The document describes suitable calibration and measurement strategies to quantify ammonia in (bio)methane around and above the  $10 \text{ mg m}^{-3}$  ( $14 \mu\text{mol mol}^{-1}$ ) level and applies to analysis within absolute pressure ranges of 1 bar – 2 bar, temperatures of  $0 \text{ }^\circ\text{C}$  –  $40 \text{ }^\circ\text{C}$  and relative humidity  $<90 \%$ .

References are also made to additional standards that are applied either to natural gas analysis or air quality measurements. In this document the matrix gas is always methane or biomethane and the measurand is the amount fraction  $\text{NH}_3$ .

NOTE 1 bar =  $0,1 \text{ MPa} = 10^5 \text{ Pa}$ ;  $1 \text{ MPa} = 1 \text{ N/mm}^2$

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 6143, *Gas analysis — Comparison methods for determining and checking the composition of calibration gas mixtures*

ISO 6145-1, *Gas analysis — Preparation of calibration gas mixtures using dynamic methods — Part 1: General aspects*

ISO 7504, *Gas analysis — Vocabulary*

ISO 9169, *Air quality — Definition and determination of performance characteristics of an automatic measuring system*

ISO 10715, *Natural gas — Gas sampling*

ISO 10723, *Natural gas — Performance evaluation for analytical systems*

ISO 14912, *Gas analysis — Conversion of gas mixture composition data*

ISO 14532, *Natural gas — Vocabulary*

## ISO/FDIS 2612:2023(E)

ISO 16664, *Gas analysis — Handling of calibration gases and gas mixtures — Guidelines*

IEC 61207-7, *Expression of performance of gas analyzers — Part 7: Tuneable semiconductor laser gas analyzers*

### 3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 6143, ISO 7504, ISO 61439169, ISO 10715, ISO 10723, ISO 14532, IEC 61207-7, ISO 9169 and the following apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

#### 3.1

##### **diode laser**

semiconductor laser which is formed from a p-n junction and powered by injected electric current

[SOURCE: IEC 61207-7:2013, 3.2]

#### 3.2

##### **ultraviolet-visible spectroscopy**

UV/vis

spectroscopy of radiation that consists of electromagnetic radiation with wavelengths in the ultraviolet and/or visible regions

[SOURCE: ISO/TS 80004-6:2021<sup>[3]</sup>, 5.6, modified — The second term was originally "UV-Vis spectroscopy".]

#### 3.3

##### **span gas**

gas or gas mixture used to adjust and check the *span point* (3.4(3.4)) on the response line of the measuring system

Note 1 to entry: This amount fraction is often chosen around 70 % to 90 % of full scale.

[SOURCE: ISO 12039<sup>[4]</sup>:2019, 3.15]

#### 3.4

##### **span point**

value of the output quantity (measured signal) of the automatic measuring system for the purpose of calibration, adjustment, etc. that represents a correct measured value generated by reference material

[SOURCE: ISO 13199<sup>[5]</sup>:2012, 3.14]

#### 3.5

##### **performance characteristic**

one of the quantities assigned to the analytical instrument in order to define its performance