

ISO/~~DIS~~**FDIS** 20304-2:~~2023(E)~~

~~2023-05-10~~

ISO/TC 281/~~AWG 3~~

Secretariat: JISC

Date: 2023-07-25

Fine bubble technology — Water treatment applications —

**Part 2:
Test methods using *Escherichia coli* as a test micro-organism**

iTeh STANDARD PREVIEW
(standards.iteh.ai)

FDIS stage

ISO FDIS 20304-2
<https://standards.iteh.ai/standards/sist/7884e481-3f65-494c-b376-4e16a002ef20/iso-fdis-20304-2>

ISO/~~DIS~~FDIS 20304-2:2023(E)

© ISO ~~2022~~2023

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
~~Email~~E-mail: copyright@iso.org
Website: www.iso.org

Published in Switzerland

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO/FDIS 20304-2

<https://standards.iteh.ai/catalog/standards/sist/7884e481-3665-494c-b376-4e16a002ef20/iso-fdis-20304-2>

Contents

Foreword.....	v
Introduction.....	vi
Part 2: Test methods using <i>Escherichia coli</i> as a test micro-organism.....	1
1 Scope.....	1
2 Normative references.....	1
3 Terms and definitions.....	1
4 Safety precautions.....	2
5 Requirements.....	2
6 Test methods.....	2
6.1 Principle.....	2
6.2 Test facility.....	2
6.3 Culture media and reagents.....	2
6.3.1 General.....	2
6.3.2 Water.....	3
6.3.3 Tryptone soya agar (TSA).....	3
6.3.4 Tryptone soya broth (TSB).....	3
6.3.5 Lysogeny broth (LB).....	3
6.3.6 Cryoprotective solution for bacterial species.....	3
6.3.7 Physiological saline.....	4
6.3.8 Tryptone sodium chloride solution (TSCS).....	4
6.4 Laboratory apparatus and glassware.....	4
6.5 Preparation of bacterial test suspension.....	5
6.5.1 Storage of strains.....	5
6.5.2 Working culture of test bacteria.....	5
6.5.3 Bacterial test suspension.....	6
6.6 Procedure.....	6
6.6.1 Choice of experimental conditions.....	7
6.6.2 Test procedure for antibacterial activity.....	7
6.7 Calculation and expression of results.....	8
Annex A (normative) Strain numbers.....	9
A.1 General.....	9
A.2 List of bacteria.....	9
Annex B (normative) Shaking method.....	10
B.1 General.....	10
B.2 Shaking by mixer.....	10
B.3 Shaking by hand.....	10
B.4 Shaking by homogenizer.....	10
Annex C (normative) Luminescence photometers and RLU.....	11

C.1	General	11
C.2	Principle of luminescence photometers	11
C.3	RLU	11
Annex D (informative)	Test report of interlaboratory tests for disinfection efficiency of fine bubbles ..	12
D.1	Test report	12
D.2	Interlaboratory tests and the results	13
D.2.1	Fine bubbles generator system and test facility (see Figure 1)	13
D.2.2	Test report-1	14
D.2.3	Test report-2	15
D.2.4	Test report-3	17
D.2.5	Conclusion of evaluation	18
Bibliography	24

Foreword — iv

Introduction — v

1 Scope — 1

2 Normative references — 1

3 Terms and definitions — 1

4 Safety precautions — 2

5 Requirements — 2

6 Test methods — 2

Annex A (informative) Strain numbers — 9 [iteh.ai/catalog/standards/sist/7884e481-3665-494c-b376-](https://standards.iteh.ai/catalog/standards/sist/7884e481-3665-494c-b376-4e16a002ef20/iso-fdis-20304-2)

Annex B (normative) Shaking method — 10 [4e16a002ef20/iso-fdis-20304-2](https://standards.iteh.ai/catalog/standards/sist/7884e481-3665-494c-b376-4e16a002ef20/iso-fdis-20304-2)

Annex C (normative) Luminescence photometers and RLU — 11

Annex D Test report of interlaboratory tests for disinfection efficiency of fine bubbles — 12

Bibliography — 21

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO/FDIS 20304-2

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of a patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 281, Fine bubble technology.

A list of all parts in the ISO 20304 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

In the last decade, fine bubble technology generated by hydrodynamic cavitation methods has been applied as part of water treatment facilities for water disinfection purposes. However, absence of ~~international standards~~International Standards in relation to the efficiency of water disinfectant properties on fine bubbles has been an obstacle to trading and technology transfer. This document describes a method for the determination of water disinfection efficiency of fine bubbles generated by a hydrodynamic cavitation.

Fine bubbles act as a disinfectant through chemical (i.e. generation of OH radicals) and physical mechanisms caused by presence of shock waves, pressure gradients and shear forces. The function as a disinfectant of fine bubbles was also found in that fine bubbles provide a more effective means for cleaning and disinfecting both, the bath and the reservoir, than traditional ultrasonic vibrator.^[1]

Properties of fine bubbles and their disinfection mechanisms differ according to applied generation processes, bubble size distributions and bubble densities. However existing studies showed that fine bubbles generated have a significant water disinfection effect. Particularly, high deactivation efficiency of *Escherichia coli* (*E. coli*) has been achieved in water disinfection by microbubbles generated using hydrodynamic cavitation as well.^{[2,3 and 4], [2]-[4]}

Generation of highly reactive free radicals and turbulence associated with collapsing micro bubbles provides great potential for water disinfection. The effect of ozone microbubble on *E. coli* has often been found more effective and with faster disinfection kinetics of *E. coli*, when the ozone gas is activated as the micro or ultrafine bubble forms.^{[5 and 6], [5], [6]}

The test methods on the antibacterial activity of textile products have been standardized by ISO/TC 38, *textiles*. The present standard operation procedure (SOP) for the evaluation of disinfection efficiency is based on the quantitative measurement by ATP luminescence method for the evaluation of antibacterial activity.

Therefore, under the test conditions ~~which are configured by ISO/TC 281~~, establishment of ~~international standards~~International Standards to measure the water disinfection efficiency of fine bubbles is essential to promote the relevant trading technology transfer.

Fine bubble technology — Water treatment applications — ~~Part 2:~~ Test methods using *Escherichia coli* as a test micro-organism

Part 2: Test methods using *Escherichia coli* as a test micro-organism

1 Scope

This document specifies a test method for assessing bactericidal viability of *Escherichia coli* as a test micro-organism, in dispersions of various fine bubbles generated by the hydrodynamic cavitation of water medium.

2 Normative references

~~The following document is referred to in the text in such a way that some or all of their content constitutes requirements of this document. For undated references, the latest edition of the referenced document (including any amendments) applies:~~

~~ISO/IEC Directives, Part 2. www.iso.org/directives~~

~~There are no normative references in this document.~~

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ~~— IEC Electropedia: available at <https://www.electropedia.org/>~~
- ~~— ISO Online browsing platform: available at <https://www.iso.org/obp>~~
- ~~— IEC Electropedia: available at <https://www.electropedia.org/>~~

3.1 hydrodynamic cavitation

fine bubble generation method used by physical hydrodynamic structures and pressures

3.2 disinfection efficiency

ratio of viable cell count to the initial number of the test bacteria

3.3 control suspension

suspension used to validate the growth condition of test bacteria and validate the test, i.e. bacterial test suspension produced without using fine bubble generation

3.4 test solution

solution samples obtained for the analysis of antibacterial activity after operating the test facility

3.5 test gas

gas mixed in the middle of operating the cavitation unit

3.6
relative light unit
RLU

unit used for adenosine triphosphate (ATP) luminescence measurement

4 Safety precautions

The test methods specified in this document ~~requires~~require the use of bacteria and conditions that promote bacterial growth. Since the bacteria can be pathogenic, the tests should be carried out by persons with training and experience in the use of microbiological techniques. The laboratory facilities should be designated as appropriate biosafety level ~~1~~², [7].

Appropriate safety precautions should be observed with due consideration given to country-specific.

5 Requirements

This document is to provide the measurement method of disinfection efficiency for fine bubbles enhanced waters by test hydrodynamic cavitation units. The disinfection efficiency is measured by the test fine bubble solution generated by each test hydrodynamic cavitation unit.

The antibacterial (bactericidal) activity shall be evaluated using the following test organism: *E. coli* (see ~~Table A.1~~Table A.1). The test tank and test facility should be cleaned by appropriate methods like triple-wash with sterilized water after chlorination before the test.

6 Test methods

6.1 Principle

The test tank is filled with a test suspension of bacteria and the test hydrodynamic cavitation (HC) unit is operated under the required initial test condition (~~15~~approximately 15 °C to 24 °C ± 1 °C). The fine bubbles-enriched water will be contacted to the test suspension of bacteria in the test tank (see ~~Figure 1~~Figure 1).

At a specific contact time, an aliquot is taken; the bactericidal action in this portion is immediately measured using the luminescence photometer (see ~~6.4.8~~Firstly 6.4.8). First, the relationship between the measurement unit of the luminescence photometer, RLU (relative light unit, RLU) (see ~~Annex C~~Annex C) and counted CFU (colonies forming unit, CFU) of the initial density of the test microorganism is measured to check the initial cell density of the bacterial test suspension.

The test is performed using *E. coli* (see ~~Table 3~~Table A.1).

6.2 Test facility

Fine bubbles are generated by circulating the liquid in a fine bubble generating system with the following characteristics: a 125-litre tank, a 0,4-kW centrifugal multistage pump made of stainless steel, PVC pipes and a cavitation unit with the test fine bubble generating unit (see ~~Figure 1~~Figure 1). Test gas can be chosen for the appropriate test purposes.

6.3 Culture media and reagents

6.3.1 General

The reagents shall be of analytical grade and/or appropriate for microbiological purposes.

NOTE—To improve reproducibility, dehydrated material should be used for the preparation of culture media. The manufacturer's instructions relating to the preparation of these products should be rigorously followed.

6.3.2 Water

Water used in tests shall be analytical-grade water for microbiological media preparation, which is freshly distilled and/or ion-exchanged and/or ultra-filtered and/or filtered with ~~RO~~ (reverse osmosis (RO)). It shall be free from substances that are toxic or inhibiting to the bacteria.

6.3.3 Tryptone Soya Agar soya agar (TSA)

TSA used in the test is for maintenance of bacterial strains and performance of viable counts.

Tryptone, pancreatic digest of casein	15,0-g
Soya peptone, papaic digest of soybean meal	5,0-g
NaCl	5,0-g
Agar	5,0-g
Water (see 6.3.2) 6.3.2)	to 1,0-l

Sterilize in the autoclave (see 6.4.1) (6.4.1). After sterilization, the pH of the medium shall be $7,2 \pm 0,2$ when measured at 20 °C.

6.3.4 Tryptone soya broth (TSB)

Tryptone, pancreatic digest of casein	17-g
Soya peptone, papaic digest of soya	3-g
Sodium chloride (NaCl)	5-g
Glucose	2,5-g
Dipotassium hydrogen phosphate	2,5-g
Water	1 000-ml

Mix well and adjust pH, $7,2 \pm 0,2$, then sterilize by autoclave (6.4.1) (6.4.1).

6.3.5 Lysogeny broth (LB)

Tryptone, pancreatic digest of casein	10-g
Yeast Yeast extract	5-g
NaCl	10-g
Water	1 000-ml

Mix well and adjust pH, $6,9 \pm 0,2$, then sterilize by autoclave (6.4.1) (6.4.1).

6.3.6 Cryoprotective solution for bacterial species

For freezing, a cryoprotective solution containing 150-g/l of glycerol or 100-g/l of dimethylsulfoxide shall be used and prepared as follows:

TSB (see 6.3.4) 6.3.4) or LB (see 6.3.5) 6.3.5):	1 000-ml
Add Glycerol: 150-g or dimethylsulfoxide:	100-g

Mix well and sterilize by autoclave (6.4.1) (6.4.1).

For solutions containing glycerol, sterilize the mixed solution by autoclave (6.4.1) (6.4.1). For solutions containing dimethylsulfoxide, sterilize the mixed solution by using a 0,22-µm membrane filter.

NOTE Any commercially available product can be used as long as it is a cryoprotective solution or preserving system that contains glycerol or dimethylsulfoxide.

ISO/FDIS 20304-2:2023(E)

6.3.7 Physiological saline

Sodium chloride (NaCl)	8,5-g
Water	1 000-ml
Mix well, then sterilize by autoclave (6.4.1) (6.4.1).	

6.3.8 Tryptone ~~Sodium Chloride Solution~~sodium chloride solution (TSCS)

Tryptone, pancreatic digest of casein	1-g
NaCl	8,5-g
Water	1 000-ml

6.4 Laboratory apparatus and glassware

The usual laboratory apparatus and, in particular, the following shall be used.

6.4.1 Autoclave, capable of sterilizing at $121\text{ }^{\circ}\text{C} \pm 2\text{ }^{\circ}\text{C}$ and $103\text{-kPa} \pm 5\text{-kPa}$.

6.4.2 Incubator, capable of being controlled at either $36\text{ }^{\circ}\text{C} \pm 1\text{ }^{\circ}\text{C}$ or $37\text{ }^{\circ}\text{C} \pm 1\text{ }^{\circ}\text{C}$. An incubator at $37\text{ }^{\circ}\text{C} \pm 1\text{ }^{\circ}\text{C}$ may be used if an incubator at $36\text{ }^{\circ}\text{C} \pm 1\text{ }^{\circ}\text{C}$ is not available.

6.4.3 pH meter, having an accuracy $\pm 0,1$ in calibration on pH units at $25\text{ }^{\circ}\text{C}$.

6.4.4 Stopwatch.

6.4.5 Container, test tubes or flasks of suitable capacity.

6.4.6 Pipettes, having the most suitable volume for each use, with a tip made of glass or plastic, and with a tolerance of 0,5-% or less.

6.4.7 Petri dishes, that have been sterilized, made of glass or plastic, in diameter sizes of 90 mm to 100 mm.

6.4.8 Luminescence photometer, with ATP kit for water test, capable of measuring ATP of 10^{-12} mol/l to 10^{-7} mol/l at 300-nm to 650-nm with a luminescence-measuring reagent.

6.4.9 Vials, 250-ml plastic bottles, with screw openings, polytetrafluoroethylene or silicone packing and caps made of polypropylene, polycarbonate or another suitable material.

6.4.10 Glass beads, with a diameter of 3-mm to 4-mm.

6.4.11 Freezers, one adjustable to a temperature below $-70\text{ }^{\circ}\text{C}$ and another to a temperature below $-20\text{ }^{\circ}\text{C}$.

6.4.12 Homogenizer, capable of speeds of ~~six~~ blows per second to ~~eight~~ blows per second, with the corresponding disposable containers.

6.4.13 Disposable plastic bags, sterile bags suitable for containing food products, to be used for one of the shaking methods of the specimens.

6.4.14 Sterilized stirring rod, with a 6-mm (± 2 -mm) diameter and a 300-mm (± 50 -mm) length.

6.6.5.6.5 Preparation of bacterial test suspension

6.6.3.6.5.1 Storage of strains

6.6.3.16.5.1.1 General

The strains shall be stored in accordance with the supplier's recommendations. The experiment instruments should be sterilized by the autoclave (6.4.1)(6.4.1) for a minimum of 15_min.

6.6.3.26.5.1.2 Preparation method

Obtain a sample of the freeze-dried bacterial strain in accordance with Annex A, Annex A, following the recommendations supplied with the culture and resuspend it in 5-ml of TSB (see 6.3.4)-6.3.4). Obtain a sample of the suspension and isolate it in a plastic bottle (6.4.10)(6.4.10) containing LB (see 6.3.5)-6.3.5). Incubate the cultures for 18-h to 24-h at 37 °C ± 2 °C.

After incubation, use the culture isolated in the Petri dish to verify the purity of the strain.

After verification, prepare the stock cultures to check the density of *E. coli* of the bacterial test suspension.

Sample 0,7-ml of the broth culture and spread it over the surface of the Petri dish containing the TSA. Incubate the culture on plates for 18-h to 24-h at 37-°C± ± 2-°C.

Add 10-ml of cryoprotective solution (see 6.3.6)(6.3.6) to the surface of the TSA plate culture and resuspend the cells in the solution using a sterile glass spreader. Sample the suspended cells from the surface of the agar, dilute them in 100 ml of cryoprotective solution and incubate for 30-min at 20 °C.

Using a pipette (6.4.6)(6.4.6), sample 1-ml of the suspension and transfer it to a cryogenic vial (6.4.9)(6.4.9) containing the beads (6.4.10)(6.4.10). Shake the vial in order to spread the suspended cells around the beads.

- —Where a cryoprotective solution containing dimethylsulfoxide is used, do not let it stand longer than 1 min at ambient temperature.
- —Where a cryoprotective solution containing glycerol is used, let it stand for 30_min at 20 °C.
- —Withdraw the excess cryoprotective solution with a sterile pipette. Place the cryogenic vials in a freezer (6.4.11)(6.4.11) set at -70 °C or lower.

Prepare 10⁻⁶ and 10⁻⁷ dilutions of the suspension using the serial dilution method. Take a 1,0-ml sample of each dilution and transfer it to separate Petri dishes. Add 12-ml to 15-ml of nutritive solution, cooled down to 45 °C ± 1 °C. Incubate for 18-h to 24-h under the conditions specified for the strain. Enumerate the plate cultures and confirm that the cell density of the suspension is 1 × 10⁵ CFU/ml to 5 × 10⁸ CFU/ml using TSCS (see 6.3.8)-6.3.8).

The other preparation details of initial bacteria stock culture are described in EN 1040^[91], [91]

6.6.4.6.5.2 Working culture of test bacteria

In order to prepare the working culture of strains, subculture from the stock culture by inoculation into LB broth and incubates. After 18-h to 24-h, prepare a second subculture from the first subculture in the same way and incubate for 18-h to 24-h. From this second subculture, a third subculture may be produced in the same way.

NOTE The second and/or third subculture are the working culture(s).

If it is not possible to prepare the second subculture on the same day, a 48 h subculture may be used for subsequent sub-culturing, provided that the subculture has been kept in the incubator during a 48 h period. In these circumstances, prepare a further 24 h subculture after proceeding. Do not take a fourth subculture.

6.6.5.3 Bacterial test suspension


Take the working cultures and transfer into the test tank and adjust the final volume of the bacterial test suspension to 50_l, after it has been diluted with water (see 6.3.2) at room temperature and stir with a stirring rod for 30_s.

Prepare 10⁻¹ and 10⁻² dilutions of the suspension using the serial dilution method. Take a 1,0-ml sample of each dilution and transfer it to separate Petri dishes. Add 12-ml to 15-ml of nutritive solution, cooled down to 45 °C ± 1 °C. Incubate for 18_h to 24_h under the conditions specified for the strain. Enumerate the plate cultures and confirm that the suspension contains more than 1 × 10⁵-CFU/ml. Measure RLU of the sample using luminescence photometer (6.4.8). The principle of RLU shall be as defined in Annex C.


Adjust the final concentration of E. coli, which is more than 1 × 10⁵-CFU/ml using luminescence photometer (6.4.8) and working cultured suspension.

6.6.6 Procedure

The capacity of the pump can be appropriately fixed to circulate the bacterial test suspension within the test facility (see Figure 1). Test fine bubble generating unit shall be operated to produce fine bubbles (see Figure 1). Bacterial test suspension shall be moved to the test tank within 1 h in accordance with Annex B. The produced fine bubble-enhanced water shall be contacted the bacterial test suspension during the operation of the test facility.



- Key**
- 1 test tank
 - 2 pump
 - 3 control panel
 - 4 gas injector



Inserted Cells