

FINAL
DRAFT

INTERNATIONAL
STANDARD

ISO/FDIS
20304-2

ISO/TC 281

Secretariat: JISC

Voting begins on:
2023-09-05

Voting terminates on:
2023-10-31

Fine bubble technology — Water treatment applications —

Part 2: Test methods using *Escherichia coli* as a test micro-organism

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[ISO/FDIS 20304-2](https://standards.iteh.ai/catalog/standards/sist/7884e481-3665-494c-b376-4e16a002ef20/iso-fdis-20304-2)

<https://standards.iteh.ai/catalog/standards/sist/7884e481-3665-494c-b376-4e16a002ef20/iso-fdis-20304-2>

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.



Reference number
ISO/FDIS 20304-2:2023(E)

© ISO 2023

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO/FDIS 20304-2

<https://standards.iteh.ai/catalog/standards/sist/7884e481-3665-494c-b376-4e16a002ef20/iso-fdis-20304-2>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2023

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword.....	iv
Introduction.....	v
1 Scope.....	1
2 Normative references.....	1
3 Terms and definitions.....	1
4 Safety precautions.....	2
5 Requirements.....	2
6 Test methods.....	2
6.1 Principle.....	2
6.2 Test facility.....	2
6.3 Culture media and reagents.....	2
6.3.1 General.....	2
6.3.2 Water.....	3
6.3.3 Tryptone soya agar (TSA).....	3
6.3.4 Tryptone soya broth (TSB).....	3
6.3.5 Lysogeny broth (LB).....	3
6.3.6 Cryoprotective solution for bacterial species.....	3
6.3.7 Physiological saline.....	4
6.3.8 Tryptone sodium chloride solution (TSCS).....	4
6.4 Laboratory apparatus and glassware.....	4
6.5 Preparation of bacterial test suspension.....	5
6.5.1 Storage of strains.....	5
6.5.2 Working culture of test bacteria.....	5
6.5.3 Bacterial test suspension.....	6
6.6 Procedure.....	6
6.6.1 Choice of experimental conditions.....	7
6.6.2 Test procedure for antibacterial activity.....	8
6.7 Calculation and expression of results.....	9
Annex A (normative) Strain numbers.....	10
Annex B (normative) Shaking method.....	11
Annex C (normative) Luminescence photometers and RLU.....	12
Annex D (informative) Test report of interlaboratory tests for disinfection efficiency of fine bubbles.....	13
Bibliography.....	21

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 281, *Fine bubble technology*.

A list of all parts in the ISO 20304 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

In the last decade, fine bubble technology generated by hydrodynamic cavitation methods has been applied as part of water treatment facilities for water disinfection purposes. However, absence of International Standards in relation to the efficiency of water disinfectant properties on fine bubbles has been an obstacle to trading and technology transfer. This document describes a method for the determination of water disinfection efficiency of fine bubbles generated by a hydrodynamic cavitation.

Fine bubbles act as a disinfectant through chemical (i.e. generation of OH radicals) and physical mechanisms caused by presence of shock waves, pressure gradients and shear forces. The function as a disinfectant of fine bubbles was also found in that fine bubbles provide a more effective means for cleaning and disinfecting both, the bath and the reservoir, than traditional ultrasonic vibrator.^[1]

Properties of fine bubbles and their disinfection mechanisms differ according to applied generation processes, bubble size distributions and bubble densities. However existing studies showed that fine bubbles generated have a significant water disinfection effect. Particularly, high deactivation efficiency of *Escherichia coli* (*E. coli*) has been achieved in water disinfection by microbubbles generated using hydrodynamic cavitation as well.^{[2]-[4]}

Generation of highly reactive free radicals and turbulence associated with collapsing micro bubbles provides great potential for water disinfection. The effect of ozone microbubble on *E. coli* has often been found more effective and with faster disinfection kinetics of *E. coli*, when the ozone gas is activated as the micro or ultrafine bubble forms.^{[5],[6]}

The test methods on the antibacterial activity of textile products have been standardized by ISO/TC 38, *textiles*. The present standard operation procedure (SOP) for the evaluation of disinfection efficiency is based on the quantitative measurement by ATP luminescence method for the evaluation of antibacterial activity.

Therefore, under the test conditions, establishment of International Standards to measure the water disinfection efficiency of fine bubbles is essential to promote the relevant trading technology transfer.

Fine bubble technology — Water treatment applications —

Part 2:

Test methods using *Escherichia coli* as a test micro-organism

1 Scope

This document specifies a test method for assessing bactericidal viability of *Escherichia coli* as a test micro-organism, in dispersions of various fine bubbles generated by the hydrodynamic cavitation of water medium.

2 Normative references

There are no normative references in this document.

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

3.1

hydrodynamic cavitation

fine bubble generation method used by physical hydrodynamic structures and pressures

3.2

disinfection efficiency

ratio of viable cell count to the initial number of the test bacteria

3.3

control suspension

suspension used to validate the growth condition of test bacteria and validate the test, i.e. bacterial test suspension produced without using fine bubble generation

3.4

test solution

solution samples obtained for the analysis of antibacterial activity after operating the test facility

3.5

test gas

gas mixed in the middle of operating the cavitation unit

3.6

relative light unit

RLU

unit used for adenosine triphosphate (ATP) luminescence measurement

4 Safety precautions

The test methods specified in this document require the use of bacteria and conditions that promote bacterial growth. Since the bacteria can be pathogenic, the tests should be carried out by persons with training and experience in the use of microbiological techniques. The laboratory facilities should be designated as appropriate biosafety level.^[7]

Appropriate safety precautions should be observed with due consideration given to country-specific.

5 Requirements

This document is to provide the measurement method of disinfection efficiency for fine bubbles enhanced waters by test hydrodynamic cavitation units. The disinfection efficiency is measured by the test fine bubble solution generated by each test hydrodynamic cavitation unit.

The antibacterial (bactericidal) activity shall be evaluated using the following test organism: *E. coli* (see [Table A.1](#)). The test tank and test facility should be cleaned by appropriate methods like triple-wash with sterilized water after chlorination before the test.

6 Test methods

6.1 Principle

The test tank is filled with a test suspension of bacteria and the test hydrodynamic cavitation (HC) unit is operated under the required initial test condition (approximately 15 °C to 24 °C ± 1 °C). The fine bubbles-enriched water will be contacted to the test suspension of bacteria in the test tank (see [Figure 1](#)).

At a specific contact time, an aliquot is taken; the bactericidal action in this portion is immediately measured using the luminescence photometer (see [6.4.8](#)). First, the relationship between the measurement unit of the luminescence photometer, relative light unit (RLU) (see [Annex C](#)) and counted colonies forming unit (CFU) of the initial density of the test microorganism is measured to check the initial cell density of the bacterial test suspension.

The test is performed using *E. coli* (see [Table A.1](#)).

6.2 Test facility

Fine bubbles are generated by circulating the liquid in a fine bubble generating system with the following characteristics: a 125-litre tank, a 0,4 kW centrifugal multistage pump made of stainless steel, PVC pipes and a cavitation unit with the test fine bubble generating unit (see [Figure 1](#)). Test gas can be chosen for the appropriate test purposes.

6.3 Culture media and reagents

6.3.1 General

The reagents shall be of analytical grade and/or appropriate for microbiological purposes.

To improve reproducibility, dehydrated material should be used for the preparation of culture media. The manufacturer's instructions relating to the preparation of these products should be rigorously followed.

6.3.2 Water

Water used in tests shall be analytical-grade water for microbiological media preparation, which is freshly distilled and/or ion-exchanged and/or ultra-filtered and/or filtered with reverse osmosis (RO). It shall be free from substances that are toxic or inhibiting to the bacteria.

6.3.3 Tryptone soya agar (TSA)

TSA used in the test is for maintenance of bacterial strains and performance of viable counts.

Tryptone, pancreatic digest of casein	15,0 g
Soya peptone, papaic digest of soybean meal	5,0 g
NaCl	5,0 g
Agar	5,0 g
Water (see 6.3.2)	to 1,0 l

Sterilize in the autoclave ([6.4.1](#)). After sterilization, the pH of the medium shall be $7,2 \pm 0,2$ when measured at 20 °C.

6.3.4 Tryptone soya broth (TSB)

Tryptone, pancreatic digest of casein	17 g
Soya peptone, papaic digest of soya	3 g
Sodium chloride (NaCl)	5 g
Glucose	2,5 g
Dipotassium hydrogen phosphate	2,5 g
Water	1 000 ml

Mix well and adjust pH, $7,2 \pm 0,2$, then sterilize by autoclave ([6.4.1](#)).

6.3.5 Lysogeny broth (LB)

Tryptone, pancreatic digest of casein	10 g
Yeast extract	5 g
NaCl	10 g
Water	1 000 ml

Mix well and adjust pH, $6,9 \pm 0,2$, then sterilize by autoclave ([6.4.1](#)).

6.3.6 Cryoprotective solution for bacterial species

For freezing, a cryoprotective solution containing 150 g/l of glycerol or 100 g/l of dimethylsulfoxide shall be used and prepared as follows:

TSB (see 6.3.4) or LB (see 6.3.5):	1 000 ml
Add Glycerol: 150 g or dimethylsulfoxide:	100 g

Mix well and sterilize by autoclave (6.4.1).

For solutions containing glycerol, sterilize the mixed solution by autoclave (6.4.1). For solutions containing dimethylsulfoxide, sterilize the mixed solution by using a 0,22 µm membrane filter.

NOTE Any commercially available product can be used as long as it is a cryoprotective solution or preserving system that contains glycerol or dimethylsulfoxide.

6.3.7 Physiological saline

Sodium chloride (NaCl)	8,5 g
Water	1 000 ml

Mix well, then sterilize by autoclave (6.4.1).

6.3.8 Tryptone sodium chloride solution (TSCS)

Tryptone, pancreatic digest of casein	1 g
NaCl	8,5 g
Water	1 000 ml

6.4 Laboratory apparatus and glassware

The usual laboratory apparatus and, in particular, the following shall be used.

6.4.1 Autoclave, capable of sterilizing at $121\text{ °C} \pm 2\text{ °C}$ and $103\text{ kPa} \pm 5\text{ kPa}$.

6.4.2 Incubator, capable of being controlled at either $36\text{ °C} \pm 1\text{ °C}$ or $37\text{ °C} \pm 1\text{ °C}$. An incubator at $37\text{ °C} \pm 1\text{ °C}$ may be used if an incubator at $36\text{ °C} \pm 1\text{ °C}$ is not available.

6.4.3 pH meter, having an accuracy $\pm 0,1$ in calibration on pH units at 25 °C .

6.4.4 Stopwatch.

6.4.5 Container, test tubes or flasks of suitable capacity.

6.4.6 Pipettes, having the most suitable volume for each use, with a tip made of glass or plastic, and with a tolerance of 0,5 % or less.

6.4.7 Petri dishes, that have been sterilized, made of glass or plastic, in diameter sizes of 90 mm to 100 mm.

6.4.8 Luminescence photometer, with ATP kit for water test, capable of measuring ATP of 10^{-12} mol/l to 10^{-7} mol/l at 300 nm to 650 nm with a luminescence-measuring reagent.

6.4.9 Vials, 250 ml plastic bottles, with screw openings, polytetrafluoroethylene or silicone packing and caps made of polypropylene, polycarbonate or another suitable material.

6.4.10 Glass beads, with a diameter of 3 mm to 4 mm.

6.4.11 Freezers, one adjustable to a temperature below -70 °C and another to a temperature below -20 °C .

6.4.12 Homogenizer, capable of speeds of six blows per second to eight blows per second, with the corresponding disposable containers.

6.4.13 Disposable plastic bags, sterile bags suitable for containing food products, to be used for one of the shaking methods of the specimens.

6.4.14 Sterilized stirring rod, with a 6 mm (± 2 mm) diameter and a 300 mm (± 50 mm) length.

6.5 Preparation of bacterial test suspension

6.5.1 Storage of strains

6.5.1.1 General

The strains shall be stored in accordance with the supplier's recommendations. The experiment instruments should be sterilized by the autoclave (6.4.1) for a minimum of 15 min.

6.5.1.2 Preparation method

Obtain a sample of the freeze-dried bacterial strain in accordance with Annex A, following the recommendations supplied with the culture and resuspend it in 5 ml of TSB (see 6.3.4). Obtain a sample of the suspension and isolate it in a plastic bottle (6.4.10) containing LB (see 6.3.5). Incubate the cultures for 18 h to 24 h at $37\text{ °C} \pm 2\text{ °C}$.

After incubation, use the culture isolated in the Petri dish to verify the purity of the strain.

After verification, prepare the stock cultures to check the density of *E. coli* of the bacterial test suspension.

Sample 0,7 ml of the broth culture and spread it over the surface of the Petri dish containing the TSA. Incubate the culture on plates for 18 h to 24 h at $37\text{ °C} \pm 2\text{ °C}$.

Add 10 ml of cryoprotective solution (see 6.3.6) to the surface of the TSA plate culture and resuspend the cells in the solution using a sterile glass spreader. Sample the suspended cells from the surface of the agar, dilute them in 100 ml of cryoprotective solution and incubate for 30 min at 20 °C .

Using a pipette (6.4.6), sample 1 ml of the suspension and transfer it to a cryogenic vial (6.4.9) containing the beads (6.4.10). Shake the vial in order to spread the suspended cells around the beads.

- Where a cryoprotective solution containing dimethylsulfoxide is used, do not let it stand longer than 1 min at ambient temperature.
- Where a cryoprotective solution containing glycerol is used, let it stand for 30 min at 20 °C .
- Withdraw the excess cryoprotective solution with a sterile pipette. Place the cryogenic vials in a freezer (6.4.11) set at -70 °C or lower.

Prepare 10^{-6} and 10^{-7} dilutions of the suspension using the serial dilution method. Take a 1,0 ml sample of each dilution and transfer it to separate Petri dishes. Add 12 ml to 15 ml of nutritive solution, cooled down to $45\text{ °C} \pm 1\text{ °C}$. Incubate for 18 h to 24 h under the conditions specified for the strain. Enumerate the plate cultures and confirm that the cell density of the suspension is 1×10^5 CFU/ml to 5×10^8 CFU/ml using TSCS (see 6.3.8).

The other preparation details of initial bacteria stock culture are described in EN 1040.^[8]

6.5.2 Working culture of test bacteria

In order to prepare the working culture of strains, subculture from the stock culture by inoculation into LB broth and incubates. After 18 h to 24 h, prepare a second subculture from the first subculture