

ISO/FDIS 6323-1:2024(E)

ISO/TC 146/SC 2 ~~N1331~~

Secretariat: ANSI

Date: 2024-03-20

**Workplace air — Determination of arsenic and arsenic compounds by electrothermal atomic absorption spectrometry—
~~Part 1: Arsenic and arsenic compounds, except arsine by ET-AAS~~**

First edition

Date: 2024-01-11

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

ISO/FDIS 6323-1

<https://standards.iteh.ai/catalog/standards/iso/c128b1b1-613a-464d-b26f-b6fd9ee5d1e6/iso-fdis-6323-1>

Part 1:
Arsenic and arsenic compounds, except arsine by ET-AAS

Air des lieux de travail — Détermination de l'arsenic et des composés d'arsenic par spectrométrie d'absorption atomique avec atomisation électrothermique —

Partie 1: Arsenic et composés d'arsenic, à l'exception de l'arsine par ET-AAS

FDIS stage

(<https://standards.iteh.ai>)
Document Preview

[ISO/FDIS 6323-1](#)

<https://standards.iteh.ai/catalog/standards/iso/c128b1b1-613a-464d-b26f-b6fd9ee5d1e6/iso-fdis-6323-1>

© ISO ~~2023~~2024

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: + 41 22 749 01 11
~~Email~~E-mail: copyright@iso.org
Website: www.iso.org

Published in Switzerland

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO/FDIS 6323-1](#)

<https://standards.iteh.ai/catalog/standards/iso/c128b1b1-613a-464d-b26f-b6fd9ee5d1e6/iso-fdis-6323-1>

Contents

Foreword	vii
Introduction.....	viii
1 Scope	1
2 Normative references.....	1
3 Terms and definitions	2
4 Principle.....	2
5 Requirement	2
6 Apparatus and equipment.....	2
6.1 Sampling equipment.....	2
6.1.1 Sampler	3
6.1.2 Filter capsule.....	3
6.1.3 Sampling pumps	3
6.1.4 Portable flowmeter.....	3
6.1.5 Silicone adapter	3
6.1.6 Ancillary equipment.....	3
6.2 Collection media.....	4
6.2.1 Filters	4
6.2.2 Recommendations for filters.....	5
6.2.3 Back pressure of filters	5
6.2.4 Weighing of filters.....	5
6.3 Equipment for the determination of dust concentration	5
6.4 Equipment for sample digestion.....	5
6.4.1 General.....	5
6.4.2 Equipment for all types of sample digestion	6
6.4.3 Laboratory equipment for open vessel hot block digestion.....	6
6.4.4 Laboratory equipment for high pressure microwave digestion	6
6.5 Equipment for analysis.....	6
6.5.1 General.....	6
6.5.2 Equipment for sample preparation	6
6.5.3 Analytical system	7
7 Reagents.....	7
7.1 General	7
7.2 Water.....	7
7.3 Chemicals for digestion	7
7.4 Chemicals for analysis.....	7
7.4.1 Stabilization and modifier	7
7.4.2 Calibration and quality control standards.....	7
7.5 Chemicals for method validation.....	8

8	Occupational exposure assessment.....	8
9	Sampling	8
9.1	Preliminary considerations	8
9.1.1	Collection characteristics and flow rate	8
9.1.2	Sampling period.....	8
9.1.3	Temperature and pressure effects.....	9
9.1.4	Sample handling	9
9.2	Sample preparation.....	9
9.2.1	Cleaning.....	9
9.2.2	Filter	9
9.2.3	Marking	9
9.2.4	Pre-weighing of filters.....	9
9.2.5	Assembly.....	9
9.3	Sampling preparation	10
9.3.1	Initial flow rate.....	10
9.3.2	Field blanks.....	10
9.3.3	Personal sampling.....	10
9.3.4	Static sampling	10
9.4	Parameters.....	10
9.5	Performing the sampling	11
9.5.1	Start of the sampling period	11
9.5.2	Sampling period.....	11
9.5.3	End of the sampling period	11
9.6	Transport and storage	12
9.6.1	General.....	12
9.6.2	Transportation	12
9.6.3	Storage.....	12
10	Analysis	12
10.1	Preparation.....	12
10.1.1	General.....	12
10.1.2	Weighing.....	12
10.1.3	Digestion	13
10.1.4	Solutions	13
10.2	Instrumental analysis.....	14
10.3	Estimation of detection and quantification limits	15
10.3.1	Estimation of the instrumental detection limit	15
10.3.2	Estimation of the method detection limit and quantification limit.....	15
10.4	Certified reference materials	16
10.5	Measurement uncertainty	16

11	Expression of results	16
11.1	Calculations	16
11.2	Method performance	17
11.3	Remarks	17
12	Test report	17
12.1	Test record.....	17
12.2	Laboratory report.....	19
Annex A (informative) Temperature and pressure correction		20
Annex B (informative) Typical operating parameters for determination of arsenic by electrothermal atomic absorption spectrometry		22
Bibliography		23

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO/FDIS 6323-1](#)

<https://standards.iteh.ai/catalog/standards/iso/c128b1b1-613a-464d-b26f-b6fd9ee5d1e6/iso-fdis-6323-1>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 146, *Air quality*, Subcommittee SC 2, *Workplace atmospheres*.

A list of all parts in the ISO 6323 series can be found on the ISO website. www.iso.org/iso/26f-b6fd9ee5d1e6/iso-fdis-6323-1

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Arsenic and arsenic compounds are toxic and are recognized as human carcinogens. In particular arsenic and arsenic compounds are a hazard to the health of workers in many industries through exposure by inhalation. Industrial hygienists and other public health professionals need to determine the effectiveness of measures taken to control workers' exposure. The collection of samples of air during a work activity and then measuring the amount of particular arsenic and arsenic compounds are often done to assess an individual's exposure, the effectiveness of workplace controls or respiratory protection measures. The air sampling can be done as stationary or personal air sampling. Electrothermal atomic absorption spectrometry (ET-AAS) analysis of particular arsenic and arsenic compounds in a sample of respirable or inhalable dust collected on a collection substrate (membrane filter) is employed in many countries to measure and estimate exposure to arsenic and arsenic compounds. ET-AAS is able to quantify arsenic and arsenic compounds.

iTeh Standards (<https://standards.iteh.ai>) Document Preview

[ISO/FDIS 6323-1](https://standards.iteh.ai/catalog/standards/iso/c128b1b1-613a-464d-b26f-b6fd9ee5d1e6/iso-fdis-6323-1)

<https://standards.iteh.ai/catalog/standards/iso/c128b1b1-613a-464d-b26f-b6fd9ee5d1e6/iso-fdis-6323-1>

Workplace air — Determination of arsenic and arsenic compounds by electrothermal atomic absorption spectrometry ~~—~~ —

Part 1:

Arsenic and arsenic compounds, except arsine by ET-AAS

1 Scope

This document specifies a method for the determination of the mass concentration of particulate arsenic and arsenic compounds in workplace air sampled on a filter (e. g. 37 mm cellulose nitrate filter), digested with acid or an acid mixture and analysed quantitatively by using electrothermal atomic absorption spectrometry (ET-AAS). The method is not suitable for determination of arsenic in the form of metal arsenides, which decompose in the presence of water or acid, or for arsenic trioxide vapour.

Many different types of sampling apparatus are used to collect respirable or inhalable dust, according to the occupational hygiene convention. This document is designed to accommodate the variety of samplers and collection substrates available to analysts. This document is intended to be used in conjunction with ISO 21832^H which promotes best practices for these analyses.

The method is applicable to the determination of masses of approximately 0,2 µg to 2 µg of arsenic per sample, for analysis of test solutions prepared using sample solution aliquots in the recommended range (see ~~10.1.3~~10.1.3 and ~~10.1.4.1~~10.1.4.1). The concentration range for arsenic in air, for which this procedure is applicable, is determined in part by the sampling procedure selected by the user.

The method is applicable to personal and stationary air sampling.

A number of transition metals can interfere with the determination of arsenic by electrothermal atomic absorption spectrometry (see ~~11.3~~11.3).

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 1042, Laboratory glassware — One-mark volumetric flasks

ISO 3696, Water for analytical laboratory use — Specification and test methods

ISO 7708, Air quality — Particle size fraction definitions for health-related sampling

ISO 13137, *Workplace atmospheres — Pumps for personal sampling of chemical and biological agents — Requirements and test methods*

ISO 18158, *Workplace air — Terminology*

ISO 20581, Workplace air — General requirements for the performance of procedures for the measurement of chemical agents

ISO 21832, Workplace air — Metals and metalloids in airborne particles — Requirements for evaluation of measuring procedures

~~ISO 7708, Air quality — Particle size fraction definitions for health-related sampling~~

~~ISO 1042, Laboratory glassware — One mark volumetric flasks~~

~~ISO 3696, Water for analytical laboratory use — Specification and test methods~~

DIN 12353, Laboratory ware made from fused quartz and fused silica; boiling flasks made from fused quartz; round bottom flasks, flat bottom flasks and conical flasks

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 18158 apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ~~—~~ISO Online browsing platform: available at <https://www.iso.org/obp>
- ~~—~~IEC Electropedia: available at <https://www.electropedia.org/>

4 Principle

A filter (~~6.2.1(6.2.1)~~; see ~~Table 1~~ ~~Table 1~~) is mounted in a sampler (~~6.1.1(6.1.1)~~) designed to collect the respirable or inhalable fraction of airborne particles. The sampling can be performed with personal or stationary fixed samplers. Before sampling is performed, the filter batch used shall be verified with regard to its metal content and consequently the suitability of the minimum requirements for the performance of measuring methods.

Currently limit values in different countries exist for arsenic and its compounds either as “total dust”, defined by the performance of a sampler, or in the inhalable size selective fraction. A suitable sampling device for the applicable particle fraction shall be used considering the existing limit value/particle fraction.

The fraction separated on the filter is analysed for arsenic after acid digestion using ET-AAS. As digestion media ~~can be used~~, nitric acid or a mixture of nitric and hydrochloric acid ~~can be used~~. The sample solution is allowed to cool and diluted to a given volume with ultrapure water (~~7.3.1(7.3.1)~~), depending on the digestion type used. A test solution is prepared by transferring an aliquot of the sample solution to a volumetric flask and dilution to volume with ultrapure water.

The atomic absorption spectrometer is equipped with an arsenic hollow cathode lamp or electrodeless charge lamp and heated electrically.

~~Adsorbance~~ **Absorbance** measurements are made at 193,7 nm or 197,2 nm, using a graphite furnace with platform and a matrix modifier (~~7.4.1.2(7.4.1.2)~~). For background compensation, Zeeman-Mode is used. Deuterium background compensation can also be used as an option. Results obtained by the analytical-curve technique or the analyte addition technique.

5 Requirement

The measuring procedure shall ~~comply with any relevant international, European or national standard, e.g. conform to~~ ISO 20581, ~~[2]~~ or ISO 21832, ~~[1]~~, which ~~specifies~~ **specify** performance requirements for procedures for measuring chemical agents in workplace air.

6 Apparatus and equipment

6.1 Sampling equipment

~~The following equipment is required for sampling.~~

6.1.26.1.1 Sampler

The performance of the sampler used shall match the criteria for respirable or inhalable dust as specified in ISO 7708. Samplers that use 37 -mm diameter filters (6.2.1(6.2.1)) as the collection substrate are required. A plastic filter capsule for filters with a diameter of 37 mm are necessary. A suitable supporting grid can be necessary.

Each sampler should be labelled with a unique number, in order to identify samplers that start to under-perform after long-term use.

Samplers shall ~~comply with~~meet the manufacturer's requirements for calibration.

NOTE For person-related or stationary sampling, filters with diameters of e.g. 70 mm up to 150 mm, can also be used with specific sampling systems, with appropriate adjustments to the digestion conditions.

NOTE In some countries, there can be exceptions due to national requirements.

6.1.36.1.2 Filter capsule

Matching plastic filter capsule with covers for the 37 mm filter (6.2.1(6.2.1)), for insertion into the sampler.

6.1.46.1.3 Sampling pumps

Sampling pumps shall ~~comply with~~conform to the requirements of ISO 13137.

If the sampling pump is used outside the range of conditions specified in ISO 13137, appropriate ~~action~~actions should be taken to ensure that the performance requirements are met.

6.1.56.1.4 Portable flowmeter

The flowmeter shall ~~comply with~~conform to the requirements of ISO 13137.

The flowmeter shall be capable of measuring the appropriate flow rate (see 9.3.1(9.3.1) and 9.4(9.4)) to within ± 5 %, and calibrated against a primary standard, i.e. a flowmeter of which the accuracy is traceable to national standards. If appropriate, the atmospheric temperature and pressure at which the flowmeter was calibrated should be recorded.

6.1.66.1.5 Silicone adapter

~~This~~The silicone adapter ~~must~~shall fit into the sampler head to connect the flowmeter (6.1.4(6.1.4)) for measuring/setting the air flow.

6.1.76.1.6 Ancillary equipment

~~Other~~The following ancillary equipment ~~required for sampling include the following~~shall be used:

- a) ~~a)~~ flexible tubing, to connect the sampler to the sampling pump (6.1.3(6.1.3));
- b) ~~b)~~ belts or harnesses to which the sampling pumps can conveniently be fixed for personal air sampling;
A; a tripod is required for person-related or stationary sampling;
- c) ~~c)~~ a means to transport the samples from the workplace to the laboratory, which minimises the possibility of accidental transfers of collected dust to or from the collection substrate (filter);
~~Transportation~~; transportation requires caps or covers for the samplers (filter capsule);