



**International
Standard**

ISO 6559

**Traditional Chinese medicine —
Sterile three-edge needle for
single use**

*Médecine traditionnelle chinoise — Aiguille stérile à trois bords à
usage unique*

**First edition
2024-07**

iteh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO 6559:2024](https://standards.iteh.ai/catalog/standards/iso/6c32cb78-2a19-4c24-89d9-27d931a9bba1/iso-6559-2024)

<https://standards.iteh.ai/catalog/standards/iso/6c32cb78-2a19-4c24-89d9-27d931a9bba1/iso-6559-2024>

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO 6559:2024](https://standards.iteh.ai/catalog/standards/iso/6c32cb78-2a19-4d24-89d9-27d931a9bba1/iso-6559-2024)

<https://standards.iteh.ai/catalog/standards/iso/6c32cb78-2a19-4d24-89d9-27d931a9bba1/iso-6559-2024>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2024

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Specifications	2
4.1 Configuration.....	2
4.2 Dimensions and parameters.....	3
4.2.1 Whole length of the three-edge needle.....	3
4.2.2 Length of the head of the three-edge needle.....	3
4.2.3 External diameter of the widest part of the head of the three-edge needle.....	3
4.2.4 External diameter of the body of the three-edge needle.....	3
4.3 Material.....	4
5 Requirements	4
5.1 Hardness.....	4
5.2 Roughness.....	4
5.3 Puncture force.....	4
5.4 Corrosion resistance.....	4
5.5 Sterility assurance.....	4
5.6 Biological compatibility.....	4
5.7 Appearance quality.....	4
6 Packaging	5
6.1 Primary package.....	5
6.2 Secondary package.....	5
7 Labelling	5
7.1 General.....	5
7.2 Primary package.....	5
7.3 Secondary package.....	6
8 Transit and storage	6
Annex A (informative) Test method for puncture force	7
Bibliography	9

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 249, *Traditional Chinese medicine*.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

<https://standards.iteh.ai>
ISO 6559:2024

<https://standards.iteh.ai/catalog/standards/iso/6c32cb78-2a19-4d24-89d9-27d931a9bba1/iso-6559-2024>

Introduction

Originating in ancient China, the three-edge needle is one of major types of acupuncture needles. The three-edge needle was originally called "Feng Zhen", which literally means "sharp-edge needle". Its clinical application for the specific indications was earliest recorded in *Huangdi's Canon of Medicine*, which is the earliest writing about traditional Chinese medicine. The therapy which applies the three-edge needle is called three-edge needle therapy.

The three-edge needle has a special needle body structure like an arrow head with three sharp edges. This structure was designed in order to meet specific clinical requirements, such as bloodletting and draining abscesses mainly by the manipulation methods including shallow pricking, deep pricking, multi-point shallow pricking, pricking in and edge cutting out, and needle scratching. Of the manipulation methods, pricking in and edge cutting out (mainly for breaking the fibrous tissue) and needle scratching (leaving marks on skin) can only be practiced by the three-edge needle, which cannot be replaced by other bloodletting medical instruments such as the blood point lancet and the syringe.

It is due to the special structure of the three-edge needle that the practitioner can bleed the clinically required blood amount by controlling the depth instead of repeated pricking which induces pain in patients. This structural advantage of the three-edge needle outweighs the blood point lancet and the syringe both in clinical function, such as draining an abscess, and in obtaining a clinically required amount of blood.

Since the three-edge needle is an invasive medical instrument, it is important to use the sterile needle only once. The quality of the three-edge needle has a direct impact on the safety of patients and the quality of the therapy. Therefore, this document specifies the technical parameters of the sterile three-edge needle for single use to ensure its quality and safety.

iTeh Standards (<https://standards.iteh.ai>) Document Preview

[ISO 6559:2024](https://standards.iteh.ai/catalog/standards/iso/6c32cb78-2a19-4d24-89d9-27d931a9bba1/iso-6559-2024)

<https://standards.iteh.ai/catalog/standards/iso/6c32cb78-2a19-4d24-89d9-27d931a9bba1/iso-6559-2024>

