

# INTERNATIONAL STANDARD

# IEC 61676

First edition  
2002-09

---

---

## Medical electrical equipment – Dosimetric instruments used for non-invasive measurement of X-ray tube voltage in diagnostic radiology

iTeh STANDARD PREVIEW

*Appareils électromédicaux –  
Instruments de dosimétrie pour la mesure  
non invasive de la tension du tube radiogène  
dans la radiologie de diagnostic*

<https://standards.iteh.ai/catalog/standards/sist/403376ec-f3f4f69-ae01-705cc0233aaa/iec-61676-2002>



Reference number  
IEC 61676:2002(E)

## Publication numbering

As from 1 January 1997 all IEC publications are issued with a designation in the 60000 series. For example, IEC 34-1 is now referred to as IEC 60034-1.

## Consolidated editions

The IEC is now publishing consolidated versions of its publications. For example, edition numbers 1.0, 1.1 and 1.2 refer, respectively, to the base publication, the base publication incorporating amendment 1 and the base publication incorporating amendments 1 and 2.

## Further information on IEC publications

The technical content of IEC publications is kept under constant review by the IEC, thus ensuring that the content reflects current technology. Information relating to this publication, including its validity, is available in the IEC Catalogue of publications (see below) in addition to new editions, amendments and corrigenda. Information on the subjects under consideration and work in progress undertaken by the technical committee which has prepared this publication, as well as the list of publications issued, is also available from the following:

- **IEC Web Site** ([www.iec.ch](http://www.iec.ch))

- **Catalogue of IEC publications**

The on-line catalogue on the IEC web site ([http://www.iec.ch/searchpub/cur\\_fut.htm](http://www.iec.ch/searchpub/cur_fut.htm)) enables you to search by a variety of criteria including text searches, technical committees and date of publication. On-line information is also available on recently issued publications, withdrawn and replaced publications, as well as corrigenda.

<https://standards.iteh.ai/catalog/standards/sist/403376ec-f3f-4f69-ac01-705cc0233aaa/iec-61676-2002>

- **IEC Just Published** [705cc0233aaa/iec-61676-2002](https://standards.iteh.ai/catalog/standards/sist/403376ec-f3f-4f69-ac01-705cc0233aaa/iec-61676-2002)

This summary of recently issued publications ([http://www.iec.ch/online\\_news/justpub/jp\\_entry.htm](http://www.iec.ch/online_news/justpub/jp_entry.htm)) is also available by email. Please contact the Customer Service Centre (see below) for further information.

- **Customer Service Centre**

If you have any questions regarding this publication or need further assistance, please contact the Customer Service Centre:

Email: [custserv@iec.ch](mailto:custserv@iec.ch)  
Tel: +41 22 919 02 11  
Fax: +41 22 919 03 00

# INTERNATIONAL STANDARD

# IEC 61676

First edition  
2002-09

---

---

## Medical electrical equipment – Dosimetric instruments used for non-invasive measurement of X-ray tube voltage in diagnostic radiology

iTeh STANDARD PREVIEW

*Appareils électromédicaux –  
Instrumente de dosimétrie pour la mesure  
non invasive de la tension du tube radiogène  
dans la radiologie de diagnostic*

<https://standards.iteh.ai/catalog/standards/sist/403376ec-f3f4f69-ac01-705cc0233aaa/iec-61676-2002>

© IEC 2002 — Copyright - all rights reserved

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Electrotechnical Commission, 3, rue de Varembé, PO Box 131, CH-1211 Geneva 20, Switzerland  
Telephone: +41 22 919 02 11 Telefax: +41 22 919 03 00 E-mail: [inmail@iec.ch](mailto:inmail@iec.ch) Web: [www.iec.ch](http://www.iec.ch)



Commission Electrotechnique Internationale  
International Electrotechnical Commission  
Международная Электротехническая Комиссия

PRICE CODE

V

*For price, see current catalogue*

## CONTENTS

FOREWORD .....	3
INTRODUCTION .....	5
1 Scope and object .....	6
2 Normative references .....	6
3 Terminology and definitions .....	7
4 General performance requirements for measurement of PRACTICAL PEAK VOLTAGE measurements .....	10
4.1 Quantity to be measured .....	10
4.2 Limits of PERFORMANCE CHARACTERISTICS .....	10
4.3 LIMITS OF VARIATION for effects of INFLUENCE QUANTITIES .....	13
4.4 Performance test procedures .....	15
5 Special instrumental requirements and marking .....	21
5.1 Requirements for the complete instruments .....	21
5.2 General .....	21
5.3 Display .....	22
5.4 Range of measurement .....	22
5.5 Connectors and cables .....	22
6 ACCOMPANYING DOCUMENTS .....	22
6.1 General .....	22
6.2 Information provided .....	22
6.3 Instrument description .....	22
6.4 Detector .....	22
6.5 Delay time .....	22
6.6 Measurement window .....	22
6.7 Data outlet .....	23
6.8 Transport and storage .....	23
Annex A (informative) Recommended performance criteria for the invasive divider .....	24
Annex B (informative) Additional information on PRACTICAL PEAK VOLTAGE .....	25
Annex C (informative) Glossary of defined terms .....	32

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

## MEDICAL ELECTRICAL EQUIPMENT –

**Dosimetric instruments used for non-invasive measurement  
of X-ray tube voltage in diagnostic radiology**

## FOREWORD

- 1) The IEC (International Electrotechnical Commission) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of the IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, the IEC publishes International Standards. Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. The IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of the IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested National Committees.
- 3) The documents produced have the form of recommendations for international use and are published in the form of standards, technical specifications, technical reports or guides and they are accepted by the National Committees in that sense.
- 4) In order to promote international unification, IEC National Committees undertake to apply IEC International Standards transparently to the maximum extent possible in their national and regional standards. Any divergence between the IEC Standard and the corresponding national or regional standard shall be clearly indicated in the latter.
- 5) The IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with one of its standards.
- 6) Attention is drawn to the possibility that some of the elements of this International Standard may be the subject of patent rights. The IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 61676 has been prepared by subcommittee SC 62C: Equipment for radiotherapy, nuclear medicine and radiation dosimetry, of IEC Technical Committee 62: Electrical equipment in medical practice.

The text of this standard is based on the following documents:

FDIS	Report on voting
62C/340/FDIS	62C/344/RVD

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 3.

Annexes A, B and C are for information only.

In this standard the following print types are used:

- requirements, compliance with which can be tested, and definitions: in roman type;
- notes, explanations, advice, general statements and exceptions: in small roman type;
- *test specifications: in italic type;*
- TERMS USED THROUGHOUT THIS STANDARD THAT HAVE BEEN DEFINED IN CLAUSE 3 OR IN IEC 60601-1 AND ITS COLLATERAL STANDARDS: IN SMALL CAPITALS.

The committee has decided that the contents of this publication will remain unchanged until 2004. At this date, the publication will be

- reconfirmed;
- withdrawn;
- replaced by a revised edition, or
- amended.

A bilingual version of this publication may be issued at a later date.

NOTE The committee is aware of the fact that this standard does not address all problems associated with non-invasive high voltage measurements. In particular one influence quantity concerning the target condition is not dealt with at all. Before this can be done, a substantial amount of measurements is still necessary to improve the physical understanding of this influence quantity. On the other hand, for the reasons described in the introduction there is an urgent need to publish this standard in order to assure that non-invasive measurements are comparable to each other within tolerable uncertainties, regardless of differences in X-RAY GENERATOR, waveform or other influence quantities (except target condition), which is not the case for the time being. The committee has decided to revise this standard as soon as sufficient knowledge on the outstanding items is available.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

IEC 61676:2002

<https://standards.iteh.ai/catalog/standards/sist/403376ec-f3f-4f69-ac01-705cc0233aaa/iec-61676-2002>

## INTRODUCTION

The result of a measurement of the X-RAY TUBE VOLTAGE by means of invasive or non-invasive instruments is normally expressed in the form of one single number for the value of the tube voltage, irrespective of whether the tube voltage is constant potential or shows a time dependent waveform. Non-invasive instruments for the measurement of the X-RAY TUBE VOLTAGE on the market usually indicate the 'mean peak voltage'. But the quantity 'mean peak voltage' is not unambiguously defined and may be any mean of all voltage peaks. It is impossible to establish test procedures for the performance requirements of non-invasive instruments for the measurement of the X-RAY TUBE VOLTAGE without the definition of the quantity under consideration. Therefore, this Standard is based on a quantity recently proposed in the literature<sup>1</sup> to be called "PRACTICAL PEAK VOLTAGE". The PRACTICAL PEAK VOLTAGE is unambiguously defined and applicable to any waveform. This quantity is related to the spectral distribution of the emitted X-RADIATION and the image properties. X-RAY GENERATORS operating at the same value of the PRACTICAL PEAK VOLTAGE will produce the same low level contrast in the RADIOGRAMS, even when the waveforms of the tube voltages are different. Detailed information on this concept is provided in Annex B. An example for the calculation of the PRACTICAL PEAK VOLTAGE in the case of a "falling load" waveform is also given in Annex B.

As a result of introducing a new quantity, the problem arises that this standard has been written for instruments which were not explicitly designed for the measurement of the PRACTICAL PEAK VOLTAGE. However, from preliminary results of a trial type test of a non-invasive instrument currently on the market, it can be expected that future instruments and most instruments on the market will be able to fulfil the requirements stated in this standard without insurmountable difficulties. For the most critical requirements on voltage waveform and frequency dependence of the RESPONSE, it turned out from these investigations that it is even easier to comply with the standard by using the PRACTICAL PEAK VOLTAGE as the measurement quantity.

### IEC 61676:2002

The calibration and adjustment of the X-RAY TUBE VOLTAGE of an X-RAY GENERATOR is generally performed by the MANUFACTURER using a direct INVASIVE MEASUREMENT. Instruments utilising NON-INVASIVE MEASUREMENTS can also be used to check the calibration or to adjust THE X-RAY TUBE VOLTAGE. These instruments are required to have uncertainties of the voltage measurement comparable with the INVASIVE MEASUREMENT. One of the most important parameters of diagnostic X-RAY EQUIPMENT is the voltage applied to the X-RAY TUBE, because both the image quality in diagnostic radiology and the DOSE received by the PATIENT undergoing radiological examinations are dependent on the X-RAY TUBE VOLTAGE. An overall uncertainty below  $\pm 5\%$  is required, and this value serves as a guide for the LIMITS OF VARIATION for the effects of INFLUENCE QUANTITIES.

---

<sup>1</sup> See annex B.

## MEDICAL ELECTRICAL EQUIPMENT –

### Dosimetric instruments used for non-invasive measurement of X-ray tube voltage in diagnostic radiology

#### 1 Scope and object

This International Standard specifies the performance requirements of instruments as used in the NON-INVASIVE MEASUREMENT of X-RAY TUBE VOLTAGE up to 150 kV and the relevant compliance tests. This standard also describes the method for calibration and gives guidance for estimating the uncertainty in measurements performed under conditions different from those during calibration.

Applications for such measurement are found in diagnostic RADIOLOGY including mammography, COMPUTED TOMOGRAPHY (CT), dental radiology and RADIOSCOPY. This standard is not concerned with the safety aspect of such instruments. The requirements for electrical safety applying to them are contained in IEC 61010-1.

#### 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

- IEC 60417 (all parts), *Graphical symbols for use on equipment*. IEC 61676:2002  
<https://standards.iso.org/standards/catalog/standards/siso/61676-2002.html>
- IEC 60788:1984, *Medical radiology – Terminology*
- IEC 61000-4-2:1995, *Electromagnetic compatibility (EMC) – Part 4: Testing and measurement techniques – Section 2: Electrostatic discharge immunity test*. Basic EMC Publication
- IEC 61000-4-3:2000, *Electromagnetic compatibility (EMC) – Part 4-3: Testing and measurement techniques – Radiated, radio-frequency, electromagnetic field immunity test*. Basic EMC Publication
- IEC 61000-4-4:1995, *Electromagnetic compatibility (EMC) – Part 4: Testing and measurement techniques – Section 4: Electrical fast transient/burst immunity test*. Basic EMC Publication
- IEC 61000-4-5:1995, *Electromagnetic compatibility (EMC) – Part 4: Testing and measurement techniques – Section 5: Surge immunity test*. Basic EMC Publication
- IEC 61000-4-6:1996, *Electromagnetic compatibility (EMC) – Part 4: Testing and measurement techniques – Section 6: Immunity to conducted disturbances, induced by radio frequency fields*. Basic EMC Publication
- IEC 61000-4-11:1994, *Electromagnetic compatibility (EMC) – Part 4: Testing and measurement techniques – Section 11: Voltage dips, short interruptions and voltage variations immunity tests*. Basic EMC Publication
- IEC 61010-1:2001, *Safety requirements for electrical equipment for measurement, control, and laboratory use – Part 1: General Requirements*



IEC 61187:1993, *Electrical and electronic measuring equipment – Documentation*

ISO:1993, *International vocabulary of basic and general terms in metrology* (ISBN 92-67-01075-1)

ISO 7000:1989, *Graphical symbols for use on equipment – Index and synopsis*

### 3 Terminology and definitions

For the purposes of this standard the following definitions apply.

The definitions given in this standard are generally in agreement with those in IEC 60788 and the ISO *International vocabulary of basic and general terms in metrology*. Any terms not defined in this subclause have the meanings defined in the above publications or are assumed to be in general scientific usage.

#### 3.1

##### **CORRECTION FACTOR**

dimensionless multiplier which corrects the INDICATED VALUE of an instrument from its value when operated under particular conditions to its value when operated under stated REFERENCE CONDITIONS

#### 3.2

##### **EFFECTIVE RANGE**

range of INDICATED VALUES for which an instrument complies with a stated performance. The maximum (minimum) effective INDICATED VALUE is the highest (lowest) in this range

iTeh STANDARD PREVIEW

(standards.iteh.ai)

<https://standards.iteh.ai/catalog/standards/sist/403376ec-f3f4f69-ac01-705cc0233aaa/iec-61676-2002>

#### 3.3

##### **INDICATED VALUE**

the value of quantity derived from the scale reading of an instrument together with any scale factors indicated on the control panel of the instrument

#### 3.4

##### **INFLUENCE QUANTITY**

any external quantity that may affect the performance of an instrument (e.g. ambient temperature etc.) and any property of the X-RAY EQUIPMENT under test that needs to be taken into account in using the instrument for NON-INVASIVE MEASUREMENT of X-RAY TUBE VOLTAGE (e.g. range of X-RAY TUBE VOLTAGE, ANODE ANGLE, anode material, TOTAL FILTRATION etc.)

#### 3.5

##### **INSTRUMENT PARAMETER**

any internal property of an instrument that may affect the performance of the instrument

#### 3.6

##### **INTRINSIC ERROR**

deviation of the MEASURED VALUE (i.e. the INDICATED VALUE, corrected to REFERENCE CONDITIONS) from the CONVENTIONAL TRUE VALUE under STANDARD TEST CONDITIONS

#### 3.7

##### **INVASIVE MEASUREMENT**

measurement of the X-RAY TUBE VOLTAGE by external connection of a suitable meter or a high resistance divider

**3.8****LIMITS OF VARIATION**

the maximum VARIATION of a PERFORMANCE CHARACTERISTIC  $y$ , permitted by this standard. If the LIMITS OF VARIATION are stated as  $\pm L$  % the VARIATION  $\Delta y / y$ , expressed as a percentage, shall remain in the range from  $-L$  % to  $+L$  %

**3.9****MAXIMUM PEAK VOLTAGE**

maximum value of the X-RAY TUBE VOLTAGE in a specified time interval. The unit of this quantity is the volt (V)

**3.10****MEAN PEAK VOLTAGE**

mean value of all X-RAY TUBE VOLTAGE peaks during a specified time interval. The unit of this quantity is the volt (V)

**3.11****MEASURED VALUE**

the best estimate of the CONVENTIONAL TRUE VALUE of a quantity, being derived from the INDICATED VALUE of an instrument together with the application of all relevant CORRECTION FACTORS

NOTE The CONVENTIONAL TRUE VALUE will usually be the value determined by the working standard with which the instrument under test is being compared

**3.12****MINIMUM EFFECTIVE RANGE**

the MINIMUM EFFECTIVE RANGE is the smallest permitted range of INDICATED VALUES for which an instrument complies with a stated performance

iTech STANDARD PREVIEW

(standards.iteh.ai)

[IEC 61676:2002](https://standards.iteh.ai/catalog/standards/sist/403376ec-f3f4f69-ac01-705cc0233aaa/iec-61676-2002)

**3.13****NON-INVASIVE MEASUREMENT**

measurement of X-RAY TUBE VOLTAGE by analysis of the emitted RADIATION

<https://standards.iteh.ai/catalog/standards/sist/403376ec-f3f4f69-ac01-705cc0233aaa/iec-61676-2002>

**3.14****PERFORMANCE CHARACTERISTIC**

one of the quantities used to define the performance of an instrument (e.g. RESPONSE)

**3.15****VOLTAGE RIPPLE**

the VOLTAGE RIPPLE at the X-RAY TUBE,  $r$ , is expressed as a percentage of the peak voltage,  $U_{\max}$ , over a specified time interval. This is expressed by the equation:

$$r = \frac{U_{\max} - U_{\min}}{U_{\max}} \cdot 100 \%$$

where  $U_{\max}$  is the highest voltage in the interval, and  $U_{\min}$  is the lowest voltage in the interval

**3.16****PRACTICAL PEAK VOLTAGE (PPV)**

The PRACTICAL PEAK VOLTAGE  $\hat{U}$  is defined as:

$$\hat{U} = \frac{\int_{U_{\min}}^{U_{\max}} p(U) \cdot w(U) \cdot U \, dU}{\int_{U_{\min}}^{U_{\max}} p(U) \cdot w(U) \, dU} \quad \text{with} \quad \int_{U_{\min}}^{U_{\max}} p(U) \, dU = 1$$

where  $p(U)$  is the distribution function for the voltage  $U$  and  $w(U)$  is a weighting function.  $U_{\max}$  is the highest voltage in the interval, and  $U_{\min}$  is the lowest voltage in the interval. The unit of the quantity PRACTICAL PEAK VOLTAGE is the volt (V)

NOTE Additional information on the PRACTICAL PEAK VOLTAGE, the weighting function  $w(U)$  and the distribution function  $p(U)$  is provided in Annex B. Using this weighting function  $w(U)$  the PRACTICAL PEAK VOLTAGE will be defined as the constant potential which produces the same AIR KERMA contrast behind a specified PHANTOM as the non-dc voltage under test.

**3.17****RATED RANGE (of use)**

the range of values of an INFLUENCE QUANTITY or INSTRUMENT PARAMETER within which the instrument will operate within the LIMITS OF VARIATION. Its limits are the maximum and minimum RATED values.

The MINIMUM RATED RANGE is the least range of an INFLUENCE QUANTITY or INSTRUMENT PARAMETER within which the instrument shall operate within the specified LIMITS OF VARIATION in order to comply with this standard

[IEC 61676:2002](https://standards.iteh.ai/catalog/standards/sist/403376ec-f3f4f69-ac01-705cc0233aaa/iec-61676-2002)

**3.18**

<https://standards.iteh.ai/catalog/standards/sist/403376ec-f3f4f69-ac01-705cc0233aaa/iec-61676-2002>

**REFERENCE CONDITIONS**

conditions under which all INFLUENCE QUANTITIES and INSTRUMENT PARAMETERS have their REFERENCE VALUES

**3.19****REFERENCE VALUE**

particular value of an INFLUENCE QUANTITY (or INSTRUMENT PARAMETER) chosen for the purposes of reference i.e. the value of an INFLUENCE QUANTITY (or INSTRUMENT PARAMETER) at which the CORRECTION FACTOR for dependence on that INFLUENCE QUANTITY (or INSTRUMENT PARAMETER) is unity

**3.20****RELATIVE INTRINSIC ERROR**

the ratio of the INTRINSIC ERROR to the CONVENTIONAL TRUE VALUE

**3.21****RESPONSE**

the quotient of the INDICATED VALUE divided by the CONVENTIONAL TRUE VALUE

**3.22****STANDARD TEST CONDITIONS**

conditions under which all INFLUENCE QUANTITIES and INSTRUMENT PARAMETERS have their STANDARD TEST VALUES

**3.23**

**STANDARD TEST VALUES**

a value, values, or a range of values of an INFLUENCE QUANTITY or INSTRUMENT PARAMETER, which is/are permitted when carrying out calibrations or tests on another INFLUENCE QUANTITY or INSTRUMENT PARAMETER

**3.24**

**VARIATION**

The relative difference  $\Delta y / y$ , between the values of a PERFORMANCE CHARACTERISTIC  $y$ , when one INFLUENCE QUANTITY (or INSTRUMENT PARAMETER) assumes successively two specified values, the other INFLUENCE QUANTITIES (and INSTRUMENT PARAMETERS) being kept constant at the STANDARD TEST VALUES (unless other values are specified)

**3.25**

**X-RAY TUBE VOLTAGE**

potential difference applied to an X-RAY TUBE between the anode and the cathode . The unit of this quantity is the volt (V)

**4 General performance requirements for measurement of PRACTICAL PEAK VOLTAGE measurements**

**4.1 Quantity to be measured**

The quantity to be measured is the PRACTICAL PEAK VOLTAGE.

NOTE Additional quantities may be displayed.

The MINIMUM EFFECTIVE RANGES of PRACTICAL PEAK VOLTAGE shall be as listed in table 1 for the relevant X-RAY applications.

[IEC 61676:2002](https://standards.iteh.ai/catalog/standards/sist/403376ec-f3f4f69-ac01-105c-105c-105c105c105c)

<https://standards.iteh.ai/catalog/standards/sist/403376ec-f3f4f69-ac01-105c-105c-105c105c105c>

**Table 1 - MINIMUM EFFECTIVE RANGES**

Application	Nominal Anode Material	MINIMUM EFFECTIVE RANGE
Mammography (20 kV to 50 kV)	Mo a)	24 kV to 35 kV
Diagnostic (40 kV to 150 kV)	W	60 kV to 120 kV
CT (80 kV to 150 kV)	W	100 kV to 140 kV
Dental (40 kV to 110 kV)	W	60 kV to 90 kV
Fluoroscopic (40 kV to 130 kV)	W	60 kV to 120 kV

a) For mammography anode materials other than Mo, the MINIMUM EFFECTIVE RANGE of PPV shall be at least 10 kV.

**4.2 Limits of PERFORMANCE CHARACTERISTICS**

**4.2.1 Limits**

All values of the limits of PERFORMANCE CHARACTERISTICS stated in this subclause do not contain the uncertainty of the test equipment.