DRAFT INTERNATIONAL STANDARD ISO/DIS 6760-1

ISO/TC 172/SC 3

Voting begins on: **2023-07-11**

Secretariat: **JISC**

Voting terminates on: 2023-10-03

Optics and photonics — Test method for temperature coefficient of refractive index of optical glasses —

Part 1: Minimum deviation method

Optique et photonique — Méthode d'essai pour déterminer le coefficient de température de l'indice de réfraction des verres optiques —

Partie 1: Méthode de la déviation minimale

iTeh STANDARD PREVIEW (standards.iteh.ai)

ICS: 37.020

ISO/DIS 6760-1

https://standards.iteh.ai/catalog/standards/sist/af90dc53-5da2-46ab-8a76-ce2df00fb24a/isodis-6760-1

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENT AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION. This document is circulated as received from the committee secretariat.



Reference number ISO/DIS 6760-1:2023(E)

iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO/DIS 6760-1

https://standards.iteh.ai/catalog/standards/sist/af90dc53-5da2-46ab-8a76-ce2df00fb24a/isodis-6760-1



COPYRIGHT PROTECTED DOCUMENT

© ISO 2023

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office CP 401 • Ch. de Blandonnet 8 CH-1214 Vernier, Geneva Phone: +41 22 749 01 11 Email: copyright@iso.org Website: www.iso.org

Published in Switzerland

Page

Contents

Forew	vord	iv
	duction	
1	Scope	
2	Normative references	
3	Terms and definitions	
4	Principle	2
5	Measuring apparatus 5.1 Goniometer 5.2 Light source 5.3 Detector 5.4 Thermal chamber	3 3 3
6	Specimen prism	4
7	Measurement7.1Measurement of apex angle.7.2Measurement of the angle of minimum deviation.	4
8	Calculation 8.1 Absolute refractive index 8.2 Temperature coefficient of absolute refractive index 8.3 Temperature coefficient of relative refractive index	5 6
9	How to express the temperature coefficient of refractive index	8
10	Test report	
Annex	x A (informative) Formula for calculating the refractive index of air	9
	x B (informative) Calculation method for obtaining the relative refractive index of glass at an arbitrary temperature, air pressure and relative humidity	
Annex	x C (normative) Half prism method	13
	x D (normative) Interpolation formula for $\Delta n/\Delta T$	
	x E (informative) Derivation and verification of $\Delta n_{rel} / \Delta T$	
	ography	
215110	·σ- ~r -γ	

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 172, *Optics and photonics*, Subcommittee SC 3, *Optical materials and components*.

A list of all parts in the ISO 6760 series can be found on the ISO website. 46ab-8a76-ce2df00fb24a/iso-

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at <u>www.iso.org/members.html</u>.

Introduction

Optical glass is widely used in optical devices such as cameras, telescopes, and microscopes, and its refractive index is measured by the minimum deviation method (ISO 21395-1) and the V-block refractometer method (ISO 21395-2^[6]). Here, when designing an optical apparatus that requires high resolution, it is necessary to consider the temperature change of the refractive index of the optical glass in the usage environment, however up until now, there is no international standard. In view of the above situation, this document proposes a method for measuring the temperature coefficient of refractive index of optical glass with high accuracy, aiming to help mutual understanding of measured value users and contribute to efficiency and fairness.

iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO/DIS 6760-1

https://standards.iteh.ai/catalog/standards/sist/af90dc53-5da2-46ab-8a76-ce2df00fb24a/isodis-6760-1

iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO/DIS 6760-1

https://standards.iteh.ai/catalog/standards/sist/af90dc53-5da2-46ab-8a76-ce2df00fb24a/isodis-6760-1

Optics and photonics — Test method for temperature coefficient of refractive index of optical glasses —

Part 1: Minimum deviation method

1 Scope

This document specifies the measurement method used for calculating the temperature coefficient of the refractive index by measuring the refractive index, which changes with the temperature of the optical glass using the minimum deviation method.

The intended temperature range of the specified measurement method is an arbitrary range.

The intended wavelength range for the specified measurement method is 365 nm to 1 014 nm.

The intended accuracy for the specified measurement method is 1×10^{-6} K⁻¹.

2 Normative references ANDARD PREVIEW

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 21395-1, Optics and photonics — Test method for refractive index of optical glasses — Part 1: Minimum deviation method

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

ISO Online browsing platform: available at https://www.iso.org/obp

— IEC Electropedia: available at <u>https://www.electropedia.org/</u>

3.1

temperature coefficient of refractive index

ratio of refractive index change to temperature change at a selected wavelength

Note 1 to entry: Similar to ISO 9802:2022^[1], 3.4.2.3, 3.4.2.4.

3.2

temperature coefficient of absolute refractive index $\Delta n_{\rm abs}/\Delta T$

ratio of refractive index change in vacuum to temperature change at a selected wavelength

[SOURCE: ISO 9802:2022^[1], 3.4.2.3]

3.3

temperature coefficient of relative refractive index

$\Delta n_{\rm rel}/\Delta T$

ratio of refractive index change at an air pressure of 1,013 3 \times 10⁵ Pa and a relative humidity of 0 % to temperature change at a selected wavelength

[SOURCE: ISO 9802:2022^[1], 3.4.2.4, modified: 1,013 3 × 10⁵ Pa and a relative humidity of 0 %.]

Note 1 to entry: This definition of $\Delta n_{rel}/\Delta T$ is for a specific pressure and humidity. $\Delta n_{rel}/\Delta T$ can be calculated for any other pressure and humidity by understanding the index of air in those conditions.

3.4


thermal chamber

chamber where the temperature of the specimen can be changed and/or maintained to the preset temperature

Principle 4

As shown in Figure 1, a specimen prism is placed in a thermal chamber. The temperature of the specimen prism is changed from t_1 to t_2 or from t_2 to t_1 , and the refractive index of the specimen prism is measured at the temperatures of t_1 and t_2 respectively, in accordance with the method described in ISO 21395-1 to find the temperature coefficient of refractive index. Figure 2 shows the concept of calculating this temperature coefficient of refractive index.

NOTE In this document the term "light" is used to describe not only optical radiation visible to the human eye but also radiation in the infrared and ultraviolet spectrum.



Key

- light source 1
- 2 collimator
- 3 incident light
- 4 goniometer containing the telescope (10) and detector (11)
- 5 window
- 8 specimen prism


thermal chamber containing the specimen prism (8)

- 9
- 10 telescope
- 6 rotating stage containing the thermal chamber (7) 12 thermometer

- transmitted light
- 11 detector

Figure 1 — Measurement set-up with thermal chamber

7



Кеу

 t_1, t_2 temperature of specimen prism

 n_1 refractive index of specimen prism at temperature t_1

 n_2 refractive index of specimen prism at temperature t_2

Figure 2 — Conceptual diagram for calculation of temperature coefficient of refractive index

standards.iteh.ai)

5 Measuring apparatus

ISO/DIS 6760-1

5.1 Goniometer discrete discre

The goniometer shall be in accordance with ISO 21395-1:2020, 5.2.

5.2 Light source

The light source shall be in accordance with ISO 21395-1:2020, 5.3.

5.3 Detector

The detector shall be in accordance with ISO 21395-1:2020, 5.4.


5.4 Thermal chamber

The thermal chamber shall follow the requirements below. An example of a thermal chamber is shown in <u>Figure 3</u>. The thermal chamber shall:

- a) have the ability to change the temperature of the specimen prism between the temperatures to be measured.
- b) have a structure that can maintain the temperature distribution in the specimen within the range of 1,0 K during raising and lowering of the temperature and a capability to realize the temperature change rate..
- c) have a thermometer to measure the temperature of the specimen prism with an accuracy of ± 0.2 K or better.
- d) have the ability to provide a vacuum with a residual pressure of less than 10 Pa for the purpose of having a negligible influence of the refractive index of air and of preventing condensation.

e) have windows made of a parallel plate of quartz glass polished on both sides. The wedge angle between the parallel polished faces shall not exceed 5 arc sec, the flatness of the parallel polished faces shall be $\lambda/10$ or better.

NOTE Quartz glass is used because it has a high transmittance over a wide wavelength range, a high durability against temperature changes, and is resistant to breakage.



Key https://standards.iteh.ai/catalog/standards/sist/af90dc53-5da2-46ab-8a76-ce2df00fb24a/iso-

- 3 incident light
- 9 outgoing light
- 5 window
- 8 specimen prism
- 12 thermometer

- 13⁰ thermal conductor specimen holder
- 14 heating and cooling unit15 three-way valve
- 16 to vacuum pump
- 17 leak inlet

Figure 3 — Example of thermal chamber

6 Specimen prism

The specimen prism shall be in accordance with ISO 21395-1:2020, 6.

7 Measurement

7.1 Measurement of apex angle

The apex angle of the specimen prism shall be measured in accordance with ISO 21395-1:2020, 8.2.

7.2 Measurement of the angle of minimum deviation

The angle of minimum deviation of the specimen prism shall be measured at two or more temperatures in accordance with ISO 21395-1:2020, 8.3.

The bisector of the apex angle (α) is parallel to the bisector of the angle (β) formed by the opposite twosurface window of the thermal chamber. (See Figure 4).

The degree of vacuum around the specimen prism shall be less than 10 Pa. The minimum deviation angle should be measured at a temperature within ± 0.5 °C with respect to the target temperature.

NOTE 1 When the allowable measurement error is smaller than 0.5×10^{-6} , the allowable angle difference between the bisectors of α and β is within 2 °; when the allowable measurement error is smaller than 0,5 × 10⁻⁵, the allowable angle difference between the bisectors of α and β is within 6 °.

The temperature to be measured is arbitrary. Allow sufficient time for the specimen prism to reach a uniform temperature throughout. In most cases, the temperatures measured are -40 °C, -20 °C, 0 °C, 20 °C, 40 °C, 60 °C and 80 °C.

Calculation 8

8.1 Absolute refractive index

The absolute refractive index at each temperature of the specimen prism shall be calculated by (1) (adaptation of ISO 21395-1:2020, 4).

$$n_{\rm abs}(t) = \frac{\sin\left(\frac{\alpha + \delta_{\rm min,vac}(t)}{2}\right)}{\prod_{\rm sin}\left(\frac{\alpha}{2}\right)}$$
(1)
ere

whe

 $n_{\rm abs}(t)$ is the absolute refractive index of specimen prism at temperature *t*;

is the apex angle of the specimen prism;

 $\frac{\alpha}{https://standar}$

 $\delta_{\min}(t)$ is the minimum deviation angle at temperature *t*;

t is the temperature (°C) of the specimen prism during the measurement (°C).

In ISO 21395-1 the measurements are performed in air, therefore the refractive index *n* obtained is NOTE the relative refractive index. In this document, the measurements are performed in vacuum, and therefore the result obtained by (1) is the absolute refractive index.

Figure 4 shows a schematic drawing of the light path through the thermal chamber windows and the specimen prism. The internal and external environments are air and vacuum respectively. As a consequence, light transmitted through a parallel window at non-normal incidence will be deflected.

Consequently the minimum angle of deflection in vacuum $\delta_{\min, vac}$ must be calculated using the correction formula (2) to the observed angle of minimum deflection in air $\delta_{min,air}$