



FINAL DRAFT International Standard

IEC/FDIS 80601-2-71

Medical electrical equipment — Part 2-71: Particular requirements for the basic safety and essential performance of functional near- infrared spectroscopy (functional NIRS) equipment

Appareils électromédicaux —

*Partie 2-71: Exigences particulières pour la sécurité de base
et les performances essentielles des appareils d'imagerie
spectroscopique proche infrarouge (NIRS)*

ISO/TC 121/SC 3

Secretariat: **ANSI**

Voting begins on:
2024-09-20

Voting terminates on:
2024-11-15

[IEC/FDIS 80601-2-71](#)

[e968-4d95-84e2-9392957f96ee/iec-fdis-80601-2-71](#)

This draft is submitted to a parallel vote in ISO and in IEC.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

IEC/FDIS 80601-2-71

<https://standards.iteh.ai/catalog/standards/iso/26ee0fe8-e968-4d95-84e2-9392957f96ee/iec-fdis-80601-2-71>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2024

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

CONTENTS

FOREWORD	4
INTRODUCTION	7
201.1 Scope, object and related standards	8
201.2 Normative references	10
201.3 Terms and definitions	10
201.4 General requirements	14
201.5 General requirements for testing ME EQUIPMENT	14
201.6 Classification of ME EQUIPMENT and ME SYSTEMS	14
201.7 ME EQUIPMENT identification, MARKING and documents	14
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	15
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	16
201.10 Protection against unwanted and excessive radiation HAZARDS	16
201.11 Protection against excessive temperatures and other HAZARDS	16
201.12 Accuracy of controls and instruments and protection against hazardous outputs	18
201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT	28
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	28
201.15 Construction of ME EQUIPMENT	28
201.16 ME SYSTEMS	29
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	29
201.101 Requirements for the FUNCTIONAL NIRS EQUIPMENT ACCESSORIES	29
202 Electromagnetic disturbances – Requirements and tests	29
206 Useability	30
Annexes	31
Annex C (informative) Guide to MARKING and labelling requirements for ME EQUIPMENT and ME SYSTEMS	32
Annex AA (informative) Particular guidance and rationale	34
ANNEX BB (normative) Evaluating ME EQUIPMENT performance using the FUNCTIONAL NIRS PHANTOM	38
ANNEX CC (informative) Skin temperature at the FUNCTIONAL NIRS PROBE	47
Annex DD (informative) Reference to the IMDRF essential principles and labelling guidance	48
Bibliography	50
Index of defined terms	54
Figure 201.101 – FULL WIDTH AT HALF MAXIMUM OF SPECTRAL POWER DISTRIBUTION	11
Figure 201.102 – Measurement of the AVERAGE OPTICAL POWER	19
Figure 201.103 – Measurement of PEAK WAVELENGTH and FWHM	20
Figure 201.104 – Measurement of the signal stability	21
Figure 201.105 – Measurement of the RESPONSE TIME	22
Figure 201.106 – Rise time and fall time in the RESPONSE TIME	23

Figure 201.107 – Measurement of the signal-to-noise ratio of the detected light intensity	24
Figure 201.108 – Measurement of signal-to-noise ratio of the PATHLENGTH-DEPENDENT HAEMOGLOBIN CHANGE	26
Figure 201.109 – Measurement of SIGNAL CROSS-TALK	27
Figure BB.1 – The FUNCTIONAL NIRS PHANTOM in two states with different detected light intensities	41
Figure BB.2 – FUNCTIONAL NIRS PHANTOM measurement using the reference system	42
Figure BB.3 – FUNCTIONAL NIRS PHANTOM measurement using the ME EQUIPMENT to be evaluated	42
Figure BB.4 – Schematic for measurement of OPTICAL LOSS	46
Table 201.101 – Distributed ESSENTIAL PERFORMANCE requirements	14
Table 201.102 – Performance tests employing the FUNCTIONAL NIRS EQUIPMENT or attenuator and the required OPTICAL LOSS	18
Table 201.C.101 – MARKING on the outside of FUNCTIONAL NIRS EQUIPMENT or their parts	32
Table 201.C.102 – ACCOMPANYING DOCUMENTS general	32
Table 201.C.103 – INSTRUCTIONS FOR USE	33
Table 201.C.104 – TECHNICAL DESCRIPTION	33
Table DD.1 – Correspondence between this document and the IMDRF essential principles	48
Table DD.2 – Correspondence between this document and the IMDRF labelling principles	49

ITeH Standards
(http://standards.iteh.ai)
Document Preview

[IEC/FDIS 80601-2-71](https://standards.iteh.ai/catalog/standards/iso/26ee0fe8-e968-4d95-84e2-9392957f96ee/iec-fdis-80601-2-71)

<https://standards.iteh.ai/catalog/standards/iso/26ee0fe8-e968-4d95-84e2-9392957f96ee/iec-fdis-80601-2-71>

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –**Part 2-71: Particular requirements for
the basic safety and essential performance of
functional near-infrared spectroscopy (NIRS) equipment****FOREWORD**

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) IEC and ISO draw attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC and ISO take no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC and ISO had not received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at <https://patents.iec.ch> and www.iso.org/patents. IEC and ISO shall not be held responsible for identifying any or all such patent rights.

IEC 80601-2-71 has been prepared by a Joint Working Group of IEC subcommittee 62D: Particular medical equipment, software, and systems, of IEC technical committee 62: Medical equipment, software, and systems, and ISO subcommittee SC3: Respiratory devices and related equipment used for patient care, of ISO technical committee 121: Anaesthetic and respiratory equipment. It is an International Standard.

This second edition cancels and replaces the first edition published in 2015. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) alignment with IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005/AMD2:2020, IEC 60601-1-8:2006, IEC 60601-1-8:2006/AMD1:2012, IEC 60601-1-8:2006/AMD2:2020, IEC 60601-1-2:2014, IEC 60601-1-2:2014/AMD1:2020, IEC 60601-1-6:2010, IEC 60601-1-6:2010/AMD1:2013 and IEC 60601-1-6:2010/AMD2:2020;
- b) added requirements for ESSENTIAL PERFORMANCE;
- c) added requirements for PRIMARY OPERATING FUNCTIONS;
- d) added requirements for protection against excessive temperatures;
- e) added requirements for the display legibility for OPERATORS wearing personal protective equipment;
- f) harmonization with ISO 20417, where appropriate.

This publication is published as a double logo standard.

The text of this International Standard is based on the following documents of IEC:

Draft	Report on voting
62D/XXX/FDIS	62D/XXX/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table. In ISO, the document was approved by XXX P members out of YYY having cast a vote.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/publications.

In this document, the following print types are used:

- requirements and definitions: roman type.
- *test specifications: italic type.*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS defined in Clause 3 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 AND IEC 60601-1:2005/AMD2:2020, IN THIS DOCUMENT OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term

- "clause" means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- "subclause" means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this document are preceded by the term "Clause" followed by the clause number. References to subclauses within this particular standard are by number only.

In this document, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this document;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- "may" is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 and IEC 80601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn, or
- revised.

NOTE The attention of the users of this document is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised IEC or ISO publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committees that the content of this publication be adopted for implementation nationally not earlier than 3 years from the date of publication.

[IEC/FDIS 80601-2-71](https://standards.iteh.ai/catalog/standards/iso/26ee0fe8-e968-4d95-84e2-9392957f96ee/iec-fdis-80601-2-71)

<https://standards.iteh.ai/catalog/standards/iso/26ee0fe8-e968-4d95-84e2-9392957f96ee/iec-fdis-80601-2-71>

INTRODUCTION

The minimum safety requirements specified in this document are considered to provide for a practical degree of safety in the operation of FUNCTIONAL NIRS EQUIPMENT.

The requirements are followed by specifications for the relevant tests.

A "Particular guidance and rationale" text giving some explanatory notes, where appropriate, about the more important requirements is included in Annex AA. It is considered that knowledge of the reasons for these requirements will not only facilitate the proper application of the standard but will, in due course, expedite any revision necessitated by changes in clinical practice or as a result of developments in technology. However, this annex does not form part of the requirements of this document.

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[IEC/FDIS 80601-2-71](https://standards.iteh.ai/catalog/standards/iso/26ee0fe8-e968-4d95-84e2-9392957f96ee/iec-fdis-80601-2-71)

<https://standards.iteh.ai/catalog/standards/iso/26ee0fe8-e968-4d95-84e2-9392957f96ee/iec-fdis-80601-2-71>

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-71: Particular requirements for the basic safety and essential performance of functional near-infrared spectroscopy (NIRS) equipment

201.1 Scope, object and related standards

Clause 1 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 applies, except as follows:

201.1.1 Scope

Replacement:

This part of IEC 80601 applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of FUNCTIONAL NIRS EQUIPMENT, as defined in 201.3.205, intended to be used by itself, or as a part of an ME SYSTEM hereinafter referred to as ME EQUIPMENT.

HAZARDS inherent in the intended physiological function of ME EQUIPMENT or ME SYSTEMS within the scope of this document are not covered by specific requirements in this document except in IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020, 7.2.13 and 8.4.1.

NOTE Additional information can be found in IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020, 4.2.

This document is not applicable to

- equipment for the measurement of oxygen saturation of the haemoglobin in the micro vessels (capillaries, arterioles and venules), i.e. tissue oximeters;
- frequency-domain and time-domain equipment for functional near-infrared spectroscopy;
- equipment for the measurement of changes in the concentration of chromophores other than oxy- and deoxy-haemoglobin;
- equipment for the measurement of changes in the concentration of oxy- and deoxy-haemoglobin in tissues other than the brain.

This document does not specify the requirements for:

- cerebral tissue oximeter equipment, which are given in ISO 80601-2-85 [1]¹; and
- pulse oximeter equipment, which are given in ISO 80601-2-61 [2].

201.1.2 Object

Replacement:

The object of this document is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for FUNCTIONAL NIRS EQUIPMENT as defined in 201.3.205.

NOTE This document has been prepared to address the relevant essential principles [3] and labelling principles [4] of the International Medical Devices Regulators Forum (IMDRF) as indicated in Annex DD.

¹ Numbers in square brackets refer to the Bibliography.

201.1.3 Collateral standards

Addition:

This document refers to those applicable collateral standards that are listed in Clause 2 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 and Clause 201.2 of this document.

IEC 60601-1-3 and IEC 60601-1-10 [5] do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

201.1.4 Particular standards

Addition:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020, including the collateral standards, as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY or ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 and applicable collateral standards.

The numbering of clauses and subclauses of this particular standard corresponds to that of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 with the prefix "201" (e.g. 201.1 in this document addresses the content of Clause 1 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020) or applicable collateral standard with the prefix "20x" where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this document addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 203.6 in this document addresses the content of Clause 6 of the IEC 60601-1-3 collateral standard, etc.). The changes to the text of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 are specified by the use of the following words:

"Replacement" means that the clause or subclause of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard is replaced completely by the text of this document.

"Addition" means that the text of this document additional to the requirements of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard.

"Amendment" means that the clause or subclause of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard is amended as indicated by the text of this document.

Subclauses, figures or tables which are additional to those of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 are numbered starting from 201.101. However due to the fact that definitions in IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 are numbered 3.1 through 3.154, additional definitions in this document are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

Where there is no corresponding clause or subclause in this document, the clause or subclause of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this document.

201.2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

Clause 2 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 applies, except as follows:

Addition:

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*
IEC 60601-1:2005/AMD1:2012
IEC 60601-1:2005/AMD2:2020

IEC 60825-1:2014, *Safety of laser products – Part 1: Equipment classification and requirements*

IEC 62471:2006, *Photobiological safety of lamps and lamp systems*

IEC 62570:2014, *Standard practice for marking medical devices and other items for safety in the magnetic resonance environment*

ISO 17664-1:2021, *Processing of health care products – Information to be provided by the medical device manufacturer for the processing of medical devices – Part 1: Critical and semi-critical medical devices*

ISO 17664-2:2021, *Processing of health care products – Information to be provided by the medical device manufacturer for the processing of medical devices – Part 2: Non-critical medical devices*

ISO 20417:2021, *Medical devices – Information to be supplied by the manufacturer*

201.3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005/AMD2:2020, ISO 17664-2:2021, ISO 20417:2021 and the following apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- IEC Electropedia: available at <https://www.electropedia.org/>
- ISO Online browsing platform: available at <https://www.iso.org/obp>

201.3.201**AVERAGE OPTICAL POWER**

temporal average power of continuous light or repeated light pulses from each discrete wavelength, from the EMITTER OPTODE connected to the FUNCTIONAL NIRS MONITOR

201.3.202**DETECTOR OPTODE**

part of the FUNCTIONAL NIRS PROBE which detects light from the living tissue that forms part of the APPLIED PART

2201.3.203**EMITTER OPTODE**

part of the FUNCTIONAL NIRS PROBE which emits light to the living tissue that forms part of the APPLIED PART

Note 1 to entry: An EMITTER OPTODE can contain two or more light sources (e.g. laser diodes or light-emitting diodes) operating at different NOMINAL wavelengths.

201.3.204**FWHM****FULL WIDTH AT HALF MAXIMUM OF SPECTRAL POWER DISTRIBUTION**

difference of the wavelength between the two points whose corresponding power values are equal and 3 dB lower than the values at each PEAK WAVELENGTH

Note 1 to entry: FWHM is a measure of the width of the spectral power distribution emitted by the EMITTER OPTODE connected to the FUNCTIONAL NIRS MONITOR. Figure 201.101 provides a visual representation. If there are more than two wavelengths where power VALUE is 3 dB lower than the values at each PEAK WAVELENGTH, FWHM shall be calculated from the difference between minimum and maximum wavelengths.




Figure 201.101 – FULL WIDTH AT HALF MAXIMUM OF SPECTRAL POWER DISTRIBUTION

201.3.205**FUNCTIONAL NIRS EQUIPMENT****FUNCTIONAL NEAR-INFRARED SPECTROSCOPY EQUIPMENT**

ME EQUIPMENT that estimates PATHLENGTH-DEPENDENT HAEMOGLOBIN CHANGES in living tissue by illuminating tissue with continuous light and detecting changes in the infrared and visible light intensity diffusively reflected from the tissue

201.3.206

FUNCTIONAL NIRS MONITOR

FUNCTIONAL NEAR-INFRARED SPECTROSCOPY MONITOR MONITOR

part of the FUNCTIONAL NIRS EQUIPMENT that encompasses the electronics, display and operator-equipment interface excluding the EMITTER OPTODES and DETECTOR OPTODES

Note 1 to entry: The FUNCTIONAL NIRS MONITOR typically encompasses the electronics, display and OPERATOR-EQUIPMENT INTERFACE. It can consist of a separate control unit, a computer and a display. Electronics or a control unit integrated in the APPLIED PART are not part of the FUNCTIONAL NIRS MONITOR.

201.3.207

FUNCTIONAL NIRS PROBE

FUNCTIONAL NEAR-INFRARED SPECTROSCOPY PROBE PROBE

part of the functional NIRS EQUIPMENT that includes the APPLIED PART

Note 1 to entry: This definition of PROBE is consistent with the definition of PROBE for cerebral tissue oximeters as given in ISO 80601-2-85.

Note 2 to entry: The PROBE contains single or multiple EMITTER OPTODES and DETECTOR OPTODES. A reflectance PROBE design is the typical configuration.

201.3.208

FUNCTIONAL NIRS PHANTOM

apparatus that simulates a PATHLENGTH-DEPENDENT HAEMOGLOBIN CHANGE by giving the ME EQUIPMENT a specified known change in attenuation to evaluate the difference between the measured VALUE of the pseudo PATHLENGTH-DEPENDENT HAEMOGLOBIN CHANGE obtained from the measurement on the phantom and the reference VALUE calculated from the attenuation change

Note 1 to entry: The FUNCTIONAL NIRS PHANTOM plays a role in determining the performance of FUNCTIONAL NIRS EQUIPMENT, especially PATHLENGTH-DEPENDENT HAEMOGLOBIN CHANGE measurement. A description of the function and specifications regarding the manufacturing of the FUNCTIONAL NIRS PHANTOM is found in Annex BB.

Note 2 to entry: A FUNCTIONAL NIRS PHANTOM is developed during design and is used at the time of inspection in manufacturing or after being placed into service.

Note 3 to entry: The change in attenuation corresponds to a change in OPTICAL LOSS of the FUNCTIONAL NIRS PHANTOM of the same magnitude.

201.3.209

MEASUREMENT CHANNEL

CHANNEL

combination of an EMITTER OPTODE and a DETECTOR OPTODE that provide an output

201.3.210

OPTICAL LOSS

ratio of the total optical power exiting the FUNCTIONAL NIRS PHANTOM or attenuator through a specified aperture, to the optical power emitted by the EMITTER OPTODE connected to the FUNCTIONAL NIRS MONITOR and placed on the entrance side of the FUNCTIONAL NIRS PHANTOM or attenuator

Note 1 to entry: The OPTICAL LOSS is denoted in dB. X dB OPTICAL LOSS is equivalent to $10^{-X/10}$.

Note 2 to entry: The optical power exiting the EMITTER OPTODE and the FUNCTIONAL NIRS PHANTOM can be measured with an optical power meter.

Note 3 to entry: For details of the measurement of the OPTICAL LOSS, refer to BB.3.3.

201.3.211

PATHLENGTH-DEPENDENT DEOXYHAEMOGLOBIN CHANGE

VALUE calculated from the changes in detected light intensities, equal to the product of the apparent change in the concentration of DEOXYHAEMOGLOBIN and the mean optical pathlength