

ISO/TC 28
ISO/DIS 3679:2022(E)

2022-06-03-08-31

ISO/TC 28/WG 9

Secretariat: NEN

Determination of flash point – Method for flash no-flash and flash point by small scale closed cup tester

Détermination du point d'éclair — ~~Méthode pour éclair sans éclair~~ Détermination de l'éclair de type passe / ne passe pas et méthode du point d'éclair par testeur de vase vase clos à petite échelle

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO/PRF 3679

<https://standards.iteh.ai/catalog/standards/sist/94b8c515-6966-4363-a317-a5360d974172/iso-prf-3679>

Style Definition: Heading 1: Indent: Left: 0 pt, First line: 0 pt, Tab stops: Not at 21.6 pt

Style Definition: Heading 2: Font: Bold, Tab stops: Not at 18 pt

Style Definition: Heading 3: Font: Bold

Style Definition: Heading 4: Font: Bold

Style Definition: Heading 5: Font: Bold

Style Definition: Heading 6: Font: Bold

Style Definition: ANNEX

Style Definition: ANNEXN: Font: Bold

Style Definition: ANNEXZ: Font: Bold

Style Definition: List Number

Style Definition: na2

Style Definition: na3

Style Definition: Bibliography1: Font: Indent: Left: 0 pt, Hanging: 18 pt, No bullets or numbering, Tab stops: 18 pt, List tab

Style Definition: AMEND Terms Heading: Font: Bold

Style Definition: AMEND Heading 1 Unnumbered: Font: Bold

Style Definition: Unresolved Mention1

Formatted: English (United Kingdom)

Formatted: Top: 39.7 pt

Formatted: English (United Kingdom)

Formatted: Right

Edited DIS - MUST BE USED FOR FINAL DRAFT

ISO 3679:2022(E)

© ISO 2022

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO ~~copyright office~~ Copyright Office

CP 401 • ~~Ch. de Blandonnet 8~~

CH-1214 Vernier, Geneva

Phone: +41 22 749 01 11

Fax: +41 22 749 09 47

Email: copyright@iso.org

Email: copyright@iso.org

Website: www.iso.org

Published in Switzerland.

Formatted: Default Paragraph Font

Formatted: Default Paragraph Font

Formatted: Indent: Left: 14.2 pt, Right: 14.2 pt, Space Before: 0 pt, No page break before

Formatted: Indent: Left: 14.2 pt, Right: 14.2 pt

Formatted: Indent: Left: 14.2 pt, First line: 0 pt, Right: 14.2 pt

Formatted: Indent: Left: 14.2 pt, First line: 0 pt, Right: 14.2 pt

Formatted: English (United Kingdom)

Formatted: English (United Kingdom)

Formatted: English (United Kingdom)

Formatted: Indent: Left: 14.2 pt, First line: 0 pt, Right: 14.2 pt

Formatted: Indent: Left: 14.2 pt, Right: 14.2 pt

Formatted: Left

Formatted: Left

~~ISO/DIS 3679:2022(E)~~

Formatted: Font: 11.5 pt

Formatted: Font: 11.5 pt

~~CONTENTS PAGE~~

iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO/PRF 3679

<https://standards.iteh.ai/catalog/standards/sist/94b8c515-6966-4363-a317-a5360d974172/iso-prf-3679>

Formatted: Right

© ISO 2022 – All rights reserved

iii

Edited DIS - MUST BE USED FOR FINAL DRAFT

ISO/DIS 3679:2022(E)

Formatted: Font: 11.5 pt

Formatted: Font: 11.5 pt

B.3 Procedure	22
B.3 Procedure	22
B.4 Evaluation of the test result	23
Annex C (informative) The use of a cup insert	26
C.1 General	26
C.2 Cup insert material	26
C.3 Procedure	26
C.3 Procedure	26
C.4 Verification	26
Annex D (normative) Sub-ambient Procedures procedures A and B testing for manual apparatus without integral test cup cooling facilities	27
D.1 Apparatus preparation	27
D.2 Procedure	27
Annex E (normative) Temperature measuring device specifications	28
E.1 Digital measuring devices	28
E.2 Liquid-in-glass thermometers	28
Annex F (informative) Flash point values of chemicals	30
Bibliography	31

Field Code Changed

Formatted: English (United States)

Formatted: English (United States)

Formatted: English (United States)

Field Code Changed

Formatted: English (United States)

Field Code Changed

Formatted: English (United States)

Formatted: English (United States)

Field Code Changed

Formatted: English (United States)

Formatted: English (United States)

Field Code Changed

Formatted: English (United States)

Formatted: English (United States)

Field Code Changed

Formatted: English (United States)

Formatted: English (United States)

Field Code Changed

Formatted: English (United States)

Field Code Changed

Formatted: English (United States)

ISO/PRF 3679

<https://standards.iteh.ai/catalog/standards/sist/94b8c515-6966-4363-a317-a5360d974172/iso-prf-3679>

Formatted: Right

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Formatted: English (United Kingdom)

Formatted: Don't adjust space between Latin and Asian text, Don't adjust space between Asian text and numbers

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Formatted: English (United Kingdom)

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Formatted: English (United Kingdom)

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

Formatted: English (United Kingdom)

This document was prepared by Technical Committee ISO/TC 28, *Petroleum products and lubricants*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 19, *Gaseous and liquid fuels, lubricants and related products of petroleum, synthetic and biological origin*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

Formatted: English (United Kingdom)

Formatted: Font: Not Italic, English (United Kingdom)

Formatted: English (United Kingdom)

Formatted: English (United Kingdom)

Formatted: Font: Not Italic, English (United Kingdom)

Formatted: English (United Kingdom)

Formatted: Default Paragraph Font

This fifth edition cancels and replaces the fourth edition (ISO 3679:2015), which has been technically revised.

The main changes are as follows:

- introduction, title and scope have been revised to present a more generic method description;
- terms and definitions in Clause 3 have been added;
- verification clause has been revised;
- new procedure C has been added ;
- Revision of Clause 13 wording has been revised and inclusion of precision for Procedure C has been included;
- the apparatus description in Annex A has been revised;
- Annex B has been revised and changed to normative;

Formatted: cite_sec

Formatted: cite_sec

Formatted: cite_sec

Formatted: cite_sec

Formatted: cite_app

Formatted: cite_app

Formatted: cite_app

Formatted: cite_app

Formatted: Left

Formatted: Left

~~ISO/DIS 3679:2022(E)~~

Formatted: Font: 11.5 pt

Formatted: Font: 11.5 pt

— the text has been editorially revised in line with the ISO/IEC Directives Part 2, 2021.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html www.iso.org/members.html.

Formatted: English (United Kingdom)

Formatted: Don't adjust space between Latin and Asian text, Don't adjust space between Asian text and numbers

Formatted: English (United Kingdom)

iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO/PRF 3679

<https://standards.iteh.ai/catalog/standards/sist/94b8c515-6966-4363-a317-a5360d974172/iso-prf-3679>

Formatted: Right

ISO 3679:2022(E)

Introduction

This document includes three procedures (A, B and C) covering determinations of flash no-flash and flash point. Rapid equilibrium procedures A and B enable the determination of the flash no-flash and flash point, respectively. Non-equilibrium procedure C uses automated test cup temperature control for flash point determination.

ISO 1516 and ISO 1523 are also closed cup equilibrium test methods that can be considered when selecting a method.

The apparatus specified in this document enables a similar test result to be determined using more rapid procedures, A or B, and a smaller test portion (2 ml or 4 ml), than those required in ISO 1516 or ISO 1523. In addition, the apparatus in this document can be made portable so that it is suitable for on-site testing, as well as its regular use in laboratories. Collaborative work^[4916] has shown that results obtained by these methods are comparable. Procedure C is based on test methods IP 534^[2018] and ASTM D7236^[4614].

The interpretation of flash point results obtained on solvent mixtures containing halogenated hydrocarbons should be considered with caution, as these mixtures can give anomalous results^[4917].

A limited study has indicated that some water borne paints can give an elevated flash point when an electric ignitor is used with ISO 3679.

Flash point is used in shipping, storage, handling, and safety regulations, as a classification property to define “flammable” and “combustible” materials. Precise definition of the classes is given in each particular regulation.

The flash point indicates the presence of highly volatile material(s) in a relatively non-volatile or non-flammable material. Flash point testing is often used as a preliminary step to other investigations into the composition of unknown materials.

It is not appropriate for flash point determinations to be carried out on potentially unstable, decomposable, or explosive materials. That is, unless it has been previously established that heating the specified quantity of such materials in contact with the metallic components of the flash point apparatus, within the temperature range required for the method, does not induce decomposition, explosion or other adverse effects.

The flash point is not a constant physical-chemical property of a material tested. It is a function of the apparatus design, the condition of the apparatus used, and the operational procedure carried out. Flash point can therefore only be defined in terms of a standard test method, and no general valid correlation can be guaranteed between results obtained by different test methods or with test apparatus different from that specified.

ISO/TR 29662 also gives useful advice in carrying out flash point tests and interpreting results.

Formatted: Default Paragraph Font

Formatted: Default Paragraph Font

Formatted: Left

Formatted: Left

Formatted: Font: 11.5 pt

Formatted: Font: 11.5 pt

Determination of flash point – Method for flash no-flash and flash point by small scale closed cup tester

Formatted: Section start: New page

WARNING — The use of this document can involve hazardous materials, operations and equipment. This document does not purport to address all of the safety problems associated with its use. It is the responsibility of users of this document to take appropriate measures to ensure the safety and health of personnel prior to the application of the standard, and to determine the applicability of any other restrictions for this purpose.

1 Scope

This document describes three procedures (A, B and C) covering determinations of flash no-flash and flash point.

Rapid equilibrium procedures A and B are applicable to flash no-flash and flash point tests of paints, including water-borne paints, varnishes, binders for paints and varnishes, adhesives, solvents, petroleum products including aviation turbine, diesel and kerosene fuels, fatty acid methyl esters and related products over the temperature range -30 °C to 300 °C. The rapid equilibrium procedures are used to determine whether a product will or will not flash at a specified temperature (flash no-flash procedure A) or the flash point of a sample (procedure B). When used in conjunction with the flash detector (A.1.6), this document is also suitable to determine the flash point of fatty acid methyl esters (FAME). The validity of the precision is given in Table 2.

Non-equilibrium procedure C is applicable to petroleum products including aviation turbine, diesel and kerosene fuels, and related petroleum products, over the temperature range -20 °C to 300 °C. The non-equilibrium procedure is automated to determine the flash point. Precision has been determined over the range 40 °C to 135 °C.

For specifications and regulations, procedures A or B are routinely used (see 10.1.1).

2 Normative references

The following documents, ~~are referred to in whole the text in such a way that some or in part, are normatively referenced in all of their content constitutes requirements of this document and are indispensable for its application.~~ For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO Guide 35, Reference materials — Guidance for characterization and assessment of homogeneity and stability

ISO 1513, *Paints and varnishes — Examination and preparation of test samples*

ISO 3170, *Petroleum liquids — Manual sampling*

ISO 3171, *Petroleum liquids — Automatic pipeline sampling*

ISO 15528, *Paints, varnishes and raw materials for paints and varnishes — Sampling*

ISO 17034, General requirements for the competence of reference material producers

Formatted: Default Paragraph Font

Formatted: RefNorm, Tab stops: 19.85 pt, Left + 39.7 pt, Left + 59.55 pt, Left + 79.4 pt, Left + 99.25 pt, Left + 119.05 pt, Left + 138.9 pt, Left + 158.75 pt, Left + 178.6 pt, Left + 198.45 pt, Left

Formatted: Font: 10 pt, French (Switzerland)

Formatted: Left, Space Before: 18 pt, After: 0 pt, Line spacing: Exactly 12 pt

Formatted: Font: 10 pt, French (Switzerland)

Formatted: French (Switzerland)

Formatted: French (Switzerland)

Formatted: Space Before: 18 pt, After: 0 pt, Line spacing: Exactly 12 pt

ISO 3679:2022(E)

4.3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses;

— ISO Online browsing platform: available at <https://www.iso.org/obp>

— IEC Electropedia: available at <https://www.electropedia.org/>

3.1 equilibrium

condition in flash point test methods where the vapour above the test portion and the test portion are at the same temperature at the time the ignition source is applied

Note 1 to entry: This condition ~~may not~~cannot be fully achieved in practice, since the temperature can be uneven throughout the test portion, and the test cover and shutter on the apparatus can be cooler or warmer.

EXAMPLE ___ Procedures A and B in this document, ISO 1516 and ISO 1523.

3.2 fatty acid methyl ester FAME

fuel comprising mono-alkyl esters of long chain fatty acids derived from vegetable oil or animal fats, designated B100 or biodiesel (100 %)

Note 1 to entry: FAME is specified in specifications such as EN 14214^[4210] and ASTM D6751^[4513].

3.3 flash no-flash

application of an ignition source at the specified temperature of the test portion, as measured in the prescribed manner, adjusted to account for variations in atmospheric pressure from 101,3 kPa, to determine whether the vapour of the test portion ignites and the flame propagates across the surface of the liquid under the specified conditions of test

3.4 flash point

lowest temperature of the test portion, adjusted to account for variations in atmospheric pressure from 101,3 kPa, at which application of an ignition source causes the vapour of the test portion to ignite and the flame to propagate across the surface of the liquid under the specified conditions of test

3.5 non-equilibrium

condition in flash point test methods where the vapour above the test portion and the test portion are not in temperature equilibrium at the time that the ignition source is applied

Note 1 to entry: This condition is primarily caused by the heating of the test portion at the constant prescribed rate with the vapour temperature lagging behind the test portion temperature.

EXAMPLE Procedure C in this document, ISO 2719 and ISO 13736.

Formatted: Don't adjust space between Latin and Asian text, Don't adjust space between Asian text and numbers

Formatted: English (United Kingdom)

Formatted: Font: Cambria, 11 pt, English (United Kingdom)

Formatted: No underline, Font color: Auto, English (United Kingdom)

Formatted: Don't adjust space between Latin and Asian text, Don't adjust space between Asian text and numbers, Tab stops: 19.85 pt, Left + 39.7 pt, Left + 59.55 pt, Left + 79.4 pt, Left + 99.25 pt, Left + 119.05 pt, Left + 138.9 pt, Left + 158.75 pt, Left + 178.6 pt, Left + 198.45 pt, Left

Formatted: English (United Kingdom)

Formatted: Default Paragraph Font, English (United Kingdom)

Formatted: No underline, Font color: Auto, English (United Kingdom)

Formatted: English (United Kingdom)

Formatted: Default Paragraph Font, English (United Kingdom)

Formatted: Example

Formatted: Example

5.4 Principle

5.4.1 Rapid equilibrium procedures A and B

A 2-ml or 4-ml test portion is introduced into the test cup that is set and maintained at the required test temperature. After a specific time, when the vapours and test portion are deemed to be in temperature equilibrium, an ignition source is applied and a determination is made (Procedure A) as to whether or not a flash occurred. In order to determine the actual flash point of the sample, further tests, with fresh test portions at different test cup temperatures, are carried out (Procedure B) until the flash point is determined. The temperature is adjusted to account for variation in atmospheric pressure from 101,3 kPa, using a formula.

5.4.2 Non-equilibrium procedure C

A 2-ml test portion is introduced into the test cup that is set and maintained at the required start temperature. The test cup is then heated at a ramp rate of 1,5 °C/min to 2,5 °C/min and the ignition source is applied at 1 °C temperature intervals until a flash point is detected. The detected flash point temperature is adjusted to account for variation in atmospheric pressure from 101,3 kPa, using a formula.

6.5 Reagents and materials

6.5.1 Cleaning solvent, for the removal of traces of sample from the test cup and cover.

The choice of solvent depends upon the previous material tested and the tenacity of the residue. Low volatility aromatic (benzene-free) solvents can be used to remove traces of oil, and mixed solvents can be effective for the removal of gum-type deposits.

6.5.2 Reference materials, for flash point, certified reference materials (CRM) and/or secondary working standards (SWS), as described in Annex B.

6.5.3 Gas for ignitor and pilot flame, not required if an electric ignitor is used. Butane, propane, coal gas, or natural gas may be used.

6.6 Apparatus

6.6.1 Flash point apparatus, as specified in Annex A.

6.6.2 Barometer, absolute pressure reading, accurate to $\pm 0,5$ kPa. Barometers pre-corrected to give sea level readings, such as those used at weather stations and airports, shall not be used.

6.6.3 Heating bath or oven (optional), for warming the sample, if required.

The bath and oven shall be suitable for use with volatile and flammable materials.

6.6.4 Cooling bath or freezer (optional), for cooling the samples, if required, and capable of cooling the sample to 10 °C below the expected flash point.

The bath and freezer shall be suitable for use with volatile and flammable materials.

6.6.5 Draught shield (optional), if required to minimize draughts, a shield fitted at the back and on two sides of the instrument.

6.6.6 Cup insert (optional), see Annex C.

ISO 3679:2022(E)

For samples that are difficult to remove, a thin metal cup insert can be used but the precision has not been determined.

6.7 Syringes

6.7.1 Syringe, capable of delivering 2,00 ml ± 0,05 ml and equipped with a nozzle suitable for the required test temperature and apparatus.

Formatted: Font: Not Bold

To enable a 4 ml test portion to be used, this syringe can be used twice.

6.7.2 Syringe, capable of delivering 4,00 ml ± 0,10 ml and equipped with a nozzle suitable for the required test temperature and apparatus.

Formatted: Font: Not Bold

8.7 Preparation of apparatus

Formatted: Indent: Left: 0 pt, First line: 0 pt

8.7.1 General

7.1.1 Select the appropriate instrument for the relevant procedure and the expected flash point temperature. Follow the manufacturer’s instructions for the correct set-up, verification (see 7.4) and operation of the apparatus, especially the operation and setting of the ignition source.

7.1.1.2 Procedure C is automated and requires automated temperature ramp control; sub-ambient testing requires integrated cooling (see A.1.5).

7.1.2.3 The use of a cup insert (6.6) for potentially adherent materials is described in Annex C.

7.1.3.4 When testing FAME (procedures A and B), use a 2 ml ± 0,05 ml test portion and a 60 s ± 2 s test time, combined with an electronic thermal flash detector (see A.1.6).

7.1.4.5 For sub-ambient test temperatures, use Annex D, unless the apparatus has integral test cup cooling facilities.

8.7.2 Location of apparatus

Support the apparatus specified in Annex A on a level and steady surface in a draught-free position.

A draught shield (6.5) ~~should be~~ used when protection from draughts is not available.

WARNING — When testing materials which can produce toxic vapours, the apparatus should be located in a fume hood with an individual control of air flow, adjusted such that the vapours are withdrawn without causing air currents around the test cup during the test.

8.7.3 Cleaning of the test cup assembly and accessories

Clean the test cup cover and its accessories with an appropriate solvent (5.1) to remove traces of gum or residue from the previous test. Wipe dry to remove all traces of solvent.

Follow the manufacturer’s instructions for the care and servicing of the instrument, especially regarding electronic ignitors and flash detectors which can be fragile.

A stream of clean dry air, such as compressed air, can be used to remove the last traces of solvent used.

The filler orifice can be cleaned using a suitable cleaning device such as a small brush.

8.7.4 Apparatus verification

7.4.1 Check the temperature measuring devices and barometer at least once a year to ensure that they are in accordance with A.1.4, Annex E and 6.2, respectively.

7.4.2 Ensure the correct operation of ignition sources, in accordance with the ~~manufacturers'~~ ~~manufacturer's~~ instructions and this test method.

7.4.3 Verify the accuracy of the apparatus at least once a year by testing a CRM (see 5.2 and Annex B). It is recommended that more frequent verification checks are made using a reference material (see 5.2 and Annex B).

7.4.4 The result of a single test obtained for a reference material shall be equal to or less than $R/\sqrt{2}$ from the certified value of the CRM or from the accepted reference value (ARV) of the SWS, where R is the reproducibility of the test procedure.

NOTE These reference materials (RM) and in-house quality control samples can also be used to monitor stability and establish statistical control limits, according to ISO 4259-4 or equivalent standard, if required.

7.4.5 The numerical values obtained during the verification check shall not be used to provide a bias statement, nor shall they be used to make any correction to the flash points subsequently determined using the apparatus.

7.4.6 If the instrument fails the verification test, it is recommended that the operator follow the ~~manufacturers'~~ ~~manufacturer's~~ instructions and check the following, and then repeat the verification check:

- a) the cover makes a vapour tight seal with the test cup;
- b) the shutter provides a light tight seal;
- c) adequate heat transfer paste surrounds the temperature measuring device inserted in the test cup block;
- d) the ignition source operates correctly;
- e) the flash detector (A.1.6) operates **correctly** (if fitted);
- f) ~~the~~ the temperature measuring device reads correctly.

9.8 Sampling

8.1 Unless otherwise specified, obtain samples in accordance with the procedures given in ISO 1513, ISO 15528, ISO 3170, or ISO 3171 or an equivalent national standard.

8.2 Place sufficient sample volume for testing in a tightly-sealed container made of material appropriate to the liquid being sampled, and for safety purposes, ensure that the sample container is only filled to between 85 % and 95 % of its capacity.

8.3 Store the samples in conditions to minimize vapour loss and pressure build up. Avoid storage of samples at temperatures in excess of 30 °C.