

ISO/~~DIS-PRF~~ 7935:2023(E)

ISO/~~TC_146/SC_1/WG_34~~

Secretariat:-BIS

Date: 2023-12-22

Stationary ~~Source~~source emissions — Determination of the mass concentration of sulfur dioxide in flue gases — Performance characteristics of automated measuring systems

~~Second edition~~

~~Date: 2023-06-16~~

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

ISO 7935:2024

<https://standards.iteh.ai/catalog/standards/iso/de5f0943-72d4-496d-bf59-7dd753d757f8/iso-7935-2024>

Émissions de sources fixes — Détermination de la concentration en masse de dioxyde de soufre —
Caractéristiques de performance des systèmes de mesurage automatiques

PROOF

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

ISO 7935:2024

<https://standards.iteh.ai/catalog/standards/iso/de5f0943-72d4-496d-bf59-7dd753d757f8/iso-7935-2024>

Edited DIS - MUST BE USED FOR FINAL DRAFT

© ISO 2023

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ~~ISO's~~ISO's member body in the country of the requester.

ISO ~~Copyright Office~~copyright office
CP 401 • ~~Ch. de Blandonnet 8~~
CH-1214 Vernier, Geneva
Phone: + 41 22 749 01 11

~~Email: copyright@iso.org~~

~~E-mail: copyright@iso.org~~

~~Website: www.iso.orgwww.iso.org~~

Published in Switzerland.

iTeh Standards (<https://standards.iteh.ai>) Document Preview

[ISO 7935:2024](#)

<https://standards.iteh.ai/catalog/standards/iso/de5f0943-72d4-496d-bf59-7dd753d757f8/iso-7935-2024>

Contents

Foreword	vii
Introduction.....	ix
1 Scope	1
2 Normative references.....	1
3 Terms and definitions.....	1
4 Symbols and abbreviated terms	5
4.1 Symbols.....	5
4.2 Abbreviated terms.....	5
5 Principle.....	6
6 Description of the automated measuring systems.....	6
6.1 Sampling and sample gas conditioning systems.....	6
6.2 Analyser equipment.....	6
7 Performance characteristics and criteria.....	6
7.1 Performance criteria.....	6
7.2 Determination of the performance characteristics.....	7
7.2.1 Performance test.....	7
7.2.2 Ongoing quality control.....	8
8 Selection and installation procedure.....	8
8.1 Choice of the measuring system.....	8
8.2 Sampling.....	9
8.2.1 Sampling location.....	9
8.2.2 Representative sampling.....	9
8.3 Calculation.....	9
8.3.1 Conversion from volume to mass concentration for SO ₂	9
9 Quality assurance and quality control procedures.....	10
9.1 General.....	10
9.2 Frequency of checks.....	10
9.3 Calibration, validation and measurement uncertainty.....	11
10 Test report.....	11
Annex A (informative) Extractive SO ₂ measurement systems.....	13
Annex B (informative) In situ SO ₂ measurement systems.....	24
Annex C (normative) Operational gases.....	30
Annex D (normative) Procedures for determination of the performance characteristics.....	31
Annex E (informative) Examples of the results for the assessment of SO ₂ AMS — SO ₂ measurement....	40
Annex F (informative) Calculation of uncertainty of measurement of SO ₂	46

Bibliography	55
--------------------	----

Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Symbols and abbreviated terms	4
5 Principle	5
6 Description of the automated measuring systems	6
6.1 Sampling and sample gas conditioning systems	6
6.2 Analyser equipment	6
7 Performance characteristics and criteria	6
7.1 Performance criteria	6
7.2 Determination of the performance characteristics	7
7.2.1 Performance test	7
7.2.2 Ongoing quality control	7
8 Selection and installation procedure	8
8.1 Choice of the measuring system	8
8.2 Sampling	8
8.2.1 Sampling location	8
8.2.2 Representative sampling	8
8.3 Calculation	9
8.3.1 Conversion from volume to mass concentration for SO ₂	9
9 Quality assurance and quality control procedures	9
9.1 General	9
9.2 Frequency of checks	9
9.3 Calibration, validation and measurement uncertainty	10
10 Test report	11
Annex A (informative)	13
Annex B (informative)	21
Annex C (normative)	26
Annex D (normative)	27
Annex E (informative)	36
Annex F (informative)	42
Bibliography	49

iTeh Standards
(<https://standards.itih.ai>)
Document Preview

[ISO 7935:2024](https://standards.itih.ai/catalog/standards/iso/de5f0943-72d4-496d-bf59-7dd753d757f8/iso-7935-2024)

<https://standards.itih.ai/catalog/standards/iso/de5f0943-72d4-496d-bf59-7dd753d757f8/iso-7935-2024>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 146, *Air quality*, Subcommittee SC 1, *Stationary source emissions*.

This second edition cancels and replaces the first edition (ISO 7935:1992), which has been technically revised.

The main changes compared to the previous edition are as follows:

- ~~the~~ the structure and the components have been updated to be similar to the latest editions of e.g. ISO 10849 (measurement of nitrogen oxides), ISO 12039 (measurement of CO, CO₂ and O₂), ISO 17179 (measurement of NH₃), ISO 13199 (measurement of total VOC), ISO 25140 (measurement of CH₄), ISO 21258 (measurement of N₂O);
- ~~Clause 3~~ Clause 3 has been revised with the addition or deletion and change in terms and definitions;
- ~~a~~ a new analytical technique has been added (laser spectroscopic technique or tunable laser spectroscopy) for measurement of SO₂;

- ~~the~~ performance characteristics and criteria as well as QA/QC procedures have been changed to harmonize with latest ISO standards;
- ~~examples~~ of performance test results and the results of uncertainty calculation have been added for SO₂ measurement.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

ISO 7935:2024

<https://standards.iteh.ai/catalog/standards/iso/de5f0943-72d4-496d-bf59-7dd753d757f8/iso-7935-2024>

Introduction

Sulfur dioxide (SO₂) can arise in considerable quantities from combustion of fossil fuels used for energy generation, industrial activities processing sulfur or sulfur containing material, and from combustion of sulfur containing waste. The waste gas from these processes, containing sulfur dioxide, is usually discharged into the ambient atmosphere, via a duct or a chimney.

For evaluating the mass concentration of sulfur dioxide present in the waste gas of stationary source emissions, a number of highly developed methods of integrated sampling and subsequent determination by chemical analysis and automated measuring systems are available.

iTeh Standards (<https://standards.itih.ai>) Document Preview

[ISO 7935:2024](https://standards.itih.ai/catalog/standards/iso/de5f0943-72d4-496d-bf59-7dd753d757f8/iso-7935-2024)

<https://standards.itih.ai/catalog/standards/iso/de5f0943-72d4-496d-bf59-7dd753d757f8/iso-7935-2024>

Stationary source emissions — Determination of the mass concentration of sulfur dioxide in flue gases — Performance characteristics of automated measuring systems

1 Scope

This document specifies a method for the determination of sulfur dioxide (SO₂) in flue gases of stationary sources and describes the fundamental structure and the key performance characteristics of automated measuring systems.

The method allows continuous monitoring with permanently installed measuring systems of SO₂ emissions.

This document describes extractive systems and in situ (non-extractive) systems in connection with a range of analysers that operate using, for example, the following principles:

- non-dispersive infrared absorption (NDIR);
- Fourier transform infrared (FTIR) spectroscopy;
- laser spectroscopic technique or tunable laser spectroscopy (TLS);
- non-dispersive ultraviolet absorption (NDUV);
- differential optical absorption spectroscopy (DOAS);

Other equivalent instrumental methods can be used provided they meet the minimum performance requirements specified in this document. The measuring system can be validated with reference materials, according to this document, or comparable methods.

Automated measuring system (AMS) based on the principles listed above has been used successfully in this application for the measuring ranges as shown in [Annex E](#).

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 9169, *Air quality — Definition and determination of performance characteristics of an automatic measuring system*

ISO 14956, *Air quality — Evaluation of the suitability of a measurement procedure by comparison with a required measurement uncertainty*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>

— IEC Electropedia: available at <https://www.electropedia.org/>

3.1

analyser

analytical part in an extractive or *in situ automated* measuring system (3.2)(3.3)

[SOURCE: ISO 12039:2019^[2], [2] 3.1]

3.2

automated measuring system

AMS

measuring system interacting with the flue gas under investigation, returning an output signal proportional to the physical unit of the *measurand* (3.9)(3.9) in unattended operation

Note_1_to_entry:- For the purposes of this document, an AMS is a system that can be attached to a duct or stack to continuously or intermittently measure the mass concentration of SO₂ passing through the duct.

[SOURCE: ISO 9169:2006, 2.1.2, modified — Note 1 to entry has been replaced.]

3.3

in situ AMS

non-extractive system that measures the concentration directly in the duct or stack

Note_1_to_entry:- In situ systems measure either across the stack or duct or at a point within the duct or stack.

3.4

parallel measurements

measurements taken on the same duct in the same sampling plane for the same period of time with the AMS under test and with the reference method at points a short distance from each other, providing pairs of measured values

Note_1_to_entry:- See 3.20.3.20.

<https://standards.iteh.ai/catalog/standards/iso/de5f0943-72d4-496d-bf59-7dd753d757f8/iso-7935-2024>

3.5

independent reading

reading that is not influenced by a previous individual reading by separating two individual readings by at least four response times

3.6

interference

cross-sensitivity

negative or positive effect upon the response of the measuring system, due to a component of the sample that is not the measurand

3.7

interferent

interfering substance

substance present in the air mass under investigation, other than the *measurand* (3.9)(3.9), that affects the response of AMS (3.2)(3.2)

3.8

lack-of-fit

systematic deviation within the range of application, between the accepted value of a reference material applied to the measuring system and the corresponding result of measurement produced by the measuring system

Note_1_to_entry:-Lack-of-fit can be a function of the result of measurement.

Note_2_to_entry:-The expression “lack-of-fit” is often replaced in everyday language for linear relations by “linearity” or “deviation from linearity”.

[SOURCE: ISO 9169:2006, 2.2.9, modified — Note 2 to entry has been removed.]

3.9

measurand

particular quantity subject to measurement

[SOURCE: ISO/IEC Guide 98-3:2008, ~~6:1~~-3:2008^[6], B.2.9, modified — Example and Note 1 to entry have been removed.]

3.10

performance characteristic

quantity assigned to equipment in order to define its performance

Note_1_to_entry:-Performance characteristics can be described by values, tolerances or ranges.

3.11

period of unattended operation

maximum interval of time for which the performance characteristics remain within a predefined range without external servicing, e.g. refill, adjustment

[SOURCE: ISO 9169:2006, 2.2.11]

<https://standards.iteh.ai/catalog/standards/iso/de5f0943-72d4-496d-bf59-7dd753d757f8/iso-7935-2024>

Note_1_to_entry:-The period of unattended operation is often called maintenance interval.

3.12

reference material

substance or mixture of substances with a known concentration within specified limits, or a device of known characteristics

Note_1_to_entry:-Normally calibration gases, gas cells, gratings or filters are used.

[SOURCE: ISO 14385-1:2014^[3], ^[3] 3.20]

3.13

reference method

measurement method taken as a reference by convention, which gives the accepted reference value of the measurand

Note_1_to_entry:-See [3.4.3.4](#).

3.14**transport time in the measuring system line**

time period for transportation of the sampled gas from the inlet of the probe to the inlet of the measurement instrument

3.15**response time**

time interval between the instant when a stimulus is subjected to bring about a specified abrupt change and the instant when the response reaches and remains within specified limits around its final stable value, determined as the sum of the lag time and the rise time in the rising mode, and the sum of the lag time and the fall time in the falling mode

[SOURCE: ISO 9169:2006, 2.2.4]

Note_1_to_entry:-Lag time, rise time and fall time are defined in ISO 9169:2006.

3.16**span gas**

gas or gas mixture used to adjust and check the span point on the response line of the measuring system

Note_1_to_entry:-The concentration is often chosen around 70 % to 90 % of full scale.

3.17**span point**

value of the output quantity (measured signal) of the automated measuring system for the purpose of calibration, adjustment, etc. that represents a correct measured value generated by reference gas

3.18**standard uncertainty**

uncertainty of the result of a measurement expressed as a standard deviation

[SOURCE: ISO/IEC Guide 98-3:2008, ^[6]3:2008^[6], 2.3.1]

3.19**uncertainty**

parameter associated with the result of a measurement, that characterizes the dispersion of the values that can reasonably be attributed to the measurand

[SOURCE: ISO/IEC Guide 98-3:2008, ^[6]3:2008^[6], 2.2.3 modified — Notes 1, 2 and 3 to entry have been removed.]

3.20**validation of an automated measuring system**

procedure to check the statistical relationship between values of the measurand indicated by the automated measuring system and the corresponding values given by parallel measurements implemented simultaneously at the same measuring point

3.21**zero gas**

gas or gas mixture used to establish the *zero point* (3.22)(3.22) on a calibration curve within a given concentration range