

FINAL
DRAFT

INTERNATIONAL
STANDARD

ISO/FDIS
18666

ISO/TC 249

Secretariat: SAC

Voting begins on:
2021-08-18

Voting terminates on:
2021-10-13

Traditional Chinese medicine — General requirements of moxibustion devices

*Médecine traditionnelle chinoise — Exigences générales concernant
les dispositifs de moxibustion*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO/FDIS 18666

<https://standards.iteh.ai/catalog/standards/sist/630364ed-6b91-47b3-bdf7-b2c74e577c7/iso-fdis-18666>

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.



Reference number
ISO/FDIS 18666:2021(E)

© ISO 2021

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO/FDIS 18666

<https://standards.iteh.ai/catalog/standards/sist/630364ed-6b91-47b3-bdf7-b2c74e577c7/iso-fdis-18666>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2021

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Configuration	2
4.1 Structure.....	2
4.1.1 Size and shape.....	2
4.1.2 Visual inspection.....	2
4.2 Expression of dimensions.....	3
5 Materials	3
5.1 General.....	3
5.2 Body of moxibustion devices.....	3
5.3 Moxibustion materials.....	3
5.3.1 Moxa floss.....	3
5.3.2 Moxa sticks.....	3
6 Moxibustion temperature	4
6.1 General.....	4
6.2 Moxibustion with tube.....	4
6.3 Moxibustion devices other than moxibustion with tube.....	4
6.4 Moxibustion devices for repeated use.....	4
7 Safety requirements	4
7.1 General.....	4
7.2 Environmental protection.....	4
8 Packaging	5
8.1 Primary packaging.....	5
8.2 Secondary packaging.....	5
9 Labelling	5
9.1 Primary label.....	5
9.2 Secondary label.....	5
10 Transport and storage	6
11 Instructions for use	6
Annex A (informative) Test methods to determine the moxibustion temperature	7
Annex B (informative) Method of artificial drying of mugwort leaves by heat	10
Bibliography	11

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 249, *Traditional Chinese medicine*.

This second edition cancels and replaces the first edition (ISO 18666:2015), of which it constitutes a minor revision. The changes to the previous edition are as follows:

- Introduction: the term 'International Standard' changed to 'document' in the first and third paragraphs; improvement of English expression to 'increasing expectations and concerns by patients' and the term 'ensure' changed to 'guide' in the second and third paragraphs;
- Scope: the term 'International Standard' changed to 'document' in the first and third paragraphs; minor editorial change of the term 'uses' to 'use' in the second paragraph;
- **Clause 3**: addition of the introductory text according to ISO/IEC Directives, Part 2;
- **3.4**: deletion of the phrase 'during treatment';
- **4.1.1**: improvement of English expression for clarity;
- minor editorial changes.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

This document specifies the general requirements for safety and quality of moxibustion devices, including the moxibustion materials used in such devices. The safety and quality of both moxibustion devices and materials are closely related to moxibustion safety and quality.

There is a wide variety of moxibustion devices and materials currently available commercially, but there are no standards guiding their manufacture and finishing. Increased interest and use of moxibustion, as well as increasing expectations and concerns by patients regarding moxibustion safety and quality, have given rise to the need to improve safety and quality of moxibustion through implementation of an International Standard.

The primary aim of this document is to guide the safety and quality of moxibustion devices and materials.

[Annex A](#) gives guidance on the methods that can be used to determine the moxibustion temperature at the human body surface during treatment using the moxibustion device.

[Annex B](#) gives guidance on the method of artificial drying of mugwort leaves by heat.

iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO/FDIS 18666

<https://standards.iteh.ai/catalog/standards/sist/630364ed-6b91-47b3-bdf7-b2c74e577c7/iso-fdis-18666>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO/FDIS 18666

<https://standards.iteh.ai/catalog/standards/sist/630364ed-6b91-47b3-bdf7-b2c74ef577c7/iso-fdis-18666>

Traditional Chinese medicine — General requirements of moxibustion devices

1 Scope

This document specifies the general requirements for configuration, materials, performance and safety requirements of moxibustion devices. It also specifies the minimum requirements for moxibustion materials used in moxibustion devices.

It is applicable across a wide range of moxibustion devices that use moxa floss as the main combustion material and can remain on or over the body throughout the moxibustion process. It is applicable to moxibustion devices for both single and repeated usage.

This document does not apply to devices that imitate moxibustion, such as electro-moxibustion and infrared moxibustion devices that do not involve the use of moxa floss. It also does not apply to moxa floss used in direct moxibustion.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 10993-1, *Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process* ISO/FDIS 18666

<https://standards.iteh.ai/catalog/standards/sist/630364ed-6b91-47b3-bdf7->

ISO 15223-1, *Medical devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General requirements* ISO 15223-1:2016

ISO/TS 10993-19, *Biological evaluation of medical devices — Part 19: Physico-chemical, morphological and topographical characterization of materials*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

3.1

moxibustion device

apparatus that uses moxa floss as the main combustion material and is intended for single or repeated usage

EXAMPLE Moxibustion with tube, such as a short moxa roll with a cardboard base and a moxa tube (made of cardboard) that is single-use and developed as an alternative to direct moxibustion.

Note 1 to entry: Moxibustion devices include those accessories as defined by the manufacturers that are necessary to enable the normal use of the moxibustion device.

3.2

moxibustion material

combustible material comprising mainly moxa floss and used in moxibustion

3.3

body of moxibustion device

part of the moxibustion device that is used to hold moxibustion materials and remains on or over the human body throughout the moxibustion process

3.4

moxibustion temperature

temperature at the human body surface when using the moxibustion device

3.5

moxa floss

cotton-like material for moxibustion made from mugwort leaves

3.6

moxa stick

round long stick made of moxa floss, also called moxa roll

Note 1 to entry: Moxa sticks can be in the form of a pure moxa stick, a medicinal moxa stick (with additives) or a smokeless moxa stick.

3.7

medicinal moxa stick

moxibustion with the moxa roll made of moxa and various substances

3.8

fineness of moxa floss

weight of the starting material (mugwort leaves) to the weight of the final product (moxa floss) presented in the form of a ratio

ITeH STANDARD PREVIEW
(standards.iteh.ai)

standards.iteh.ai/catalog/standards/sist/630364ed-6b91-47b3-bdf7-b2c74e577c7/iso-fdis-18666

4 Configuration

4.1 Structure

4.1.1 Size and shape

The size and shape of moxibustion devices can vary according to the intended use, such as on a specific point or an area of the human body surface.

4.1.2 Visual inspection

When inspected by normal or corrected-to-normal vision:

- a) the surface of the moxibustion device shall not have any obvious defects such as breakages, protrusions, dents or rust that can cause accidental cuts or injuries to the user;
- b) moxibustion devices for repeated use shall have a temperature-adjusting mechanism or other appropriate arrangement to maintain moxibustion at an appropriate temperature;
- c) except for moxibustion with tube, the body of the moxibustion device shall ensure that the removal of ash is easy and shall be well ventilated to facilitate combustion; it should also contain:
 - 1) a mounting base or support that can hold the stick (s) firmly;
 - 2) a barrier or other safety arrangement to prevent ash or ember from falling onto the surface of the human body.

4.2 Expression of dimensions

The moxibustion device shall be easy to manipulate in size and applicable to various parts of the human body. The dimensions of the moxibustion device shall be expressed in centimetres and specified as length × breadth × height for a rectangular device, or diameter × height for a circular device.

EXAMPLE 1 For a rectangular device, 15 cm × 10 cm × 8 cm.

EXAMPLE 2 For a circular device, 4 cm × 6 cm.

5 Materials

5.1 General

The materials of the moxibustion device in contact with the human body surface shall be assessed for their biological safety and physico-chemical, morphological and topographical (PMT) characteristics in accordance with ISO 10993-1 and ISO/TS 10993-19.

5.2 Body of moxibustion devices

The frame of moxibustion devices can be made of materials such as wood, plastic, metal or cardboard. It shall have properties of low flammability and be fire-retardant to prevent fire hazards during moxibustion. The material of the component of the moxibustion device that comes into contact with the human body surface shall have low thermal conductivity and shall not cause skin sensitivity or irritation.

5.3 Moxibustion materials

5.3.1 Moxa floss

<https://standards.iteh.ai/catalog/standards/sist/630364ed-6b91-47b3-bdf7-b2c74e577c7/iso-fdis-18666>

5.3.1.1 Moxa floss shall be processed from mugwort leaves of the *Artemisia* genus, including *Artemisia argyi*, *Artemisia princeps*, *Artemisia vulgaris* and other *Artemisia* species that have proven to be equally suitable.

5.3.1.2 Moxa leaves shall be stored under a dry and well-ventilated environment for at least 1 year or stored for 3 months and be artificially dried by heating before processing into moxa floss.

5.3.1.3 The minimum fineness of moxa floss should be in the ratio of 3:1, where 3 kg of mugwort leaves are processed into 1 kg of moxa floss.

5.3.1.4 Moxa floss shall not contain any foreign matter, such as metal shavings, which may contaminate the moxa floss during production and processing.

5.3.1.5 The quality and safety of moxa floss can be determined by visual or physical inspection and the use of other inspection equipment, such as a metal detector.

5.3.2 Moxa sticks

5.3.2.1 The external wrapping used for moxa sticks made using moxa floss shall be material manufactured from mulberry leaves or other combustible paper.

5.3.2.2 The external wrapping used for moxa sticks made using moxa floss shall:

a) have a similar burning rate to moxa floss;

- b) not emit any harmful substances during combustion;
- c) not be torn easily.

5.3.2.3 The adhesive material used in moxa sticks shall not be toxic or produce harmful substances during combustion.

6 Moxibustion temperature

6.1 General

The test methods to determine the moxibustion temperature are stated in [Annex A](#).

6.2 Moxibustion with tube

The maximum temperature shall not exceed 60 °C.

6.3 Moxibustion devices other than moxibustion with tube

Except for moxibustion with tube, the moxibustion temperature of the moxibustion device shall be adjustable and the maximum temperature achievable during use of the device shall not exceed 51 °C.

6.4 Moxibustion devices for repeated use

For moxibustion devices for repeated use, there shall be a temperature-adjusting mechanism to maintain the temperature under the user's thermal pain threshold.

7 Safety requirements

7.1 General

7.1.1 Moxibustion devices for repeated use shall have a barrier or other safety arrangement to prevent any ash, ember or spark from falling onto the human body surface.

7.1.2 The body of the moxibustion device shall be able to remain on or over the human body surface in a stable manner throughout the moxibustion process to prevent the device from falling over or being dislodged.

7.1.3 The material of the part of the moxibustion device that comes into contact with the human body surface shall have poor thermal conductivity to ensure the temperature is within the acceptance limits of the moxibustion recipient.

7.1.4 Moxibustion devices that remain on the human body surface shall be either for single use or repeated use. For repeated use, they shall have a disposable bottom or protective lining on the areas that come into contact with the human body surface to prevent any form of direct-contact infection or transmission.

7.2 Environmental protection

Necessary measures that can protect the environment shall be identified and employed. Moxibustion devices can include a protective installation, use better quality moxa floss or have additives added to the moxa floss to reduce the amount of smoke.