

INTERNATIONAL STANDARD

IEC 61779-5

First edition
1998-04

Electrical apparatus for the detection and measurement of flammable gases –

Part 5:

Performance requirements for group II apparatus indicating a volume fraction up to 100 % gas

*Appareils électriques de détection et de mesure
des gaz combustibles –*

Partie 5:

*Règles de performance des appareils du groupe II pouvant indiquer
une fraction volumique jusqu'à 100 % de gaz*

IEC 61779-5:1998

<https://standards.iteh.ai/catalog/standards/iec/61779-5:1998>



Reference number
IEC 61779-5:1998(E)

Numbering

As from 1 January 1997 all IEC publications are issued with a designation in the 60000 series.

Consolidated publications

Consolidated versions of some IEC publications including amendments are available. For example, edition numbers 1.0, 1.1 and 1.2 refer, respectively, to the base publication, the base publication incorporating amendment 1 and the base publication incorporating amendments 1 and 2.

Validity of this publication

The technical content of IEC publications is kept under constant review by the IEC, thus ensuring that the content reflects current technology.

Information relating to the date of the reconfirmation of the publication is available in the IEC catalogue.

Information on the subjects under consideration and work in progress undertaken by the technical committee which has prepared this publication, as well as the list of publications issued, is to be found at the following IEC sources:

- **IEC web site***
- **Catalogue of IEC publications**
Published yearly with regular updates
(On-line catalogue)*
- **IEC Bulletin**
Available both at the IEC web site* and as a printed periodical

Terminology, graphical and letter symbols

For general terminology, readers are referred to IEC 60050: *International Electrotechnical Vocabulary (IEV)*.

For graphical symbols, and letter symbols and signs approved by the IEC for general use, readers are referred to publications IEC 60027: *Letter symbols to be used in electrical technology*, IEC 60417: *Graphical symbols for use on equipment. Index, survey and compilation of the single sheets* and IEC 60617: *Graphical symbols for diagrams*.

<https://standards.iteh.ai/catalog/standards-iec/61779-5-1998>

<https://standards.iteh.ai/catalog/standards-iec/61779-5-1998>

* See web site address on title page.

INTERNATIONAL STANDARD

IEC 61779-5

First edition
1998-04

Electrical apparatus for the detection and measurement of flammable gases – Part 5: Performance requirements for group II apparatus indicating a volume fraction up to 100 % gas

*Appareils électriques de détection et de mesure
des gaz combustibles –*

*Partie 5:
Règles de performance des appareils du groupe II pouvant indiquer
une fraction volumique jusqu'à 100 % de gaz*

IEC 61779-5:1998

<https://standards.iteh.ai/catalog/standards-iec/61779-5-1998>

© IEC 1998 — Copyright - all rights reserved

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Electrotechnical Commission
Telefax: +41 22 919 0300

3, rue de Varembeé Geneva, Switzerland
e-mail: inmail@iec.ch IEC web site <http://www.iec.ch>



Commission Electrotechnique Internationale
International Electrotechnical Commission
Международная Электротехническая Комиссия

PRICE CODE

G

*Pour prix, voir catalogue en vigueur
For price, see current catalogue*

INTERNATIONAL ELECTROTECHNICAL COMMISSION

ELECTRICAL APPARATUS FOR THE DETECTION AND MEASUREMENT OF FLAMMABLE GASES –

Part 5: Performance requirements for group II apparatus indicating a volume fraction up to 100 % gas

FOREWORD

- 1) The IEC (International Electrotechnical Commission) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of the IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, the IEC publishes International Standards. Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. The IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of the IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested National Committees.
- 3) The documents produced have the form of recommendations for international use and are published in the form of standards, technical reports or guides and they are accepted by the National Committees in that sense.
- 4) In order to promote international unification, IEC National Committees undertake to apply IEC International Standards transparently to the maximum extent possible in their national and regional standards. Any divergence between the IEC Standard and the corresponding national or regional standard shall be clearly indicated in the latter.
- 5) The IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with one of its standards.
- 6) Attention is drawn to the possibility that some of the elements of this International Standard may be the subject of patent rights. The IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 61779-5 has been prepared by subcommittee 31L: Electrical apparatus for the detection of flammable gases, of IEC technical committee 31: Electrical apparatus for explosive atmospheres.

This standard should be read in conjunction with IEC 61779-1.

The text of this standard is based on the following documents:

FDIS	Report on voting
31L/51/FDIS	31L/56/RVD

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

A bilingual version of this standard may be issued at a later date.

ELECTRICAL APPARATUS FOR THE DETECTION AND MEASUREMENT OF FLAMMABLE GASES –

Part 5: Performance requirements for group II apparatus indicating a volume fraction up to 100 % gas

1 Scope

1.1 This part of IEC 61779 specifies requirements for group II (as defined in part 1) portable, transportable and fixed apparatus for the detection and measurement of combustible gas or vapour concentrations with air. The apparatus, or parts thereof, may be installed or used in potentially explosive atmospheres, other than mines susceptible to firedamp (i.e. group I). The requirements and test methods applicable to the apparatus covered by this standard are specified in part 1.

1.2 This standard is restricted to apparatus intended for the detection and measurement of volume ratios of combustible gas or vapour in air in volume fractions from 0 % to 100 % .

NOTE 1 – Apparatus covered by this standard will normally be intended to operate in volume ratios greater than 100 % LEL.

NOTE 2 – Although apparatus of the types covered by this standard may be suitable for detecting a wide range of combustible gases, particular gases (e.g. methane or propane) are specified in part 1 as the components of the test gases for the purpose of practical convenience. The performance requirements specified in this standard should therefore be considered with caution when the apparatus is used to detect other combustible gases, as some parameters – such as time of response – will be modified.

2 Definitions

For the purpose of this part of IEC 61779, the definitions given in part 1 apply.

3 General requirements

The apparatus shall comply with the general requirements specified in part 1 and with the performance requirements specified in clause 4 of this standard.

Compliance shall be determined in accordance with the appropriate test requirements and methods, including initial calibration, specified in part 1.

It shall be verified that the contents of the manufacturer's instruction manual are in accordance with the requirements specified in part 1.

4 Performance requirements

4.1 General

The normal conditions for tests are specified in 4.3 of part 1. Compliance shall be determined in accordance with the test methods specified in 4.4 of part 1.

4.2 Unpowered storage

After being submitted to the conditions specified in 4.4.2 of part 1, the apparatus shall meet the requirements specified in 4.3 to clause 5 of this standard.

4.3 Calibration curve (not applicable to alarm-only apparatus)

4.3.1 Calibration curve

After initial adjustment with the standard test gas, each individual indication in the three sets of indications (after correction using the manufacturer's calibration curves, if necessary) obtained for each of the four gas volume ratios distributed over the measuring range, shall not differ from these volume ratios by more than $\pm 5\%$ of the measuring range or $\pm 10\%$ of the indication, whichever is the greater.

4.3.2 Response to gases other than the test gas

The apparatus indications (after correction using the manufacturer's calibration curves, if necessary) obtained for each of the three gas volume ratios of each gas tested shall not differ from these volume ratios by more than $\pm 7\%$ of the measuring range or $\pm 15\%$ of the indication, whichever is the greater.

4.4 Stability (continuous duty apparatus)

Continuous duty apparatus shall comply with the following requirements:

a) short-term stability

The short-term variation shall not exceed $\pm 3\%$ of the measuring range or $\pm 10\%$ of the indication, whichever is the greater.

b) long-term stability (fixed and transportable apparatus)

The long-term variation shall not exceed $\pm 7\%$ of the measuring range or $\pm 15\%$ of the indication, whichever is the greater.

c) long-term stability (portable apparatus)

The long-term variation shall not exceed $\pm 5\%$ of the measuring range or $\pm 10\%$ of the indication, whichever is the greater.

4.5 Stability (spot-reading apparatus)

The variation of the indication shall not exceed $\pm 3\%$ of the measuring range or $\pm 10\%$ of the indication, whichever is the greater.

4.6 Alarm

The alarm(s) shall operate during every cycle of the test. If a latching alarm is provided, the manual reset action shall be checked.

4.7 Temperature

4.7.1 Apparatus where the control unit and sensors are used in the same environment

The variation of the indication from that at $20\text{ }^{\circ}\text{C}$

- a) shall not, at $-10\text{ }^{\circ}\text{C}$, exceed $\pm 7\%$ of the measuring range or $\pm 10\%$ of the indication, whichever is the greater.
- b) shall not, at $+40\text{ }^{\circ}\text{C}$, exceed $\pm 5\%$ of the measuring range or $\pm 10\%$ of the indication, whichever is the greater.