



Designation: ~~D6906 – 11a~~ D6906 – 12

## Standard Test Method for Determination of Titanium Treatment Weight on Metal Substrates by Wavelength Dispersive X-Ray Fluorescence<sup>1</sup>

This standard is issued under the fixed designation D6906; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

### 1. Scope

1.1 This test method covers the use of wavelength dispersive X-ray fluorescence (WDXRF) techniques for determination of the coating weight of titanium treatments on metal substrates. These techniques are applicable for determination of the coating weight as titanium or total coating weight of a titanium containing treatment, or both, on a variety of metal substrates.

1.2 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

### 2. Referenced Documents

2.1 *ASTM Standards:*<sup>2</sup>

[E177 Practice for Use of the Terms Precision and Bias in ASTM Test Methods](#)

[E691 Practice for Conducting an Interlaboratory Study to Determine the Precision of a Test Method](#)

### 3. Summary of Practice

3.1 *Excitation*—The measurement of titanium treatment coating weights by WDXRF methods is based on the combined interaction of the titanium coating and the substrate with an intense beam of primary radiation. Since each element fluoresces at an energy characteristic of the particular element, this interaction results in the generation of X-rays of defined energy. The primary radiation may be generated by an X-ray tube or derived from a radioisotope.

3.2 *Detection*—The secondary beam (fluorescent X-rays of the elements and scattered radiation) is read by a detector that can discriminate between the energy levels of fluorescing radiations in the secondary beam. The detection system includes the radiation detector with electronics for pulse amplification and pulse counting.

3.3 *Basic Principle:* [iteh.ai/catalog/standards/sist/d003801c-1eda-412c-9504-8d8299e6e417/astm-d6906-12](http://iteh.ai/catalog/standards/sist/d003801c-1eda-412c-9504-8d8299e6e417/astm-d6906-12)

3.3.1 A relationship exists between the treatment coating weight and secondary radiation intensity. This relationship is usually linear within the desired coating weights of the titanium treatments on metal substrates. The measurements are based on primary standards of known coating weights and instrument calibration that correlates the secondary radiation intensity with the coating weight quantitatively.

3.3.2 The coating weight is determined by measurement of the fluorescent X-rays of the coating. The detection system is set to count the number of X-rays in an energy region that is characteristic of X-rays from the element of interest. The element of interest in this practice is titanium.

3.3.3 If a linear relationship exists, the coating weight and number of counts of X-rays of a titanium treatment on a particular substrate can be expressed by a conversion factor that represents the number of counts for a particular coating weight unit/unit area. This is usually expressed in mg/ft<sup>2</sup> or mg/m<sup>2</sup> of titanium or total coating weight.

3.3.4 The exact relationship between the measured number of counts and the corresponding coating weight must be established for each individual combination of substrate and titanium-containing treatment. Usually determined by the treatment supplier, this relationship is established by using primary standards having known amounts of the same treatment applied to the same substrate composition as the specimens to be measured.

<sup>1</sup> This test method is under the jurisdiction of ASTM Committee D01 on Paint and Related Coatings, Materials, and Applications and is the direct responsibility of Subcommittee D01.53 on Coil Coated Metal.

Current edition approved June 1, 2011; July 15, 2012. Published June 2011; January 2013. Originally approved in 2003. Last previous edition approved in 2011 as D6906 – 11; D6906 – 11A. DOI: 10.1520/D6906-11A; 10.1520/D6906-12.

<sup>2</sup> For referenced ASTM standards, visit the ASTM website, [www.astm.org](http://www.astm.org), or contact ASTM Customer Service at [service@astm.org](mailto:service@astm.org). For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

3.3.5 Some X-ray apparatuses have a data handling system whereby a coating weight versus X-ray counts curve may be established within the system for the direct readout of coating weight. If such apparatus does not permit the entry of a conversion factor as described in 2.3.33.3.3, it is calibrated using a bare, untreated specimen and a minimum of three specimens with known coating weights of the treatment and substrate combination of interest. The coating weight to be measured must be within the range of these known coating weights. More than three known specimens must be used if the relationship of X-ray counts to coating weight is not linear over the range to be measured. The treatment supplier should be consulted for recommendations for establishing the curve in the instrument for the particular treatment and substrate combination of interest.

**4. Significance and Use**

4.1 The procedure described in this test method is designed to provide a method by which the coating weight of titanium treatments on metal substrates may be determined.

4.2 This test method is applicable for determination of the total coating weight and the titanium coating weight of a titanium-containing treatment.

**5. Apparatus and Materials**

5.1 *Measuring Instrument*, which is capable of determining the coating weights of titanium-containing treatments on metal substrates by X-ray fluorescence is required. The treatment supplier should be consulted for the suitability of the instrumentation to be used

5.2 *Calibration Standard*, necessary to calibrate the instrument. The count value of this standard must be specified by the treatment supplier.

5.3 *Treated Coupon*, on which the coating weight is to be determined must be cut to the required size for the instrument from the treated substrate.

5.4 *Blank (Bare and Untreated) Coupon* should be a sample of the same metal substrate on which the treatment coating weight is to be determined. It may be necessary to prepare a blank coupon from a treated sample if an untreated coupon is not available. To best imitate a bare, untreated blank, abrade a treated coupon that is from the same metal specimen as the test specimen using a small abrasive pad.

5.4.1 The first abrading is made parallel with the rolling direction of the metal, the second abrading is made perpendicular to the rolling direction of the metal, and the third abrading is made parallel with the rolling direction of the metal. This procedure should be repeated until constant readings are obtained. Always use the same side of the metal substrate from which the readings of the treated coupon will be taken.

**6. Test Specimens**

6.1 All test specimens must be flat in the area of measurement and free of burrs and distortions that would prevent proper seating in the specimen holder.

6.2 The treatment on the substrate must be uniform in the area of measurement.

6.3 The area of measurement must be maintained free of foreign materials. The specimen must be handled only by the edges that are outside of the area to be measured.

6.4 The coated area of the specimen must be larger than the measuring area.

**7. Procedure**

7.1 Operate the instrument in accordance with the manufacturer’s instructions.

7.2 Set the instrument settings as follows:

Dial and arm	titanium position
Seconds indicator	pretreatment supplier
Multiplier switch	pretreatment supplier
Response switch	pretreatment supplier
Range	pretreatment supplier
Milliamps	adjust for calibration of output
	pretreatment supplier

7.3 All specimens must be seated firmly and securely over the measuring opening. The distance between the measuring apparatus and specimen must be maintained the same as that during the calibration. The blank and treated specimens must be placed in the holder so that the rolling direction of the metal is in the same orientation. Whenever a sample tray holder is a part of the apparatus, the same opening of the slide must be used for the blank and treated specimen unless the openings have been determined to produce equivalent results. If it is necessary to use a backer to hold the test specimen firmly against the window, make sure that the backer is of untreated coupons of the same metal as the specimen. The same backer must be used for each set of measurements.