

ISO/TC 121/SC 3

Secretariat: ANSI

Voting begins on:
2023-01-26

Voting terminates on:
2023-03-23

Medical electrical equipment —

Part 2-72:

Particular requirements for basic safety and essential performance of home healthcare environment ventilators for ventilator-dependent patients

Appareils électromédicaux —

Partie 2-72: Exigences particulières pour la sécurité de base et les performances essentielles des ventilateurs utilisés dans l'environnement des soins à domicile pour les patients ventilodépendants

Member bodies are requested to consult relevant national interests in IEC/SC 62D before casting their ballot to the e-Balloting application.

ISO/CEN PARALLEL PROCESSING

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.



Reference number
ISO/FDIS 80601-2-72:2023(E)

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[ISO/FDIS 80601-2-72](https://standards.iteh.ai/catalog/standards/sist/7dd10752-d386-4733-8c89-fe2e9eb758ca/iso-fdis-80601-2-72)
[https://standards.iteh.ai/catalog/standards/sist/7dd10752-d386-4733-8c89-
fe2e9eb758ca/iso-fdis-80601-2-72](https://standards.iteh.ai/catalog/standards/sist/7dd10752-d386-4733-8c89-fe2e9eb758ca/iso-fdis-80601-2-72)



COPYRIGHT PROTECTED DOCUMENT

© ISO 2023

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Foreword	v
Introduction.....	vii
201.1 Scope, object, and related standards.....	1
201.2 Normative references	3
201.3 Terms and definitions	5
201.4 General requirements	21
201.5 General requirements for testing of <i>ME equipment</i>	25
201.6 Classification of <i>ME equipment</i> and <i>ME systems</i>	26
201.7 <i>ME equipment</i> identification, marking, and documents.....	26
201.8 Protection against electrical <i>hazards</i> from <i>ME equipment</i>	35
201.9 Protection against mechanical hazards of <i>ME equipment</i> and <i>ME systems</i>	35
201.10 Protection against unwanted and excessive radiation <i>hazards</i>	37
201.11 Protection against excessive temperatures and other <i>hazards</i>	37
201.12 Accuracy of controls and instruments and protection against hazardous outputs	40
201.13 <i>Hazardous situations</i> and fault conditions.....	59
201.14 Programmable electrical medical systems (PEMS)	61
201.15 Construction of <i>ME equipment</i>	61
201.16 <i>ME systems</i>	62
201.17 Electromagnetic compatibility of <i>ME equipment</i> and <i>ME systems</i>	63
201.101 Gas connections	63
201.102 Requirements for the <i>VBS</i> and <i>accessories</i>	66
201.103 Spontaneous breathing during loss of power supply.....	67
201.104 Indication of duration of operation	68
201.105 Functional connection	68
201.106 Display loops.....	69
201.107 <i>Ventilator</i> security	69
201.108 Oxygen inlet port.....	70
202 Electromagnetic disturbances — Requirements and tests.....	70
206 Usability.....	71
208 General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems.....	73
211 Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment.....	75

Annex C (informative) Guide to marking and labelling requirements for ME equipment and ME systems.....	78
Annex D (informative) Symbols on marking.....	85
Annex AA (informative) Particular guidance and rationale.....	87
Annex BB (informative) Data interface requirements	113
Annex CC (informative) Reference to the IMDRF essential principles and labelling guidances.....	121
Annex DD (informative) Reference to the essential principles.....	126
Annex EE (informative) Reference to the general safety and performance requirements.....	130
Bibliography	134

iTeh STANDARD PREVIEW
(standards.iteh.ai)

<https://standards.iteh.ai/catalog/standards/sist/7dd10752-d386-4733-8c89-fe2e9eb758ca/iso-fdis-80601-2-72>

Foreword

ISO (the International Organization for Standardization) and IEC (the International Electrotechnical Commission) form the specialized system for worldwide standardization. National bodies that are members of ISO or IEC participate in the development of International Standards through technical committees established by the respective organization to deal with particular fields of technical activity. ISO and IEC technical committees collaborate in fields of mutual interest. Other international organizations, governmental and non-governmental, in liaison with ISO and IEC, also take part in the work.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives or www.iec.ch/members_experts/refdocs).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO and IEC shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents) or the IEC list of patent declarations received (see <https://patents.iec.ch>).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html. In the IEC, see www.iec.ch/understanding-standards.

This document was prepared by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 3, *Respiratory devices and related equipment used for patient care* and Technical Committee IEC/TC 62, *Electrical equipment in medical practice*, Subcommittee SC 62D, *Electric equipment*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 215, *Respiratory and anaesthetic equipment*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 80601-2-72:2015), which has been technically revised. The main changes compared to the previous edition are as follows:

The main changes are as follows:

- added requirements for display during calibration of gas monitors;
- clarified *maximum limited pressure* requirements;
- clarified high *airway pressure alarm condition* requirements;
- added requirements for *ventilator system recovery*;
- added requirements for response to an increase in set oxygen (O₂) concentration; and
- harmonization with ISO 20417, where appropriate.

ISO/FDIS 80601-2-72:2023(E)

A list of all parts in the ISO 80601 series and the IEC 80601 series can be found on the ISO and IEC websites.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html and www.iec.ch/national-committees.

iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO/FDIS 80601-2-72

<https://standards.iteh.ai/catalog/standards/sist/7dd10752-d386-4733-8c89-fe2e9eb758ca/iso-fdis-80601-2-72>

Introduction

This document specifies requirements for *lung ventilators* that are intended for use in the *home healthcare environment* for *patients* who are dependent for *ventilation* for their life support. These *ventilators* are frequently used in locations where the *supply mains* driving the *ventilator* is not reliable. These *ventilators* are often supervised by non-healthcare personnel (*lay operators*) with varying levels of training. *Lung ventilators* conforming with this standard can be used elsewhere (i.e. in healthcare facilities).

In referring to the structure of this document,

- “clause” means one of the 5 numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 201 includes 201.7, 201.8, etc.), and
- “subclause” means a numbered subdivision of a clause (e.g. 201.7, 201.8 and 201.9 are all subclauses of Clause 201).

References to clauses within this document are preceded by the term “Clause” followed by the clause number. References to subclauses within this document are by number only.

In this document, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

For the purposes of this document, the auxiliary verb:

- “shall” indicates a requirement;
- “should” indicates a recommendation;
- “may” indicates a permission;
- “can” is used to describe a possibility or capability; and
- “must” is used to express an external constraint.

Annex C contains a guide to the *marking* and labelling requirements in this document.

Annex D contains a summary of the *symbols* referenced in this document.

Requirements in this document have been decomposed so that each requirement is uniquely delineated. This is done to support automated requirements tracking.

Medical electrical equipment —

Part 2-72:

Particular requirements for basic safety and essential performance of home healthcare environment ventilators for ventilator-dependent patients

201.1 Scope, object, and related standards

IEC 60601-1:2005+AMD1:2012+AMD2:2020, Clause 1 applies, except as follows:

201.1.1 Scope

Replacement:

NOTE 1 There is guidance or rationale for this subclause contained in Clause AA.2.

This document applies to the *basic safety* and *essential performance* of a *ventilator* in combination with its *accessories*, hereafter referred to as *ME equipment*:

— intended for use in the *home healthcare environment*;

NOTE 2 In the *home healthcare environment*, the *supply mains* driving the *ventilator* is often not reliable.

NOTE 3 Such *ventilators* can also be used in non-critical care applications of *professional healthcare facilities*.

— intended for use by a *lay operator*; and

— intended for those *patients* who need differing levels of support from *artificial ventilation* including for *ventilator-dependent patients*.

A *ventilator* is not considered to use a *physiologic closed-loop control system* unless it uses a physiological *patient* variable to adjust the *ventilation* therapy settings.

This document is also applicable to those *accessories* intended by their *manufacturer* to be connected to a *ventilator breathing system* or to a *ventilator* where the characteristics of those *accessories* can affect the *basic safety* or *essential performance* of the *ventilator*.

EXAMPLE Breathing tubes, *connectors*, water traps, expiratory valve, *humidifier*, *breathing system filter*, external electrical power source, and *distributed alarm system*.

NOTE 4 If a clause or subclause is specifically intended to be applicable to *ME equipment* only or to *ME systems* only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to *ME equipment* and to *ME systems*, as relevant.

Hazards inherent in the intended physiological function of *ME equipment* or *ME systems* within the scope of this document are not covered by specific requirements in this document except for the requirements specified in IEC 60601-1:2005+AMD1:2012+AMD2:2020, 7.2.13 and 8.4.1.

NOTE 5 Additional information can be found in IEC 60601-1:2005+AMD1:2012+AMD2:2020, 4.2.

ISO/FDIS 80601-2-72:2023(E)

This document does not specify the requirements for:

- *ventilators* or *accessories* intended for critical care applications, which are given in ISO 80601-2-12;
- *ventilators* or *accessories* intended for anaesthetic applications, which are given in ISO 80601-2-13;
- *ventilators* or *accessories* intended for emergency and transport which are given in ISO 80601-2-84;
- *ventilators* or *accessories* intended for homecare ventilatory support equipment (intended only to augment the *ventilation* of spontaneously breathing *patients*), which are given in ISO 80601-2-79 and ISO 80601-2-80;
- obstructive sleep apnoea therapy *ME equipment*, which are given in ISO 80601-2-70;
- high-frequency *ventilators*, which are given in ISO 80601-2-87.
- respiratory high-flow therapy equipment, which are given in ISO 80601-2-90;

NOTE 6 An ISO 80601-2-72 *ventilator* can incorporate high-flow therapy operational mode, but such a mode is only for spontaneously breathing *patients*.

- user-powered resuscitators, which are given in ISO 10651-4;
- gas-powered emergency resuscitators, which are given in ISO 10651-5;
- oxygen therapy constant flow *ME equipment*; and
- cuirass and “iron-lung” *ventilators*.

201.1.2 Object

Replacement:

The object of this document is to establish particular *basic safety* and *essential performance* requirements for a *ventilator*, as defined in 201.3.217, and its *accessories*.

Accessories are included because the combination of the *ventilator* and the *accessories* needs to be adequately safe. *Accessories* can have a significant impact on the *basic safety* or *essential performance* of a *ventilator*.

NOTE 1 This document has been prepared to address the relevant *essential principles*^[31] and labelling^[32] guidance of the International Medical Devices Regulators Forum (IMDRF) as indicated in Annex CC.

NOTE 2 This document has been prepared to address the relevant *essential principles of safety and performance* of ISO 16142-1:2016 as indicated in Annex DD.

NOTE 3 This document has been prepared to address the relevant general safety and performance requirements of European regulation (EU) 2017/745^[33] as indicated in Annex EE.

201.1.3 Collateral standards

Amendment (add after existing text):

This document refers to those applicable collateral standards that are listed in Clause 2 of the general standard and in 201.2 of this document.

NOTE The general standard is IEC 60601-1:2005+AMD1:2012+AMD2:2020.

IEC 60601-1-2:2014+AMD1:2020, IEC 60601-1-6:2010+AMD1:2013+AMD2:2020, IEC 60601-1-8:2016+AMD1:2012+AMD2:2020 and IEC 60601-1-11:2015+AMD1:2020 apply as modified in Clauses 202, 206, 208 and 211 respectively. IEC 60601-1-3, IEC 60601-1-9 and IEC 60601-1-12 do not apply.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards can modify, replace, or delete requirements contained in the general standard, including the collateral standards, as appropriate for the particular *ME equipment* under consideration, and may add other *basic safety* or *essential performance* requirements.

A requirement of a particular standard takes priority over IEC 60601-1:2005+AMD1:2012+AMD2:2020 or the collateral standards.

For brevity, IEC 60601-1:2005+AMD1:2012+AMD2:2020 is referred to in this document as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this document corresponds to those of the general standard with the prefix “201” (e.g. 201.1 in this document addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix “2xx” where xx is the final digits of the collateral standard document number (e.g. 202.4 addresses the content of IEC 60601-1-2, Clause 4 collateral standard, 208.4 addresses the content of IEC 60601-1-8, Clause 4 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

- “Replacement” means that the clause or subclause of IEC 60601-1:2005+AMD1:2012+AMD2:2020 or the applicable collateral standard is replaced completely by the text of this document.
- “Addition” means that the text of this particular standard is additional to the requirements of IEC 60601-1:2005+AMD1:2012+AMD2:2020 or the applicable collateral standard.
- “Amendment” means that the clause or subclause of IEC 60601-1:2005+AMD1:2012+AMD2:2020 or the applicable collateral standard is amended as indicated by the text of this document.

Subclauses or figures that are additional to those of the general standard are numbered starting from 201.101. However, due to the fact that definitions in the general standard are numbered 3.1 through 3.154, additional definitions in this document are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses or figures that are additional to those of a collateral standard are numbered starting from 20x, where “x” is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term “this document” is used to make reference to IEC 60601-1:2005+AMD1:2012+AMD2:2020, any applicable collateral standards, and this document taken together.

Where there is no corresponding clause or subclause in this document, the clause or subclause of IEC 60601-1:2005+AMD1:2012+AMD2:2020 or the applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of IEC 60601-1:2005+AMD1:2012+AMD2:2020 or the applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this document.

201.2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

- ISO 32:1977, *Gas cylinders for medical use — Marking for identification of content*
- ISO 3744:2010, *Acoustics — Determination of sound power levels and sound energy levels of noise sources using sound pressure — Engineering methods for an essentially free field over a reflecting plane*
- ISO 4871:1996, *Acoustics — Declaration and verification of noise emission values of machinery and equipment*
- ISO 5356-1:2015, *Anaesthetic and respiratory equipment — Conical connectors — Part 1: Cones and sockets*
- ISO 5359:2014+AMD1:2017, *Low-pressure hose assemblies for use with medical gases*
- ISO 5367:2014, *Anaesthetic and respiratory equipment — Breathing sets and connectors*
- ISO 7396-1:2016+AMD1:2017, *Medical gas pipeline systems — Part 1: Pipeline systems for compressed medical gases and vacuum*
- ISO 9360-1:2000, *Anaesthetic and respiratory equipment — Heat and moisture exchangers (HMEs) for humidifying respired gases in humans — Part 1: HMEs for use with minimum tidal volumes of 250 ml*
- ISO 9360-2:2001, *Anaesthetic and respiratory equipment — Heat and moisture exchangers (HMEs) for humidifying respired gases in humans — Part 2: HMEs for use with tracheostomized patients having minimum tidal volumes of 250 ml*
- ISO 14937:2009, *Sterilization of health care products — General requirements for characterization of a sterilizing agent and the development, validation and routine control of a sterilization process for medical devices*
- ISO 17664-1:2021, *Processing of health care products — Information to be provided by the medical device manufacturer for the processing of medical devices — Part 1: Critical and semi-critical medical devices*
- ISO 17664-2:2021, *Processing of health care products — Information to be provided by the medical device manufacturer for the processing of medical devices — Part 2: Non-critical medical devices*
- ISO 18562-1:2017, *Biocompatibility evaluation of breathing gas pathways in healthcare applications — Part 1: Evaluation and testing within a risk management process*
- ISO 20417:2021, *Medical devices — Information to be supplied by the manufacturer*
- ISO 23328-1:2003, *Breathing system filters for anaesthetic and respiratory use — Part 1: Salt test method to assess filtration performance*
- ISO 23328-2:2002, *Breathing system filters for anaesthetic and respiratory use — Part 2: Non-filtration aspects*
- ISO 80369-1:2018, *Small-bore connectors for liquids and gases in healthcare applications — Part 1: General requirements*
- ISO 80601-2-55:2018, *Medical electrical equipment — Part 2-55: Particular requirements for the basic safety and essential performance of respiratory gas monitors*
- ISO 80601-2-74:2021, *Medical electrical equipment — Part 2-74: Particular requirements for basic safety and essential performance of respiratory humidifying equipment*
- IEC 60601-1:2005+AMD1:2012+AMD2:2020, *Medical electrical equipment — Part 1: General requirements for basic safety and essential performance*
- IEC 61672-1:2013, *Electroacoustics — Sound level meters — Part 1: Specifications*
- IEC 62304:2006+AMD1:2015, *Medical device software - Software life cycle processes*
- IEC 62570:2014, *Standard practice for marking medical devices and other items for safety in the magnetic resonance environment*

IEC 81001-5-1:2021, *Health software and health IT systems safety, effectiveness and security — Part 5-1: Security — Activities in the product life cycle*

IEC Guide 115:2021, *Application of uncertainty of measurement to conformity assessment activities in the electrotechnical sector*

201.3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 60601-1:2005+AMD1:2012+AMD2:2020, and the following apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

201.3.201

accompanying information

information accompanying or *marked* on a medical device or *accessory* for the user or those accountable for the installation, use, *processing*, maintenance, decommissioning and disposal of the medical device or *accessory*, particularly regarding safe use

Note 1 to entry: The *accompanying information* shall be regarded as part of the medical device or *accessory*.

Note 2 to entry: The *accompanying information* can consist of the label, *marking*, *instructions for use*, *technical description*, installation manual, quick reference guide, etc.

Note 3 to entry: *Accompanying information* is not necessarily a written or printed document but could involve auditory, visual, or tactile materials and multiple media types (e.g. CD/DVD-ROM, USB stick, website).

[SOURCE: ISO 20417:2021, 3.2, modified — deleted note 4.]

201.3.202

acknowledged

state of an *alarm system* initiated by *operator* action, where the auditory *alarm signal* associated with a currently active *alarm condition* is inactivated until the *alarm condition* no longer exists or until a predetermined time interval has elapsed

Note 1 to entry: *Acknowledged* only affects *alarm signals* that are active at the time of the *operator* action.

[SOURCE: IEC 60601-1-8:2006+AMD1:2012+AMD2:2020, 3.37]

201.3.203

airway device

device intended to provide a *gas pathway* to and from the *patient's* airway

[SOURCE: ISO 4135:2022, 3.8.1.2]

201.3.204

airway pressure

P_{aw}

pressure at the *patient-connection port* or at the distal *outlet* of the equipment where there is no *patient-connection port*

Note 1 to entry: The *airway pressure* can be derived from pressure measurements made anywhere within the equipment.

[SOURCE: ISO 4135:2022, 3.1.4.41.1]

201.3.205

alarm condition delay

time from the occurrence of a triggering event either in the *patient*, for *physiological alarm conditions*, or in the equipment, for *technical alarm conditions*, to when the *alarm system* determines that an *alarm condition* exists

[SOURCE: IEC 60601-1-8:2006+AMD1:2012+AMD2:2020, 3.2]

201.3.206

alarm limit

threshold used by an *alarm system* to determine an *alarm condition*

[SOURCE: IEC 60601-1-8:2006+AMD1:2012+AMD2:2020, 3.3]

201.3.207

alarm off

state of indefinite duration in which an *alarm system* or part of an *alarm system* does not generate *alarm signals*

[SOURCE: IEC 60601-1-8:2006+AMD1:2012+AMD2:2020, 3.4]

201.3.208

alarm paused

state of limited duration in which the *alarm system* or part of the *alarm system* does not generate *alarm signals*

[SOURCE: IEC 60601-1-8:2006+AMD1:2012+AMD2:2020, 3.5]

201.3.209

alarm setting

alarm system configuration, including but not limited to:

- *alarm limits*;
- the characteristics of any *alarm signal* inactivation states; and
- the values of variables or parameters that determine the function of the *alarm system*

Note 1 to entry: Some algorithmically-determined *alarm settings* can require time to be determined or re-determined.

[SOURCE: IEC 60601-1-8:2006+AMD1:2012+AMD2:2020, 3.8]

201.3.210

alarm signal generation delay

time from the onset of an *alarm condition* to the generation of its *alarm signal(s)*

[SOURCE: IEC 60601-1-8:2006+AMD1:2012+AMD2:2020, 3.10]

201.3.211

artificial ventilation

intermittent elevation of the pressure in the *patient's airway* relative to that in the *lungs* by external means with the intention of augmenting, or totally controlling, the *ventilation* of a *patient*

EXAMPLE Means used to provide *artificial ventilation* are manual resuscitation; mouth-to-mouth resuscitation; automatic *ventilation*; mechanical *ventilation*.

Note 1 to entry: Common classifications of areas of application of *artificial ventilation* are: emergency; transport; home-care; anaesthesia; critical care; rehabilitation.

Note 2 to entry: Classifications used to denote means used for *artificial ventilation* include: positive-pressure; negative-pressure; gas-powered; *operator*-powered; electrically-powered.

Note 3 to entry: Negative-pressure *ventilation* elevates the relative pressure in the airway by intermittently lowering the pressure in the *lungs*.

[SOURCE: ISO 19223:2019, 3.1.10]

201.3.212

attack

attempt to destroy, expose, alter, disable, steal or gain unauthorized access to or make unauthorized use of an asset

[SOURCE: IEC 81001-5-1:2021, 3.5]

201.3.213

audio off

state of indefinite duration in which the *alarm system* or part of the *alarm system* does not generate an auditory *alarm signal*

[SOURCE: IEC 60601-1-8:2006+AMD1:2012+AMD2:2020, 3.12]

201.3.214

audio paused

state of limited duration in which the *alarm system* or part of the *alarm system* does not generate an auditory *alarm signal*

[SOURCE: IEC 60601-1-8:2006+AMD1:2012+AMD2:2020, 3.13]

201.3.215

BAP

quantity by which the baseline *airway pressure* is set to be positively offset from the ambient pressure

[SOURCE: ISO 19223:2019, 3.10.2, modified — deleted notes.]

201.3.216

bias flow

flow that passes through the *ventilator breathing system* to the *exhaust port* but is not intended to contribute to the work of *lung ventilation*

[SOURCE: ISO 19223:2019, 3.7.7, modified — deleted notes.]

201.3.217

biocompatibility

ability to be in contact with a living system without producing an unacceptable adverse effect

Note 1 to entry: Medical devices may produce some level of adverse effect, but that level may be determined to be acceptable when considering the benefits provided by the medical device.

[SOURCE: ISO 18562-1:2017, 3.2]