

ISO 80601-2-55:2018/PRF Amd 1

ISO/TC 121/SC 1

Secretariat: DIN

Date: 2023-10-18

Medical electrical equipment —

Part 2-55: Particular requirements for the basic safety and essential performance of respiratory gas monitors

AMENDMENT 1

Appareils électromédicaux —

Partie 2-55: Exigences particulières relatives à la sécurité de base et aux performances essentielles des moniteurs de gaz respiratoires

AMENDEMENT 1

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

ISO 80601-2-55:2018/PRF Amd 1

<https://standards.iteh.ai/catalog/standards/sist/ee121c37-96d8-460f-98ff-fadbf81373ca/iso-80601-2-55-2018-prf-amd-1>

PROOF/ÉPREUVE

ISO 80601-2-55:2018/PRF Amd 1(E)

© ISO ~~20XX~~2018

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO 80601-2-55:2018/PRF Amd 1](https://standards.iteh.ai/catalog/standards/sist/ee121c37-96d8-460f-98ff-fadbf81373ca/iso-80601-2-55-2018-prf-amd-1)

<https://standards.iteh.ai/catalog/standards/sist/ee121c37-96d8-460f-98ff-fadbf81373ca/iso-80601-2-55-2018-prf-amd-1>

ISO 80601-2-55:2018/PRF Amd 1(E)

Table of Contents

Foreword	4
201.1	5
201.1.1	5
201.1.2	5
201.1.3	5
201.1.4	5
201.2	5
201.3	5
201.4	5
201.4.3	6
201.5	6
201.6	6
201.7	6
201.7.2.3	6
201.7.4.3	6
201.7.9.2	6
201.7.9.2.2	6
201.7.9.2.2.101	6
201.7.9.2.5	7
201.7.9.2.8	7
201.7.9.2.9	7
201.7.9.2.9.101	7
201.7.9.2.13	7
201.7.9.2.14	7
201.7.9.2.15	7
201.7.9.3	7
201.8	7
201.9	7
201.10	7
201.11	8
201.11.6.6	8
201.11.6.7	8
201.12	8
201.12.1	8
201.12.1.101.2	8
201.12.1.101.4.1	8
201.13	9
201.14	9
201.15	9

ISO 80601-2-55:2018/PRF Amd 1(E)

<i>201.15.3.5</i>	9
<i>201.15.3.5.101.1</i>	9
<i>201.15.3.5.101.2</i>	9
<i>201.16</i>	9
<i>201.17</i>	9
<i>202</i>	9
<i>206</i>	9
<i>208</i>	10
<i>208.6.6.2</i>	10
<i>208.6.8.5</i>	10
<i>211</i>	10
<i>212</i>	10
<i>Annex D</i>	10
<i>Annex AA</i>	10
<i>201.11.9.101.2</i>	10
<i>Annex DD</i>	10

ITen Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO 80601-2-55:2018/PRF Amd 1](https://standards.iteh.ai/catalog/standards/sist/ee121c37-96d8-460f-98ff-fadbf81373ca/iso-80601-2-55-2018-prf-amd-1)

<https://standards.iteh.ai/catalog/standards/sist/ee121c37-96d8-460f-98ff-fadbf81373ca/iso-80601-2-55-2018-prf-amd-1>

ISO 80601-2-55:2018/PRF Amd 1(E)

Foreword

ISO (the International Organization for Standardization) and IEC (the International Electrotechnical Commission) form the specialized system for worldwide standardization. National bodies that are members of ISO or IEC participate in the development of International Standards through technical committees established by the respective organization to deal with particular fields of technical activity. ISO and IEC technical committees collaborate in fields of mutual interest. Other international organizations, governmental and non-governmental, in liaison with ISO and IEC, also take part in the work.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives or www.iec.ch/members_experts/refdocs).

Field Code Changed

ISO and IEC draw attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO and IEC take no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO and IEC had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents and <https://patents.iec.ch>. ISO and IEC shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html. In the IEC, see www.iec.ch/understanding-standards.

Field Code Changed

This document was prepared jointly by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 1, *Breathing attachments and anaesthetic machines*, and in collaboration with Technical Committee IEC/TC 62, *Electrical equipment in medical practice*, Subcommittee SC D, *Electrical equipment*, in collaboration and with the European Committee for Standardization (CEN) Technical Committee CEN/TC 215, *Respiratory and anaesthetic equipment*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

A list of all parts in the ISO/IEC 80601 series can be found on the ISO and IEC websites.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html and www.iec.ch/national-committees.

ISO 80601-2-55:2018/PRF Amd 1(E)

www.iso.org/members.html and www.iec.ch/national-committees.

~~Medical electrical equipment — Part 2-55: Particular requirements for the basic safety and essential performance of respiratory gas monitors~~

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO 80601-2-55:2018/PRF Amd 1](https://standards.iteh.ai/catalog/standards/sist/ee121c37-96d8-460f-98ff-fadbf81373ca/iso-80601-2-55-2018-prf-amd-1)

<https://standards.iteh.ai/catalog/standards/sist/ee121c37-96d8-460f-98ff-fadbf81373ca/iso-80601-2-55-2018-prf-amd-1>