

ISO/TC 147/SC 4

Secretariat: DIN

Voting begins on:
2023-10-30

Voting terminates on:
2023-12-25

Water quality — Detection and quantification of *Legionella* spp. and/or *Legionella pneumophila* by concentration and genic amplification by quantitative polymerase chain reaction (qPCR) —

Part 2:

On-site methods

*Qualité de l'eau — Détection et quantification de *Legionella* spp. et/ou *Legionella pneumophila* par concentration et amplification génique par réaction de polymérisation en chaîne quantitative (qPCR) —*

Partie 2: Méthodes sur site

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.



Reference number
ISO/DTS 12869-2:2023(E)

© ISO 2023

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO/DTS 12869-2](https://standards.iteh.ai/catalog/standards/sist/0014d6b6-6880-47ce-8985-03e4ba10469d/iso-dts-12869-2)

<https://standards.iteh.ai/catalog/standards/sist/0014d6b6-6880-47ce-8985-03e4ba10469d/iso-dts-12869-2>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2023

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword.....	v
Introduction.....	vi
1 Scope.....	1
2 Normative references.....	1
3 Terms, definitions, symbols and abbreviated terms.....	2
3.1 Terms and definitions.....	2
3.2 Symbols and abbreviated terms.....	3
4 Principle.....	4
5 Sampling.....	4
6 General testing conditions.....	5
6.1 General.....	5
6.2 End users.....	5
6.2.1 General.....	5
6.2.2 Manufacturer's instructions.....	5
6.2.3 Proficiency.....	6
6.2.4 Usability validation (human factors testing [HFT]).....	7
6.3 Premises.....	7
6.3.1 Manufacturer premises.....	7
6.3.2 End user premises.....	7
6.4 Apparatus and consumables (excluding reagents).....	7
6.4.1 General.....	7
6.4.2 Safety.....	7
6.4.3 Concentration.....	8
6.4.4 PCR (detection and quantification).....	8
6.5 Reagents.....	8
6.6 Decontamination of equipment and premises.....	9
6.7 Maintenance and calibration.....	9
6.8 Treatment and elimination of waste.....	9
7 Procedure.....	9
7.1 Concentration.....	9
7.2 Bacterial concentration and bacterial recovery.....	9
7.2.1 General.....	9
7.2.2 Protocols.....	10
7.2.3 Stability of bacterial eluates and DNA.....	10
7.3 DNA amplification by PCR.....	10
7.3.1 General.....	10
7.3.2 Target sequences, primers and probes.....	10
7.3.3 Amplification mix preparation.....	10
7.4 Quantitative detection.....	11
7.4.1 General.....	11
7.4.2 Protocol.....	12
7.5 Qualitative detection.....	13
8 Expression of the results.....	13
9 Technical protocol for the characterization and the validation of the method.....	13
9.1 General.....	13
9.2 Inclusivity and exclusivity of probes and primers.....	13
9.3 Verification of the calibration function of the quantitative PCR phase.....	13
9.4 Verification of the PCR limit of detection, L_{DqPCR}	13
9.5 Verification of the PCR limit of quantification, L_{QqPCR}	13
9.5.1 Principle.....	13
9.5.2 Experimental design.....	14

9.5.3	Analysis of results.....	14
9.5.4	Theoretical limit of quantification of the whole method.....	14
9.6	Recovery/accuracy.....	14
9.6.1	Principle.....	14
9.6.2	Protocol for preparation of bacteria.....	14
9.7	Precision.....	15
9.7.1	General.....	15
9.7.2	Reproducibility.....	15
9.7.3	Intermediate precision.....	15
9.8	Robustness.....	16
9.9	Measurement uncertainty of the whole method.....	16
9.10	On-site verification by end user.....	16
10	Quality control.....	16
10.1	General.....	16
10.2	Connecting the calibration solution and the reference material to the primary standard.....	17
10.2.1	Principle.....	17
10.2.2	Protocol.....	17
10.2.3	Data analysis.....	17
10.3	Monitoring performance.....	17
10.4	Positive and negative controls of the method.....	17
10.4.1	Positive and negative controls performed by the manufacturer.....	17
10.4.2	Positive and negative controls performed by the end user.....	17
10.5	No template control.....	18
10.6	Inhibition control.....	18
10.6.1	General.....	18
10.6.2	Inhibition control is the target.....	18
10.6.3	Inhibition control is either a plasmid or an oligonucleotide.....	18
11	Test report.....	19
Annex A (normative) Responsibilities of the manufacturer and the end user.....		20
Annex B (informative) Usability validation protocol (human factor testing).....		21
Annex C (informative) Example on-site system verification protocol.....		23
Bibliography.....		25

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 147, *Water quality*, Subcommittee SC 4, *Microbiological methods*.

A list of all parts in the ISO 12869 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

ISO/TS 12869 provides the guidelines, minimum requirements and performance characteristics intended to guarantee that the quantification of *L. pneumophila* or *Legionella* spp. by amplification of specific DNA sequences (PCR) and real time detection of specific of specific DNA sequences (PCR) and real-time detection of specific fluorophores is reproducible between methodologies completed by different laboratories.

Similar to ISO/TS 12869, this document specifies a method to determine recovery of the bacteria and subsequent DNA amplification (lysis efficiency is not estimated).

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO/DTS 12869-2](https://standards.iteh.ai/catalog/standards/sist/0014d6b6-6880-47ce-8985-03e4ba10469d/iso-dts-12869-2)

<https://standards.iteh.ai/catalog/standards/sist/0014d6b6-6880-47ce-8985-03e4ba10469d/iso-dts-12869-2>

Water quality — Detection and quantification of *Legionella* spp. and/or *Legionella pneumophila* by concentration and genic amplification by quantitative polymerase chain reaction (qPCR) —

Part 2: On-site methods

1 Scope

This document provides the guidelines, minimum requirements and performance characteristics intended to guarantee that manufactured systems intended for on-site/field use (i.e. outside the laboratory) provide reliable and reproducible results.

This document specifies the requirements for technologies that enable on-site detection and quantification of *Legionella* spp. and *L. pneumophila* using a quantitative polymerase chain reaction assay (qPCR). It specifies general methodological requirements, performance evaluation requirements and quality control requirements. This document is intended to be used by manufacturers of these technologies so that they produce detection systems that end users can operate safely and effectively. End users will be guided by this document to adhere to manufacturer's instructions, to ensure user competency and to perform the necessary controls.

Technical details specified in this document are given for information only. Any other technical solutions complying with the performance requirements are suitable.

NOTE For validation and performance requirements, see [Clause 9](#).

This document is intended to be applied in the bacteriological investigation of all types of water (hot or cold water, cooling tower water, etc.), unless the nature and/or content of suspended matter and/or background microorganisms interfere with the determination. This interference can result in an adverse effect on both the detection limit and the quantification limit.

The results are expressed as the number of genome units of *Legionella* spp. and/or *L. pneumophila* per millilitre (or litre) of sample.

Although the method described in this document is applicable to all types of water, some additives, such as chemicals used for water treatment, can interfere with and/or affect the sensitivity of the method.

The qPCR methods do not give any information about the physiological state of the *Legionella*. However, there are on-site qPCR methodologies which are able to distinguish intact bacteria from free DNA.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 5667-1, *Water quality — Sampling — Part 1: Guidance on the design of sampling programmes and sampling techniques*

ISO 19458, *Water quality — Sampling for microbiological analysis*

ISO/TS 12869:2019, *Water quality — Detection and quantification of Legionella spp. and/or Legionella pneumophila by concentration and genic amplification by quantitative polymerase chain reaction (qPCR)*

ISO 11731, *Water quality — Enumeration of Legionella*

3 Terms, definitions, symbols and abbreviated terms

3.1 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO/TS 12869 and the following apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

3.1.1

Legionella spp.

several species of *Legionella*, including *L. pneumophila*

3.1.2

PCR inhibition control

materials and processes used to assess if the sample DNA extract contains (an) inhibitor(s)

Note 1 to entry: The control can be a plasmid, an oligonucleotide or the *L. pneumophila* genomic DNA. A specific probe shall be used to detect the inhibition control.

3.1.3

bacterial recovery

evaluation of the reported quantity of bacteria by the *on-site qPCR* (3.1.7) system when a known quantity of reference material is tested

3.1.4

working calibration solution

L. pneumophila DNA calibrated solutions, derived from a standard solution, for which accuracy is determined by an independent method (e.g. digital droplet PCR). Used to establish the calibration curve

3.1.5

negative control of the method

control for monitoring the whole process in this method (from filtration to extraction to qPCR)

3.1.6

no template control

NTC

control for monitoring qPCR reagent amplification

3.1.7

on-site qPCR

qPCR testing that can occur immediately after sample collection, such that sample preservation is not required (e.g. sodium thiosulfate)

Note 1 to entry: On-site qPCR is validated for use by non-laboratory personnel that have been trained in the procedure.

3.1.8**concentration device**

device that prepares a water sample for qPCR amplification

Note 1 to entry: This kind of device is designed such that it can be used safely and effectively by non-laboratory trained personnel.

3.1.9**threshold cycle** C_t

Note 1 to entry: number of PCR cycles (denaturation and amplification) required to replicate the DNA copies originally present in the sample, so that the concentration of DNA exceeds the detection limit

Note 2 to entry: The C_t value is the intercept of the line that represents the DNA concentration of a sample with fluorescent base line.

3.1.10**genome unit****GU**

Note 1 to entry: unit representing a single copy of bacterial genomic DNA

3.1.11**graphical user interface****GUI**

on-screen controls of the testing equipment which can describe sample concentration and analysis procedure

Note 1 to entry: The interface is designed such that it can be used and understood by non-laboratory personnel that have been trained in the procedure.

3.1.12**end user**

operator

individual who performs the assay on the test system

3.1.13**critical task**

step in the on-site test workflow that can lead to a hazardous situation, such as an incorrect test result and/or injury to the test operator (3.1.13), if performed incorrectly

3.1.14**batch**

manufacturing production run used to generate one or more lots of finished goods

3.2 Symbols and abbreviated terms

C_t	threshold cycle
$D_{\text{opt}600}$	optical density at 600 nm
$L_{D\text{qPCR}}$	(detection limit of the qPCR) lowest number of genome units that give a positive result in the qPCR with 90 % confidence
$L_{Q\text{qPCR}}$	(quantification limit of the qPCR) lowest number of genome units that can be quantified with an accuracy less than or equal to $0,20 \log_{10}$ unit
V_b	volume of the bacterial sample in the reaction
V_f	final volume of the reaction

V_w	volume of water in the reaction
BCYE	buffered charcoal yeast extract
BSA	bovine serum albumin
DMSO	dimethyl sulfoxide
GMP	good manufacturing practice
GU	genome unit
GUI	graphical user interface
HFT	human factors testing
NTC	no template control
PPE	personal protective equipment
QC	quality control
UNG	uracil-DNA N-glycosylase

4 Principle

The detection and quantification of *Legionella* spp. or *L. pneumophila* by on-site qPCR are carried out in three phases:

- concentration of bacteria from water samples by the concentration device;
- recovery of the bacteria from the concentration device and transfer of the bacteria to a vessel or apparatus in which bacterial lysis and DNA extraction occur; PCR can then proceed in the same apparatus or an additional step can be required to transfer the extracted DNA to a PCR reaction;
- amplification, detection and quantification of one or more specific DNA sequences belonging to the *Legionella* genus and/or *L. pneumophila* species by real-time PCR.

5 Sampling

Sampling shall be in accordance with ISO 19458, however one of the main advantages of on-site testing is the ability to test samples immediately after collection, thus mitigating the known effects of time on sample quality (see Reference [1]). The manufacturer shall indicate to the end user the acceptable holding times between sample collection and analysis. These time intervals can vary between water sources (e.g. potable water without biocides present versus cooling tower samples that contain biocides). The manufacturer shall validate these holding times, which will be provided in the instructions (see 6.2.2).

Sampling conditions (e.g. water treatment, temperature, turbidity, time that water was run prior to sampling) shall be indicated on the test report if they are known. Manufacturers will validate conditions, including temperature and commonly used chemicals (e.g. biocides, neutralizing agents, anti-corrosives) in intended sample types, that are compatible with the testing system. Manufacturers will indicate the compatible sample conditions to end users. Samples shall not be exposed to conditions that the manufacturer has not validated.

Biocides (bactericides or bacteriostatics) are sometimes used, in particular in cooling tower circuits. The presence of biocides, however, which can lead to under quantification of the analyte, therefore the presence of biocides shall thus be declared and indicated on the test report if it is known. When inhibition of PCR sufficient to result in under quantification is detected, the test result shall be suppressed and a warning message specifying that interference was encountered shall be provided to

the end user. Where appropriate, sample containers shall contain a suitable neutralizing agent (refer to ISO 19458). As it is not always possible to neutralize these products, minimizing the elapsed time between sample collection and analysis is recommended.

Manufacturer's shall indicate to end users the need for a sampling plan and refer users to ISO 5667-1 for guidance.

6 General testing conditions

6.1 General

PCR is a sensitive detection method, the results of which can be affected by aerosols, dust and other particles which can contain contaminating DNA. It is therefore essential to physically separate the different stages of the analysis. The on-site qPCR concentration device shall be designed in such a way to prevent this type of contamination.

The principles to be applied are as follows:

- single use concentration device and qPCR reagent;
- procedures for eliminating DNA traces and amplicons shall be described to the user in the event of accidental contamination of the premises or apparatus;
- regular manufacturing quality controls shall be used to demonstrate the effectiveness of the concentration device and qPCR reagent production procedures with the objective of ensuring that there is no contaminating *Legionella* DNA or PCR products/ amplicons (see [10.4](#)).

The manufacturer and the end users shall fulfil the responsibilities listed in [Annex A](#).

6.2 End users

6.2.1 General

All personnel who perform on-site qPCR shall be provided with instructions to operate the system, as well as training materials as needed. Instructions shall be provided as a physical copy of the instructions, a training video or interactive instructions provided by a GUI (graphical user interface).

The test operators shall wear personal protective equipment required for sample collection as per jurisdictional guidelines. Gloves are required. They shall be disposable and talc-free.

As the qPCR results shall be analysed and interpreted by software and expressed to the user via the GUI in appropriate units (e.g. GU/ml or GU/l), the operator shall not require additional advanced training or experience in PCR data analysis. Likewise, the presence of inhibition shall be determined, via PCR inhibition control, by automated analysis of the data by the software.

6.2.2 Manufacturer's instructions

Instructions shall be provided to end users by manufacturers. The instructions shall include clear and specific information necessary to safely and effectively perform tests on-site. The following topic shall be included in the manufacturer's instructions to end users.

- Intended use: Statement of the test system input material(s) and result output(s) and how the results may be used. The latter shall include a statement notifying the end user to be aware of jurisdictional requirements and how they can affect how the results are used. The intended end user shall also be provided.
- Warnings and precautions: Description of safety measures required to avoid any risk of harm to the operator or other individuals when the test is performed. Risks considered shall include the risk of exposure to aerosolized *Legionella* as well as any critical concerns that can lead to incorrect results