
**Plastics collapsible containers
for human blood and blood
components —**

**Part 1:
Conventional containers**

AMENDMENT 1

*Poches en plastique souple pour le sang et les composants du sang —
Partie 1: Poches conventionnelles*

AMENDMENT 1 md 1:2023

<https://standards.itih.ai/catalog/standards/sist/28044400-4b88-4003-843c-7aef615fd00/iso-3826-1-2019-amd-1-2023>



iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO 3826-1:2019/Amd 1:2023

<https://standards.iteh.ai/catalog/standards/sist/28044400-4b88-4003-843c-7aee615fd00/iso-3826-1-2019-amd-1-2023>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2023

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 76, *Transfusion, infusion and injection, and blood processing equipment for medical and pharmaceutical use*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 205, *Non-active medical devices*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

A list of all parts in the ISO 3826 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Plastics collapsible containers for human blood and blood components —

Part 1: Conventional containers

AMENDMENT 1

6.2.5

Replace the text with the following:

6.2.5.1 General

The instructions for use shall indicate if the plastics container is intended for freezing and/or irradiation applications.

The user shall be aware of particular requirements from any other regulatory authority (e.g. EDQM Guide to the preparation, use and quality assurance of blood components).

6.2.5.2 Freezing

This requirement refers primarily to plasma-freezing bags.

The plastics container, filled to half of its nominal capacity with water as specified in ISO 3696, shall withstand a slow freezing to and storage at -80 °C for 24 h, subsequent immersion in water at $(37 \pm 2)\text{ °C}$ for 60 min, and returning to a temperature of $(23 \pm 2)\text{ °C}$. The plastics container shall meet the requirements of 5.6.3, 5.9, 6.2.7, and 6.2.8.

Plastics containers intended to be shock-frozen (blast frozen) shall be validated for this application.

If a refrigerant solution is used, the plastics container may be enclosed in a protective bag to avoid direct contact between the refrigerant solution and the plastics container.

6.2.5.3 Ionizing irradiation

This requirement refers primarily to containers intended to store irradiated blood components.

The plastics container, filled to nominal capacity with water (for containers with a nominal capacity of greater 350 ml the maximum filling volume shall not exceed 350 ml), shall withstand a maximum irradiation dose of $50\text{ m}^2\cdot\text{s}^{-2}$ (Gy) using validated irradiation equipment.

The plastics container following irradiation shall meet the requirements of 5.6.3, 5.9, 6.2.4, 6.2.7 and 6.2.8.

The integrity of plastics containers intended to be irradiated shall be validated for this specific application.