ISO TC 30/SC 05/WG 15

Secretariat: SNV

Date: 2023-09-09xx

Measurement of fluid flow in closed conduits — Clamp-on ultrasonic transit-time meters for liquids and gases.

<u>Mesurage du débit des fluides dans les conduites fermées — Débitmètres non-intrusifs à ultrasons</u> <u>à temps de transit pour les liquides et les gaz</u> Style Definition: Heading 1: Indent: Left: 0 pt, First

line: 0 pt, Tab stops: Not at 21.6 pt

Style Definition: Heading 2: Font: Bold, Tab stops: Not at 18 pt

Style Definition: Heading 3: Font: Bold

Style Definition: Heading 4: Font: Bold

Style Definition: Heading 5: Font: Bold **Style Definition:** Heading 6: Font: Bold

Style Definition: ANNEX

Style Definition: AMEND Terms Heading: Font: Bold

Style Definition: AMEND Heading 1 Unnumbered:

Font: Bold

Formatted: French (Switzerland)

Formatted: French (Switzerland)

Formatted: French (Switzerland)

Formatted: Font: Cambria

iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO/FDIS 24062

© ISO 2023

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office

CP 401 • Ch. de Blandonnet 8

CH-1214 Vernier, Geneva

Phone: +41 22 749 01 11

Email: copyright@iso.org

Website: www.iso.org

Published in Switzerland

iTeh STANDARD PREVIEW (standards.iteh.ai)

<u>ISO/FDIS 24062</u>

Contents

Forew	ordvi		
Introd	luctionvii		
1	Scope1		
2	Normative references1		
3	Terms and definitions		
3.1	Quantities1		
3.1.1	volume flowrate1		
3.1.2	mean flow velocity1		
3.1.3	path velocity2		
3.1.4	Reynolds number		
3.2	Meter design		
3.2.1	ultrasonic path2		
3.2.2	electronic unit		
3.3	Thermodynamic conditions		
3.3.1	metering conditions2		
3.3.2	standard conditions2		
3.4	Statistics3	7	
3.4.1	error3		
3.4.2	repeatability (of results of measurements)3		
3.4.3	reproducibility (of results of measurements)		
3.4.4	resolution4		
3.4.5	zero flow4		
3.4.6	uncertainty (of measurement)4		
3.4.7	standard uncertainty4		
3.4.8	expanded uncertainty 4		
3.4.9	coverage factor5	5-5	
3.5	Calibration		
3.5.1	flow calibration 5		
3.6	Symbols and subscripts6		
4	Principle of measurement		
4.1	General		
4.2	Generic description7		
4.2.1	Transducers9		
4.2.2	Ultrasonic path configurations		
4.2.3	Electronic unit functionality		
4.3	Special considerations		
4.3.1	Gas		
4.3.2	Extreme temperatures		
4.4	Additional measurements		
4.4.1	Standard volume flow		
4.4.2	Heat flow		
4.4.3	Mass flow		
5	Test and calibration13		
5.1	General		
5.2	Individual testing — Flow calibration under flowing conditions		
5.2.1	General		
5.2.2	Laboratory flow calibration13		

6	Installation	15
6.1	General	
6.2	Installation location	15
6.2.1	Upstream and downstream straight pipe length requirements and path	
	configuration	15
6.3	Transmitter programming	16
6.4	Pressure and temperature inputs	16
6.5	Transducer installation	17
6.5.1	Transducer type	17
6.5.2	Transducer orientation	
6.5.3	Coupling	
6.5.4	Transducer mounting	
6.6	Gas considerations	
6.7	Cable	
6.8	Additional installation effects	
6.8.1	Non-steady flow	
6.8.2	Contamination by secondary phase	20
6.8.3	Vibration	
6.8.4	Thermal insulation	
6.8.5	Stratification	
6.8.6	Wall roughness signal considerations	
6.8.7	Non-Newtonian fluids	
6.9	Operational diagnostics	
6.9.1	Fluid speed of sound	
6.9.2	Gain	22
6.9.3	Signal to noise ratio	22
7	Maintenance	22
•		
8	Uncertainty in measurement	
8.1	Mathematical model	
8.2	Evaluation of the contributory variances	
8.2.1	Uncertainty of the cross-sectional area, $u(A)$	
8.2.2	Uncertainty in the velocity profile, $u(K_p)$	
8.2.3	Uncertainty of the path-geometry factor, $u(K_g)$ and time measurement	26
8.2.4	Uncertainty of time measurement	26
8.2.5	Uncertainty of the delay time, $u(t_0)$	27
8.2.6	Effect of multiple paths	27
8.3	Additional effects on measurement uncertainty	27
8.3.1	Effect of pipe material on sound speed	27
8.3.2	Effect of fluid sound speed	
8.3.3	Reproducibility of transmitter mounting	
8.3.4	External influences	
Annov	A (informative) Example of uncertainty calculation	
	Seneral	
	Mathematical model	
	<u>Installation conditions</u>	
	<u>Evaluation of the contributory variances</u>	
A.1.3.1	Uncertainty of the cross-sectional area, u(1)	29
A.1.3.2	! Uncertainty in the velocity profile, u(K _p)	30
A.1.3.3	Uncertainty of the path-geometry factor, u(K _e)	30

Formatted: Default Paragraph Font, Font: Bold, English (United Kingdom)

A.1.3.4 Uncertainty of the time measurement	30
A.1.3.5 Uncertainty of the delay time, $u(t_0)$	30
A.1.3.6 Effect of multiple paths	32
A.1.4 Numerical examples for combined standard uncertainty	32
A.1.5 Propagation of Uncertainty	35
Pibliography	36

iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO/FDIS 24062

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives/.

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents.. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 30, *Measurement of fluid flow in closed conduits*, Subcommittee SC 5, *Velocity and mass methods*.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Non-intrusive (clamp-on) ultrasonic meters (USMs) have become one of the accepted flow measurement technologies for a wide range of applications, including process and control measurements. Non-intrusiveness also brings characteristics relevant for economics, safety, and environment. Ultrasonic technology has inherent features such as no pressure loss and wide rangeability. USMs can deliver diagnostic information through which it may be possible to demonstrate that an ultrasonic flowmeter is performing in accordance with specification.

This document provides a description of the non-intrusive (clamp-on) meter, typical application areas, the measures which should be used in assessing the likely performance in terms of error, repeatability and reproducibility when used under ideal and non-ideal operational conditions.

iTeh STANDARD PREVIEW (standards.iteh.ai)

<u>ISO/FDIS 24062</u> https://standards.iteh.ai/catalog/standards/sist/8359db50-44af-4aa5-9<mark>4a0-d07f0c8b74f3/iso</mark>

iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO/FDIS 24062

Measurement of fluid flow in closed conduits — Clamp-on ultrasonic transit-time meters for liquids and gases

1 Scope

This document specifies requirements and recommendations for non-intrusive (clamp-on) ultrasonic flowmeters (USMs), which use the transit time of ultrasonic signals to measure the volumetric flowrate in closed conduits. Transit time flowmeters are predominantly used on single-phase fluids (liquid and gases) but can also be used where small quantities of additional phases are present.

This document specifies performance, calibration, and output characteristics, and deals with installation conditions.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 4006, Measurement of fluid flow in closed conduits — Vocabulary and symbols

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 4006 and the following apply. ISO and IEC maintain terminology databases for use in standardization at the following addresses:

— ISO Online browsing platform: available at https://www.iso.org/obp

__ IEC Electropedia: available at https://www.electropedia.org/

3.1 Quantities

3.1.1 volume flowrate

 q_{\downarrow}

$$\frac{q_V - \frac{dV}{dt}}{dt}$$

$$q_V = \frac{\mathrm{d}V}{\mathrm{d}t}$$

where

V is volume;

t is time

Formatted: English (United States)

Formatted: Adjust space between Latin and Asian text, Adjust space between Asian text and numbers, Tab stops: Not at 19.85 pt + 39.7 pt + 59.55 pt + 79.4 pt + 99.25 pt + 119.05 pt + 138.9 pt + 158.75 pt + 178.6 pt + 198.45 pt

Formatted: English (United States)

Formatted: Font: Italic

Field Code Changed

[SOURCE: ISO 80000-4:2006, 4-31[2]]

3.1.2

mean flow velocity

volume flowrate (3.1.1) divided by the cross-sectional area of the pipe

path velocity

average fluid velocity on an ultrasonic path

3.1.4

Reynolds number

 Re_D

dimensionless parameter expressing the ratio between the inertia and viscous forces in the fluid

$$Re_{D} = \frac{\rho v_{a}D}{\mu} = \frac{v_{a}D}{v_{kv}}$$

$$Re_D = \frac{\rho v_A D}{\mu} = \frac{v_A D}{v_{kv}}$$

Field Code Changed

where

is density; ρ

 $v_e v_A$ is the mean flow velocity;

D

is the pipe internal diameter;

is the dynamic viscosity; Is.iteh.ai/catalog/standards/sist/8359db50-44af-4aa5-94a0-d07f0c8b74f3/isoμ

iTeh STANDARD PREVI

is the kinematic viscosity

3.2 Meter design

3.2.1

ultrasonic path

path travelled by an ultrasonic signal between a pair of ultrasonic transducers

3.2.2

electronic unit

part of the meter that controls the transducers, processes the signals into a flow-rate and provides outputs (see subclause 4.2.3)

3.3 Thermodynamic conditions

3.3.1

metering conditions

conditions, at the point of measurement, of the fluid of which the volume flow is to be measured

Note 1 to entry: Also known as operating conditions or actual conditions.

3.3.2

standard conditions

defined temperature and pressure conditions used in the measurement of fluid quantity so that the standard volume is the volume that would be occupied by a quantity of fluid if it were at standard temperature and pressure

Note 1 to entry: Standard conditions may be defined by regulation or contract.

Note 2 to entry: Not preferred alternatives: reference conditions, base conditions, normal conditions, etc (see ISO-91:2017[3]).

Note 3 to entry: Metering and standard conditions relate only to the volume of the fluid to be measured or indicated and should not be confused with rated operating conditions or reference conditions (see ISO/IEC Guide 99:2007, 4.9 and $4.11^{[4]}$), which refer to influence quantities (see ISO/IEC Guide 99:2007, $2.52^{[4]}$).

3.4 Statistics

3.4.1

error

measured quantity value minus a reference quantity value

[SOURCE: ISO/IEC Guide 99:2007, 2.16[4]]

3.4.2

repeatability (of results of measurements)

closeness of the agreement between the results of successive measurements of the same measurand carried out under the same conditions of measurement

Note 1 to entry: These conditions are called repeatability conditions.

Note 2 to entry: Repeatability conditions include:

ISO/FDIS 24062

- the same measurement procedure; ai/catalog/standards/sist/8359db50-44af-4aa5-94a0-d07f0c8b74f3/iso-
- the same observer;
- the same measuring instrument, used under the same conditions;
- the same location;
- repetition over a short period of time.

Note 3 to entry: Repeatability may be expressed quantitatively in terms of the dispersion characteristics of the results.

[SOURCE: ISO/IEC Guide 98-3:2008, B.2.15^[5]]

3.4.3

reproducibility (of results of measurements)

closeness of the agreement between the results of measurements of the same measurand carried out under changed conditions of measurement

Note 1 to entry: A valid statement of reproducibility requires specification of the conditions changed.

Note 2 to entry: The changed conditions may include:

principle of measurement;

_	method	οf	measurement:

observer:

measuring instrument;

reference standard:

location;

conditions of use;

time.

Note 3 to entry: Reproducibility may be expressed quantitatively in terms of the dispersion characteristics of the

Note 4 to entry:—Results are here usually understood to be corrected results.

[SOURCE: ISO/IEC Guide 98-3:2008, B.2.16[5]]

resolution

smallest difference between indications of a meter that can be meaningfully distinguished

meter reading when the fluid is at rest, i.e. both axial and non-axial velocity components are essentially zero

3.4.6

uncertainty (of measurement)

parameter, associated with the result of measurement, that characterizes the dispersion of the values about 100 measurement, that characterizes the dispersion of the values about 100 measurement, that characterizes the dispersion of the values about 100 measurement, that characterizes the dispersion of the values about 100 measurement, that characterizes the dispersion of the values about 100 measurement, that characterizes the dispersion of the values about 100 measurement, that characterizes the dispersion of the values about 100 measurement, that characterizes the dispersion of the values about 100 measurement, that characterizes the dispersion of the values about 100 measurement, that characterizes the dispersion of the values about 100 measurement, that characterizes the dispersion of the values about 100 measurement, that characterizes the dispersion of the values about 100 measurement 100 m that could reasonably be attributed to the measurand

Note 1 to entry: The parameter may be, for example, a standard deviation (or a given multiple of it), or the halfwidth of an interval having a stated level of confidence.

Note 2 to entry: Uncertainty of measurement comprises, in general, many components. Some of these components may be evaluated from the statistical distribution of the results of a series of measurements and can be characterized by experimental standard deviations. The other components, which can also be characterized by standard deviations, are evaluated from assumed probability distributions based on experience or other information.

Note 3 to entry: It is understood that the result of the measurement is the best estimate of the value of the measurand, and that all components of uncertainty, including those arising from systematic effects, such as components associated with corrections and reference standards, contribute to the dispersion.

[SOURCE: ISO/IEC Guide 98-3:2008, B.2.18^[5]]

3.4.7

standard uncertainty

uncertainty of the result of a measurement expressed as a standard deviation

[SOURCE: ISO/IEC Guide 98-3:2008, 2.3.1^[5]]

3.4.8

expanded uncertainty

quantity defining an interval about the result of a measurement that may be expected to encompass a large fraction of the distribution of values that could reasonably be attributed to the measurand

[SOURCE: ISO/IEC Guide 98-3:2008, 2.3.5[5]]

Note 1 to entry: The large fraction is normally 95 % and is generally associated with a coverage factor k = 1,96.

Note 2 to entry: The expanded uncertainty is often referred to as the uncertainty.

coverage factor

numerical factor used as a multiplier of the standard uncertainty in order to obtain an expanded uncertainty (3.4.98)

[SOURCE: ISO/IEC Guide 98-3:2008, 2.3.6^[5]]

3.5 Calibration

3.5.1

flow calibration
calibration of a meter against a reference using fluid flowing through the meter

3.6 Symbols and subscripts

The symbols and subscripts used in this document are given in Table 1 and Table 2.

Table 1 — Symbols O/FDIS 24062

https://standa _{Quantity} h.ai/catalog/standa	Symbol	Dimension ^a	SI unit	94a0-d07f0c8b74f3/iso-
Cross-sectional area of pipe	1d _A s-240	L_2	m ²	
Speed of sound in fluid	С	LT ⁻¹	m/s	
Internal pipe diameter	$D_{\rm i}$	L	m	
External pipe diameter	D _e	L	m	
Integers (1.2.3. ,,,)	i,j,n	_	1	
Calibration factor	K	_	1	
Path-geometry factor	K_{g}	L ^b or LT ^{-1 c}	m ^b or m/s ^c	
Velocity profile correction factor	$K_{\rm p}$	_	1	
Minimum distance to a specified upstream flow disturbance	l_{\min}	L	m	
Path length	l_{p}	L	m	
Absolute pressure	р	ML-1T -2	Pa	
Volume flowrate	$q_{\underline{\nu}}$	L3T-1	m³/s	Formatted: Font: Italic
Pipe Reynolds number	Rep	_	1	Formatted: Font: Italic
Transit time	$t_{ m tr}$	Т	s	
Time delay	t_0	T	S	