



Designation: ~~D6035~~—08 D6035/D6035M – 13

Standard Test Method for Determining the Effect of Freeze-Thaw on Hydraulic Conductivity of Compacted or Intact Soil Specimens Using a Flexible Wall Permeameter¹

This standard is issued under the fixed designation ~~D6035~~; D6035/D6035M; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope*

1.1 This test method covers laboratory measurement of the effect of freeze-thaw on the hydraulic conductivity of compacted or intact soil specimens using Test Method D5084 and a flexible wall permeameter to determine hydraulic conductivity. This test method does not provide steps to perform sampling of, or testing ~~on~~of, in situ soils that have already been subjected to freeze-thaw conditions.

1.2 This test method may be used with intact specimens (block or thin-walled) or laboratory compacted specimens and shall be used for soils that have an initial hydraulic conductivity less than or equal to $1\text{E-}5$ m/s [$3.94\text{E-}4$ in./s] ($1\text{E-}3$ cm/s) (**Note 1**).

NOTE 1—The maximum initial hydraulic conductivity is given as $1\text{E-}3$ cm/s, $\text{E-}5$ m/s [$3.94\text{E-}4$ in./s]. This should also apply to the final hydraulic conductivity. It is expected that if the initial hydraulic conductivity is $1\text{E-}3$ cm/s, $\text{E-}5$ m/s ($3.94\text{E-}4$ in./s), then the final hydraulic conductivity will not change (increase) significantly (that is, greater than $1\text{E-}3$ cm/s), $\text{E-}5$ m/s ($3.94\text{E-}4$ in./s).

1.3 Soil specimens tested using this test method can be subjected to three-dimensional freeze-thaw (herein referred to as 3-d) or one-dimensional freeze-thaw (herein referred to as 1-d). (For a discussion of one-dimensional freezing versus three-dimensional freezing, refer to Zimmie² or Othman.³)

1.4 Soil specimens tested using this test method can be tested in a closed system (that is, no access to an external supply of water during freezing) or an open system.

1.5 All observed and calculated values shall conform to the guidelines for significant digits and rounding established in Practice D6026.

1.6 The values stated in SI units or inch-pound units (presented in brackets) are to be regarded as the standard, unless other units are specifically given. By tradition, it is U.S. practice to report hydraulic conductivity in centimetres per second, although the common SI units for hydraulic conductivity are metres per second. The values are to be calculated and reported in accordance with Practices separately as standard. The values stated in each system may not be exact equivalents; therefore, each system shall be used independently of the other. Combining values from the two systems may result in non-conformance with the standard. Reporting of test results in units other than SI shall not ~~D6026~~ be regarded as conformance with this test method.

1.7 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 ASTM Standards:⁴

D653 Terminology Relating to Soil, Rock, and Contained Fluids

¹ This test method is under the jurisdiction of ASTM Committee D18 on Soil and Rock and is the direct responsibility of Subcommittee D18.19 on Frozen Soils and Rock. Current edition approved Jan. 1, 2008 Aug. 15, 2013. Published February 2008 September 2013. Originally approved in 1996. Last previous edition approved in 2002 2008 as ~~D6035-02~~ D6035-08. DOI: 10.1520/D6035-08-10.1520/D6035_D6035M-13.

² Zimmie, T. F., and La Plante, C., "The Effect of Freeze/Thaw Cycles on the Permeability of a Fine-Grained Soil," *Hazardous and Industrial Wastes, Proceedings of the Twenty-Second Mid-Atlantic Industrial Waste Conference*, Joseph P. Martin, Shi-Chieh Cheng, and Mary Ann Susavidge, eds., Drexel University, 1990, pp. 580-593.

³ Othman, M. A., Benson, C. H., Chamberlain, E. J., and Zimmie, T. F., "Laboratory Testing to Evaluate Changes in Hydraulic Conductivity of Compacted Clays Caused by Freeze-Thaw: State-of-the-Art," *Hydraulic Conductivity and Waste Contaminant Transport in Soils, ASTM STP 1142*, David E. Daniel, and Stephen J. Trautwein, eds., American Society for Testing and Materials, Conshohocken, PA, pp. 227-254. Othman, M. A., Benson, C. H., Chamberlain, E. J., and Zimmie, T. F., "Laboratory Testing to Evaluate Changes in Hydraulic Conductivity of Compacted Clays Caused by Freeze-Thaw: State-of-the-Art," *Hydraulic Conductivity and Waste Contaminant Transport in Soils, ASTM STP 1142*, David E. Daniel, and Stephen J. Trautwein, eds., American Society for Testing and Materials, Conshohocken, PA, pp. 227-254.

⁴ For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

*A Summary of Changes section appears at the end of this standard



- D1587 Practice for Thin-Walled Tube Sampling of Soils for Geotechnical Purposes
- D2113 Practice for Rock Core Drilling and Sampling of Rock for Site Investigation
- D2216 Test Methods for Laboratory Determination of Water (Moisture) Content of Soil and Rock by Mass
- D3740 Practice for Minimum Requirements for Agencies Engaged in Testing and/or Inspection of Soil and Rock as Used in Engineering Design and Construction
- D4220 Practices for Preserving and Transporting Soil Samples
- D4753 Guide for Evaluating, Selecting, and Specifying Balances and Standard Masses for Use in Soil, Rock, and Construction Materials Testing
- D5084 Test Methods for Measurement of Hydraulic Conductivity of Saturated Porous Materials Using a Flexible Wall Permeameter
- D6026 Practice for Using Significant Digits in Geotechnical Data
- E145 Specification for Gravity-Convection and Forced-Ventilation Ovens

3. Terminology

3.1 Definitions: Definitions

3.1.1 For common definitions of other terms in this standard, see Terminology D653.—For

3.1.2 ~~Hydraulic conductivity~~—Refer to Terminology D653 for standard definition of this term. common definitions of other terms in this standard, see Terminology D653, including hydraulic conductivity.

3.2 Definitions of Terms Specific to This Standard:

3.2.1 *freeze-thaw cycle, n*—a loop from room temperature to the ambient temperature of the freezing cabinet, and back to room temperature.

3.2.2 *freezing, closed system, n*—freezing that occurs under conditions that preclude the gain or loss of any water in the system.

3.2.3 *freezing, open system, n*—freezing that occurs under conditions that allow the gain or loss of water in the system by movement of pore water from or to an external source to growing ice lenses.

4. Significance and Use

4.1 This test method identifies the changes in hydraulic conductivity as a result of freeze-thaw on natural soils only.

4.2 It is the user's responsibility when using this test method to determine the appropriate moisturewater content of the laboratory-compacted specimens (that is, dry, wet, or at optimum moisturewater content) (Note 2).

NOTE 2—It is common practice to construct clay liners and covers at optimum or greater than optimum moisturewater content. Specimens compacted dry of optimum moisturewater content typically do not contain larger pore sizes as a result of freeze-thaw because the effects of freeze-thaw are minimized by the lack of water in the sample. Therefore, the effect of freeze-thaw on the hydraulic conductivity is minimal, or the hydraulic conductivity may increase slightly.³

4.3 The requestor must provide information regarding the effective stresses to be applied during testing, especially for determining the final hydraulic conductivity. Using high effective stresses (that is, 35 kPa (5 psi) [5 psi] as allowed by Test Method D5084) can decrease an already increased hydraulic conductivity resulting in lower final hydraulic conductivity values. The long-term effect of freeze-thaw on the hydraulic conductivity of compacted soils is unknown. The increased hydraulic conductivity caused by freeze-thaw may be temporary. For example, the overburden pressure imparted by the waste placed on a soil liner in a landfill after being subjected to freeze-thaw may reduce the size of the cracks and pores that cause the increase in hydraulic conductivity. It is not known if the pressure would overcome the macroscopically increased hydraulic conductivity sufficiently to return the soil to its original hydraulic conductivity (prior to freeze-thaw). For cases such as landfill covers, where the overburden pressure is low, the increase in hydraulic conductivity due to freeze-thaw will likely be permanent. Thus, the requestor must take the application of the test method into account when establishing the effective stress.

4.4 The specimen shall be frozen to -15°C [5°F] unless the requestor specifically dictates otherwise. It has been documented in the literature that the initial (that is, 0 to -15°C) to -15°C [32°F to 5°F] freezing condition causes the most significant effects³ in hydraulic conductivity. Freezing rate and ultimate temperature should mimic the field conditions. It has been shown that superfreezing (that is, freezing the specimen at very cold temperatures and very short time periods) produces erroneous results.

4.5 The thawed specimen temperature and thaw rate shall mimic field conditions. Thawing specimens in an oven (that is, overheating) will produce erroneous results.

4.6 Literature relating to this subject indicates that the effects of freeze-thaw usually occur by Cycle 10, thus it is recommended that at least 10 freeze-thaw cycles shall be performed to ensure that the full effects of freeze-thaw are measured. If the hydraulic conductivity values are still increasing after 10 freeze-thaw cycles, the test method shall be continued (that is, more freeze-thaw cycles shall be performed).

NOTE 3—The quality of the result produced by this standard is dependent on the competence of the personnel performing it, and the suitability of the equipment and facilities used. Agencies that meet the criteria of Practice D3740 are generally considered capable of competent and objective testing/sampling/inspection/etc. Users of this standard are cautioned that compliance with Practice D3740 does not in itself assure reliable results. Reliable



results depend on many factors; Practice D3740 provides a means of evaluating some of those factors.

5. Apparatus

- 5.1 *Freezing Cabinet*, capable of maintaining at least $-15 \pm 1^{\circ}\text{C}$; $\pm 1^{\circ}\text{C}$ [$5 \pm 0.3^{\circ}\text{F}$].
- 5.2 The apparatus listed in Test Method D5084 (see 5.1 through section 5.18).

6. Reagents

6.1 *Deaired Water*—To aid in removing as much air from the specimen as possible during the hydraulic conductivity portion of the test, deaired water shall be used.

6.2 *Optional*—If the specimen is frozen/thawed in the flexiwall permeameter, a mixture of propylene glycol and tap water can be placed in the flexi-wall permeameter cell. The compatibility of the mixture and membrane used shall be determined. Membranes may degrade and cause cell leakage. This mixture should have a freezing point lower than the ambient temperature of the freezing cabinet. The specimen will freeze, but the cell fluid will not. This allows the total stress on the specimen to remain unchanged during the freeze-thaw procedures.

7. Test Specimens

7.1 *Size*—The size of the specimen depends on the type and size of permeameter being used for the hydraulic conductivity testing. The specimen shall have a minimum diameter of 71 ± 7.1 mm (2.80 ± 0.28 in.) [2.8 in.] or greater and a minimum height of 71 ± 7.1 mm. The height and diameter of the specimen shall be measured to the nearest 0.3 mm (0.01 ± 0.01 in.) [0.01 in.] or better. The length and diameter shall not vary by more than $\pm 5\%$. For specimen diameters of 71 ± 7.1 mm, clods less than 12.7 mm (0.5 ± 0.5 in.) [0.5 in.] shall not be reduced. For specimen diameters of 101.6 mm (4.00 ± 0.40 in.) [4.00 in.], clod sizes less than 17.0 mm (0.67 ± 0.67 in.) [0.67 in.] shall not be reduced. The surface of the test specimen may be uneven, but indentations must not be so deep that the length or diameter of the specimen varies by more than $\pm 5\%$. The diameter and height of the specimen shall each be at least 6 times greater than the largest particle size within the specimen. After completion of the test method, if oversized particles are found, it should be noted in the report.

7.2 Intact or laboratory-compacted specimens can be used with this methodology. Test Method D5084 shall be followed to prepare intact or laboratory-compacted specimens.

7.3 *Number of Test Specimens*—This test method provides Test Method A and Test Method B for performing the test.

7.3.1 *Test Method A*—A specimen shall be prepared for each hydraulic conductivity determination performed. For example, if the hydraulic conductivity is performed initially, after 5 cycles and finally after 10 freeze-thaw cycles, a total of 3 specimens would be required. One specimen would be used for the initial hydraulic conductivity, the second specimen hydraulic conductivity would be determined after subjecting the specimen to 5 freeze-thaw cycles and the third specimen hydraulic conductivity would be determined after subjecting the specimen to 10 freeze-thaw cycles. This test method requires similar specimen preparation methods to ensure that representative samples are used (see Note 4).

7.3.2 *Test Method B*—One specimen can be used for the entire test method. This is not recommended for specimens with initial moisturewater contents significantly lower than 100 % saturation of the soil (Note 4).

NOTE 4—Using more than one specimen offers the advantage of comparison of hydraulic conductivities at an unchanging moisturewater content. When using one specimen for the entire test, the initial hydraulic conductivity test saturates the specimen producing a specimen for the freeze-thaw cycles that has a moisturewater content increased from the original compacted moisturewater content. This test method allows either procedure; however cautions the user about the moisture content water content conditions. The results should not be significantly different if the initial moisturewater content is almost at saturation, which is the case if soils are compacted well above optimum moisturewater content. Using more than one specimen for the test method, specifically undisturbed specimens, has the disadvantage of potential nonrepresentative specimens and test results that are not comparable. In addition, considerations of duplicate specimens are encouraged to manage the inherent variability in the testing.

7.3.3 *Intact Specimens*—Intact test specimens shall be prepared from a representative portion of intact specimens secured in accordance with Practice D1587 or Practice D2113, and preserved and transported in accordance with requirements for Group C materials in Practice D4220 (refer to Test Method D5084 for further discussion).

7.3.4 *Laboratory Compacted Specimens*—Refer to Test Method D5084. Specimens shall be compacted in a mold or piece of PVC pipe or any other suitable apparatus so that the samples are confined during the freeze-thaw cycles. The user of the test method can perform 1-d freeze-thaw by placing insulation around the pipe and using a control sample with a thermocouple (see Note 5) or can perform 3-d freezing by placing the pipe in the freezer without insulation.

7.4 Other preparation test methods can be followed if specified and if identified in the report.

7.5 Determine the water content (Test Method D2216), height, diameter, and mass of each test specimen and calculate the dry unit weight. The initial degree of saturation can be estimated.

8. Procedure

- 8.1 Determine the initial hydraulic conductivity of the specimen following Test Method D5084.