



Designation: E961 – 97 (Reapproved 2013)

Standard Specification for Blood Sedimentation Tube, Wintrobe, Glass, Reusable¹

This standard is issued under the fixed designation E961; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ε) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This specification covers reusable blood sedimentation tubes suitable for determining sedimentation rates and the volume of packed red blood cells.

1.2 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.3 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 *ASTM Standards:*²

E438 Specification for Glasses in Laboratory Apparatus

E920 Specification for Commercially Packaged Laboratory Apparatus

E921 Specification for Export Packaged Laboratory Apparatus

E1133 Practice for Performance Testing of Packaged Laboratory Apparatus for United States Government Procurements

E1157 Specification for Sampling and Testing of Reusable Laboratory Glassware

3. Terminology

3.1 *Definitions of Terms Specific to This Standard:*

3.1.1 *reusable*—Capable of being used again.

3.1.2 *Wintrobe*—The surname of the individual responsible for the design of the Wintrobe tube and the method of use.³

¹ This specification is under the jurisdiction of ASTM Committee E41 on Laboratory Apparatus and is the direct responsibility of Subcommittee E41.01 on Apparatus.

Current edition approved Nov. 1, 2013. Published December 2013. Originally approved in 1983. Last previous edition approved in 2008 as E961 – 97 (2008). DOI: 10.1520/E0961-97R13.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ Wintrobe, Maxwell M., "Laboratory Evaluation of Erythrocytes," *Clinical Hematology*, Seventh Ed., 1974, pp. 109–134.

4. Classification

4.1 This specification covers a tube that is intended to be used until it is no longer considered a functional device for the purpose intended.

5. Materials

5.1 *Glass*—The tubes made to this specification shall be fabricated from borosilicate glass, Type I, Class B, or soda-lime glass, Type II, in accordance with Specification E438.

6. Dimensions and Graduations

6.1 *Dimensions*—The tube shall be made of tubing with an outside diameter (O.D.) of 7.0 to 8.0 mm with an inside diameter (I.D.) of 2.9 to 3.3 mm. The uniformity of the bore shall be ± 0.1 mm throughout the tube. The tube shall be 110 to 117 mm long and have a graduated scale of 105 ± 0.25 mm from the inside bottom of the tube. The tube shall be legibly marked with the manufacturer's or vendor's name or mark and possess a frosted area for marking purposes.

6.2 *Graduation Scale*—The tube shall be graduated 105 ± 0.25 mm in 1-mm divisions and numbered every 1 cm with two sets of numerals. One set of graduation numerals shall be from 0 to 9 cm down the left side of the graduation scale and the other set of g (20 to 25°C) for 15 min. Remove tube from the solution and thoroughly rinse in tap water followed by distilled water. Dry the tube by rubbing vigorously, 5 to 10 strokes, with a laboratory cloth or tissue. This appearance of the markings should be the same as before the test, when judged by eye under normal room lighting.

6.3 *Resistance to Centrifugal Force Test*—Fill the tube with water to the top graduation line and place in a centrifuge. The speed and dimensions of centrifuge head shall be such that the inside bottom of the tube is subjected to a relative centrifugal force of (RCF) not less than 2 500 gravities. Calculate the relative centrifugal force as follows:

$$\begin{aligned} RCF &= \text{relative centrifugal force in gravities,} \\ RCF &= 00001118 \times r \times N^2 \text{ gravities,} \\ r &= \text{rotating radius to inside or outside of tube in centimetre,} \\ &\text{and} \\ N &= \text{rotating speed in revolutions per minute.} \end{aligned}$$

6.4 For additional sampling and testing data, see Specification E1157.



7. Packaging

7.1 For packaging, select from either Specification E920, Specification E921, or Practice E1133.

8. Keywords

8.1 blood; glass; reusable; sedimentation rate; wintrobe

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org). Permission rights to photocopy the standard may also be secured from the ASTM website (www.astm.org/COPYRIGHT/).

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[ASTM E961-97\(2013\)](#)

[/catalog/standards/astm/b871fb9e-5298-4611-bd25-67c5302377cb/astm-e961-97-2013](#)