

FINAL DRAFT International **Standard**

ISO/FDIS 21135 IULTCS IUC 442

IULTCS

Voting begins on: 2024-02-16

2024-04-12

Voting terminates on:

Chemicals for the leather tanning industry — Determination of the total content of certain bisphenols

Produits chimiques pour l'industrie du tannage du cuir — 2000 2001 Détermination de la teneur totale en certains bisphénols

Document Preview

https://standards.iteh.ai/catalog/standards/iso/2fdaf211-358e-41ff-93c3-b6dd02b08183/iso-fdis-21135

ISO/CEN PARALLEL PROCESSING

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNO-LOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

ISO/FDIS 21135:2024(en)

iTeh Standards (https://standards.iteh.ai) Document Preview

ISO/FDIS 21135

https://standards.iteh.ai/catalog/standards/iso/2fdaf211-358e-41f8-93c3-b6dd02b08183/iso-fdis-21135



COPYRIGHT PROTECTED DOCUMENT

© ISO 2024

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office CP 401 • Ch. de Blandonnet 8 CH-1214 Vernier, Geneva Phone: +41 22 749 01 11 Email: copyright@iso.org

Website: <u>www.iso.org</u> Published in Switzerland

Con	tents	Page			
Forev	vord	iv			
Intro	duction	v			
1	Scope	1			
2	Normative references	1			
3	Terms and definitions	1			
4	Principle				
5	Apparatus				
6	Reagents				
7	Sampling and sample preparation				
8	Procedure 8.1 Extraction 8.2 Instrumental analysis	4			
9	Expression of results 9.1 Calculation without internal standard 9.2 Calculation with internal standard 9.3 Calculation of the results as a sum	4 5			
10	Precision	5			
11	Test report Tab Standards	5			
Anne	A (informative) Chromatographic analysis operating parameters for LC-MS/MS	7			
	B (informative) Chromatographic analysis operating parameters for LC-MS				
	C (informative) Chromatographic analysis operating parameters for LC-UV, LC-LC-FLD	DAD or			
Anne	x D (informative) Precision: reliability of the method	11			
	graphy				
	incilistandards iteh ailoatalogistandardsiisoi /tdaf/TLL35Xe_4TtX_Y3c3_h6dd(T/h(IXTX3/iso.				

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

IULTCS, originally formed in 1897, is a world-wide organization of professional leather societies to further the advancement of leather science and technology. IULTCS has three Commissions, which are responsible for establishing international methods for the sampling and testing of leather. ISO recognizes IULTCS as an international standardizing body for the preparation of test methods for leather.

This document was prepared by the Chemical Test Commission of the International Union of Leather Technologists and Chemists Societies (IUC Commission, IULTCS), in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 289, Leather, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

This document includes a procedure for analysing certain bisphenols in leather tanning chemicals using liquid chromatography (LC) equipment. With this analytical method, bisphenol A, bisphenol AF, bisphenol B, bisphenol F and bisphenol S can be determined.

In the leather industry, bisphenol F can be an impurity in synthetic tanning agents. Bisphenol S is a monomer that is used to manufacture synthetic tanning agents, which can lead to residues in the final product.

Bisphenol A is a synthetic organic chemical primarily used as a monomer in the manufacture of high-performance plastics, other polymers, such as resins, and in the colour developer for thermoprint paper. Bisphenol AF is a fluorinated organic compound that is an analogue of bisphenol A in which the two methyl groups are replaced with trifluoromethyl groups. Bisphenol B is similar to bisphenol A and is used in the manufacture of plastics and resins.

At present, the official European Chemicals Agency (ECHA) classification recognized in the European Union (EU) is the following:

- bisphenol A as toxic to reproduction, skin sensitizer and endocrine disruptor;
- bisphenol B as endocrine disruptor;
- bisphenol S as toxic to reproduction and endocrine disruptor. [3]

iTeh Standards (https://standards.iteh.ai) Document Preview

ISO/FDIS 21135

https://standards.iteh.ai/catalog/standards/iso/2fdaf211-358e-41f8-93c3-h6dd02h08183/iso-fdis-21135

iTeh Standards (https://standards.iteh.ai) Document Preview

ISO/FDIS 21135

https://standards.iteh.ai/catalog/standards/iso/2fdaf211-358e-41f8-93c3-b6dd02b08183/iso-fdis-21135

Chemicals for the leather tanning industry — Determination of the total content of certain bisphenols

Scope

This document specifies a method for determining the total content (solvent extractable) of the following bisphenols in chemicals for the leather tanning industry:

- bisphenol A;
- bisphenol AF;
- bisphenol B;
- bisphenol F;
- bisphenol S.

This method requires the use of liquid chromatography (LC) with either a single quadrupole mass spectrometer (MS), a triple quadrupole mass spectrometer (MS/MS), an ultraviolet (UV) detector, a diode array detector (DAD) or a fluorescence detector (FLD) to identify and quantify the bisphenols.

- NOTE 1 This method can also be used for other bisphenols if they are validated by the laboratory.
- Bisphenol S cannot be detected with FLD. NOTE 2

Normative references

There are no normative references in this documentals 21135

Terms and definitions 3

No terms and definitions are listed in this document.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at https://www.iso.org/obp
- IEC Electropedia: available at https://www.electropedia.org/

Principle

The sample of the chemical is extracted in methanol using an ultrasonic bath. Subsequently, an aliquot of the solution can be directly analysed, using LC-MS, LC-MS/MS or LC with a UV detector (LC-UV), DAD (LC-DAD) or FLD (LC-FLD).

Apparatus

The usual laboratory apparatus and, in particular, the following shall be used:

5.1 **Ultrasonic bath**, with controllable heating capable of maintaining a temperature of (60 ± 5) °C.

- **5.2 Glass container with a screw cap**, for example volume of 20 ml.
- **5.3 Suitable syringe membrane filters**, for example PTFE with pore size 0,2 μ m or PTFE-with pore size 0,45 μ m.
- **5.4 Volumetric flasks**, for example volume of 10 ml and 100 ml.
- **5.5 LC vials, with cap**, for example volume of 2 ml.
- **5.6 Analytical balance**, with a resolution of 0.1 mg.
- **5.7 Pipettes**, various sizes, for example volume of 1 ml to 20 ml.
- **5.8** Instrumental equipment, LC-MS/MS.
- **5.9 Alternative instrumental equipment**, LC-MS, LC-UV, LC-DAD or LC-FLD.

NOTE If two detectors are used, they can be arranged in series on the same LC system.

6 Reagents

If not otherwise specified, analytical reagent grade chemicals shall be used.

- **6.1 Methanol,** CAS Registry Number[®] (CAS RN)¹⁾ 67-56-1. For LC-MS/MS it is necessary to have LC-MS quality. HPLC quality methanol is suitable for LC-UV, LC-DAD or LC-FLD.
- **6.2 Water,** deionised or distilled, ultra-pure quality for LC-MS and LC-MS/MS, HPLC grade for LC-DAD, UV and FLD.
- **6.3 Bisphenol A,** CAS RN 80-05-7, minimum 98,0 %.
- **6.4** Bisphenol AF, CAS RN 1478-61-1, minimum 98,0 %.58e-41f8-93c3-b6dd02b08183/iso-fdis-21135
- **6.5 Bisphenol B,** CAS RN 77-40-7, minimum 98,0 %.
- **6.6 Bisphenol F,** CAS RN 620-92-8, minimum 98,0 %.
- **6.7 Bisphenol S,** CAS RN 80-09-1, minimum 98,0 %.
- **6.8** Stock solutions of a mix of bisphenol A, AF, B, F and S, $\rho = 1 \text{ mg/l}$, 10 mg/l and 50 mg/l.

EXAMPLE 100 mg of each of the respective bisphenols, A $(\underline{6.3})$, AF $(\underline{6.4})$ B $(\underline{6.5})$, F $(\underline{6.6})$ and S $(\underline{6.7})$, is dissolved in separate 100 ml volumetric flasks $(\underline{5.4})$ with methanol $(\underline{6.1})$. Mixed stock solutions are prepared to obtain, respectively, concentrations of 1 mg/l, 10 mg/l and 50 mg/l in methanol.

6.9 Internal standard, $\rho = 50 \text{ mg/l}$.

When using LC-MS or LC-MS/MS, the use of internal standards for each type of bisphenol is highly recommended to avoid matrix effects.

Examples of suitable mass-labelled internal standards:

¹⁾ Chemical Abstracts Service (CAS) Registry Number $^{\circledR}$ is a trademark of the American Chemical Society (ACS). This information is given for the convenience of users of this document and does not constitute an endorsement by ISO of the product named. Equivalent products may be used if they can be shown to lead to the same results.

for bisphenol A: bisphenol A-D8, CAS RN 92739-58-7;

bisphenol A-D16, CAS RN 96210-87-6;

for bisphenol B: bisphenol B-D8, CAS RN to be assigned;

for bisphenol AF: bisphenol AF-13C12, CAS RN to be assigned;

for bisphenol F: bisphenol F-D10, CAS RN 1794786-93-8;

for bisphenol S: bisphenol S-D8, CAS RN 2483831-28-1.

Prepare a 50 mg/l solution of the internal standard by diluting the commercial solution with methanol.

6.10 Calibration solutions of bisphenols.

For LC-MS/MS or LC/MS techniques, prepare at least four calibration solutions of ρ = 0,05 µg/ml to ρ = 1 µg/ml of bisphenols using the stock solutions (6.8), see <u>Table 1</u>. For LC-DAD or UV or FLD techniques, prepare at least four calibration solutions of ρ = 0,5 µg/ml to ρ = 20 µg/ml using the stock solutions (6.8), see <u>Table 2</u>.

Table 1 — Example of calibration solutions for LC-MS/MS or LC/MS

Concentration	Volume methanol (6.1)	Volume of mix of bisphenols 1 mg/l	Volume of mix of bisphenols 10 mg/l	Volume of internal standard at 50 mg/l
	(0.1)	(6.8)	(6.8)	(6.9) (only for MS detection)
μg/ml	μl	h Standar	°ds µl	μl
0,05	930	50	0	20
0,1	880	Stan 100 ards	.itenoai)	20
0,2	960	0, 1	20	20
0,5	930	interot Fre	VIEW ₅₀	20
1	880	0	100	20

Table 2 — Example of calibration solutions for LC-DAD or UV or FLD and the solutions for LC-DAD or UV or FLD and t

Concentration	Volume methanol (6.1)	Volume of mix of bisphenols 10 mg/l (6.8)	Volume of mix of bisphenols 50 mg/l (6.8)
μg/ml	μl	μl	μl
0,5	950	50	0
2	800	200	0
10	0	1 000	0
20	600	0	400

7 Sampling and sample preparation

The sample should be thoroughly mixed to get a representative test portion for analysis.

In the case of a powdered sample, if the particles are heterogeneous, manually or mechanically mill to homogenize the size of the particles.

8 Procedure

8.1 Extraction

Accurately weigh (0.2 ± 0.050) g of the sample measured to the nearest 0.001 g with an analytical balance $(\underline{5.6})$ in a screw-top glass container $(\underline{5.2})$ and add 20 ml methanol $(\underline{6.1})$. Close the container and place it for (60 ± 5) min in a pre-heated ultrasonic bath $(\underline{5.1})$ at (60 ± 5) °C.

After cooling down to room temperature, if the extraction solution appears opalescent or contains particles, centrifugate it for (5 ± 1) min at $(4\ 000 \pm 400)$ rpm, then filter an aliquot of extraction solution (5.3) into a LC sample vial (5.5). The aliquot is now ready for the LC-UV, LC-DAD or LC-FLD analysis.

For LC-MS or LC-MS/MS, take an aliquot of 200 μ l of filtered extraction solution (5.3) into a LC sample vial (5.5). Add 780 μ l of methanol (6.1) and 20 μ l of internal standard (6.9). The aliquot is now ready for the LC-MS or LC-MS/MS analysis.

NOTE Due to the composition of the matrix, it is possible that at the end of the extraction procedure the solution will be opalescent, due to the low solubility in methanol of the other components of the chemical product analysed. This does not affect the extraction efficiency of bisphenols as they are very soluble in the extraction solution.

8.2 Instrumental analysis

The detection of the bisphenols is made using LC-MS/MS (5.8) or, alternatively, LC-MS, LC-UV, LC-DAD or LC-FLD (5.9). Examples of suitable chromatographic conditions are given in <u>Annex A</u> (for LC-MS/MS), <u>Annex B</u> (for LC-UV, LC-DAD and LC-FLD) and <u>Annex C</u> (for LC-MS).

If the concentration of bisphenols is out of the range of the calibration, make a suitable dilution and inject the new aliquot.

9 Expression of results

9.1 Calculation without internal standard

The content of each bisphenol is calculated as the mass fraction, w, in milligrams per kilogram (mg/kg) of the chemical sample according to Formula (1):

$$w = \frac{(A_s - b) \cdot V}{a \cdot m} \tag{1}$$

where

- $A_{\rm s}$ is the peak area of each bisphenol in the extraction solution;
- *b* is the intercept of the calibration graph;
- *a* is the slope of the calibration graph;
- *V* is the final volume used (20 ml);
- *m* is the mass of the chemical sample in grams (g).

Any detail shall be noted in the test report.