



Designation: ~~E473 – 11a~~ **E473 – 14**

Standard Terminology Relating to Thermal Analysis and Rheology¹

This standard is issued under the fixed designation E473; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reappraisal. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reappraisal.

1. Scope

1.1 This terminology is a compilation of definitions of terms used in ASTM documents relating to thermal analysis and rheology. This terminology includes only those terms for which ASTM either has standards or is contemplating some action. It is not intended to be an all-inclusive listing of terms related to thermal analysis and rheology.

1.2 This terminology specifically supports the single-word form for terms using thermo as a prefix, such as thermoanalytical or thermomagnetometry, while recognizing that for some terms a two-word form can be used, such as thermal analysis. This terminology does not support, nor does it recommend, use of the grammatically incorrect, single-word form using thermal as a prefix, such as, thermalanalytical or thermalmagnetometry.

1.3 A definition is a single sentence with additional information included in a *Discussion* area. It is reviewed every five years.

2. Terminology

adiabatic, *adj*—no heat exchange with the surroundings.

calorimeter, *n*—apparatus for measuring quantities of absorbed or evolved heat.

combined, *adj*—the application of two or more techniques to different samples at the same time.

controlled-rate thermal analysis (CRTA), *n*—a family of techniques that monitors the temperature versus time profile needed to maintain a chosen, fixed rate of change of a property of a substance.

¹ This terminology is under the jurisdiction of ASTM Committee E37 on Thermal Measurements and are the direct responsibility of Subcommittee E37.03 on Nomenclature and Definitions.

Current edition approved June 15, 2011; Aug. 15, 2014. Published June 2011; September 2014. Originally approved in 1973. Last previous edition approved in 2011 as E473 – 11; E473 – 11a. DOI: 10.1520/E0473-11A; 10.1520/E0473-14.

<https://standards.iteh.ai/catalog/standards/sist/f963a15a-f2e4-4d5e-8b3f-14cbf12acc56/astm-e473-14>

DISCUSSION—

Compared to controlled-temperature experiments, where the reaction rate tends to increase exponentially and the rate can become limited by heat or mass transfer, CRTA experiments are more likely to involve the chemical reaction as the limiting step. This technique can also improve the resolution of multiple reactions. For example, in controlled rate experiments, power to the furnace is controlled to ensure a fixed rate of mass loss (or gain).

controlled-temperature program, *n*—the temperature history experienced by a sample during the course of a thermal analysis experiment.

DISCUSSION—

In contrast to controlled-rate experiments, power to the furnace is controlled to ensure a fixed rate of temperature change for controlled-temperature experiments. The program may include heating or cooling segments in which the temperature is changed at a fixed rate, isothermal segments in which time becomes the explicit independent variable, or any sequence of these individual segments. If the atmosphere (or vacuum) around the sample is changed by some external action (depending on the independent variable only—temperature or time) during the course of the experiment, that too becomes part of the controlled-temperature program.

curve, thermal, *n*—the plot of a dependent parameter against an independent parameter such as temperature or time.

dielectric analysis (DEA), *n*—a technique in which the dielectric constant (permittivity or capacitance) and dielectric loss (conductance) of a substance under oscillating electric field are measured as a function of temperature or time while the substance is subjected to a controlled-temperature program in a specified atmosphere.

derivative, *adj*—pertaining to the first derivative (mathematical) of any curve with respect to temperature or time.

differential, *adj*—pertaining to a difference in measured or measurable quantities usually between a substance and some reference or standard material.

differential scanning calorimetry (DSC), *n*—a technique in which the heat flow difference into a substance and a reference material is measured as a function of temperature while the substance and reference material are subjected to a controlled-temperature program.

DISCUSSION—

The record is the differential scanning calorimetric or DSC curve. Two modes, power compensation differential scanning calorimetry, and heat flux differential scanning calorimetry can be distinguished, depending on the method of measurement used.

DISCUSSION—

Two conventions exist in thermal analysis. In the physicist's convention, exothermic behavior increases downward on the thermal curve. In the chemist's convention, exothermic behavior increases upward on the thermal curve. Committee E37 takes no position on which convention shall be used. To aid the user, the direction of exothermic (or conversely, endothermic) behavior shall be indicated on each thermal curve.

differential thermal analysis (DTA), *n*—a technique in which the temperature difference between the substance and a reference material is measured as a function of temperature, while the substance and reference material are subjected to a controlled-temperature program.

DISCUSSION—

The term *quantitative differential thermal analysis* covers those uses of DTA where the equipment is designed to produce quantitative results.

dilatometry, *n*—see **thermodilatometry**.

dynamic mechanical analysis (DMA), *n*—a technique in which the storage modulus (elastic response) and loss modulus (viscous response) of a substance under oscillatory load is measured as a function of temperature, time, or frequency of oscillation, while the substance is subjected to a controlled-temperature program in a specified atmosphere.

endotherm, *n*—In thermal analysis, the thermal record of a transition where heat is absorbed by the specimen.

evolved gas analysis (EGA), *n*—a technique in which the nature or amount, or both, of gas or vapor evolved by a substance is subjected to a controlled-temperature program.

DISCUSSION—

Some specific forms of EGA have become established for investigating different aspects of catalysis, such as reduction, oxidation, or desorption. In this context, EGA in a hydrogen atmosphere is known as temperature-programmed reduction (TPR); EGA in an oxygen atmosphere is temperature-programmed oxidation (TPO); and EGA in the absence of decomposition, in an inert atmosphere or vacuum, is temperature-programmed desorption (TPD). For each technique the method used for gas identification and quantification should always be clearly stated.

evolved gas detection (EGD), *n*—see **evolved gas analysis**.

extrapolated onset value, *n*—the value of the independent parameter found by extrapolating the dependent parameter baseline prior to the event and a tangent constructed at the inflection point on the leading edge to their intersection.

first-deviation-from baseline, *n*—the value of the independent parameter at which a deflection is first observed from the established dependent parameter baseline prior to the event.

high-pressure (HP...), *adj*—a prefix for different thermoanalytical techniques in which the pressure in the apparatus is above ambient.

DISCUSSION—

As an example, high-pressure thermogravimetric analysis is designated HPTGA.

isoperibol, *adj*—to maintain constant surroundings.

DISCUSSION—