



Designation: **D4951 – 09 D4951 – 14**

Standard Test Method for Determination of Additive Elements in Lubricating Oils by Inductively Coupled Plasma Atomic Emission Spectrometry¹

This standard is issued under the fixed designation D4951; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

This standard has been approved for use by agencies of the U.S. Department of Defense.

1. Scope*

1.1 This test method covers the quantitative determination of barium, boron, calcium, copper, magnesium, molybdenum, phosphorus, sulfur, and zinc in unused lubricating oils and additive packages.

1.2 The precision statements are valid for dilutions in which the mass % sample in solvent is held constant in the range of 1 to 5 mass % oil.

1.3 The precision tables define the concentration ranges covered in the interlaboratory study. However, both lower and higher concentrations can be determined by this test method. The low concentration limits are dependent on the sensitivity of the ICP instrument and the dilution factor. The high concentration limits are determined by the product of the maximum concentration defined by the linear calibration curve and the sample dilution factor.

1.4 Sulfur can be determined if the instrument can operate at a wavelength of 180 nm.

1.5 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.6 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 *ASTM Standards:*²

[D1552 Test Method for Sulfur in Petroleum Products \(High-Temperature Method\)](#)

[D4057 Practice for Manual Sampling of Petroleum and Petroleum Products](#)

[D4307 Practice for Preparation of Liquid Blends for Use as Analytical Standards](#) <https://standards.kenia.gov/standards/astm-d4951-14>

[D4628 Test Method for Analysis of Barium, Calcium, Magnesium, and Zinc in Unused Lubricating Oils by Atomic Absorption Spectrometry](#)

[D4927 Test Methods for Elemental Analysis of Lubricant and Additive Components—Barium, Calcium, Phosphorus, Sulfur, and Zinc by Wavelength-Dispersive X-Ray Fluorescence Spectroscopy](#)

[D5185 Test Method for Multielement Determination of Used and Unused Lubricating Oils and Base Oils by Inductively Coupled Plasma Atomic Emission Spectrometry \(ICP-AES\)](#)

[D6299 Practice for Applying Statistical Quality Assurance and Control Charting Techniques to Evaluate Analytical Measurement System Performance](#)

3. Summary of Test Method

3.1 A sample portion is weighed and diluted by mass with mixed xylenes or other solvent. An internal standard, which is required, is either weighed separately into the test solution or is previously combined with the dilution solvent. Calibration standards are prepared similarly. The solutions are introduced to the ICP instrument by free aspiration or an optional peristaltic

¹ This test method is under the jurisdiction of ASTM Committee D02 on Petroleum Products, Liquid Fuels, and Lubricants and is the direct responsibility of Subcommittee D02.03 on Elemental Analysis.

Current edition approved April 15, 2009; Oct. 1, 2014. Published May 2009; October 2014. Originally approved in 1989. Last previous edition approved in 2006; 2009 as D4951D4951 – 09, –06. DOI: 10.1520/D4951-09; 10.1520/D4951-14.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

*A Summary of Changes section appears at the end of this standard

TABLE 1 Lubricants and Additive Materials

Element	Compound	Performance
<u>Barium</u>	<u>Sulfonates, phenates</u>	<u>Detergent inhibitors, corrosion inhibitors, detergents, rust inhibitors, and ATF</u>
<u>Boron</u>	<u>Inorganic borates and borate esters, borated dispersants and detergents</u>	<u>Anti-wear agents, anti-oxidant, deodorant, cutting oils, and brake fluids</u>
<u>Calcium</u>	<u>Sulfonates, phenates</u>	<u>Detergent inhibitors and dispersants</u>
<u>Copper</u>	<u>Copper dithiophosphates</u>	<u>Anti-wear agent and anti-oxidant</u>
<u>Magnesium</u>	<u>Sulfonates, phenates</u>	<u>Detergent inhibitors</u>
<u>Molybdenum</u>	<u>Dialkylthiophosphate, dialkyldithiocarbamate, other molybdenum complexes</u>	<u>Friction modifier additives</u>
<u>Phosphorus</u>	<u>Dithiophosphates, phosphates, and phosphites</u>	<u>Anti-rusting agents, extreme pressure additives, and anti-wear</u>
<u>Sulfur</u>	<u>Base oils, sulfonates, thiophosphates, polysulfides, and other sulfurized components</u>	<u>Detergents, extreme pressure additives, and antiwear</u>
<u>Zinc</u>	<u>Dialkyldithiophosphates, dithiocarbamates, phenolates, carboxylates</u>	<u>Anti-oxidant, corrosion inhibitors, anti-wear additives, detergents, crankcase oils, hypoid gear lubricants, aircraft piston engine oils, turbine oils, ATF, railroad diesel engine oils, and brake lubricants</u>

pump. By comparing emission intensities of elements in the test specimen with emission intensities measured with the calibration standards and by applying the appropriate internal standard correction, the concentrations of elements in the sample are calculable.

4. Significance and Use

4.1 This test method usually requires several minutes per sample. This test method covers eight elements and thus provides more elemental composition data than Test Method **D4628** or Test Methods **D4927**. In addition, this test method provides more accurate results than Test Method **D5185**, which is intended for used lubricating oils and base oils.

4.2 Additive packages are blends of individual additives, which can act as detergents, antioxidants, antiwear agents, and so forth. Many additives contain one or more elements covered by this test method. Additive package specifications are based, in part, on elemental composition. Lubricating oils are typically blends of additive packages, and their specifications are also determined, in part, by elemental composition. This test method can be used to determine if additive packages and unused lubricating oils meet specifications with respect to elemental composition. [ASTM D4951-14](#)

4.3 Several additive elements and their compounds are added to the lubricating oils to give beneficial performance (**Table 1**).

5. Interferences

5.1 *Spectral*—There are no known spectral interferences between elements covered by this test method when using the spectral lines listed in **Table 12**. However, if spectral interferences exist because of other interfering elements or selection of other spectral lines, correct for the interference using the technique described in Test Method **D5185**.

5.2 *Viscosity Index Improver Effect*—*Effect*—Viscosity index improvers, which can be present in multigrade lubricating oils, can bias measurements. However, the biases can be reduced to negligible proportion by using the specified solvent-to-sample dilution and an internal standard.

6. Apparatus

6.1 *Inductively-Coupled Plasma Atomic Emission Spectrometer*—Either a sequential or simultaneous spectrometer is suitable, if equipped with a quartz ICP torch and RF generator to form and sustain the plasma.

6.2 *Analytical Balance*, capable of weighing to 0.001 g or 0.0001 g, with sufficient capacity to weigh prepared solutions.

6.3 *Peristaltic Pump, (Recommended)*—*(Recommended)*—A peristaltic pump is strongly recommended to provide a constant flow of solution. The pumping speed must be in the range 0.5 to 3 mL/min. The pump tubing must be able to withstand at least 6 h exposure to the dilution solvent. Fluoroelastomer copolymer³ tubing is recommended.

6.4 *Solvent Dispenser, (Optional)*—*(Optional)*—A solvent dispenser calibrated to deliver the required weight of diluent can be advantageous. Ensure that solvent drip does not affect accuracy.

6.5 *Specimen Solution Containers, Containers*, of appropriate size, glass or polyolefin vials or bottles, with screw caps.

³ Fluoroelastomer copolymer is manufactured as Viton, a trademark owned by E. I. duPont de Nemours.

TABLE 12 Elements Determined and Suggested Wavelengths^A

Element	Wavelength, nm
Barium	233.53, 455.40, 493.41
Boron ^B	182.59, 249.68
Calcium	315.88, 317.93, 364.4, 422.67
Copper	324.75
Magnesium	279.08, 279.55, 285.21
Molybdenum	202.03, 281.62
Phosphorus ^B	177.51, 178.29, 213.62, 214.91, 253.40
Sulfur ^B	180.73, 182.04, 182.62
Zinc	202.55, 206.20, 213.86, 334.58, 481.05

^A These wavelengths are only suggested and do not represent all possible choices.

^B Wavelengths for boron, phosphorus, and sulfur below 190 nm require that a vacuum or inert gas purged optical path be used.

6.6 *Vortexer, (Optional)*—Vortex the sample plus diluent mixture until the sample is completely dissolved.

6.7 *Ultrasonic Homogenizer, Optional*—A bath-type or probe-type ultrasonic homogenizer can be used to homogenize the test specimen.

7. Reagents and Materials

7.1 *Purity of Reagents*—Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society where such specifications are available.⁴

7.2 *Base Oil*, U.S.P. white oil, or a lubricating base oil that is free of analytes, having a viscosity at room temperature as close as possible to that of the samples to be analyzed. (**Warning**—Lubricating base oils can contain sulfur. For preparation of sulfur standards and blending of additive packages, white oil is recommended.)

7.3 *Internal Standard, (Required)*—(*Required*)—An oil-soluble internal standard element is required. The following internal standards were successfully used in the interlaboratory study on precision: Ag, Be, Cd, Co (most common), La, Mn, Pb, Y.

7.4 *Organometallic Standards*—Multi-element standards, containing known concentrations (approximately 0.1 mass %) of each element, can be prepared from the individual metal concentrates. Refer to Practice D4307 for a procedure for preparation of multicomponent liquid blends. When preparing multi-element standards, be certain that proper mixing is achieved. Commercially available multi-element blends (with known concentrations of each element at approximately 0.1 mass %) are also satisfactory.

7.4.1 More than one multi-element standard can be necessary to cover all elements, and the user of this test method can select the combination of elements and their concentrations in the multi-element standards. It can be advantageous to select concentrations that are typical of unused oils. However, it is imperative that concentrations are selected such that the emission intensities measured with the working standards can be measured precisely (that is, the emission intensities are significantly greater than background) and that these standards represent the linear region of the calibration curve. Frequently, the instrument manufacturer publishes guidelines for determining linear range.

7.4.2 Some commercially available organometallic standards are prepared from metal sulfonates and therefore contain sulfur. For sulfur determinations, a separate sulfur standard can be required. A sulfur standard can be prepared by blending NIST SRM 1622 with white oil.

7.4.3 Metal sulfonates can be used as a sulfur standard if the sulfur content is known or determined by an appropriate test method such as Test Method D1552.

7.4.4 Petroleum additives can also be used as organometallic standards if their use does not adversely affect precision nor introduce significant bias.

7.5 *Dilution Solvent*—Mixed xylenes, o-xylene, and kerosine were successfully used in the interlaboratory study on precision.

8. Internal Standardization (Required)

8.1 The internal standard procedure requires that every test solution (sample and standard) have the same concentration (or a known concentration) of an internal standard element that is not present in the original sample. The internal standard is usually combined with the dilution solvent. Internal standard compensation is typically handled in one of two different ways, which can be summarized as follows.

⁴ *Reagent Chemicals, American Chemical Society Specifications—Specifications*, American Chemical Society, Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.

8.1.1 Calibration curves are based on the measured intensity of each analyte divided (that is, scaled) by the measured intensity of the internal standard per unit internal standard element concentration. Concentrations for each analyte in the test specimen solution are read directly from these calibration curves.

8.1.2 For each analyte and the internal standard element, calibration curves are based on measured (unscaled) intensities. Uncorrected concentrations for each analyte in the test specimen solution are read from these calibration curves. Corrected analyte concentrations are calculated by multiplying the uncorrected concentrations by a factor equal to the actual internal standard concentration divided by the uncorrected internal standard concentration determined by analysis.

8.2 Dissolve the organometallic compound representing the internal standard in dilution solvent and transfer to a dispensing vessel. The stability of this solution must be monitored and prepared fresh (typically weekly) when the concentration of the internal standard element changes significantly. The concentration of internal standard element shall be at least 100 times its detection limit. A concentration in the range of 10 to 200 mg/kg is typical.

NOTE 1—This test method specifies that the internal standard is combined with the dilution solvent because this technique is common and efficient when preparing many samples. However, the internal standard can be added separately from the dilution solvent as long as the internal standard concentration is constant or accurately known.

9. Sampling

9.1 The objective of sampling is to obtain a test specimen that is representative of the entire quantity. Thus, take lab samples in accordance with the instructions in Practice D4057. The specific sampling technique can affect the accuracy of this test method.

10. Preparation of Apparatus

10.1 *Instrument*—Design differences between instruments, ICP excitation sources, and different selected analytical wavelengths for individual spectrometers make it impractical to detail the operating conditions. Consult the manufacturer's instructions for operating the instrument with organic solvents. Set up the instrument for use with the particular dilution solvent chosen.

10.2 *Peristaltic Pump*—If a peristaltic pump is used, inspect the pump tubing and replace it, if necessary, before starting each day. Verify the solution uptake rate and adjust it to the desired rate.

10.3 *ICP Excitation Source*—Initiate the plasma source at least 30 min before performing an analysis. During this warm up period, nebulize dilution solvent. Inspect the torch for carbon buildup during the warm up period. If carbon buildup occurs, replace the torch immediately and consult the manufacturer's operating guide to take proper steps to remedy the situation.

NOTE 2—Carbon that accumulates on the tip of the torch injector tube can be removed by using nebulizer gas that consists of approximately 1 % oxygen in argon.

10.3.1 Generally, carbon buildup can be minimized by increasing the intermediate argon flow rate or lowering the torch, or both, relative to the load coil.

NOTE 3—Some manufacturers recommend even longer warm up periods to minimize changes in the slopes of the calibration curves.

10.4 *Wavelength Profiling*—Perform any wavelength profiling that is specified in the normal operation of the instrument.

10.5 *Operating Parameters*—Assign the appropriate operating parameters to the instrument task file so that the desired elements can be determined. Parameters to be included are element, wavelength, background correction points (optional), interelement correction factors (refer to 5.1), integration time, and internal standard compensation (required). Multiple integrations (typically three) are required for each measurement. A typical integration time is 10 s.

11. Preparation of Test Specimens

11.1 *Diluent*—Diluent refers to the dilution solvent containing the internal standard (refer to 8.2).

11.2 Test specimen solutions are prepared in the same way that calibration standards are prepared (refer to 12.2). The mass % oil in diluent must be the same for calibration standards and test specimen solutions.

11.2.1 *Lubricating Oil Specimens*—Weigh appropriate amount of the test specimen to the nearest 0.001 g. The weight of the test specimen taken will vary depending upon the metal concentration of the specimen. Dilute by mass with the diluent. Mix well.

11.2.2 *Additive Packages*—The concentrations of additive elements in additive packages are typically ten times the concentrations in lubricating oils. Therefore, additive packages are first blended with base oil before adding diluent.

11.2.2.1 Weigh appropriate amount of the additive package to the nearest 0.001 g. The weight of the test specimen taken will vary depending upon the metal concentration in the specimen. Add approximately ten times this amount of base oil, weighed to the nearest 0.001 g. Dilute this mixture by mass with diluent. Mix well.

11.3 Record all weights and calculate dilution factors by dividing the sum of the weights of the diluent, sample, and base oil (if any) by the weight of the sample.

12. Preparation of Calibration Standards and Check Standards

12.1 *Diluent*—Diluent refers to the dilution solvent containing the internal standard (refer to 8.2).