

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Expression of performance of electrochemical analyzers –
Part 3: Electrolytic conductivity
(standards.iteh.ai)

Expression des qualités de fonctionnement des analyseurs électrochimiques –
Partie 3: Conductivité électrolytique
<https://standards.iteh.ai/catalog/standards/sist/7a5e6c87-ca15-4e9f-948b-acf306e59e68/iec-60746-3-2002>



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2002 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester.

If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de la CEI ou du Comité national de la CEI du pays du demandeur.

Si vous avez des questions sur le copyright de la CEI ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de la CEI de votre pays de résidence.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
Fax: +41 22 919 03 00
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

Useful links:

IEC publications search - www.iec.ch/searchpub

The advanced search enables you to find IEC publications by a variety of criteria (reference number, text, technical committee,...).

It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available on-line and also once a month by email.

Electropedia - www.electropedia.org

The world's leading online dictionary of electronic and electrical terms containing more than 30 000 terms and definitions in English and French, with equivalent terms in additional languages. Also known as the International Electrotechnical Vocabulary (IEV) on-line.

Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: csc@iec.ch.

A propos de la CEI

La Commission Electrotechnique Internationale (CEI) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications CEI

Le contenu technique des publications de la CEI est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Liens utiles:

Recherche de publications CEI - www.iec.ch/searchpub

La recherche avancée vous permet de trouver des publications CEI en utilisant différents critères (numéro de référence, texte, comité d'études,...).

Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

Just Published CEI - webstore.iec.ch/justpublished

Restez informé sur les nouvelles publications de la CEI. Just Published détaille les nouvelles publications parues. Disponible en ligne et aussi une fois par mois par email.

Electropedia - www.electropedia.org

Le premier dictionnaire en ligne au monde de termes électroniques et électriques. Il contient plus de 30 000 termes et définitions en anglais et en français, ainsi que les termes équivalents dans les langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (VEI) en ligne.

Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: csc@iec.ch.

INTERNATIONAL STANDARD

NORME INTERNATIONALE

**Expression of performance of electrochemical analyzers –
Part 3: Electrolytic conductivity**
(standards.iteh.ai)

**Expression des qualités de fonctionnement des analyseurs électrochimiques –
Partie 3: Conductivité électrolytique**
(standards.iteh.ai)

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

PRICE CODE
CODE PRIX



ICS 17.020; 71.040

ISBN 978-2-83220-536-5

**Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

CONTENTS

FOREWORD.....	3
1 Scope.....	5
2 Normative references	5
3 Definitions	5
4 Procedure for specification	8
4.1 Additional statements on sensor units	8
4.2 Additional statements on electronic units.....	8
4.3 Additional statements on complete analyzers	8
5 Recommended standard values and ranges of influence quantities affecting the performance of electronic units.....	8
6 Verification of values	8
6.1 General aspects	9
6.2 Calibration.....	9
6.3 Test solutions.....	9
6.4 Test procedures	10
Annex A (informative) Electrolytic conductivity values of potassium chloride calibration solutions and pure water	12
Annex B (informative) Electrolytic conductivity values of aqueous sodium chloride solutions	13
Annex C (normative) Alternative procedures for measuring response times: delay (T_{10}), rise (fall) (T_r , T_f) and 90 % (T_{90}) times	15
Annex D (informative) Conductivity cells	16
Bibliography.....	17
Table A.1 – Electrolytic conductivity values.....	12
Table A.2 – Electrolytic conductivity of pure water	12
Table B.1 – Conductivity of sodium chloride solutions at 18 °C	13
Table B.2 – Temperature coefficients for low-concentration sodium chloride solutions	14
Table B.3 – Tentative corrections to sodium chloride solution temperature coefficients.....	14

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**EXPRESSION OF PERFORMANCE
OF ELECTROCHEMICAL ANALYZERS –**
Part 3: Electrolytic conductivity

FOREWORD

- 1) The IEC (International Electrotechnical Commission) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of the IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, the IEC publishes International Standards. Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. The IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of the IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested National Committees.
- 3) The documents produced have the form of recommendations for international use and are published in the form of standards, technical specifications, technical reports or guides and they are accepted by the National Committees in that sense.
- 4) In order to promote international unification, IEC National Committees undertake to apply IEC International Standards transparently to the maximum extent possible in their national and regional standards. Any divergence between the IEC Standard and the corresponding national or regional standard shall be clearly indicated in the latter.
- 5) The IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with one of its standards.
- 6) Attention is drawn to the possibility that some of the elements of this International Standard may be the subject of patent rights. The IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 60746-3 has been prepared by subcommittee 65D: Analyzing equipment, of IEC technical committee 65: Industrial-process measurement and control.

This second edition cancels and replaces the first edition, published in 1985, and constitutes a technical revision.

This standard shall be used in conjunction with IEC 60746-1.

This bilingual version (2012-12) corresponds to the monolingual English version, published in 2002-06.

The text of this standard is based on the following documents:

FDIS	Report on voting
65D/85/FDIS	65D/87/RVD

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

The French version of this standard has not been voted upon.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 3.

Annex C forms an integral part of this standard.

Annexes A, B and D are for information only.

IEC 60746 consists of the following parts, under the general title *Expression of performance of electrochemical analyzers*:

Part 1: General

Part 2: pH value

Part 3: Electrolytic conductivity

Part 4: Dissolved oxygen in water measured by membrane covered amperometric sensors

Part 5: Oxidation-reduction potential or redox potential

Part 6: Conductivity effect of foreign ions in ultrapure waters, from combined conductivity and pH¹⁾

The committee has decided that the contents of this publication will remain unchanged until 2007. At this date, the publication will be

- reconfirmed;
- withdrawn;
- replaced by a revised edition, or
- amended.

The contents of the corrigendum of January 2003 have been included in this copy.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[IEC 60746-3:2002](https://standards.iteh.ai/catalog/standards/sist/7a5e6c87-ca15-4e9f-948b-acf306e59e68/iec-60746-3-2002)

<https://standards.iteh.ai/catalog/standards/sist/7a5e6c87-ca15-4e9f-948b-acf306e59e68/iec-60746-3-2002>

¹⁾ Under consideration.

EXPRESSION OF PERFORMANCE OF ELECTROCHEMICAL ANALYZERS –

Part 3: Electrolytic conductivity

1 Scope

This part of IEC 60746 is intended

- to specify terminology, definitions and requirements for statements by manufacturers for analyzers, sensor units, and electronic units used for the determination of the electrolytic conductivity of aqueous solutions;
- to establish performance tests for such analyzers, sensor units and electronic units;
- to provide basic documents to support the applications of quality assurance standards.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60746-1, *Expression of performance of electrochemical analyzers – Part 1: General*

[IEC 60746-3:2002](https://standards.iteh.ai/catalog/standards/sist/7a5e6c87-ca15-4e9f-948b-acf306e59e68/iec-60746-3-2002)

3 Definitions <https://standards.iteh.ai/catalog/standards/sist/7a5e6c87-ca15-4e9f-948b-acf306e59e68/iec-60746-3-2002>

For the purpose of this part of IEC 60746, the definitions of IEC 60746-1 apply, together with the following definitions.

3.1

electrolytic conductance

current divided by the potential difference in the case of ionic charge transport within an electrolytic solution filling a conductivity cell:

$$G = \frac{I}{U}$$

where

I is the current through the electrolyte, in amperes (A);

U is the potential difference applied across the electrodes, in volts (V);

G is the electrolytic conductance, in siemens (S).

Electrolytic resistance is the reciprocal of electrolytic conductance with the ohm (Ω) as the unit of measurement

3.2 electrolytic conductivity

formerly called specific conductance, is defined by the equation:

$$\kappa = \frac{j}{E}$$

where

j is the electric current density, in $A \cdot m^{-2}$;

E is the electric field strength, in $V \cdot m^{-1}$.

The unit of electrolytic conductivity, κ , is siemens per metre ($S \cdot m^{-1}$). Electrolytic resistivity is the reciprocal of electrolytic conductivity with the unit of ohmmeter ($\Omega \cdot m$)

NOTE In practical use, the most commonly employed conductivity unit is microsiemens per centimetre ($\mu S \cdot cm^{-1}$) or the corresponding resistivity unit, megohm per centimetre ($M\Omega \cdot cm$)

$$1 \mu S \cdot cm^{-1} = 10^{-4} S \cdot m^{-1} = 1 M\Omega \cdot cm.$$

3.3 cell constant of the sensor unit

an electrolytic conductor of a uniform cross-section X and length L is defined by the equation:

$$K_{\text{cell}} = \frac{L}{X}$$

where K_{cell} is the cell constant, in m^{-1} (see Note 1).

It is usual to measure electrolytic conductivity by means of cells without a uniform cross-section, in which case the K_{cell} should be determined by means of a reference solution of known electrolytic conductivity.

The relationship between electrolytic conductance and electrolytic conductivity is defined by the equation:

$$\kappa = K_{\text{cell}} \times G$$

where

κ is the electrolytic conductivity, in $S \cdot m^{-1}$;

G is the electrolytic conductance, in S ;

K_{cell} is the cell constant, in m^{-1} .

NOTE 1 In practical use, K_{cell} is generally expressed in cm^{-1} , κ in $\mu S \cdot cm^{-1}$ and G in μS (see 3.2).

NOTE 2 The cell constant will normally have a constant value over a stated range (see 4.1). Outside this range, it should be expected that polarization or other effects will produce errors (see 3.4 and annex D).

3.4 polarization

effect which occurs at electrode surfaces in an electrolytic solution when the current between the electrodes is such as to produce electrolysis and consequent partial insulation of the electrode surface

To avoid this uncertainty, different measuring methods can be applied (see 3.7, 3.8 and annex D):

- a) a.c. measurements with a frequency high enough to avoid polarization effects;
- b) four or six electrode measurements with separated current transporting and potential measuring electrodes;
- c) inductive or capacitive measurements by coupling between the electrolytic conductor and the electrical measuring circuit through non-conductive media.

In each case, the relationship between the electrolytic conductivity and the measured output quantity is established by the cell constant.

3.5 temperature coefficient

relative increase (or decrease) of the electrolytic conductivity of a solution per kelvin temperature change. The temperature coefficient is dependent on the reference temperature and the nature of the solution.

The following approximate equation can be applied for strong electrolyte solutions where $\kappa > 10^{-4} \text{ S}\cdot\text{m}^{-1}$ ($1 \mu\text{S}\cdot\text{cm}^{-1}$)

$$\kappa_t = \kappa_{tr} \times (1 + \alpha \Delta t)$$

where

κ_t is the electrolytic conductivity at temperature t ;

κ_{tr} is the electrolytic conductivity at reference temperature tr ;

Δt is the temperature difference $t - tr$;

α is the temperature coefficient.

In practice, this formula is sufficiently accurate over a small temperature range. For large temperature ranges, it is usually necessary to add higher terms of a polynomial series (such as $\beta(\Delta t)^2 + \gamma(\Delta t)^3 + \dots$) to the above equation, to obtain sufficient accuracy. The percentage temperature coefficient, which is the percentage relative deviation per kelvin from the reference value κ_{tr} , is often used so that

$$\alpha (\%) = 100 \alpha$$

(standards.iteh.ai)

NOTE The manufacturer's literature should be consulted for details of sample temperature compensation technique(s) applied.

[IEC 60746-3:2002](https://standards.iteh.ai/catalog/standards/sist/7a5e6c87-ca15-4e9f-948b-acf306e59e68/iec-60746-3-2002)

<https://standards.iteh.ai/catalog/standards/sist/7a5e6c87-ca15-4e9f-948b-acf306e59e68/iec-60746-3-2002>

3.6 simulator

a series of non-inductive resistors (preferably step-variable, e.g., a decade resistance box), used for the performance tests of conductimetric electronic units, simulating two- and three-electrode sensors

NOTE The minimum step should preferably be $0,01 R$, where R is the reciprocal value of the nominal full range conductivity value. Analogously, the highest resistance value should correspond to the lower limit of the measuring range: if the range begins at zero, the resistance value should be at least $10 R$ for testing at 10 % of the range.

For multi-electrode sensor simulator design, the manufacturer must be consulted.

The temperature sensor may be simulated by another variable precision resistor, e.g., a variable decade resistance box.

3.7 cell capacitance

produced by the electrostatic field existing between the sensor's measuring electrodes due to the high dielectric constant of water. Its value is inversely proportional to the cell constant and expressed by the approximate relationship:

$$C_{\text{cell}} \Psi \frac{7}{K_{\text{cell}}}$$

where

C_{cell} is the cell capacitance, in picofarads (pF: $1 \text{ pF} = 10^{-12} \text{ F}$);

K_{cell} is the cell constant, in cm^{-1} ;

C_{cell} may disturb low conductivity measurements made with two- or three-electrode cells when too high a frequency is used. The effect can be reduced by means of phase discrimination within the electronic unit.

3.8

leakage currents

a.c. currents at the measuring frequency flowing from the cell electrodes to local conductive parts in contact with the sample solution.

They effectively alter the cell constant causing measuring uncertainties and arise principally in symmetrical cells, i.e., with similar electrodes side by side.

4 Procedure for specification

See Clause 5 of IEC 60746-1, plus the following.

4.1 Additional statements on sensor units

- a) Type, i.e., flow-through, dip or insertion unit, number of electrodes, if electrodeless whether inductive or capacitive cell (common types of cell are described in annex D).
- b) Cell constant, tolerance and corresponding range of measurement (see 4.3a)).
- c) Type of temperature compensator (for example, Pt100).
- d) Sensor dimensions, mounting and connection details.

4.2 Additional statements on electronic units

- a) Measuring frequency/frequencies.
- b) Cell constant adjustment range.
- c) Type of temperature compensator sensor to which the electronic unit can be connected and maximum permitted resistance of compensator plus connection leads.
- d) Reference temperature adjustment range, if fixed, state temperature.
- e) Range of temperature coefficient adjustment and details of sample temperature compensation that may be applied. If fixed, state value.
- f) Installation details.

4.3 Additional statements on complete analyzers

- a) Measuring ranges (rated and effective).

NOTE Some analyzers employ concentration units, for example, mass % NaCl, g NaOH per litre, etc. For such analyzers, the rated range should be specified on the measurement unit as well as the corresponding conductivity at the rated reference temperature.

- b) Reference temperature for the measurement.
- c) Installation details.

5 Recommended standard values and ranges of influence quantities affecting the performance of electronic units

See annex A of IEC 60746-1.

6 Verification of values

See clause 6 of IEC 60746-1

6.1 General aspects

The parameters required to be set in the electronic unit for the specific combination of the electronic unit and sensor shall be established. These shall include

- the range to be measured and the units for display (if a display is fitted);
- the range and type of any transmitted output;
- the type of sample, i.e., a flow-through sample and its flow rate, or a static sample into which the sensor is immersed and its minimum depth;
- the type and range of sample temperature compensation for which verification is required, if applicable.

6.2 Calibration

For accurate calibration of a conductivity analyzer the following parameters are required for adjustment on the electronic unit:

- a) the cell constant (see 3.3 and 4.2 b));
- b) the reference temperature (see 4.2d));
- c) the temperature coefficient or appropriate algorithm (see 4.2e)).

NOTE The exact value of the cell constant can be determined using an appropriate calibration solution (see annex A).

iTeh STANDARD PREVIEW
(standards.iteh.ai)

6.3 Test solutions

Test solutions shall be applied in a manner suited to the design of the sensor.

[IEC 60746-3:2002](#)

For flow-through sensors, the solutions shall be applied at a flow rate within the manufacturer's stated rated range.

For sensors which can be immersed into test solutions, it is essential that the sensor unit is rinsed several times with water of negligible conductivity (in comparison with the range to be tested) after immersion in one solution prior to immersion in a fresh solution. A good procedure is to keep a second container of each test solution concentration to be used for the final rinse prior to immersion in each respective accurate test solution.

The immersion of sensor probes into containers of test solution exposed to air is not appropriate for measurements below $100 \mu\text{S}\cdot\text{cm}^{-1}$.

NOTE De-ionized water in an open container absorbs CO_2 ; a typical equilibrium conductivity of approximately $0,9 \mu\text{S}\cdot\text{cm}^{-1}$ is eventually reached.

Examples of test solutions are tabulated in annex B.

For low-conductivity solutions below about $100 \mu\text{S}\cdot\text{cm}^{-1}$ (at $25 \text{ }^\circ\text{C}$), it is essential that flowing solutions of appropriate conductivities are generated by continuous injection of, e.g., NaCl solutions into a pure water stream at a controlled flow rate. Required concentrations may be determined by extrapolation of values in annex B.

Pure water can only be generated by a circulatory de-ionization system: standard test solutions shall be generated from such water.

6.4 Test procedures

The following test procedures shall be carried out as described in IEC 60746-1, using the appropriate test solutions.

a) Intrinsic uncertainty

b) Linearity uncertainty

NOTE 1 Deviations from linearity of the sensor unit may be produced at high conductivities by polarization with too low a measuring frequency, and at low conductivities by cell capacitance with too high a frequency. Conductivity analyzers that incorporate frequency selection permit simple verification. For analyzers without frequency selection, polarization errors can be revealed by using test solutions at the upper range (approximately 80 % – 100 %). Capacitance uncertainties may be demonstrated by parallel connection of a capacitor of approximately the value given by the equation in 3.7 to the electronic unit input, i.e., to the cell electrodes (for four- and six-electrode cells, the inner voltage electrodes): if an increase of displayed value is observed, the capacitance uncertainty is about the same.

c) Repeatability

d) Output fluctuation

e) Warm-up time

f) Drift

NOTE 2 Drift is generally reported as a linear regression in two ways, short term over a period in the range of 1 h to 24 h and for a longer period in the range of 30 days to 100 days.

g) Response times

NOTE 3 These are given in detail in annex C, where method A is the preferred and the only method appropriate to flow-through cells.

h) Sample temperature

Uncertainties caused by variation of the sample temperature shall be determined at two points near the lower and higher limits of the measuring range. Measurements shall be made at the reference sample temperature, then at the lowest temperature within the compensation range (see 4.2d) and 4.2e)) and repeated at the highest temperature within that range.

NOTE 4 For flowing samples, this may be achieved by passage of the solution through a stainless-steel heat exchanger prior to the sensor.

i) Primary influence quantities

Response to the following influence quantities will generally need to be determined using two test solutions near the higher and lower ends of the range. Influence quantities should first be applied at the reference value then the two limits (upper and lower) of the rated range. Final measurement should be made when the quantity is returned to the reference value.

Variations in electrical supply characteristics usually affect the electronics unit only and may be tested with only one solution at the mid-point of the span. These tests may be carried out on the electronics unit alone, using resistors to simulate the mid-point of the range for conductivity and the temperature sensor.

- Ambient temperature
- Humidity
- Sample flow-rate
- Sample pressure
- Sample outlet pressure (if different from above)
- Vibration
- Supply voltage
- AC supply frequency, or
- DC supply ripple and impedance
- Electromagnetic compatibility