



Designation: **E2124 – 10** **E2124 – 15**

Standard Practice for Specification for Equipment and Supplies in Sexual Assault Investigations¹

This standard is issued under the fixed designation E2124; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This practice describes the basic instruments used for the medical-legal examination of victims or suspects, or both, in sexual assault investigations. These specifications are designed to be used in conjunction with other specifications, guides, and practices associated with sexual assault examinations that are listed in Section 2.

1.2 Additional information can be found in Terminology E1732, Guide E1843, Specification E2057, and Practice E2123.

1.3 This practice offers a set of instructions for performing one or more specific operations. This standard cannot replace knowledge, skill, or ability acquired through appropriate education, training, and experience and should be used in conjunction with sound professional judgment.

1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 *ASTM Standards:*²

E1732 Terminology Relating to Forensic Science

E1843 Guide for Sexual Assault Investigation, Examination, and Evidence Collection

E2057 Specifications for Preparation of Laboratory Analysis Requests in Sexual Assault Investigations

E2123 Practice for Transmittal of Evidence in Sexual Assault Investigation

3. Summary of Practice

3.1 These specifications outline considerations when choosing equipment to be used in the medical-legal examination of sexual assaults. The types of equipment listed should be included in any examination facility. Considerations for choosing a variation or optional equipment are also given.

4. Instrumentation and Supplies

4.1 The following instrumentation should be available at examining facilities where sexual assault cases are received:

4.1.1 *Speculum*—The speculum is used to view the vaginal walls and cervix. Clear plastic disposable specula are preferred because they allow the examiner to see through them and visualize the vaginal walls. Metal specula do not allow this visibility and also may reflect light when taking photographs. Speculum illuminators which attach to the handle should be considered as an aid (as they assist) in visualizing the vaginal walls and cervical area.

4.1.2 *Colposcope*—The colposcope is used to magnify and provide increased visibility, especially in the vaginal area where structures include many folds and grooves of tissue. The examiner must deliberately examine each structure to determine the presence or absence of injury or foreign material, or both. The use of colposcopy has been shown to be critical in both pediatric and adult sexual assault examinations.

4.1.2.1 There are several varieties of colposcopes manufactured and each program should review numerous types to ascertain the best instrument for their facility in view of existing equipment and physical layout of the examination environment.

¹ This practice is under the jurisdiction of ASTM Committee E30 on Forensic Sciences and is the direct responsibility of Subcommittee E30.01 on Criminalistics. Current edition approved June 1, 2010/Jan. 1, 2015. Published August 2010/January 2015. Originally approved in 2001. Last previous edition approved in 2009/2010 as E2124 – 09/E2124 – 10. DOI: 10.1520/E2124-10.10.1520/E2124-15.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For Annual Book of ASTM Standards volume information, refer to the standard's Document Summary page on the ASTM website.