



Designation: ~~D3444 – 00 (Reapproved 2010)~~<sup>ε1</sup> D3444 – 00 (Reapproved 2015)

## Standard Test Method for Total Acid Number of Trichlorotrifluoroethane<sup>1</sup>

This standard is issued under the fixed designation D3444; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

---

<sup>ε1</sup> NOTE—Section 5.1 was updated in June 2010.

---

### 1. Scope

1.1 This test method covers the determination of total acidity in trichlorotrifluoroethane and other halocarbons that are liquid at room temperature.

1.2 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.3 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.* Specific precautionary statements are given in Section 6.

### 2. Summary of Test Method

2.1 The total acidity of the solvent is determined by titration with standard alcoholic potassium hydroxide solution using phenolphthalein solution as the indicator.

### 3. Significance and Use

3.1 This test method determines the total acidity of trichlorotrifluoroethane by titration with alcoholic potassium hydroxide using phenolphthalein for an indicator.

3.2 This test method can be used to develop production and purchasing specifications.

### 4. Apparatus

4.1 *Buret*, 2-mL in 0.01-mL subdivisions, self-leveling, with 1000-mL reservoir.

4.2 *Vinyl Tubing*.<sup>22</sup>

4.3 *Two Indicating Carbon Dioxide Absorbent Tubes, 30–50 mesh*.

### 5. Reagents

5.1 *Purity of Reagents*—Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society, where such specifications are available.<sup>3</sup> Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

5.2 *Purity of Water*—Unless otherwise indicated, references to water shall be understood to mean halide-free distilled water.

5.3 *Isopropyl Alcohol*.

5.4 *Phenolphthalein Indicator Solution* (0.05 g in 100 mL of isopropyl alcohol).

5.5 *Potassium Acid Phthalate*, U.S. National Bureau of Standards primary standard. Dry this chemical in a 120°C oven for several hours prior to use. Store the dry chemical in a desiccator.

---

<sup>1</sup> This test method is under the jurisdiction of ASTM Committee D26 on Halogenated Organic Solvents and Fire Extinguishing Agents and is the direct responsibility of Subcommittee D26.04 on Test Methods.

Current edition approved June 1, 2010/June 1, 2015. Published July 2010/June 2015. Originally approved in 1975. Last previous edition approved in 2005/2010 as D3444 – 00(2005)-(2010)<sup>ε1</sup>. DOI: 10.1520/D3444-00R10E1-10.1520/D3444-00R15.

<sup>2</sup> Tygon has been found satisfactory. An equivalent may be used.

<sup>3</sup> *Reagent Chemicals, American Chemical Society Specifications*, American Chemical Society, Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.