



Designation: D7165 – 10 (Reapproved 2015)

# Standard Practice for Gas Chromatograph Based On-line/At-line Analysis for Sulfur Content of Gaseous Fuels<sup>1</sup>

This standard is issued under the fixed designation D7165; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

## 1. Scope

1.1 This practice is for the determination of volatile sulfur-containing compounds in high methane content gaseous fuels such as natural gas using on-line/at-line instrumentation, and continuous fuel monitors (CFMS). It has been successfully applied to other types of gaseous samples including air, digester, landfill, and refinery fuel gas. The detection range for sulfur compounds, reported as picograms sulfur, based upon the analysis of a 1 cc sample, is one hundred (100) to one million (1,000,000). This is equivalent to 0.1 to 1,000 mg/m<sup>3</sup>.

1.2 This practice does not purport to measure all sulfur species in a sample. Only volatile compounds that are transported to an instrument under the measurement conditions selected are measured.

1.3 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.4 *This practice does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this practice to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

## 2. Referenced Documents

### 2.1 ASTM Standards:<sup>2</sup>

- D1072 Test Method for Total Sulfur in Fuel Gases by Combustion and Barium Chloride Titration
- D1945 Test Method for Analysis of Natural Gas by Gas Chromatography
- D3606 Test Method for Determination of Benzene and Toluene in Finished Motor and Aviation Gasoline by Gas Chromatography

<sup>1</sup> This practice is under the jurisdiction of ASTM Committee D03 on Gaseous Fuels and is the direct responsibility of Subcommittee D03.12 on On-Line/At-Line Analysis of Gaseous Fuels.

Current edition approved June 1, 2015. Published July 2015. Originally approved in 2006. Last previous edition approved in 2010 as D7165-10. DOI: 10.1520/D7165-10R15.

<sup>2</sup> For referenced ASTM standards, visit the ASTM website, [www.astm.org](http://www.astm.org), or contact ASTM Customer Service at [service@astm.org](mailto:service@astm.org). For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

- D3764 Practice for Validation of the Performance of Process Stream Analyzer Systems
  - D4084 Test Method for Analysis of Hydrogen Sulfide in Gaseous Fuels (Lead Acetate Reaction Rate Method)
  - D4468 Test Method for Total Sulfur in Gaseous Fuels by Hydrogenolysis and Rateometric Colorimetry
  - D4626 Practice for Calculation of Gas Chromatographic Response Factors
  - D4810 Test Method for Hydrogen Sulfide in Natural Gas Using Length-of-Stain Detector Tubes
  - D5504 Test Method for Determination of Sulfur Compounds in Natural Gas and Gaseous Fuels by Gas Chromatography and Chemiluminescence
  - D6621 Practice for Performance Testing of Process Analyzers for Aromatic Hydrocarbon Materials
  - D6122 Practice for Validation of the Performance of Multi-variate Online, At-Line, and Laboratory Infrared Spectrophotometer Based Analyzer Systems
  - D6228 Test Method for Determination of Sulfur Compounds in Natural Gas and Gaseous Fuels by Gas Chromatography and Flame Photometric Detection
  - E594 Practice for Testing Flame Ionization Detectors Used in Gas or Supercritical Fluid Chromatography
- 2.2 ISO Standards<sup>3</sup>
- ISO 7504 Gas Analysis-Vocabulary

## 3. Terminology

### 3.1 Definitions:

3.1.1 *calibration gas mixture, n*—a certified gas mixture with known composition used for the calibration of a measuring instrument or for the validation of a measurement or gas analytical method.

3.1.1.1 *Discussion*—Calibration Gas Mixtures are the analogues of measurement standards in physical metrology (reference ISO 7504 paragraph 4.1).

3.1.2 *direct sampling*—Sampling where there is no direct connection between the medium to be sampled and the analytical unit.

<sup>3</sup> Available from International Organization for Standardization (ISO), 1, ch. de la Voie-Creuse, Case postale 56, CH-1211, Geneva 20, Switzerland, <http://www.iso.ch>.

3.1.3 *in-line instrument*—Instrument whose active element is installed in the pipeline and measures at pipeline conditions.

3.1.4 *on-line instrument*—Automated instrument that samples gas directly from the pipeline, but is installed externally.

3.1.5 *at-line instrument*—instrument requiring operator interaction to sample gas directly from the pipeline.

3.1.6 *continuous fuel monitor (CFM)*—Instrument that samples gas directly from the pipeline on a continuous or semi-continuous basis.

3.1.7 *total reduced sulfur (TRS)*—Summation of sulfur species where the sulfur oxidation number is  $-2$ , excluding sulfur dioxide, sulfones, and other inorganic sulfur compounds. This includes but is not limited to mercaptans, sulfides, and disulfides.

3.1.8 *near-real time monitoring systems*—Monitoring system where measurement occurs soon after sample flow through the system or soon after sample extraction. The definition of a near real time monitoring system can be application specific.

3.2 *reference gas mixture, n*—a certified gas mixture with known composition used as a reference standard from which other compositional data are derived.

3.2.1 *Discussion*—Reference Gas Mixtures are the analogues of measurement standards of reference standards (reference ISO 7504 paragraph 4.1.1).

## 4. Summary of Practice

4.1 A representative sample of the gaseous fuel is extracted from a process pipe or pipeline and is transferred in a timely manner to an analyzer inlet system. The sample is conditioned with minimum impact on sulfur content. A precisely measured volume of sample is injected into the analyzer. Excess process or pipeline sample is vented or is returned to the process stream dependant upon application and regulatory requirements.

4.2 Sample containing carrier gas is fed to a gas chromatograph where the components are separated using either a packed or capillary column. Measurement is performed using a suitable sulfur detection system.

4.3 Calibration, precision, calibration error, performance audit tests, maintenance methodology and miscellaneous quality assurance procedures are conducted to determine analyzer performance characteristics and validate both the operation and the quality of generated results.

## 5. Significance and Use

5.1 On-line, at-line, in-line, CFMS, and other near-real time monitoring systems that measure fuel gas characteristics, such as the sulfur content, are prevalent in the natural gas and fuel gas industries. The installation and operation of particular systems vary on the specific objectives, contractual obligations, process type, regulatory requirements, and internal performance requirements needed by the user. This standard is intended to provide guidelines for standardized start-up procedures, operating procedures, and quality assurance practices for on-line, at-line, in-line, CFMS, and other near-real time gas chromatographic based sulfur monitoring systems

used to determine fuel gas sulfur content. For measurement of gaseous fuel properties using laboratory based methods the user is referred to Test Methods [D1072](#), [D1945](#), [D4084](#), [D4468](#), [D4810](#) and Practices [D4626](#), [E594](#).

## 6. Apparatus

6.1 *Instrument*—Any gas chromatographic based instrument of standard manufacture, with hardware necessary for interfacing to a natural gas or other fuel gas pipeline and containing all features necessary for the intended application(s) can be used.

6.1.1 The chromatographic parameters must be capable of obtaining retention time repeatability of 0.05 min. (3 sec.). Instrumentation must meet the performance characteristics for repeatability and precision without encountering unacceptable interference or bias. The components coming in contact with sample, such as tubing and valving, must be passivated or constructed of inert materials to ensure an accurate sulfur gas measurement.

6.2 *Sample Inlet System*—A sample inlet system capable of operating continuously above the maximum column temperature is necessary. A variety of sample inlet configurations can be used including but not limited to on-column systems and split/splitless injection system capable of splitless operation and split control from 10:1 up to 50:1. An automated gas sampling valve is required for many applications. The inlet system must be constructed of inert material and evaluated frequently for compatibility with reactive sulfur compounds. The sampling inlet system is heated as necessary so as to prevent condensation. All wetted sampling system components must be constructed of inert or passivated materials. Sample delivered to the inlet system should be in the gas phase free of particulate or fluidic matter.

6.2.1 *Carrier and Detector Gas Control*—Constant flow control of carrier and detector gases is critical for optimum and consistent analytical performance. Control is achieved by use of pressure regulators and fixed flow restrictors. The gas flow is measured by appropriate means and adjusted, as required, to the desired value. Mass flow controllers, capable of maintaining a gas flow constant to within  $\pm 1\%$  at the flow rates necessary for optimal instrument performance can be used.

6.2.2 *Detector*—Sulfur compounds can be measured using a variety of detectors including but not limited to: sulfur chemiluminescence, flame photometric, electrochemical cell, oxidative cell and reductive cells. In selecting a detector, the user should consider the linearity, sensitivity, and selectivity of particular detection systems prior to installation. The user should also consider interference from substances in the gas stream that could result in inaccurate sulfur gas measurement due to effects such as quenching.

6.3 *Columns*—A variety of columns can be used to separate the sulfur compounds in the sample. Typically, a 60 m  $\times$  0.53 mm ID fused silica open tubular column containing a 5  $\mu$ m film thickness of bonded methyl silicone liquid phase is used. The selected column must provide retention and resolution characteristics that satisfy the intended application. The column must be inert towards sulfur compounds. The column must also demonstrate a sufficiently low liquid phase bleed at high

temperature such that a loss of the instrument response is not encountered while operating the column at elevated temperatures.

6.4 *Data Acquisition*—Data acquisition and storage can be accomplished using a number of devices and media. Following are some examples.

6.4.1 *Recorder*—As an example, a 0 to 1 mV range recording potentiometer or equivalent, with a full-scale response time of 2 s or less can be used. A 4-20 mA range recorder can also be used.

6.4.2 *Integrator*—An electronic integrating device or computer can be used. For GC based systems, it is suggested that the device and software have the following capabilities:

6.4.2.1 Graphic presentation of chromatograms.

6.4.2.2 Digital display of chromatographic peak areas.

6.4.2.3 Identification of peaks by retention time or relative retention time, or both.

6.4.2.4 Calculation and use of response factors.

6.4.2.5 External standard calculation and data presentation.

6.4.3 *Distributed Control Systems (DCS)*—Depending on the site requirements, the analytical results are sometimes fed to a distributed control system. The information is then used to make the appropriate adjustments to the process. Signal isolation between the analyzer and the distributed control network is most often required. Communications protocols with the DCS will dictate the required signal output requirements for the analyzer.

6.4.4 *Data Management Systems*—Data management systems or other data and data processing repositories are sometimes used to collect and process the results from a wide variety of instrumentation at a single facility. The information is then available for rapid dissemination within the organization of the operating facility. Communications protocols with the data management system will dictate the required signal output requirements for the analyzer.

## 7. Reagents and Materials

**NOTE 1—Warning:** Sulfur compounds contained in permeation tubes or compressed gas cylinders may be flammable and harmful or fatal if ingested or inhaled. Permeation tubes, which emit their contents continuously, and compressed gas standards should only be handled in well ventilated locations away from sparks and flames. Improper handling of compressed gas cylinders containing air, hydrogen, argon, nitrogen or helium can result in an explosion or in creating oxygen deficient atmospheres. Rapid release of argon, nitrogen or helium can result in asphyxiation. Compressed air supports combustion.

7.1 *Sulfur Standards*—Accurate sulfur standards are required for the quantitation of the sulfur content of natural gas. Permeation and compressed gas standards should be stable, and of the highest available accuracy and purity.

7.1.1 *Permeation Devices*—Sulfur standards can be produced on demand using permeation tubes, one for each selected sulfur species, gravimetrically calibrated and certified at a convenient operating temperature. With constant temperature, calibration gases covering a wide range of concentration can be generated by varying and accurately measuring the flow rate of diluent gas passing over the tubes. Permeation devices delivering calibrant at a known high purity must be used since contaminants will adversely impact the calculation of analyte concentration due to error in permeation

rate calculated from differential weight measurements of these devices. It is suggested that certified permeation devices be used whenever available.

7.1.1.1 *Permeation System Temperature Control*—Permeation devices are maintained at the calibration temperature within 0.1 °C.

7.1.1.2 *Permeation System Flow Control*—The permeation flow system measures diluent gas flow over the permeation tubes within  $\pm 2$  percent.

7.1.1.3 Permeation tube emission rates are expressed in units of mass of the emitted sulfur compound contained inside per unit time, i.e. nanograms of methyl mercaptan per minute. The sulfur emission rate is calculated knowing the molecular formula of the sulfur compound used in the permeation tube.

7.1.1.4 Permeation tubes are inspected and weighed to the nearest 0.01 mg on at least a monthly basis using a balance calibrated against NIST traceable “S” class weights or the equivalent. Analyte concentration is calculated by weight loss and dilution gas flow rate as per Practice **D3606**. These devices are discarded when the liquid contents are reduced to less than ten (10) percent of the initial volume or when the permeation surface is unusually discolored or otherwise compromised.

7.1.1.5 Permeation tubes must be stored in accordance with the manufacturer’s recommendation. Improper storage can result in damage and/or a change in the characteristics of the permeation membrane. Such damage and/or characteristic change results in an actual permeation rate that differs from the certified permeation rate.

7.2 *Compressed Gas Standards*—Alternatively, blended gaseous sulfur standards in nitrogen, helium or methane base gas may be used. Care must be exercised in the use of compressed gas standards since they can introduce errors in measurement due to lack of uniformity in their manufacture or instability in their storage and use. Standards should be blended such that components will not condense under storage or while the standard is in use. The protocol for compressed gas standards contained in the appendix can be used to ensure uniformity in compressed gas standard manufacture and provide for traceability to a NIST or NMi (Nederlands Meetinstituut) reference material.

7.2.1 Compressed gas standard regulators must be appropriate for the delivery of sulfur gases and attached fittings must be passivated or inert to sulfur gases.

7.2.2 All compressed gas standards must be re-certified as recommended by the manufacturer or as needed to insure accuracy.

7.3 The following sulfur compounds, including the molecular formula and the CAS number, are commonly found or are added to natural gas and related fuel gases and may be useful as calibrants for on-line and at-line monitors:

7.3.1 Hydrogen sulfide (H<sub>2</sub>S) (7783-06-4)

7.3.2 Methyl mercaptan (CH<sub>3</sub>SH) (74-93-1)

7.3.3 Ethyl mercaptan (CH<sub>3</sub>CH<sub>2</sub>SH) (75-08-1)

7.3.4 1-propanethiol (CH<sub>3</sub>CH<sub>2</sub>CH<sub>2</sub>SH) (107-03-9)

7.3.5 2-propanethiol (CH<sub>3</sub>CHSHCH<sub>3</sub>) (75-33-2)

7.3.6 Dimethyl sulfide (CH<sub>3</sub>SCH<sub>3</sub>) (75-18-3)

7.3.7 Dimethyl disulfide (CH<sub>3</sub>SSCH<sub>3</sub>) (624-92-0)

7.3.8 Tetrahydrothiophene (THT) (110-01-0)