



Designation: D4606 – 15

Standard Test Method for Determination of Arsenic and Selenium in Coal by the Hydride Generation/Atomic Absorption Method¹

This standard is issued under the fixed designation D4606; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This test method² covers the determination of total arsenic and selenium in coal.

1.2 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.3 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 *ASTM Standards:*³

[D346 Practice for Collection and Preparation of Coke Samples for Laboratory Analysis](#)

[D2013 Practice for Preparing Coal Samples for Analysis](#)

[D3173 Test Method for Moisture in the Analysis Sample of Coal and Coke](#)

[D3180 Practice for Calculating Coal and Coke Analyses from As-Determined to Different Bases](#)

[D7582 Test Methods for Proximate Analysis of Coal and Coke by Macro Thermogravimetric Analysis](#)

3. Summary of Test Method

3.1 Arsenic and selenium are determined by mixing a weighed coal sample with Eschka mixture and igniting at

¹ This test method is under the jurisdiction of ASTM Committee D05 on Coal and Coke and is the direct responsibility of Subcommittee D05.29 on Major Elements in Ash and Trace Elements of Coal.

Current edition approved Nov. 1, 2015. Published December 2015. Originally approved in 1986. Last previous edition approved in 2007 as D4606 – 03(2007). DOI: 10.1520/D4606-15.

² For information concerning experimental work on which this test method is based see: Bosshart, R. E., Price, A. A., and Ford, C. T., "Evaluation of the Effect of Coal Cleaning on Fugitive Elements, Phase II Final Report, Part II Analytical Methods," *ERDA Report No. C00-44727-35*, 1980, pp. 94–102; Fernandez, F. J., "Atomic Absorption Determination of Gaseous Hydrides Utilizing Sodium Borohydride Reduction," *Atomic Absorption Newsletter*, Vol 12, No. 4, 1973, pp. 93–97; and Brodie, K. G., "A Comparative Study—Determining Arsenic and Selenium by AAS," *American Laboratory*, March 1977, pp. 73–78.

³ For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

750 °C. The mixture is dissolved in hydrochloric acid and the gaseous hydride of each element is generated from the appropriate oxidation state and determined by atomic absorption spectrophotometry.

4. Significance and Use

4.1 This test method permits measurement of the total arsenic and selenium content of coal for the purpose of evaluating these elements where they can be of concern, for example, in coal combustion. When coal samples are prepared for analysis in accordance with this test method, the arsenic and selenium are quantitatively retained and are representative of the total amounts in the coal.

5. Apparatus

5.1 *Atomic Absorption Spectrophotometer*, with background correction system and peak profile recording device.

5.2 *Hydride Generation Apparatus*, for producing the hydrides of arsenic and selenium.

5.3 *Burner or Heated Quartz Cell*, for thermal decomposition of the hydrides.

5.4 *Hotplate*, capable of maintaining a temperature of a solution at 60 °C to 90 °C.

5.5 *Ignition Crucibles*—Porcelain crucible of 30 mL capacity. Do not use a porcelain crucible in which the glaze is flaked.

5.6 *Analytical Balance*, capable of weighing to a resolution of 0.0001 g.

6. Reagents

6.1 *Purity of Reagents*—Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society, where such specifications are available.⁴

⁴ *Reagent Chemicals, American Chemical Society Specifications*, American Chemical Society, Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.

6.2 *Purity of Water*—Use high-purity, conductivity water, prepared by passing distilled water (or equivalent) through an ion exchange resin.

6.3 *Eschka Mixture*—Prepare (with thorough mixing) a mixture that has a mass fraction of 67 % light calcined magnesium oxide (MgO) and 33 % anhydrous sodium carbonate (Na₂CO₃). The mixture shall be as free as possible from arsenic and selenium.

6.4 *Hydrochloric Acid Concentrated* (sp gr 1.19)—Concentrated hydrochloric acid (HCl).

6.5 *Hydrochloric Acid (1 + 4)*—Prepare a solution of dilute HCl in water with a volume fraction of 20 % of concentrated hydrochloric acid (HCl, sp gr 1.19).

6.6 *Potassium Iodide Solution* 0.2 g/mL—Dissolve 20 g of potassium iodide (KI) in 100 mL of water.

6.7 *Sodium Borohydride (Pellets or Solution)* 0.03 g/mL—To dilute sodium hydroxide (NaOH) solution (0.01 g/mL), add sodium borohydride (NaBH₄) to give a solution that is 0.03 g/mL sodium borohydride (NaBH₄). Prepare fresh daily.

6.8 *Arsenic Stock Solution* (1000 mg/L Arsenic)—Certified commercially available standard solution or prepared from primary standard arsenic trioxide (As₂O₃).

6.9 *Selenium Stock Solution* (1000 mg/L Selenium)—Certified commercially available standard solution or prepared from selenium metal (99.99 %).

7. Analysis Sample

7.1 Obtain the sample for coal in accordance with Method D2013 and the sample for coke in accordance with Practice D346. Prepare the analysis sample by pulverizing the material to pass a 250 µm (No. 60) U.S.A. standard sieve.

7.2 A separate portion of the analysis sample shall be analyzed concurrently for moisture content in accordance with Test Method D3173 or Test Methods D7582.

7.3 Use certified reference materials, such as the National Institute of Standards and Technology (NIST) Standard Reference Material (SRM) coals. SRM 1632d and SRM 163a5 are two different coal samples that have certified values for both arsenic and selenium. Results obtained by analyzing these coals for arsenic and selenium using the test methods described herein can be used for checking analytical technique and test method accuracy. At least one SRM coal sample should be analyzed as a control when a set of coal samples are analyzed.

8. Sample Preparation and Solution

8.1 To minimize the potential for contamination, clean all labware thoroughly with dilute HCl solution (1 + 9) and rinse with water. The hydride generation apparatus shall be kept equally clean.

8.2 Weigh approximately 1 g of coal into a preweighed 30 mL crucible. Record the mass to the nearest 0.0001 g. Thoroughly mix the sample with 1.5 g of Eschka mixture and cover the mixture with an additional 1.5 g of Eschka mixture. Place the sample in a cold muffle furnace. Set the temperature

control to 500 °C. Heat the sample for 1 h. Increase the furnace temperature to 750 °C. Continue heating at this temperature for 3 h. Remove the sample and allow it to cool to room temperature.

8.3 Add 20 mL to 30 mL of hot water to a 150 mL beaker. Transfer the contents of the crucible from 8.2 to the beaker. Add 5 mL of concentrated HCl to the crucible. Then slowly and carefully, with the aid of a stirring rod, transfer the HCl to the beaker. Rinse any remaining material in the crucible into the beaker with water. Add 15 mL concentrated HCl in three 5 mL portions to the crucible and transfer to the beaker. Swirl the contents until the Eschka has dissolved (Note 1). Allow the solution to cool to room temperature and transfer it to a 100 mL polypropylene volumetric flask. Dilute to volume with water.

NOTE 1—Some residue can remain.

9. Blank

9.1 Weigh 15 g of Eschka mixture into a crucible. Heat as described in 8.2. Add 100 mL to 150 mL hot deionized water to a 500 mL beaker. Transfer the contents of the crucible to the beaker. Add 25 mL concentrated HCl to the crucible. Then slowly and carefully, with the aid of a stirring rod, transfer the HCl to the beaker. Rinse any remaining material in the crucible into the beaker with water. Add 75 mL concentrated HCl in three 25 mL portions to the crucible and transfer to the beaker. Swirl the contents until the Eschka has dissolved (Note 1). Allow the solution to cool to room temperature and transfer to a 500 mL polypropylene volumetric flask. Dilute to volume with water.

10. Procedure for Arsenic

10.1 The solutions and preparations described are typical for North American coals. Different solution concentrations can be required to establish suitable analytical results for those elements with mass fractions outside the typical range. Each analyst shall determine the sensitivity and optimum method of calibration of their own equipment and choose standards with concentration ranges compatible with the samples and instruments specific to their work. Twenty millilitres is a normal working volume for batch hydride generation systems. If a hydride generation system is designed to accommodate other than 20 mL (sample + acid), a different volume can be used and all reagents and sample volumes can be ratioed accordingly.

10.2 Prepare an intermediate arsenic standard 1 mL to 10 µg/mL from the stock solution. Prepare fresh daily. Prepare a working standard (Note 2) by adding 1 mL of the intermediate standard to a 100 mL polypropylene volumetric flask. Dilute to volume with blank (Section 9).

NOTE 2—Five millilitres of the working standard should produce a signal approximately in the middle of the linear dynamic range for the atomic absorption operating conditions chosen.

10.3 Prepare three calibration standards by combining volumes of the working standard and blank solution that yield a total volume equivalent to the analytical volume (A_v) (see Annex A1) in a reaction flask. Dilute to 20 mL with HCl (1 + 4).