



Designation: **E973–10 (Reapproved 2015) E973 – 15**

Standard Test Method for Determination of the Spectral Mismatch Parameter Between a Photovoltaic Device and a Photovoltaic Reference Cell ¹

This standard is issued under the fixed designation E973; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This test method ~~covers~~provides a procedure for the determination of a spectral mismatch parameter used in performance testing of photovoltaic devices.

1.2 The spectral mismatch parameter is a measure of the ~~error, error~~ introduced in the testing of a photovoltaic ~~device, device~~ that is caused by ~~mismatch between the spectral responses of the photovoltaic device the photovoltaic device under test and the photovoltaic reference cell, cell having non-identical quantum efficiencies,~~ as well as mismatch between the test light source and the reference spectral irradiance distribution to which the photovoltaic reference cell was calibrated. ~~Examples of reference spectral irradiance distributions are Tables E490 or G173.~~

1.2.1 Examples of reference spectral irradiance distributions are Tables E490 or G173.

1.3 The spectral mismatch parameter can be used to correct photovoltaic performance data for spectral mismatch error.

1.4 Temperature-dependent quantum efficiencies are used to quantify the effects of temperature differences between test conditions and reporting conditions.

1.5 This test method is intended for use with linear photovoltaic ~~devices, devices in which short-circuit is directly proportional to incident irradiance.~~

1.6 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.7 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 *ASTM Standards:*²

[E490 Standard Solar Constant and Zero Air Mass Solar Spectral Irradiance Tables](#)

[E772 Terminology of Solar Energy Conversion](#)

[E948 Test Method for Electrical Performance of Photovoltaic Cells Using Reference Cells Under Simulated Sunlight](#)

[E1021 Test Method for Spectral Responsivity Measurements of Photovoltaic Devices](#)

[E1036 Test Methods for Electrical Performance of Nonconcentrator Terrestrial Photovoltaic Modules and Arrays Using Reference Cells](#)

~~[E1039 Test Method for Calibration of Silicon Non-Concentrator Photovoltaic Primary Reference Cells Under Global Irradiation \(Withdrawn 2004\)](#)~~³

[E1125 Test Method for Calibration of Primary Non-Concentrator Terrestrial Photovoltaic Reference Cells Using a Tabular Spectrum](#)

~~[E1328 Terminology Relating to Photovoltaic Solar Energy Conversion \(Withdrawn 2012\)](#)~~³

[E1362 Test Method for Calibration of Non-Concentrator Photovoltaic Secondary Reference Cells](#)

[G138 Test Method for Calibration of a Spectroradiometer Using a Standard Source of Irradiance](#)

[G173 Tables for Reference Solar Spectral Irradiances: Direct Normal and Hemispherical on 37° Tilted Surface](#)

¹ This test method is under the jurisdiction of ASTM Committee E44 on Solar, Geothermal and Other Alternative Energy Sources and is the direct responsibility of Subcommittee E44.09 on Photovoltaic Electric Power Conversion.

Current edition approved March 1, 2015; Dec. 1, 2015. Published April 2015/January 2016. Originally approved in 1983. Last previous edition approved in 2010 as E973–10, –10(2015). DOI: 10.1520/E0973-10R15; 10.1520/E0973-15.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

SI10 Standard for Use of the International System of Units (SI): The Modern Metric System
3. Terminology

3.1 *Definitions*—Definitions of terms used in this test method may be found in Terminology [E772](#) and Terminology [E1328](#).

3.2 *Definitions of Terms Specific to This Standard:*

3.2.1 *test light source, n*—a source of illumination whose spectral irradiance will be used for the spectral mismatch calculation.

The light source may be natural sunlight or a solar simulator.

3.3 *Symbols*—The following symbols and units are used in this test method:

M —spectral mismatch parameter;

ϵ —measurement error in short-circuit current;

λ —wavelength, μm or nm ;

$R_r(\lambda)$ —spectral response of reference cell, AW^{-1} ;

$R_f(\lambda)$ —spectral response of photovoltaic device, AW^{-1} ;

E —irradiance, Wm^{-2} ;

$E(\lambda)$ —spectral irradiance, $\text{Wm}^{-2}\mu\text{m}^{-1}$ or $\text{Wm}^{-2}\text{nm}^{-1}$, and

$E_0(\lambda)$ —reference spectral irradiance, $\text{Wm}^{-2}\mu\text{m}^{-1}$ or $\text{Wm}^{-2}\text{nm}^{-1}$.

NOTE 1—Following normal SI rules for compound units (see Practice [SI10](#)), the units for spectral irradiance, the derivative of irradiance with respect to wavelength $dE/d(\lambda)$, would be Wm^{-3} . However, to avoid possible confusion with a volumetric power density unit and for convenience in numerical calculations, it is common practice to separate the wavelength in the compound unit. This compound unit is also used in Tables [G173](#).

3.3 Symbols: The following symbols and units are used in this test method:

3.3.1 λ —wavelength (μm or nm).

3.3.2 D —as a subscript, refers to the device to be tested.

3.3.3 R —as a subscript, refers to the reference cell.

3.3.4 S —as a subscript, refers to the test light source.

3.3.5 O —as a subscript, refers to the reference spectral irradiance distribution.

3.3.6 A —active area, (m^2).

3.3.7 E —irradiance ($\text{W}\cdot\text{m}^{-2}$).

3.3.8 $E_S(\lambda)$ —spectral irradiance, test light source ($\text{W}\cdot\text{m}^{-2}\cdot\mu\text{m}^{-1}$ or $\text{W}\cdot\text{m}^{-2}\cdot\text{nm}^{-1}$).

3.3.9 $E_0(\lambda)$ —spectral irradiance, to which the reference cell is calibrated ($\text{W}\cdot\text{m}^{-2}\cdot\mu\text{m}^{-1}$ or $\text{W}\cdot\text{m}^{-2}\cdot\text{nm}^{-1}$).

3.3.9.1 Discussion—

[ASTM E973-15](#)

<https://standards.iteh.ai/catalog/standards/sist/fb31fe74-90d6-4974-9020-eca12d6af4de/astm-e973-15>

Following normal SI rules for compound units (see Practice [SI10](#)), the units for spectral irradiance, the derivative of irradiance, with respect to wavelength, $dE/d\lambda$, would be $\text{W}\cdot\text{m}^{-3}$. However, to avoid possible confusion with a volumetric power density unit and for convenience in numerical calculations, it is common practice to separate the wavelength in the compound unit. This compound unit is also used in Tables [G173](#).

3.3.10 I —short-circuit current (A).

3.3.11 J_L —light-generated photocurrent density ($\text{A}\cdot\text{m}^{-2}$).

3.3.12 M —spectral mismatch parameter (dimensionless).

3.3.13 $Q(\lambda, T)$ —quantum efficiency (electrons per photon or %).

3.3.14 $\Theta(\lambda)$ —partial derivative of quantum efficiency with respect to temperature (electrons per photon $\cdot^\circ\text{C}^{-1}$ or $\%\cdot^\circ\text{C}^{-1}$).

3.3.15 $R(\lambda)$ —spectral responsivity ($\text{A}\cdot\text{W}^{-1}$).

3.3.16 T —temperature ($^\circ\text{C}$).

3.3.17 T_{R0} —temperature, at which the reference cell is calibrated ($^\circ\text{C}$).

3.3.18 T_{D0} —temperature, to which the short-circuit current of the device to be tested will be reported ($^\circ\text{C}$).

3.3.18.1 Discussion—

When reporting photovoltaic performance to Standard Reporting Conditions (SRC), it is common for $T_{R0} = T_{D0} = 25^\circ\text{C}$.

3.3.19 q —electron charge (C).

3.3.20 h —Planck constant (J $\cdot\text{s}$).

3.3.21 c —speed of light ($\text{m}\cdot\text{s}^{-1}$).

3.3.22 ΔT —temperature difference ($^{\circ}\text{C}$).

3.3.23 ε —measurement error in short-circuit current (dimensionless).

4. Summary of Test Method

4.1 Spectral mismatch error occurs when a calibrated reference cell is used to measure total irradiance of a test light source (such as a solar simulator) during a photovoltaic device performance measurement, and the incident spectral irradiance of the test light source differs from the reference spectral irradiance distribution to which the reference cell is calibrated.

4.2 The magnitude of the error depends on how the quantum efficiencies of the photovoltaic reference cell and the device to be tested differ from one another; these quantum efficiencies vary with temperature.

4.3 Determination of the spectral mismatch parameter M requires the spectral response characteristics of the photovoltaic device and the spectral irradiance distribution of the test light source, along with the spectral response and the reference spectral irradiance distribution used for the reference cell calibration: six spectral quantities.

4.3.1 The spectral irradiance distribution of the test light source $E_S(\lambda)$.

4.3.2 The reference spectral irradiance distribution to which the photovoltaic reference cell was calibrated $E_0(\lambda)$.

4.3.3 *Photovoltaic Reference Cell:*

4.3.3.1 The quantum efficiency at the temperature corresponding to its calibration constant, $Q_R(\lambda T_0)$

4.3.3.2 The partial derivative of the quantum efficiency with respect to temperature, $\Theta_R(\lambda) = \partial Q_R / \partial T(\lambda)$.

4.3.4 *Device to be Tested:*

4.3.4.1 The quantum efficiency at the temperature at which its performance will be reported, $Q_D(\lambda, T_{D0})$.

4.3.4.2 The derivative of the quantum efficiency with respect to temperature, $\Theta_D(\lambda) = \partial Q_D / \partial T(\lambda)$

4.4 Because all four spectral quantities appear in both the numerator and the denominator in the calculation of the spectral mismatch parameter (see Temperatures of both devices are measured, 8.1), multiplicative calibration errors cancel, and therefore M only relative quantities are calculated using Eq 1 needed, although absolute spectral quantities may be used if available and numerical integration.

5. Significance and Use

5.1 The calculated error in the photovoltaic device current determined from the spectral mismatch parameter can be used to determine if a measurement will be within specified limits before the actual measurement is performed.

5.2 The spectral mismatch parameter also provides a means of correcting the error in the measured device current due to spectral mismatch.

5.2.1 The spectral mismatch parameter is formulated as the fractional error in the short-circuit current due to spectral and temperature differences:

5.2.2 Error due to spectral mismatch can be corrected by dividing the measured photovoltaic cell multiplying a reference cell's measured short-circuit current by M , a procedure technique used in Test Methods E948 and E1036.

5.3 Because all spectral quantities appear in both the numerator and the denominator in the calculation of the spectral mismatch parameter (see 8.1), multiplicative calibration errors cancel, and therefore only relative quantities are needed (although absolute spectral quantities may be used if available).

5.4 Temperature-dependent spectral mismatch is a more accurate method to correct photovoltaic current measurements compared with fixed-value temperature coefficients.³

6. Apparatus

6.1 *Quantum Efficiency Measurement Apparatus*—As required by Test Method E1021 for spectral responsivity measurements.

6.2 *Spectral Irradiance Measurement Equipment*—In addition to the apparatus A spectroradiometer as defined and required by Test Methods Method E1021 G138, the following apparatus is required and calibrated according to Test Method G138.

6.2.1 The wavelength resolution shall be no greater than 10 nm.

6.2.2 It is recommended that the wavelength pass-bandwidth be no greater than 6 nm.

6.2.3 The wavelength range should be wide enough to include the quantum efficiencies of both the photovoltaic device to be tested and the photovoltaic reference cell.

6.2.4 *Spectral Irradiance Measurement Instrument*—A spectroradiometer, defined in Test Method The spectroradiometer must be able to scan the required wavelength range in a time G138, calibrated according to Test Method period short enough such that the G138: spectral irradiance at any wavelength does not vary more than $\pm 5\%$ during the entire scan.

³The last approved version of this historical standard is referenced on www.astm.org.

⁴Seaman, C., "Calibration of Solar Cells by the Reference Cell Method—The Spectral Mismatch Problem," *Solar Energy*, Vol 29, 1982, pp. 291–298.

⁵Ostervald, C. R., "Translation of Device Performance Measurements to Reference Conditions," Campanelli, M., Moriarty, T., Emery, K. A., and Williams, R., "Temperature-Dependent Spectral Mismatch Corrections," *Solar Cells, IEEE Journal of Photovoltaics*, Vol 48, 1986, pp. 269–279-5, No. 6, November 2015, pp. 1692–1697. DOI:10.1109/JPHOTOV.2015.2459914

- 6.1.1.1 The wavelength resolution shall be no greater than 10 nm.
- 6.1.1.2 The wavelength pass-bandwidth shall be no greater than 6 nm.
- 6.1.1.3 The wavelength range shall be wide enough to include the spectral response of the photovoltaic device and the photovoltaic reference cell.
- 6.1.1.4 The spectral irradiance measurement instrument must be able to scan the required wavelength range in a time period short enough such that the spectral irradiance at any wavelength does not vary more than $\pm 5\%$ during the entire scan.
- 6.2.5 Test Methods [E948](#), [E1036](#), and [E1125](#) provide additional guidance for spectral irradiance measurements.
- 6.3 *Temperature Measurement Equipment*—As required by Test Method [E948](#) or Test Methods [E1036](#).

7. Procedure

7.1 ~~Determine~~Obtain the reference spectral response irradiance distribution, $RE_{t0}(\lambda)$ of the photovoltaic device using Test Methods ~~(λ)~~, to which the photovoltaic reference cell is calibrated, such as Tables ~~E1021~~[E490](#) or [G173](#).

7.2 Obtain the spectral response quantum efficiency ~~Rof the~~ $r(\lambda)$ of the photovoltaic reference cell. photovoltaic reference cell at its calibration temperature, $Q_R(\lambda, T_{R0})$.

7.2.1 An expression that converts spectral responsivity to quantum efficiency is provided in Test Methods [E1021](#).

NOTE 1—Test Methods ~~E1039~~, [E1125](#), and [E1362](#) require the spectral response responsivity to be provided as part of the reference cell calibration certificate.

7.3 Obtain the partial derivative of quantum efficiency with respect to temperature, $\Theta_R(\lambda)$, for the photovoltaic reference cell (see [8.1](#)).

7.3.1 If $\Theta_R(\lambda)$ is not provided with the calibration certificate of the photovoltaic reference cell, the derivatiave function must be calculated from a series of quantum efficiency measurements at several temperatures. An acceptable procedure is given in [Annex A1](#).

7.4 Measure the quantum efficiency of the photovoltaic device to be tested at the temperature to which its performance will be reported, $Q_D(\lambda, T_{D0})$, and its partial derivative of quantum efficiency with respect to temperature, $\Theta_D(\lambda)$, using the procedure given in [Annex A1](#) (see also [8.1](#)).

7.5 Measure the spectral irradiance, $E_S(\lambda)(\lambda)$, of the test light source, using the spectral irradiance measurement instrument equipment (see [6.1.4](#) [6.2.1](#)).

7.6 ~~Obtain the reference spectral irradiance distribution~~ Measure the temperature $E_{\sigma}(\lambda)$ that corresponds to the calibration of the photovoltaic reference cell, *such* T_{R} , as Tables using the [E490](#) or temperature measurement [G173](#) equipment.

7.7 Measure the temperature of the photovoltaic device to be tested, T_D , using the temperature measurement equipment.

8. Calculation of Results

8.1 Calculate the spectral mismatch parameter with:³

$$M = \frac{\int_{\lambda_1}^{\lambda_2} E(\lambda) R_i(\lambda) d\lambda}{\int_{\lambda_3}^{\lambda_4} E(\lambda) R_r(\lambda) d\lambda} \times \frac{\int_{\lambda_3}^{\lambda_4} E_o(\lambda) R_r(\lambda) d\lambda}{\int_{\lambda_1}^{\lambda_2} E_o(\lambda) R_i(\lambda) d\lambda} \quad (1)$$

$$M = \frac{\int_{\lambda_1}^{\lambda_2} \lambda Q_D(\lambda, T_{D0}) E_S(\lambda) d\lambda + \Delta T_D \int_{\lambda_1}^{\lambda_2} \lambda \Theta_D(\lambda) E_S(\lambda) d\lambda}{\int_{\lambda_3}^{\lambda_4} \lambda Q_R(\lambda, T_{R0}) E_S(\lambda) d\lambda + \Delta T_R \int_{\lambda_3}^{\lambda_4} \lambda \Theta_R(\lambda) E_S(\lambda) d\lambda} \times \frac{\int_{\lambda_3}^{\lambda_4} \lambda Q_R(\lambda, T_{R0}) E_o(\lambda) d\lambda}{\int_{\lambda_1}^{\lambda_2} \lambda Q_D(\lambda, T_{D0}) E_o(\lambda) d\lambda}, \quad (1)$$

using ΔT_R suitable $= T_R - T_{R0}$ and $\Delta T_D = T_D - T_{D0}$. Use an appropriate numerical integration scheme such as those that described in Tables [G173](#). [Appendix X1](#) provides the derivation of [Eq 1](#). If $|\Delta T_R| \leq 0.5^\circ\text{C}$ and $|\Delta T_D| \leq 0.5^\circ\text{C}$, then $\Theta_R(\lambda)$ and $\Theta_D(\lambda)$ may be omitted and [Eq 1](#) simplified to:

$$M = \frac{\int_{\lambda_1}^{\lambda_2} \lambda Q_D(\lambda, T_{D0}) E_S(\lambda) d\lambda}{\int_{\lambda_3}^{\lambda_4} \lambda Q_R(\lambda, T_{R0}) E_S(\lambda) d\lambda} \times \frac{\int_{\lambda_3}^{\lambda_4} \lambda Q_R(\lambda, T_{R0}) E_o(\lambda) d\lambda}{\int_{\lambda_1}^{\lambda_2} \lambda Q_D(\lambda, T_{D0}) E_o(\lambda) d\lambda}, \quad (2)$$

- 8.1.1 The wavelength integration limits λ_1 and λ_2 shall correspond to the spectral response limits of the photovoltaic device.
- 8.1.2 The wavelength integration limits λ_3 and λ_4 shall correspond to the spectral response limits of the photovoltaic reference cell.

8.2 *Optional*—Calculate the measurement error due to spectral mismatch using:

$$\varepsilon = |M - 1| \quad (3)$$