This document is not an ASTM standard and is intended only to provide the user of an ASTM standard an indication of what changes have been made to the previous version. Because it may not be technically possible to adequately depict all changes accurately, ASTM recommends that users consult prior editions as appropriate. In all cases only the current version of the standard as published by ASTM is to be considered the official document.



Designation: D7962 - 15 D7962 - 15a

Standard Practice for Determination of Minimum Immersion Depth and Assessment of Temperature Sensor Measurement Drift¹

This standard is issued under the fixed designation D7962; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ε) indicates an editorial change since the last revision or reapproval.

1. Scope Scope*

1.1 This practice describes a two procedures for use with temperature measurement devices. Methodology is described for determining minimum immersion depth for thermal sensors, in particular RTDs or similar temperature sensors. Included is a procedure for consistently preparing a reference bath for the purpose of monitoring measurement drift of thermal sensors such as liquid-in-glass or digital contact thermometers.

1.2 This practice focuses on temperature measurement drift in a laboratory. If the user requires greater measurement accuracy, then they should follow the instructions in Practice E563.

1.3 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.4 This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.

2. Referenced Documents

2.1 ASTM Standards:²

E563 Practice for Preparation and Use of an Ice-Point Bath as a Reference Temperature

3. Terminology

3.1 Definitions:

Document Preview

3.1.1 *digital contact thermometer (DCT), n*—an electronic device consisting of a digital display and associated temperature sensing probe.

<u>ASTM D7962-15a</u>

¹ This practice is under the jurisdiction of ASTM Committee D02 on Petroleum Products, Liquid Fuels, and Lubricants and is the direct responsibility of Subcommittee D02.91 on Coordinating Subcommittee on Thermometry.

Current edition approved April 1, 2015 Dec. 1, 2015. Published April 2015 February 2016. Originally approved in 2015. Last previous edition approved in 2015 as <u>D7962 - 15.</u> DOI: 10.1520/D7962-15.10.1520/D7962-15A.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For Annual Book of ASTM Standards volume information, refer to the standard's Document Summary page on the ASTM website.

3.1.1.1 Discussion—

This device consists of a temperature sensor connected to a measuring instrument; this instrument measures the temperaturedependent quantity of the sensor, computes the temperature from the measured quantity, and provides a digital output. This digital output goes to a digital display and/or recording device that may be internal or external to the device. These devices are sometimes referred to as a "digital thermometer".

3.1.1.2 Discussion-

Portable electronic thermometers (PET) is an acronym sometimes used to refer to a subset of the devices covered by this definition. 3.1.2 *ice-point bath*, *n*—physical system containing ice and water assembled to realize the ice point as a reference temperature, or to establish a constant temperature near 0 °C.

3.1.3 *minimum immersion length*, *n*—depth that a thermometer should be immersed, in a uniform temperature environment, such that further immersion does not produce a change in indicated temperature greater than the specified tolerance.

*A Summary of Changes section appears at the end of this standard

Copyright © ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959. United States

4. Summary of Practice

4.1 This practice describes a procedure for consistently preparing an ice bath that is an intimate mixture of crushed ice or ice particles and water in a thermally insulating vessel open to the atmosphere.

4.1.1 Caution—when the ice bath is not made from distilled water, its temperature will differ from the natural fixed-point temperature by a consistent amount, typically less than $0.02 \,^{\circ}$ C. If the user needs a more accurate ice point, then they should use Practice E563 to prepare the ice bath.

4.2 This practice includes a procedure for determining the minimum immersion depth of the temperature sensor using an ice bath.

4.2.1 This procedure determines minimum immersion depth with a 25 °C differential between sensed temperature and ambient temperature. If the probe is subjected to a greater differential temperature, a larger immersion depth may be required to correctly measure the temperature.

5. Significance and Use

5.1 This practice provides a means for the users of ASTM Committee D02 standards to monitor the drift in sensed temperature of liquid-in-glass thermometer (LiG), and digital contact thermometers (DCT). Digital contact thermometers are sometimes referred to as portable electronic thermometers (PET) or simply digital thermometers.

5.2 This practice is not suitable for determining the accuracy or calibration of a temperature measuring device as the error in the ice bath temperature can be greater than 0.02 °C. For greater accuracy, the user should use Practice E563 to prepare the ice bath.

5.3 The ice point is a common practical industrial reference point of thermometry. The ice point is relatively simple to realize and provides a readily available natural fixed-point reference temperature.

5.4 This practice only checks the measurement drift at a single temperature. It will not detect a change in measurement response with change in temperature. Temperature measuring devices should be recalibrated at set intervals. See device supplier for recommendations.

5.5 This practice provides a technique to determine minimum immersion depth of the sensing probe of the thermometer using an ice bath. The minimum immersion depth determined by this practice may change when the differential temperature differs significantly from the conditions described. A greater differential will likely increase the minimum immersion depth.

6. Reagents and Equipment

6.1 *Water*—Good quality drinking water from a consistent source.

6.2 *Ice*—Crushed ice prepared from the water in 6.1. TM D7962-15a

6.3 *Insulated container*—an insulated container large enough to hold the thermometer and provide several centimeters of water-ice slush around the measuring portion of the thermometer.

7. Procedure—Drift in Temperature Measurement

7.1 Preparation of the Ice-Point Bath:

7.1.1 Determine the water source to be used each time this practice is followed in your laboratory to determine measurement drift and note the water source with the measurement. Altering the water source between uses of the practice when monitoring a temperature sensor may result in observing a false drift.

7.1.2 Ensure that all equipment is clean and free of any oily residue. Rinse the equipment with the type of water used for the ice-point bath medium. Use clean, "powder free" laboratory gloves to handle the ice and equipment.

7.1.3 Prepare ice by using the water from the source determined in 7.1.1. Chill a quantity of the water to near 0 $^{\circ}$ C in a flask then shake vigorously to aerate the water before freezing in small cubes or thin sheets.

7.1.3.1 Prepare finely divided ice by shaving or crushing. Shaved ice resembling snow is preferred, but crushed ice is acceptable if the particles are small (approximately 2 mm to 6 mm in diameter) and there is a large distribution in size.

7.1.4 Prepare the bath in a clean, thermally insulated vessel, preferably a wide-mouthed Dewar vacuum flask fitted with an insulating closure such as a stopper. The vessel should be large enough that its size does not affect the water-ice equilibrium temperature and of such diameter and depth that in thermal equilibrium the test objects will not significantly modify the temperature of the bath over the region to which the ice point is to be applied. For usual applications, a diameter of 70 mm and a depth of 300 mm may be adequate.

7.1.5 Alternately, add shaved ice and chilled water to the vessel, using enough water to saturate the ice but not enough to float it. As the vessel fills, compress the ice-water mixture to force out excess water. The objective is to surround each particle of ice with water, filling all voids, but to keep the ice particles as close together as possible. Continue adding ice and water and compressing until the vessel is filled to the required level. Decant or siphon off excess water.

7.1.6 Cover the ice-point bath leaving just enough open space to access the thermometer probe and agitate the mix. Allow the bath and vessel to equilibrate for at least 30 min before using.