This document is not an ASTM standard and is intended only to provide the user of an ASTM standard an indication of what changes have been made to the previous version. Because it may not be technically possible to adequately depict all changes accurately, ASTM recommends that users consult prior editions as appropriate. In all cases only the current version of the standard as published by ASTM is to be considered the official document.



Designation: <del>D7597 – 09<sup>ε2</sup> D7597 – 16</del>

# Standard Test Method for Determination of Diisopropyl Methylphosphonate, Ethyl Hydrogen Dimethylamidophosphate, Ethyl Methylphosphonic Acid, Isopropyl Methylphosphonic Acid, Methylphosphonic Acid and Pinacolyl Methylphosphonic Acid in Water by Liquid Chromatography/Tandem Mass Spectrometry<sup>1</sup>

This standard is issued under the fixed designation D7597; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\varepsilon$ ) indicates an editorial change since the last revision or reapproval.

 $\epsilon^1$  NOTE—This test method was changed editorially in February 2012.  $\epsilon^2$  NOTE—Added research report footnote to Section 16 editorially in June 2013.

#### 1. Scope

1.1 This procedure covers the determination of diisopropyl methylphosphonate (DIMP), ethyl hydrogen dimethylamidophosphate (EHDMAP), ethyl methylphosphonic acid (EMPA), isopropyl methylphosphonic acid (IMPA), methylphosphonic acid (MPA) and pinacolyl methylphosphonic acid (PMPA) (referred to collectively as organophosphonates in this test method) in surface water by direct injection using liquid chromatography (LC) and detected with tandem mass spectrometry (MS/MS) using electrospray ionization (ESI). These analytes are qualitatively and quantitatively determined by this method. This method adheres to single reaction monitoring (SRM) mass spectrometry.

1.2 This test method has been developed by US EPA Region 5 Chicago Regional Laboratory (CRL).

1.3 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.4 The detection verification level (DVL) and reporting range for the organophosphonates are listed in Table 1.

1.4.1 The DVL is required to be at a concentration at least three times below the reporting limit (RL) and have a signal/noise ratio greater than 3:1. Fig. 1 displays the signal/noise ratios at the DVLs for the organophosphonates in the ESI positive mode and Fig. 2 in the ESI negative mode. <u>ASTM D7597-16</u>

1.4.2 The reporting limit is the concentration of the Level 1 calibration standard as shown in Table 2 for the organophosphonates except for MPA in the ESI negative mode which is at Level 2 due to not meeting the DVL criteria at the lower concentration level. The DVL for MPA in the ESI negative mode is at 20  $\mu$ g/L, which forces a raised reporting limit. However, the multi-laboratory validation required a spike of all target analytes at Level 1 concentrations. The mean recovery for MPA in the ESI negative mode at this level was 98.7 % as shown in Table 3. If your instrument's sensitivity can meet the requirements in this test method, MPA may have a 50  $\mu$ g/L reporting limit.

1.5 This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.

#### 2. Referenced Documents

2.1 ASTM Standards:<sup>2</sup>

- D1129 Terminology Relating to Water
- D1193 Specification for Reagent Water

D2777 Practice for Determination of Precision and Bias of Applicable Test Methods of Committee D19 on Water

<sup>&</sup>lt;sup>1</sup> This test method is under the jurisdiction of ASTM Committee D19 on Water and is the direct responsibility of Subcommittee D19.06 on Methods for Analysis for Organic Substances in Water.

Current edition approved Dec. 1, 2009Feb. 1, 2016. Published January 2010May 2016. Originally approved in 2009. Last previous edition approved in 2009 as D7597-09e<sup>1</sup>. DOI: 10.1520/D7597-09E02:10.1520/D7597-16.

<sup>&</sup>lt;sup>2</sup> For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For Annual Book of ASTM Standards volume information, refer to the standard's Document Summary page on the ASTM website.



**TABLE 1** Detection Verification Level and Reporting Range

			-	
Analyte	ESI Mode	DVL (µg/L)	Reporting Range (µg/L)	
Diisopropyl methylphosphonate	Positive	1	5–150	
Ethyl hydrogen dimethylamidophosphate	Negative	0.25	5–150	
Ethyl hydrogen dimethylamidophosphate	Positive	0.25	5–150	
Ethyl methylphosphonic acid	Negative	5	50-1500	
Ethyl methylphosphonic acid	Positive	5	50-1500	
Isopropyl methylphosphonic acid	Negative	10	50-1500	
Isopropyl Methylphosphonic acid	Positive	5	50-1500	
Methylphosphonic acid	Negative	20	100–1500	
Methylphosphonic acid	Positive	10	50-1500	
Pinacolyl methylphosphonic acid	Negative	5	50-1500	
Pinacolyl methylphosphonic acid	Positive	5	50-1500	



FIG. 1 Example ESI Positive Mode SRM Chromatograms Signal/Noise Ratios

D3856 Guide for Management Systems in Laboratories Engaged in Analysis of Water

D3694 Practices for Preparation of Sample Containers and for Preservation of Organic Constituents

D5847 Practice for Writing Quality Control Specifications for Standard Test Methods for Water Analysis

E2554 Practice for Estimating and Monitoring the Uncertainty of Test Results of a Test Method Using Control Chart Techniques 2.2 *Other Documents:*<sup>3</sup>

EPA publicationPublication SW-846 Test Methods for Evaluating Solid Waste, Physical/Chemical Methods

### 3. Terminology

3.1 Definitions:

3.1.1 For definitions of terms used in this standard, refer to Terminology D1129.

3.2 *Definitions: Definitions of Terms Specific to This Standard:* 

3.2.1 *detection verification level (DVL), level, DVL, n*—a concentration that has a signal/noise ratio greater than 3:1 and is at least 3 times below the reporting limit (RL).

<sup>&</sup>lt;sup>3</sup> Available from United States Environmental Protection Agency (EPA), Ariel Rios Bldg., 1200 Pennsylvania Ave., NW, Washington, DC 20460, http://www.epa.gov.

🕅 D7597 – 16



3.2.2 reporting limit (RL), independent reference material, IRM, n—the concentration of the lowest-level calibration standard used for quantification. a material of known purity and concentration obtained either from the National Institute of Standards and Technology (NIST) or other reputable supplier. The IRM shall be obtained from a different lot of material than is used for calibration.

100

175

3.2.3 organophosphonates, n-in this test method, diisopropyl methylphosphonate (DIMP), ethyl hydrogen dimethylamidophosphate (EHDMAP), ethyl methylphosphonic acid (EMPA), isopropyl methylphosphonic acid (IMPA), methylphosphonic acid (MPA) and pinacolyl methylphosphonic acid (PMPA).

3.3 Abbreviations: Acronyms:

3.3.1 CCC, n-Continuing Calibration Check

3.3.2 IC, n-Initial Calibration

3.3.3 LC, n-Liquid Chromatography

3.3.4 LCS/LCSD, n—Laboratory Control Sample/Laboratory Control Sample Duplicate

3.3.5 MDL, n-Method Detection Limit

3.3.6 MeOH, n-Methanol

3.3.7 *mM*, *n*—millimolar,  $1 \times 10^{-3}$  moles/L

3.3.8 MRM, n-Multiple Reaction Monitoring

3.3.9 MS/MSD, n-Matrix Spike/Matrix Spike Duplicate

3.3.10 NA, adj-Not Available



TABLE 3 MULT-LADOTATORY RECOVERY Data III Reagent Wate	TABLE	3 M	lulti-L	.aboratory	Recovery	Data	in	Reagent	Wate
--	-------	-----	---------	------------	----------	------	----	---------	------

Analyte E		Spike Conc.	# Results	# Labs	Bias			Precision			
	ESI				Mean	Min	Max	Overall SD	Pooled	Overall	Pooled
	Mode				Recovery	Recovery	Recovery	(%)	within-lab	RSD (%)	within-lab
		(ppp)			(%)	(%)	(%)		SD (%)		RSD (%)
DIMP	Pos	5	16	4	95.1	65.8	136.0	17.6	19.4	21.3	22.4
DIMP	Pos	10	19	5	98.2	80.0	121.0	6.1	5.5	13.6	13.8
DIMP	Pos	25	26	6	102.9	74.4	128.0	5.6	5.7	14.5	14.1
DIMP	Pos	125	22	5	96.6	80.4	120.0	4.4	4.5	11.0	11.4
DIMP-D14	Pos	25	86	6	102.6	54.8	127.6	9.8	9.5	11.2	10.9
EHDMAP	Neg	5	12	3	57.5	0.0	220.0	31.4	22.3	71.1	123.6
EHDMAP	Neg	10	16	4	47.1	0.0	178.0	13.3	10.8	66.8	142.0
EHDMAP	Neg	25	22	5	84.1	54.0	141.2	6.9	6.6	22.9	27.3
EHDMAP	Neg	125	18	4	87.4	64.2	141.6	5.0	5.1	25.4	29.1
EHDMAP	Pos	5	16	4	77.0	0.0	134.2	7.4	8.0	51.4	66.8
EHDMAP	Pos	10	20	5	70.0	0.0	143.0	7.7	10.0	53.6	76.5
EHDMAP	Pos	25	26	6	89.6	60.0	128.8	5.2	6.2	23.5	26.2
EHDMAP	Pos	125	22	5	87.5	59.0	123.2	5.2	6.0	23.9	27.3
EMPA	Neg	50	16	4	110.1	74.8	170.6	22.4	17.2	25.6	23.3
EMPA	Neg	100	20	5	108.3	87.7	175.0	11.1	8.7	24.8	22.9
EMPA	Neg	250	26	6	104.8	82.0	122.6	6.1	5.5	11.8	11.3
EMPA	Neg	1250	22	5	101.5	87.2	126.4	8.4	8.1	11.2	11.0
EMPA	Pos	50	16	4	95.4	77.6	122.8	12.9	13.3	13.1	13.7
EMPA	Pos	100	20	5	96.0	61.4	132.5	9.5	9.5	15.9	16.6
EMPA	Pos	250	26	6	99.7	70.0	133.2	5.9	5.4	18.2	18.2
EMPA	Pos	1250	21	5	93.9	84.0	108.4	2.7	3.0	7.7	8.2
IMPA	Neg	50	16	4	88.0	56.6	140.4	23.7	26.0	23.5	26.7
IMPA	Neg	100	20	5	88.0	68.5	118.0	12.9	14.3	13.4	15.2
IMPA	Neg	250	26	6	98.1	72.8	144.0	13.1	11.9	19.2	19.6
IMPA	Neg	1250	21	5	90.7	73.1	103.0	5.5	6.1	8.7	9.6
IMPA	Pos	50	16	4	98.3	47.8	139.6	19.2	20.5	27.2	27.7
IMPA	Pos	100	19	5	95.4	72.3	120.5	9.8	10.3	12.4	13.0
IMPA	Pos	250	26	6	97.0	79.2	188.4	7.4	7.6	10.9	11.2
IMPA	Pos	1250	21	5	91.3	70.4	115.5	5.1	5.2	11.4	12.5
MPA	Neg	50	16	4	98.7	3.3	175.0	14.2	25.3	60.5	61.3
MPA	Neg	100	20	5	100.0	41.9	142.0	8.9	9.2	30.8	30.8
MPA	Neg	250	26	6	99.5	66.0	124.5	7.6	7.6	14.5	14.5
MPA	Neg	1250	22	5	102.7	81.8	130.5	10.5	9.9	12.4	12.1
MPA	Pos	50	16	4	68.3	9.8	139.6	13.4	20.3	36.6	53.6
MPA	Pos	100	20	5	80.5	48.4	149.7	14.0	12.6	26.8	33.3
MPA	Pos	250	26	6	91.7	33.9	153.7	8.0	7.8	31.8	34.7
MPA	Pos	1250	22	5	95.8	31.8	208.2	12.6	8.3	43.4	45.3
MPA-D3	Neg	250	84	6	111.2	57.2	190.8	16.2	12.5	30.0	26.9
MPA-D3	Pos	250	68	5	104.4	58.4	151.8	14.3	14.0	18.0	17.3
PMPA	Neg	50	15	4	87.8 <u>AS</u>	77.6 D/5	124.4	8.4	8.2	13.8	15.7
PMPA	Neg	100	19	5	91.6	83.6	98.2	2.8	3.0 01-11-64	4.4	4.8 7 1 6
PMPA PMPA	Neg	250	26	6	101.0	77.2	123.8	5.0020-10	4.8 4.8	12.1	12.0
PMPA	Neg	1250	22	5	99.2	84.8	126.5	4.9	4.8	12.3	12.4
PMPA	Pos	50	16	4	90.8	60.8	148.8	15.8	16.1	25.6	28.2
PMPA	Pos	100	15	4	95.2	86.8	114.0	3.8	4.0	7.9	8.3
PMPA	Pos	250	20	5	103.8	85.2	136.1	4.5	3.8	15.3	14.7
PMPA	Pos	1250	12	3	99.8	88.8	117.5	5.2	5.0	7.5	7.5
PMPA-13C6	Neg	250	83	6	99.5	74.8	128.3	9.4	9.2	11.1	11.1

- 3.3.11 ND-<u>ND, n</u>-non-detect
- 3.3.12 P&A, n-Precision and Accuracy

3.3.13 PPB, n-parts per billion

- 3.3.14 PPT, n-parts per trillion
- 3.3.15 QA, adj-Quality Assurance
- 3.3.16 QC, adj-Quality Control
- 3.3.17 RL, n-Reporting Limit
- 3.3.18 RSD, n-Relative Standard Deviation
- 3.3.19 RT, n-Retention Time
- 3.3.20 SDS, n-Safety Data Sheets
- 3.3.21 SRM, n-Single Reaction Monitoring
- 3.3.22 SS, n—Surrogate Standard
- 3.3.23 TC, n-Target Compound



3.3.24  $\mu M$ , *n*—micromolar, 1 × 10<sup>-6</sup> moles/L

3.3.25 VOA, n-Volatile Organic Analysis

## 4. Summary of Test Methods

4.1 This is a performance-based method and modifications are allowed to improve performance.

4.2 For organophosphonate analysis, samples are shipped to the lab between 0°C and 6°C and analyzed within 1 day of eollection. In the lab, the samples are spiked with surrogate, filtered using a syringe-driven Millex HV PVDF filter unit and analyzed directly by LC/MS/MS.

4.3 The organophosphonates and the surrogates; diisopropyl methylphosphonate- $D_{14}$ , pinacolyl methylphosphonic acid- $1^{13}C_6$ and methylphosphonic acid- $D_3$  are identified by retention time and one SRM transition. The target analytes and surrogates are quantitated using the SRM transitions utilizing an external calibration. The final report issued for each sample lists the concentration of each organophosphonate target compound and each surrogate recovery.

## 4. Summary of Test Method

4.1 This is a performance-based method and modifications are allowed to improve performance.

4.2 For organophosphonate analysis, samples are shipped to the lab between 0°C and 6°C and analyzed within 1 day of collection. In the lab, the samples are spiked with surrogate, filtered using a syringe-driven filter unit and analyzed directly by LC/MS/MS.

<u>4.3 The organophosphonates and the surrogates; diisopropyl methylphosphonate- $D_{14}$ , pinacolyl methylphosphonic acid- ${}^{13}C_6$ and methylphosphonic acid- $D_3$  are identified by retention time and one SRM transition. The target analytes and surrogates are quantitated using the SRM transitions utilizing an external calibration. The final report issued for each sample lists the concentration of each organophosphonate target compound and each surrogate recovery.</u>

#### 5. Significance and Use

5.1 Organophosphate pesticides affect the nervous system by disrupting the enzyme that regulates acetylcholine, a neurotransmitter. They were developed during the early 19th century, but their effects on insects, which are similar to their effects on humans, were discovered in 1932. Some are poisonous and were used as chemical weapon agents. Organophosphate pesticides are usually not persistent in the environment.<sup>4,5</sup>

5.2 This test method is for the analysis of selected organophosphorous-based chemical weapon agent degradation products from Sarin (GB), Soman (GD), Tabun (GA) and VX. This method has been investigated for use with reagent and surface water.

#### 6. Interferences

# ASTM D7597-16

6.1 Method interferences may be caused by contaminants in solvents, reagents, glassware and other apparatus producing discrete artifacts or elevated baselines. All of these materials are demonstrated to be free from interferences by analyzing laboratory reagent blanks under the same conditions as samples.

6.2 All glassware is washed in hot water with a detergent, rinsed in hot water followed by distilled water. The glassware is then dried and heated in an oven at 250°C for 15 to 30 minutes. All glassware is subsequently cleaned with acetone, then methanol.

6.3 All reagents and solvents should be pesticide residue purity or higher to minimize interference problems.

6.4 Matrix interferences may be caused by contaminants that are co-extracted from the sample. The extent of matrix interferences can vary considerably from sample source depending on variations of the sample matrix.

# 7. Apparatus

7.1 LC/MS/MS <del>System</del>System:

7.1.1 *Liquid Chromatography (LC) System*—A complete LC system is needed in order to analyze samples.<sup>6</sup> <u>This should include</u> a sample injection system, a solvent pumping system capable of mixing solvents, a sample compartment capable of maintaining required temperature and a temperature controlled column compartment. A system that is capable of performing at the flows, pressures, controlled temperatures, sample volumes and requirements of the standard may be used.

7.1.2 Analytical <u>Column-Waters—Column<sup>7</sup>—Atlantis dC18, 150 mm × 2.1 mm, 3  $\mu$ m particle size, or equivalent. A reverse phase C18 particle column was used to develop this test method. Any column that achieves adequate resolution may be used. The retention times and order of elution may change depending on the column that is used and need to be monitored.</u>

<sup>&</sup>lt;sup>4</sup> Additional information about organophosphate pesticides is available on the Internet at http://www.epa.gov (2009).

<sup>&</sup>lt;sup>5</sup> Additional information about chemical weapon agents is available on the Internet at http://www.opcw.org (2009).

<sup>&</sup>lt;sup>6</sup> A Waters Alliance (a trademark of the Waters Corporation, Milford, MA) High Performance Liquid Chromatography (HPLC) System was used to develop this test method. System, or equivalent, was found suitable for use. The multi-laboratory study included Agilent and Waters LC systems.

 $<sup>^{7}</sup>$  A Waters Atlantis (a trademark of the Waters Corporation, Milford, MA) dC18, 150 mm × 2.1 mm, 3  $\mu$ m particle size, or equivalent, has been found suitable for use.



7.1.3 Tandem Mass Spectrometer (MS/MS) System—A MS/MS system capable of MRM analysis.<sup>8</sup> A system that is capable of performing at the requirements in this standard may be used.

#### 7.2 Filtration DeviceDevice:

7.2.1 Hypodermic syringe—Syringe—A lockluer-lock tip glass syringe capable of holding a Millex HV Syringe Driven Filter Unit PVDF 0.45 µm (Millipore Corporation, Catalog # SLHV033NS) or similar may be used syringe-driven filter unit.

7.2.1.1 A 25-mL lock tip glass syringe size is recommended since a 25-mL sample size is used in this test method.

7.2.2 Filter-Filter Unit9-Millex HV Syringe Driven Filter Unit PVDF 0.45 µm (Millipore Corporation, Catalog # SLHV033NS) or similar may be used. A PVDF filter units were used to filter the samples.

#### 8. Reagents and Materials

8.1 Purity of Reagents—High-performance liquid chromatography (HPLC) pesticide residue analysis and spectrophotometry grade chemicals shall be used in all tests. Unless indicated otherwise, it is intended that all reagents shall conform to the Committee on Analytical Reagents of the American Chemical Society.<sup>10</sup> Other reagent grades may be used provided they are first determined they are of sufficiently high purity to permit their use without affecting the accuracy of the measurements.

8.2 Purity of Water—Unless otherwise indicated, references to water shall be understood to mean reagent water conforming to Type 1 of Specification D1193. It must be demonstrated that this water does not contain contaminants at concentrations sufficient to interfere with the analysis.

8.3 Gases—Ultrapure nitrogen and argon.

- 8.4 Acetonitrile (CAS # 75-05-8).
- 8.5 Methanol (CAS # 67-56-1).
- 8.6 Acetone (CAS # 67-64-1).

8.7 Formic acid (<del>≥95%,</del> (≥95 %, CAS # 64-18-6).

8.8 Diisopropyl methylphosphonate (DIMP, CAS # 1445-75-6).

8.9 Ethyl hydrogen dimethylamidophosphate (EHDMAP, CAS # 2632-86-2).

8.10 Ethyl methylphosphonic acid (EMPA, CAS # 1832-53-7).

8.11 Isopropyl methylphosphonic acid (IMPA, CAS # 1832-54-8).

8.12 Methylphosphonic acid (MPA, CAS # 993-13-5).

8.13 Pinacolyl methylphosphonic acid (PMPA, CAS # 616-52-4).

8.14 Diisopropyl methylphosphonate-D<sub>14</sub> (DIMP-D14, Unlabeled CAS # 1445-75-6).

8.14.1 DIMP-D14 represents deuterium labeled diisopropyl methylphosphonate where the two isopropyl moieties contain all 2H.

8.15 Methylphosphonic acid-D<sub>3</sub> (MPA-D3, Unlabeled CAS # 993-13-5).

8.15.1 MPA-D3 represents deuterium labeled methylphosphonic acid where the methyl moiety contains all <sup>2</sup>H.

8.16 Pinacolyl methylphosphonic acid- ${}^{13}C_6$  (PMPA-13C6, Unlabeled CAS # 616-52-4). 8.16.1 PMPA-13C6 represents  ${}^{13}C$  labeled pinacolyl methylphosphonic where all the trimethylpropyl carbon atoms are uniformly labeled <sup>13</sup>C.

## 9. Hazards

9.1 Normal laboratory safety applies to this method. Analysts should wear safety glasses, gloves, and lab coats when working in the lab. Analysts should review the Material-Safety Data Sheets (MSDS) (SDS) for all reagents used in this method.

## **10.** Sampling

10.1 Sampling—Grab samples must be collected in ≥25-mL pre-cleaned amber glass bottles with Teflon-lined caps demonstrated to be free of interferences. This test method requires a 25-mL sample size per analysis. Conventional sampling practices should be followed. Refer to Guide D3856 and Practices D3694.

<sup>&</sup>lt;sup>8</sup> A Waters Quattro (a trademark of the Waters Corporation, Milford, MA) micro API mass spectrometer was used to develop this test method.spectrometer, or equivalent, was found suitable for use. The multi-laboratory study included Applied Biosystems and Waters mass spectrometers.

PA Millex (a trademark of Merck KGAA, Darmstadt, Germany) HV Syringe Driven Filter Unit PVDF 0.45 μm (Millipore Corporation, Catalog # SLHV033NS) has been found suitable for use for this method, any filter unit may be used that meets the performance of this method may be used.

<sup>&</sup>lt;sup>10</sup> Reagent Chemicals, American Chemical Society Specifications, American Chemical Society, Washington, DC. For Suggestions on the testing of reagents not listed by the American Chemical Society, see Annual Standards for Laboratory Chemicals, BDH Ltd., Poole, Dorset, U.K., and the United States Pharmacopeia and National Formulary, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.