



Designation: **F2605–08^{ε1} F2605 – 16**

Standard Test Method for Determining the Molar Mass of Sodium Alginate by Size Exclusion Chromatography with Multi-angle Light Scattering Detection (SEC-MALS)¹

This standard is issued under the fixed designation F2605; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reappraisal. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reappraisal.

^{ε1} NOTE—Subsection 6.1.5 was editorially corrected in September 2008.

1. Scope

1.1 This test method covers the determination of the molar mass (typically expressed as grams/mole) of sodium alginate intended for use in biomedical and pharmaceutical applications as well as in ~~tissue-engineered-tissue-engineered~~ medical products (TEMPs) by size exclusion chromatography with multi-angle laser light scattering detection (SEC-MALS). A guide for the characterization of alginate has been published as Guide F2064.

1.2 Alginate used in TEMPs should be well characterized, including the molar mass and polydispersity (molar mass distribution) in order to ensure uniformity and correct functionality in the final product. This test method will assist end users in choosing the correct alginate for their particular application. Alginate may have utility as a scaffold or matrix material for TEMPs, in cell and tissue encapsulation applications, and in drug delivery formulations.

1.3 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 *ASTM Standards:*²

F2064 Guide for Characterization and Testing of Alginates as Starting Materials Intended for Use in Biomedical and Tissue Engineered Medical Product Applications

F2315 Guide for Immobilization or Encapsulation of Living Cells or Tissue in Alginate Gels

2.2 *United States Pharmacopeia/National Formulary:*³

<621> Chromatography

2.3 *National Institute of Standards and Technology:*⁴

NIST SP811 Special Publication: Guide for the Use of the International System of Units

2.4 *ISO Standards:*⁵

ISO 31-8 Quantities and units- Part 8: Physical chemistry and molecular physics

3. Terminology

3.1 *Definitions:*

¹ This test method is under the jurisdiction of ASTM Committee F04 on Medical and Surgical Materials and Devices and is the direct responsibility of Subcommittee F04.42 on Biomaterials and Biomolecules for TEMPs.

Current edition approved Feb. 1, 2008; May 1, 2016. Published May 2008; June 2016. Originally approved in 2008. Last previous edition approved in 2008 as F2605 – 08^{ε1}. DOI: 10.1520/F2605-08E01; 10.1520/F2605-16.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ Available from United States Pharmacopeia and National Formulary, U.S. Pharmaceutical Convention, Inc. (USPC), Rockville, MD.

⁴ Available from National Institute of Standards and Technology (NIST), 100 Bureau Dr., Stop 1070, Gaithersburg, MD 20899-1070, <http://physics.nist.gov/cuu/Units/bibliography.html>.

⁵ Available from International Organization for Standardization (ISO), ISO Central Secretariat, BIBC II, Chemin de Blandonnet 8, CP 401, 1214 Vernier, Geneva, Switzerland, <http://www.iso.org>.

3.1.1 *alginate, n*—a polysaccharide substance extracted from brown algae, mainly occurring in the cell walls and intercellular spaces of brown seaweed and kelp. Its main function is to contribute to the strength and flexibility of the seaweed plant. Sodium alginate, and in particular calcium cross-linked alginate gels are used in ~~tissue-engineered~~ tissue-engineered medical products (TEMPs) as biomedical scaffolds and matrices, for immobilizing living cells (see Guide F2315), and in drug delivery systems.

3.1.2 *molar mass average, n*—the given molar mass (M_w) of an alginate will always represent an average of all of the molecules in the population. The most common ways to express the molar mass are as the *number average* (M_n) and the *mass average* (M_w). The two averages are defined by the following equations:

$$\bar{M}_n = \frac{\sum_i N_i M_i}{\sum_i N_i} \quad \text{and} \quad \bar{M}_w = \frac{\sum_i w_i M_i}{\sum_i w_i} = \frac{\sum_i N_i M_i^2}{\sum_i N_i M_i} \quad (1)$$

where:

N_i = number of molecules having a specific molar mass M_i , and

w_i = mass of molecules having a specific molar mass M_i .

3.1.2.1 Discussion—

In a polydisperse molecular population the relation $M_w^- > M_n^-$ is always valid. The coefficient M_w^- / M_n^- is referred to as the polydispersity index, and will typically be in the range 1.5 to 3.0 for commercial alginates.

NOTE 1—The term molecular weight (abbreviated MW) is obsolete and should be replaced by the SI (Système Internationale) equivalent of either relative molecular mass (M_r), which reflects the dimensionless ratio of the mass of a single molecule to an atomic mass unit (see ISO 31-8), or molar mass (M), which refers to the mass of a mole of a substance and is typically expressed as grams/mole. For polymers and other macromolecules, use of the symbols M_w , M_n , and M_z continue, referring to mass-average molar mass, number-average molar mass, and z-average molar mass, respectively. For more information regarding proper utilization of SI units, see NIST SP811.

4. Significance and Use

4.1 The composition and sequential structure of alginate, as well as the molar mass and molar mass distribution, determines the functionality of alginate in an application. For instance, the gelling properties of an alginate are highly dependent upon the composition and molar mass of the polymer.

4.2 Light scattering is one of very few methods available for the determination of absolute molar mass and structure, and it is applicable over the broadest range of molar masses of any method. Combining light scattering detection with size exclusion chromatography (SEC), which sorts molecules according to size, gives the ability to analyze polydisperse samples, as well as ~~obtaining to obtain~~ information on branching and molecular conformation. This means that both the number-average and mass-average values for molar mass and size may be obtained for most samples. Furthermore, one has the ability to calculate the distributions of the molar masses and sizes.

4.3 Multi-angle laser light scattering (MALS) is a technique where measurements are made simultaneously over a range of different ~~angles~~ angles and used to determine the scattering at 0° , which directly relates to molecular weight. MALS detection can be used to obtain information on molecular size, since this parameter is determined by the angular variation of the scattered light. ~~Molar mass may in principle~~ This can be related to branching, aggregation, and molecular conformation. Molar mass can also be determined by detecting scattered light at a single low angle (LALLS). ~~However, advantages with MALS as compared to LALLS are: (LALS) (1) less noise at larger angles, (2) the precision of measurements are greatly improved by detecting at several angles; and (3) assuming that the ability to detect angular variation allows determination of size, branching, aggregation, and molecular conformation; this is not significantly different from the scattering at 0° .~~

4.4 Size exclusion chromatography uses columns, which are typically packed with polymer particles containing a network of uniform pores into which solute and solvent molecules can diffuse. While in the pores, molecules are effectively trapped and removed from the flow of the mobile phase. The average residence time in the pores depends upon the size of the solute molecules. Molecules that are larger than the average pore size of the packing are excluded and experience virtually no retention; these are eluted first, in the void volume of the column. ~~Molecules; Molecules which may penetrate the pores will have a larger volume available for diffusion; they will suffer retention depending on diffusion; their retention will depend on their molecular size, with the smaller molecules eluting last.~~

4.5 For polyelectrolytes, dialysis against the elution buffer has been suggested, in order to eliminate Donnan-type artifacts in the molar mass determination by light scattering (1, 2).⁶ However, in the present method, the size exclusion chromatography step preceding the light scatter detection is an efficient substitute for a dialysis step. The sample is separated on SEC columns with large excess of elution buffer for 30 to 40 min, and it is therefore in full equilibrium with the elution buffer when it reaches the MALS detector.

⁶ The boldface numbers in parentheses refer to a list of references at the end of this standard.

5. Materials

5.1 Chemicals:

- 5.1.1 Alginate sample.
- 5.1.2 Deionized water (Milli-Q Plus or equivalent; conductivity < 10 $\mu\text{S}/\text{cm}$).
- 5.1.3 Na_2SO_4 (sodium sulfate).
- 5.1.4 EDTA (ethylene diamine tetraacetic acid).
- 5.1.5 NaOH (1 mol/L).
- 5.1.6 Pullulan standards. See **Note 2**.

NOTE 2—A series of linear homopolysaccharides with sufficiently narrow dispersity to be suitable for utilization as molar mass calibration standards in aqueous eluent.

5.2 The Mobile Phase:

5.2.1 For SEC-MALS of alginate, a mobile phase of 0.05 mol/L Na_2SO_4 /0.01 mol/L EDTA in deionized water is used. Adjust pH to 6.0 using 1 mol/L NaOH.

5.2.2 Mobile phase should be prepared as a stock solution of 0.10 mol/L Na_2SO_4 /0.02 mol/L EDTA in deionized water, which can be stored cool (3 to 8°C) for 6 months. Before use as a mobile phase, the stock solution is diluted 1:1 (v/v) with deionized water and passed through a 0.22 μm filter.

5.3 Instruments:

- 5.3.1 Analytical balance (0.1 mg).
- 5.3.2 Shaking device.
- 5.3.3 pH meter.
- 5.3.4 HPLC—High Performance Liquid Chromatography (HPLC) system with injector, pump, degassing unit.
- 5.3.5 Size exclusion columns: TSK-Gel PW_{XL} columns from Tosoh Biosep., for example, PW_{XL} -guard column + G6000 PW_{XL} + G5000 PW_{XL} + G3000 PW_{XL} (last in the series), or equivalent.
- 5.3.6 Refractive Index (RI) detector, with a known calibration constant (dn/dV).
- 5.3.7 Multiple Angle Laser Light Scattering (MALS) detector, with known calibration constant.
- 5.3.8 Computer with suitable software.

6. Procedure

6.1 Preparation of Standards and Alginate Samples for SEC-MALS:

- 6.1.1 Samples are prepared at a concentration suitable for injection of 200 μL of sample.
- 6.1.2 Dissolve all samples in deionized water at twice the required concentration for molar mass determination by shaking at about 100 min^{-1} overnight at cool temperature (3 to 8°C).
- 6.1.3 Dilute samples 1+1 with stock solution of mobile phase and shake gently for a few seconds.
- 6.1.4 Pass all samples through a 0.45 μm filter, and transfer to HPLC vials.
- 6.1.5 Final concentration of pullulan standards of known M_w^- values of approximately 11 800 to 47 300, 112 000, 212 000, and 404 000 g/mol should be approximately 4, 3, 2, and 1.5 mg/mL, respectively.
- 6.1.6 Guidelines for the final concentration of alginates for molar mass determination are given in **Table 1**. If SEC-MALS data display poor reproducibility with respect to replicates, this might be an indication of column overload. In this case, less sample should be injected.


6.2 Chromatography and Data Collection:

- 6.2.1 The complete experimental setup of the SEC-MALS system is shown in **Fig. 1**. The refractive index detector is placed at the end of the solvent/sample line as it is highly sensitive to pressure changes.
- 6.2.2 Pullulan standards should be injected and analyzed with 2 replicates before and after all alginate samples (total of 4 replicates). 3 replicates should be injected for alginates.
- 6.2.3 A procedure for setting up the chromatography run and collecting the data is given below:

TABLE 1 Suggestions for Concentration and Injected Mass of Alginate Samples for SEC-MALS

M_w^- (g/mol)	Apparent Viscosity (mPas)	Concentration for Injection (mg/mL)	Injected Mass ^A (mg)
<50 000	<10	4	0.8
50 000–75 000	10–20	3	0.6
75 000–100 000	20–40	2	0.4
100 000–150 000	40–100	1.5	0.3
150 000–250 000	100–300	1	0.2
>250 000	>300	0.5	0.1

^A Injected mass = Concentration*200 μL .



NOTE 1—Solid lines indicate solvent/sample flow, dashed lines indicate cabling for data transfer.

FIG. 1 Complete SEC-MALS Set-Up

iTeh Standards

(<https://standards.itih.ai>)

6.2.3.1 Use a flow rate of 0.5 mL/min.

6.2.3.2 Purge the injector with mobile phase before the sample set is run.

6.2.3.3 Purge the RI-detector refractive index (RI)-detector for at least 30 min (at 0.5 mL/min) before the start of the run.

6.2.3.4 Confirm that both the MALS detector and RI detector has a have stable and low baseline level-levels.

6.2.3.5 Define the collection set-up as follows:

(1) Inject 200 μ L of sample.

(2) After a collection delay of 10 mL (20 min), data should be collected from both detectors every 2 seconds for 40 mL (80 min).

(3) Use $dn/dc = 0.148$ mL/g and 0.150 mL/g for pullulans and alginates, respectively (relevant only for calculations).

(4) Use a second virial coefficient of 2×10^{-4} mol.mL.g⁻² and 5×10^{-3} mol.mL.g⁻² for pullulans and alginates, respectively (relevant only for calculations).

6.2.4 After all samples have been run, purge the injector with deionized water to wash off remaining salt from the valves.

6.3 Data Analysis:

6.3.1 Data analysis follows closely recommended procedures for SEC-MALS data. Generally, the chromatograms are divided into a number of volume elements, defined by the peak width, the rate of data collection and the flow rate. ~~Concentration-~~The concentration of sample in each volume element (c_i) is determined from the RI-detector response using known values of dn/dc and dn/dV (the RI-detector calibration constant). Furthermore, ~~LS-detector~~MALS-detector response is divided by c_i , the molar mass in each volume element (M_i) is considered monodisperse, and the mass is determined from a Zimm representation of a Debye plot by extrapolation to zero angle (which is essentially a solution to Eq X2.1 in X2.2). Once the values of c_i and M_i are known, calculation of the various average molar masses is straightforward.

6.3.2 In detail, the above procedure consists of the following operations to be performed in-a-using suitable software:

6.3.2.1 Define baselines for signals from both detectors.

6.3.2.2 Calculate inter-detector delay volume using a monodisperse low-molar mass pullulan standard.

6.3.2.3 Define the peak area of interest.

6.3.2.4 Normalize LS-detector responses to correct for different sensitivity at different angles. Normalization is performed on an isotropic scatterer (low molar mass compound) in the sample set, and is saved with the data file. For the other samples, one reads the normalization performed on an isotropic scatterer from file.

6.3.2.5 Check the goodness-of-fit of the LS-detectors using a 3D-representation of the data or a Debye-plot (in Zimm representation). Do not use LS-detector responses that are clearly non-linear.

6.3.2.6 Perform the required calculations for determination of M_n^- , M_w^- and M_w^-/M_n^- , using a Zimm representation of the Debye plot (that is, a plot of $K \cdot c/R(\theta)$ versus $\sin^2(\theta/2)$) for solving Eq X2.1.