

TECHNICAL REPORT

IEC TR 61010-3-032

First edition
2000-01

**Safety requirements for electrical equipment
for measurement, control, and laboratory use –**

Part 3-032:

**Conformity verification report for IEC 61010-2-032:1994,
Particular requirements for hand-held current clamps
for electrical measurement and test**

*Règles de sécurité pour appareils électriques de mesure,
de régulation et de laboratoire –*

Partie 3-032:

Rapports de vérification de la conformité de la CEI 61010-2-032



Reference number
IEC/TR 61010-3-032:2000(E)

Numbering

As from 1 January 1997 all IEC publications are issued with a designation in the 60000 series.

Consolidated publications

Consolidated versions of some IEC publications including amendments are available. For example, edition numbers 1.0, 1.1 and 1.2 refer, respectively, to the base publication, the base publication incorporating amendment 1 and the base publication incorporating amendments 1 and 2.

Validity of this publication

The technical content of IEC publications is kept under constant review by the IEC, thus ensuring that the content reflects current technology.

Information relating to the date of the reconfirmation of the publication is available in the IEC catalogue.

Information on the subjects under consideration and work in progress undertaken by the technical committee which has prepared this publication, as well as the list of publications issued, is to be found at the following IEC sources:

- **IEC web site***
- **Catalogue of IEC publications**
Published yearly with regular updates
(On-line catalogue)*
- **IEC Bulletin**
Available both at the IEC web site* and as a printed periodical

Terminology, graphical and letter symbols

For general terminology, readers are referred to IEC 60050: *International Electrotechnical Vocabulary* (IEV).

For graphical symbols, and letter symbols and signs approved by the IEC for general use, readers are referred to publications IEC 60027: *Letter symbols to be used in electrical technology*, IEC 60417: *Graphical symbols for use on equipment. Index, survey and compilation of the single sheets* and IEC 60617: *Graphical symbols for diagrams*.

* See web site address on title page.

TECHNICAL REPORT

IEC TR 61010-3-032

First edition
2000-01

Safety requirements for electrical equipment for measurement, control, and laboratory use –

Part 3-032:

Conformity verification report for IEC 61010-2-032:1994, Particular requirements for hand-held current clamps for electrical measurement and test

*Règles de sécurité pour appareils électriques de mesure,
de régulation et de laboratoire –*

Partie 3-032:

Rapports de vérification de la conformité de la CEI 61010-2-032

© IEC 2000 — Copyright - all rights reserved

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Electrotechnical Commission
Telefax: +41 22 919 0300

3, rue de Varembé Geneva, Switzerland
e-mail: inmail@iec.ch IEC web site <http://www.iec.ch>



Commission Electrotechnique Internationale
International Electrotechnical Commission
Международная Электротехническая Комиссия

PRICE CODE

X

For price, see current catalogue

CONTENTS

	Page
FOREWORD	3
Conformity verification report IEC 61010-3-032	5
Clause	
5 Marking and documentation.....	10
6 Protection against electric shock	13
7 Protection against mechanical hazards	17
8 Mechanical resistance to shock, vibration and impact (AM 1 only)..... Mechanical resistance to shock and impact (AM 2)	18
9 Equipment temperature limits and protection against the spread of fire.....	19
10 Resistance to heat	19
11 Resistance to moisture and liquids (AM 1 only)..... Protection against hazards from fluids (AM 2)	20
12 Protection against radiation, including laser sources, and against sonic and ultrasonic pressure.....	20
13 Protection against liberated gases, explosion and implosion	20
14 Components.....	21
15 Protection by interlocks	22
16 Measuring circuits (AM 2).....	22
Table 1 – Documents attached to this report	7
Table 2 – Test equipment list.....	8
Table 3 – List of components relied on for safety	9
Form A.1 to Form A.8, Form A.10 to Form A.20, Form A.24 and Form A.29.....	23 to 42

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT
FOR MEASUREMENT, CONTROL, AND LABORATORY USE –****Part 3-032: Conformity verification report for IEC 61010-2-032:1994,
Particular requirements for hand-held current clamps
for electrical measurement and test**

FOREWORD

- 1) The IEC (International Electrotechnical Commission) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of the IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, the IEC publishes International Standards. Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. The IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of the IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested National Committees.
- 3) The documents produced have the form of recommendations for international use and are published in the form of standards, technical reports or guides and they are accepted by the National Committees in that sense.
- 4) In order to promote international unification, IEC National Committees undertake to apply IEC International Standards transparently to the maximum extent possible in their national and regional standards. Any divergence between the IEC Standard and the corresponding national or regional standard shall be clearly indicated in the latter.
- 5) The IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with one of its standards.
- 6) Attention is drawn to the possibility that some of the elements of this International Standard may be the subject of patent rights. The IEC shall not be held responsible for identifying any or all such patent rights.

The main task of IEC technical committees is to prepare International Standards. However, a technical committee may propose the publication of a technical report when it has collected data of a different kind from that which is normally published as an International Standard, for example "state of the art".

Technical reports do not necessarily have to be reviewed until the data they provide are considered to be no longer valid or useful by the maintenance team.

IEC 61010-3-032, which is a technical report, has been prepared by IEC technical committee 66: Safety of measuring, control, and laboratory equipment.

The text of this technical report is based on the following documents:

Enquiry draft	Report on voting
66/192/CDV	66/216/RVC

Full information on the voting for the approval of this technical report can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 3.

This report is a Technical Report and is of a purely informative nature and is therefore by itself not to be regarded as an International Standard. It is for use by testhouses and other users to assist them with determining and recording verification of conformity of the equipment under test with the requirements of

IEC 61010-2-032:1997,
and
IEC 61010-1:1990 + amendment 1:1992

or

IEC 61010-1:1990 + amendment 1:1992 + amendment 2:1995

The protocol for completion of this report is contained in publication IEC 61010-3:1997. Requirements which apply only to amendment 2 to IEC 61010-1 are indicated by “(AM 2)” in the first column. Requirements deleted by amendment 2 to IEC 61010-1 are indicated by “(AM 1 only)” in the second column.

Reference to IEC 61010-1 clauses and subclauses which have no applicability for IEC 61010-2-032 have been identified in the comments column of the report as ‘Not applicable’. The Forms A.9, A.21, A.22, A.23, A.25, A.26, A.27 and A.28, which are normally associated with these clauses/subclauses, are also excluded from this report.

The IEC sells read-only PDF files as a general rule. In the present instance, and quite exceptionally, to enable the user to fill in the forms, a revisable file is included in a pocket affixed to the back cover of this publication.

This publication can be downloaded from the Web as a PDF file. There is, however, at the end of the document, a revisable file containing the forms. Please use the zip/unzip function.

IEC TR 61010-3-032:2000

<https://standards.iteh.ai/catalog/standards/iec/09936d4f-2366-43cb-9663-8b6c104fe44b/iec-tr-61010-3-032-2000>

<p align="center">Conformity Verification Report IEC 61010-2-032: Safety requirements for electrical equipment for measurement, control, and laboratory use Particular requirements for hand-held current clamps for electrical measurement and test</p>	
Report reference No : Compiled by (+ signature) : Approved by (+ signature) : Date of issue :	
Testing organization : Address : Testing location :	
Applicant : Address :	
Standard : IEC 61010-2-032:1994 : and : IEC 61010-1:1990 + Amendment 1:1992 : or : IEC 61010-1:1990 + Amendment 1:1992 + Amendment 2:1995 Copyright blank test report : This report has been prepared by IEC/TC 66, which retains responsibility for any changes or corrections required.	
Test procedure : Procedure deviation : Non-standard test method :	
Type of item tested : <input type="checkbox"/> Measurement Trademark : Model/type référence : Manufacturer : Rating :	
Copy of rating plate:	

Description of equipment function:

INSTALLATION/OVERVOLTAGE CATEGORY:

POLLUTION DEGREE:

Environmental rating: ☐ Standard ☐ Other (specify):

Equipment mobility: ☐ Hand-held

Operating conditions: ☐ Continuous ☐ Short-time ☐ Intermittent

Overall size of the equipment (Length × Width × Height) cm:

Mass of the equipment (kg):

Marked degree of protection to IEC 60529: IP _____

Accessories and detachable parts included in the evaluation:

Options:

NOTE "(see Form A.X)" refers to a form appended to the report.

Table 1 – Documents attached to this report

[illegible]

Table 2 – Test equipment list

[illegible]

Table 3 – List of components relied on for safety

Unique component reference or location (including drawing reference if required)	Application/Function	Manufacturer and part number (note 1)	RATING (note 2)	Licence number, file number or other documentary evidence of acceptance
NOTE 1 List all manufacturers concerned.				
NOTE 2 Electrical, mechanical, flammability, etc.				

Clause Subclause	Requirement	Result	Comments
5	Marking and documentation		
5.1.1	General Required equipment markings are		
	a) visible		
	– from the exterior		
	or		
	– after removing a cover		
	b) not put on parts which can be removed by an OPERATOR		
	c) Letter symbols (IEC 60027) used		
	d) Graphic symbols (IEC 61010-1, Table 1) used		
5.1.101	RATING		
	– RATED circuit-to-earth voltage		
	– Nature of voltage		
	– Installation category		
	– Value and nature of maximum current		
5.1.2	Identification Equipment is identified by		
	– name or trademark of manufacturer or supplier		
	– model number, name or other means		
	– clamp designated for use with specific equipment		
	– model indicated on clamp		
	or		
	– symbol 14 used and information in documentation		
5.1.4	Fuses OPERATOR replaceable fuse marking (see also 5.4.5)		
5.1.5	Measuring circuit TERMINALS		
(AM 2)	RATED maximum working voltage or current marked		