



Designation: **E854 – 14 E854 – 14^{ε1}**

Standard Test Method for Application and Analysis of Solid State Track Recorder (SSTR) Monitors for Reactor Surveillance, E706(HIB) Surveillance¹

This standard is issued under the fixed designation E854; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reappraisal. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reappraisal.

^{ε1} NOTE—The title of this test method and the Referenced Documents were updated editorially in May 2017.

1. Scope

1.1 This test method describes the use of solid-state track recorders (SSTRs) for neutron dosimetry in light-water reactor (LWR) applications. These applications extend from low neutron fluence to high neutron fluence, including high power pressure vessel surveillance and test reactor irradiations as well as low power benchmark field measurement. **(1)**² This test method replaces Method **E418**. This test method is more detailed and special attention is given to the use of state-of-the-art manual and automated track counting methods to attain high absolute accuracies. In-situ dosimetry in actual high fluence-high temperature LWR applications is emphasized.

1.2 This test method includes SSTR analysis by both manual and automated methods. To attain a desired accuracy, the track scanning method selected places limits on the allowable track density. Typically good results are obtained in the range of 5 to 800 000 tracks/cm² and accurate results at higher track densities have been demonstrated for some cases. **(2)** Track density and other factors place limits on the applicability of the SSTR method at high fluences. Special care must be exerted when measuring neutron fluences ($E > 1\text{MeV}$) above 10^{16} n/cm² **(3)**.

1.3 Low fluence and high fluence limitations exist. These limitations are discussed in detail in Sections **13** and **14** and in Refs **(3-5)**.

1.4 SSTR observations provide time-integrated reaction rates. Therefore, SSTR are truly passive-fluence detectors. They provide permanent records of dosimetry experiments without the need for time-dependent corrections, such as decay factors that arise with radiometric monitors.

1.5 Since SSTR provide a spatial record of the time-integrated reaction rate at a microscopic level, they can be used for “fine-structure” measurements. For example, spatial distributions of isotopic fission rates can be obtained at very high resolution with SSTR.

1.6 *This standard does not purport to address the safety problems associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

1.7 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced Documents

2.1 ASTM Standards:³

E418 Test Method for Fast-Neutron Flux Measurements by Track-Etch Techniques (Withdrawn 1984)⁴

E844 Guide for Sensor Set Design and Irradiation for Reactor Surveillance

¹ This test method is under the jurisdiction of ASTM Committee **E10** on Nuclear Technology and Applications and is the direct responsibility of Subcommittee **E10.05** on Nuclear Radiation Metrology.

Current edition approved July 1, 2014. Published October 2014. Originally approved in 1981. Last previous edition approved in 2009 as E854 – 03(2009). DOI: 10.1520/E0854-14.10.1520/E0854-14E01.

² The boldface numbers in parentheses refer to the list of references appended to this test method.

³ For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard’s Document Summary page on the ASTM website.

⁴ The last approved version of this historical standard is referenced on www.astm.org.

3. Summary of Test Method

3.1 SSTR are usually placed in firm surface contact with a fissionable nuclide that has been deposited on a pure nonfissionable metal substrate (backing). This typical SSTR geometry is depicted in Fig. 1. Neutron-induced fission produces latent fission-fragment tracks in the SSTR. These tracks may be developed by chemical etching to a size that is observable with an optical microscope. Microphotographs of etched fission tracks in mica, quartz glass, and natural quartz crystals can be seen in Fig. 2.

3.1.1 While the conventional SSTR geometry depicted in Fig. 1 is not mandatory, it does possess distinct advantages for dosimetry applications. In particular, it provides the highest efficiency and sensitivity while maintaining a fixed and easily reproducible geometry.

3.1.2 The track density (that is, the number of tracks per unit area) is proportional to the fission density (that is, the number of fissions per unit area). The fission density is, in turn, proportional to the exposure fluence experienced by the SSTR. The existence of nonuniformity in the fission deposit or the presence of neutron fluence rate gradients can produce non-uniform track density. Conversely, with fission deposits of proven uniformity, gradients of the neutron field can be investigated with very high spatial resolution.

3.2 The total uncertainty of SSTR fission rates is comprised of two independent sources. These two error components arise from track counting uncertainties and fission-deposit mass uncertainties. For work at the highest accuracy levels, fission-deposit mass assay should be performed both before and after the SSTR irradiation. In this way, it can be ascertained that no significant removal of fission deposit material arose in the course of the experiment.

4. Significance and Use

4.1 The SSTR method provides for the measurement of absolute-fission density per unit mass. Absolute-neutron fluence can then be inferred from these SSTR-based absolute fission rate observations if an appropriate neutron spectrum average fission cross section is known. This method is highly discriminatory against other components of the in-core radiation field. Gamma rays, beta rays, and other lightly ionizing particles do not produce observable tracks in appropriate LWR SSTR candidate materials. However, photofission can contribute to the observed fission track density and should therefore be accounted for when nonnegligible. For a more detailed discussion of photofission effects, see 14.4.

4.2 In this test method, SSTR are placed in surface contact with fissionable deposits and record neutron-induced fission fragments. By variation of the surface mass density ($\mu\text{g}/\text{cm}^2$) of the fissionable deposit as well as employing the allowable range of track densities (from roughly 1 event/ cm^2 up to 10^5 events/ cm^2 for manual scanning), a range of total fluence sensitivity covering at least 16 orders of magnitude is possible, from roughly 10^2 n/ cm^2 up to 5×10^{18} n/ cm^2 . The allowable range of fission track densities is broader than the track density range for high accuracy manual scanning work with optical microscopy cited in 1.2. In particular, automated and semi-automated methods exist that broaden the customary track density range available with manual optical microscopy. In this broader track density region, effects of reduced counting statistics at very low track densities and track pile-up corrections at very high track densities can present inherent limitations for work of high accuracy. Automated scanning techniques are described in Section 11.

4.3 For dosimetry applications, different energy regions of the neutron spectrum can be selectively emphasized by changing the nuclide used for the fission deposit.

4.4 It is possible to use SSTR directly for neutron dosimetry as described in 4.1 or to obtain a composite neutron detection efficiency by exposure in a benchmark neutron field. The fluence and spectrum-averaged cross section in this benchmark field must be known. Furthermore, application in other neutron fields may require adjustments due to spectral deviation from the benchmark field spectrum used for calibration. In any event, it must be stressed that the SSTR-fission density measurements can be carried out completely independent of any cross-section standards (6). Therefore, for certain applications, the independent nature of this




FIG. 1 Typical Geometrical Configuration Used for SSTR Neutron Dosimetry



NOTE 1—The track designated by the arrow in the mica SSTR is a fossil fission track that has been enlarged by suitable pre-irradiation etching.

FIG. 2 Microphotograph of Fission Fragment Tracks in Mica

ASTM E854-14e1

<https://standards.iteh.ai/catalog/standards/sist/97bed729-0ca7-4531-b139-fc1b374a06a8/astm-e854-14e1>

test method should not be compromised. On the other hand, many practical applications exist wherein this factor is of no consequence so that benchmark field calibration would be entirely appropriate.

5. Apparatus

5.1 *Optical Microscopes*, with a magnification of 200× or higher, employing a graduated mechanical stage with position readout to the nearest 1 μm and similar repositioning accuracy. A calibrated stage micrometer and eyepiece scanning grids are also required.

5.2 *Constant-Temperature Bath*, for etching, with temperature control to 0.1°C.

5.3 *Analytical Weighing Balance*, for preparation of etching bath solutions, with a capacity of at least 1000 g and an accuracy of at least 1 mg.

6. Reagents and Materials

6.1 *Purity of Reagents*—Distilled or demineralized water and analytical grade reagents should be used at all times. For high fluence measurements, quartz-distilled water and ultra-pure reagents are necessary in order to reduce background fission tracks from natural uranium and thorium impurities. This is particularly important if any pre-irradiation etching is performed (see 8.2).

6.2 Reagents:

6.2.1 *Hydrofluoric Acid (HF)*, weight 49 %.

6.2.2 *Sodium Hydroxide Solution (NaOH)*, 6.2 N.

6.2.3 *Distilled or Demineralized Water*.

6.2.4 *Potassium Hydroxide Solution (KOH)*, 6.2 N.

6.2.5 *Sodium Hydroxide Solution (NaOH)*, weight 65 %.

6.3 Materials:




FIG. 2 Quartz Glass (continued)

ASTM E854-14e1

6.3.1 Glass Microscope Slides.

6.3.2 Slide Cover Glasses.

7. SSTR Materials for Reactor Applications

7.1 *Required Properties*—SSTR materials for reactor applications should be transparent dielectrics with a relatively high ionization threshold, so as to discriminate against lightly ionizing particles. The materials that meet these prerequisites most closely are the minerals mica, quartz glass, and quartz crystals. Selected characteristics for these SSTR are summarized in **Table 1**. Other minerals such as apatite, sphene, and zircon are also suitable, but are not used due to inferior etching properties compared to mica and quartz. These alternative SSTR candidates often possess either higher imperfection density or poorer contrast and clarity for scanning by optical microscopy. Mica and particularly quartz can be found with the additional advantageous property of low natural uranium and thorium content. These heavy elements are undesirable in neutron-dosimetry work, since such impurities lead to background track densities when SSTR are exposed to high neutron fluence. In the case of older mineral samples, a background of fossil fission track arises due mainly to the spontaneous fission decay of ^{238}U . Glasses (and particularly phosphate glasses) are less suitable than mica and quartz due to higher uranium and thorium content. Also, the track-etching characteristics of many glasses are inferior, in that these glasses possess higher bulk etch rate and lower registration efficiency. Other SSTR materials, such as Lexan⁵ and Makrofol⁶ are also used, but are less convenient in many reactor applications due to the presence of neutron-induced recoil tracks from elements such as carbon and oxygen present in the SSTR. These detectors are also more sensitive (in the form of increased bulk etch rate) to the β and γ components of the reactor radiation field (**13**). Also, they are more sensitive to high temperatures, since the onset of track annealing occurs at a much lower temperature for plastic SSTR materials.

7.2 *Limitations of SSTR in LWR Environments:*

⁵ Lexan is a registered trademark of the General Electric Co., Pittsfield, MA.

⁶ Makrofol is a registered trademark of Farbenfabriken Bayer AG, U. S. representative Naftone, Inc., New York, NY.



FIG. 2 Quartz Crystal (001 Plane) (continued)

[ASTM E854-14e1](https://standards.itih.ai/catalog/standards/sist/97bed729-0ca7-4531-b139-fc1b374a06a8/astm-e854-14e1)

<https://standards.itih.ai/catalog/standards/sist/97bed729-0ca7-4531-b139-fc1b374a06a8/astm-e854-14e1>

7.2.1 *Thermal Annealing*—High temperatures result in the erasure of tracks due to thermal annealing. Natural quartz crystal is least affected by high temperatures, followed by mica. Lexan and Makrofol are subject to annealing at much lower temperatures. An example of the use of natural quartz crystal SSTRs for high-temperature neutron dosimetry measurements is the work described in Ref (14).

7.2.2 *Radiation Damage*—Lexan and Makrofol are highly sensitive to other components of the radiation field. As mentioned in 7.1, the bulk-etch rates of plastic SSTR are increased by exposure to β and γ radiation. Quartz has been observed to have a higher bulk etch rate after irradiation with a fluence of 4×10^{21} neutrons/cm², but both quartz and mica are very insensitive to radiation damage at lower fluences ($<10^{21}$ neutrons/cm²).

7.2.3 *Background Tracks*—Plastic track detectors will register recoil carbon and oxygen ions resulting from neutron scattering on carbon and oxygen atoms in the plastic. These fast neutron-induced recoils can produce a background of short tracks. Quartz and mica will not register such light ions and are not subject to such background tracks.

7.2.4 *Thermal Stability of Fissionable Material Foils*—Uranium foils have been observed to completely convert to oxide during high temperature irradiation.

8. SSTR Pre- and Post-Irradiation Processing

8.1 Pre-Irradiation Annealing:

8.1.1 In the case of mica SSTR, a pre-annealing procedure designed to remove fossil track damage is advisable for work at low neutron fluences. The standard procedure is annealing for 6 h at 600°C (longer time periods may result in dehydration). Fossil track densities are so low in good Brazilian quartz crystals that pre-annealing is not generally necessary. Annealing is not advised for plastic SSTR because of the possibility of thermal degradation of the polymer or altered composition, both of which could affect track registration properties of the plastic.

8.2 Pre-Irradiation Etching:

TABLE 1 Characteristics of SSTR Candidates for LWR Reactor Applications

SSTR	Optical Efficiency, %	Asymptotic ^A Sensitivity	Conditions Under Which Accurate Annealing Corrections Can Be Made	Track Reduction, %
Muscovite mica	0.9875 ± 0.0085 ^B	(1.144 ± 0.018) × 10 ¹⁹ ²³⁸ U atoms/cm ^{2B}	501°C, 146.5 h ^C	0 ^C
Makrofol N	95.2 ± 0.53 ^D
Quartz glass	~70 ^E	...	402°C, 8 h ^C	73 ^C
Natural quartz Crystal	~80 ^E	...	857°C, 1 h ^F	20 ^F

^A Needs to be known only if used with asymptotically thick sources.

^B Etched 90 min in 49 % HF (6, 7, 8).

^C Data from Ref (9).

^D Etched ~20 h in 6.2 N KOH solution at room temperature (6).

^E Quartz glass etched 5 min in 48 % HF at room temperature. Quartz crystal etched in boiling 65 % NaOH solution for 25 min (10, 11).

^F Data from Ref (12).

8.2.1 *Mica*—Unannealed fossil tracks in mica are easily distinguished from induced tracks by pre-etching for a time that is long compared to the post-etching conditions. In the case of mica, a 6-h etch in 48 % HF at room temperature results in large diamond-shaped tracks that are easily distinguished from the much smaller induced tracks revealed by a 90-min post-etch (see Fig. 2)).

8.2.2 *Quartz Crystals*—Pre-etching is needed to chemically polish the surface. Polish a crystal mechanically on the 001 or 100 plane so that it appears smooth under microscopical examination, etch for 10 min in 49 % HF at room temperature, then boil in 65 % NaOH solution for 25 min. Examine the crystal surface microscopically. If it is sufficiently free of pits, select it for use as an SSTR.

8.2.3 *Quartz Glass*—If the glass has been polished mechanically, or has a smooth surface, then pre-etch in 49 % HF for 5 min at room temperature. Upon microscopical examination a few etch pits may be present even in good-quality quartz glass. If so, they will be larger than tracks due to fission fragments revealed in the post-etch, and readily distinguished from them.

8.2.4 *Plastic-Track Recorders*—If handled properly, background from natural sources, such as radon, will be negligible. Consequently, both pre-annealing and pre-etching should be unnecessary.

8.3 Post-Irradiation Etching:

8.3.1 *Mica*—Customary etching is for 90 min in 49 % HF at room temperature. Both the etch time and temperature may be varied to give optimum track sizes for the particular type of mica used. Except for work at the highest accuracy levels, precise control of the temperature is not necessary due to the zero bulk etch rate of the mica perpendicular to the cleavage planes. In the event that precise etching control is necessary, a technique has been demonstrated for mica that permits highly reproducible and standardized track size distributions (10).

8.3.2 *Quartz Crystals*—Etch for 25 min in boiling 65 % NaOH solution. Minimize evaporation by covering the nickel or platinum crucible in which the solution is heated. If left open, condense evaporated water and return to the solution. The value of the optical efficiency is dependent on the etching conditions (since the bulk etch rate is not zero), so both the concentration of the NaOH solution and the etching temperature must be controlled.

8.3.3 *Quartz Glass*—Etch for 5 min in 48 % HF at room temperature. Temperature control is essential because of the high bulk etch rate.

8.3.4 *Lexan*,⁵ or *Makrofol*,⁶ *N*—Various time temperature combinations in 6.2 N NaOH or KOH solution have proved satisfactory, depending upon the desired purpose. Examples of appropriate conditions are: (1) 50 h in 6.2N NaOH solution at 20°C, (2) 24 h in 6.2 N KOH solution at 20°C, and (3) 30 min in 6.2 N KOH solution at 50°C.

9. SSTR Fissionable Deposits

9.1 Properties:

9.1.1 *Fission Deposit Characteristics*—Perhaps the most critical factor in attaining high accuracy in SSTR neutron dosimetry is the quality of the fission deposit. High quality SSTR fission deposits possess the following characteristics: (6-17)

9.1.1.1 Accurately known total mass and mass density. The overall accuracy of the mass calibration must be consistent with the desired overall accuracy of the measurement.

9.1.1.2 Accurately known isotopic composition. Possible interfering isotopes must be minimized and the overall fission rate must be corrected for contributions from interfering isotopes.

9.1.1.3 *Negligible Impurities*—Impurities that contribute to the measured fission rate must be minimized (< 1 % contribution) and the overall fission rate must be corrected for contributions from impurities.

9.1.1.4 High uniformity is recommended. An independent measurement is required which verifies the uniformity of the deposit to an uncertainty commensurate with the desired accuracy of subsequent measurements using the deposit. Conversely, use of nonuniform deposits entails scanning of the entire SSTR surface to attain accurate results.

9.1.2 As has already been stated in 3.2, the accuracy of fission deposit characterization provides a fundamental limitation for the accuracy of the SSTR method. Fission-deposit mass assay as well as uniformity are important. Dosimetry goal accuracies provide bounds for the acceptable quality of SSTR fission deposits. For work at the highest accuracy levels, fission deposits can be prepared at close to or better than 1 % mass assay. Less accurate SSTR dosimetry can, however, be performed at a lower cost with less stringent requirements for fission deposit characterization. The deposit backing should contribute negligible background and the deposit should be flat, rigid, and capable of maintaining good contact with the SSTR. The deposit should be firmly adherent to the backing. The appropriate mass density for a particular LWR application may be calculated from:

$$\phi t \times W = \frac{\rho M}{\eta N_o \bar{\sigma} I} \quad (1)$$

where:

- ϕt = the expected fluence,
- W = the mass density of the deposit, g/cm²,
- ρ = the track density (the optimum track density for most manual scanning is about 5×10^4 tracks/cm²),
- I = the isotopic abundance (atomic fraction),
- η = the optical efficiency of the SSTR,
- σ = the spectral average fission cross section,
- M = the average atomic weight of the isotopic mixture used, and
- N_o = Avogadro's number (6.022×10^{23}).

9.1.3 In Eq 1, the assumption is made that the thickness (mass density) of the deposit is much less than the range of a fission fragment in the deposit material. Under these conditions, self-absorption is negligible and sensitivity depends linearly on W . For deposit thicknesses greater than about 100 μ /cm², self-absorption of fission fragments by the deposit becomes increasingly important. For deposit thicknesses greater than twice the range of a fission fragment in the deposit material, the effective thickness may be represented by a constant value. This constant value is referred to as the asymptotic sensitivity, s_∞ . It can be analytically shown (6) for a uniform deposit with no fluence rate depression that the asymptotic sensitivity is approximately given by:

$$s_\infty \approx \eta \frac{\langle R \rangle}{2} \quad (2)$$

where:

- $\langle R \rangle$ = the mean fission fragment range in the deposit.

In the case of uranium metal, an asymptotic sensitivity of 4.522 ± 0.070 mg/cm² has been measured (6,8). Thicknesses in the approximate range from 0.1 to 30 mg/cm² should be avoided due to problems arising from self-absorption of fission fragments in the source. While it is possible to work in this range, additional error will be incurred due to the need to correct for self-absorption. In the region beyond 30 mg/cm², one should use the asymptotic sensitivity.

9.2 *Isotopes Required*—In general, when performing reaction rate measurements for a particular isotope, contributions to the fission rate from other isotopes must be either negligible or corrected with sufficient accuracy. For example, use of the threshold reaction ²³⁸U (n,f) in a neutron field where the thermal fluence rate is appreciable requires highly depleted uranium in order to minimize contributions from ²³⁵U (n,f). Similarly chemical purity must be taken into account. When measuring the reaction rate for an even-even nuclide such as ²⁴⁰Pu, the abundance of the fissionable even-odd isotopic neighbors ²³⁹Pu and ²⁴¹Pu must be minimized. For low fluence rate measurements, contributions from spontaneously fissioning nuclides must be minimized and if necessary spontaneous fission track contributions must be subtracted.

9.3 Source Preparation:

9.3.1 Electrodeposition and vacuum deposition are the most frequently used and the most effective techniques. The latter method normally results in more uniform deposits, but economy of material and convenience may favor the former. In both cases, actinide deposits are produced more easily in the oxide than in the metallic form. Adherence of the deposit to the backing material can often be accomplished by heating the deposit to red heat in an inert atmosphere. Uniformity can be demonstrated by α -autoradiography using an α -sensitive SSTR such as cellulose nitrate or by fission track radiography with uniform neutron field irradiations.

9.3.2 Metallic backing for the fission deposit should be chosen to meet a number of requirements. For dosimetry purposes the backing should only be thick enough to ensure firm contact between the track recorder and the deposit (see Fig. 1). Furthermore, since it is preferable that no foreign elements be introduced into the radiation environment, backing materials should be chosen wherever possible from constituent elements that already exist in the radiation environment. Neutron field perturbations due to the backing are considered in Section 12. For high-fluence measurements, extremely pure-backing materials are required in order to reduce background fission tracks from natural uranium and thorium impurities. The surface of the backing material must be smooth and preferably possess a mirror finish.

9.4 Mass Assay:

9.4.1 *Absolute Disintegration Rate*—Mass assay may be accomplished by absolute α -counting using a low geometry α -counter (6). In many cases, the alpha decay constant is known to an accuracy of better than 1 %. In fact, the uncertainty of the alpha decay constant provides a fundamental limitation in this mass-assay method. Relative masses of several sources of the same isotope may be established to better than 1 % by α -counting in a 2π proportional counter. (See Table 2 for a summary of alpha decay constants of the actinide elements (15).)

9.4.2 *Mass Spectrometry*—Mass spectrometry combined with isotopic dilution techniques is a potentially useful method for mass assay of deposits. Mass spectrometry is particularly useful for low specific activity isotopes or isotopes with decay constants that have not been measured to an accuracy of 1 %. While mass spectrometry can provide accuracies of better than 1 %, it suffers from an inherent disadvantage, namely the need for destructive analysis.

9.4.3 *Isotopic Spikes*—High specific activity isotopes may be used as a tracer to indicate target mass. Alpha active isotopes such as ^{230}Th , ^{236}Pu , and ^{238}Pu as well as γ -emitting isotopes such as ^{237}U and ^{239}Np are useful for relative mass determinations. When using isotopic spikes, care must be taken to ensure that the source isotope and the spike are chemically equivalent. Also, the fission rate of the isotopic spike and its daughter products should be kept negligible compared to the fission rate of the isotope of interest. The use of isotopic spikes that feed complex decay chains (such as ^{228}Th and ^{232}U) should be avoided.

9.4.4 *Less Frequently Used Methods*—Ion, X-ray, and Auger microprobe analysis, X-ray fluorescence, neutron activation analysis, and wet chemical analysis methods may be useful for specific applications, but rarely attain an accuracy comparable to previously mentioned methods.

9.5 *Ultra Low-mass Deposits*—Methods for producing and calibrating ultra low-mass fissionable deposits are described in reference (3). Because of the low masses involved, typically 10^{-14} to 10^{-9} grams, care must be taken to avoid contamination of the deposits. Therefore, the deposits must be made under clean conditions using high-purity materials and chemical reagents.

9.5.1 *Mass Calibration*—Isotopic spiking methods (see 9.4.3) are used, and often the limitation on the amount of spike isotope that can be added is the extent of the contribution of either impurity isotopes or daughter isotopes to the overall fission rate of the deposit. For the case when short-lived ^{239}Np is used as a tracer for ^{237}Np , the eventual decay of the spike to ^{239}Pu must be considered as it will contribute to the overall fission rate of the deposit. Therefore, the $^{239}\text{Np}/^{237}\text{Np}$ ratio must be kept small enough to ensure that the resultant $^{239}\text{Pu}/^{237}\text{Np}$ fission rate ratio in the measured neutron spectrum will be small (typically less than 0.5 %). After the fission rate measurements are performed, the spike contribution to the fission rate must be confirmed to be small by calculating the fission rate due to the known amount of ^{239}Pu from the spike using the measured fission rate from a ^{239}Pu deposit exposed in the same dosimetry location.

9.5.2 *Ultra Low-Mass Deposit Calibration Uncertainties*—Additional uncertainties exist in the calibration of ultra low-mass deposits because of the additional steps necessary in the overall calibration. When isotopic spiking methods are used to determine the relative mass scale for a set of fissionable deposits, the uncertainty in the measurement of the relative radioactivity must be taken into account. For example, when short-lived ^{237}U is used as a tracer for either ^{235}U or ^{238}U , all of the uncertainties inherent in the measurements of the relative ^{237}U gamma decay rates must be taken into account. Among these uncertainties are the precision of the source to detector geometry and the Poisson statistics of the number of gamma ray counts recorded for each deposit. In order to determine an absolute mass scale, a measurement of gamma decay rate to absolute mass must be performed. Often this measurement corresponds to a relative gamma decay rate to absolute alpha decay rate measurement for a sample where both rates can be measured with sufficient accuracy. When an alpha emitting spike is used, such as ^{236}Pu to measure relative ^{239}Pu masses, only the relative alpha peak intensities need be measured. However, the uncertainties in the alpha decay constants (half lives) of both the spike isotope and the fissionable deposit isotope contribute to the overall uncertainty. For short-lived spikes such as ^{237}U (6.75 d) or ^{239}Np (2.34 d), decay corrections must be made. An alternative method (3) which eliminates the uncertainties

TABLE 2 Decay Constants and Associated Uncertainties Used in Actinide Mass Quantification

Nuclide	$t_{1/2}$ (years)	λ (s ⁻¹)	Uncertainty, %	Reference (15)		
				Vol	Year	Page
^{230}Th	$(7.538 \pm 0.030) \times 10^4$	2.914×10^{-13}	0.40	77	1996	433
^{232}Th	$(1.405 \pm 0.006) \times 10^{10}$	1.563×10^{-18}	0.43	80	1997	723
^{233}U	$(1.592 \pm 0.020) \times 10^5$	1.380×10^{-13}	1.26	109	2008	2657
^{234}U	$(2.455 \pm 0.006) \times 10^5$	8.947×10^{-14}	0.24	113	2012	2113
^{235}U	$(7.038 \pm 0.005) \times 10^8$	3.121×10^{-17}	0.07	114	2013	751
$^{237}\text{U}^A$	6.75 \pm 0.01 days	1.189×10^{-6}	0.15	107	2006	3323
^{238}U	$(4.468 \pm 0.003) \times 10^9$	4.916×10^{-18}	0.07	108	2007	681
^{237}Np	$(2.144 \pm 0.007) \times 10^6$	1.024×10^{-14}	0.33	105	2005	109
$^{239}\text{Np}^A$	2.356 \pm 0.003 days	3.405×10^{-6}	0.13	98	2003	665
$^{236}\text{Pu}^A$	2.858 \pm 0.008	7.685×10^{-9}	0.28	107	2006	2579
^{238}Pu	$(8.770 \pm 0.010) \times 10^1$	2.505×10^{-10}	0.11	108	2007	681
^{239}Pu	$(2.411 \pm 0.003) \times 10^4$	9.110×10^{-13}	0.12	98	2003	665
^{240}Pu	$(6.561 \pm 0.007) \times 10^3$	3.348×10^{-12}	0.11	107	2006	2649
$^{241}\text{Pu}^B$	$(1.429 \pm 0.0006) \times 10^1$	1.537×10^{-9}	0.04	106	2005	89
^{242}Pu	$(3.735 \pm 0.011) \times 10^5$	5.881×10^{-14}	0.29	97	2002	129

^A Tracer materials used for quantification of low mass primary deposits (may be α or β/γ emitters, or both).

^B The branching ratio for alpha emission is $(2.46 \pm 0.01) \times 10^{-3}$ %. The partial half-life for alpha decay is 5.79×10^5 years (± 3.2 %).

contributed by the decay corrections is to use multiple detectors which are operated in parallel. Relative gamma decay rates for ^{237}U can be determined with a set of ten thin-window proportional counters setting aside one counter for a standard that is also a fissionable deposit. In each set of ten counts, the decay rate of nine deposits is measured relative to the standard that is following the same radioactive half life. However, corrections must be made for small efficiency differences in a set of ten “identical” detectors as well as for detector cross-link and detector background, and the uncertainties in these corrections all contribute to the overall uncertainty. A useful strategy in ultra low-mass deposit calibration is to ensure that the additional uncertainties added by the addition of the spiking step are kept smaller than 0.5 % by the design of the spiking procedures.

9.5.3 Independent Mass Calibration Verification—Because of the added complexities of the production and calibration of the ultra low-mass deposits used in reactor cavity neutron dosimetry (2-5), deposits made for this application have been subjected to independent mass calibration accuracy verification through irradiations in standard reference neutron fields at NIST and elsewhere (16). Typically, one deposit from each ultra low-mass electroplating series is subjected to a benchmark irradiation, although, in some cases, multiple deposits from a series have been irradiated. These irradiations and NIST comparisons are consistent with the expected uncertainty of 2 % for the spike measurement mass scales and show that the absolute mass scales are consistent to 5 %. Because ultra low-mass deposits are made by electroplating methods, uniformity is more difficult to control than for vacuum-evaporated or sputtered deposits, but the uncertainty contribution of this non-uniformity is less than 2 %. The overall uniformity does contribute to the fluence limit that can be obtained as discussed subsequently in Section 11.4.2.1.

10. Manual Track-Scanning Procedures

10.1 Equipment and Calibration:

10.1.1 For manual scanning, a good research quality binocular microscope is required, having a stage equipped with two dials or micrometers that make it possible to estimate the x and y position of the stage to the nearest micrometer. One eyepiece should contain a square grid (one with 36 squares has been found to be highly satisfactory). The grid should cover a large fraction of the field of view. Take care to adjust the microscopes so that good Kohler illumination and adequate image contrast is obtained. This is especially true when asymptotically thick deposits are used (since many of the tracks are short and possess lower optical contrasts).

10.1.2 Calibrate the width of the grid for each lens combination with a stage micrometer and estimate to the nearest 1 μm . The linearity and accuracy of the dials or micrometers must also be checked and calibrated with the stage micrometer.

10.1.3 It is important that the instructions in the microscope manual be studied and followed to optimize contrast and resolution. If transmitted bright field illumination is used (highly satisfactory for mica and Makrofol N⁶ or Lexan⁵), contrast and resolution may be improved by using oblique instead of axial illumination, if available. Especially good contrast is obtained in quartz glass when reflected light is used.

10.2 Manual Track Counting Procedure:

10.2.1 Two situations need to be considered: (1) When it is essential to count all of the fission tracks in the SSTR, which can arise when the fission deposit is not sufficiently uniform for the desired accuracy, and (2) when only a fraction of the tracks need be counted to obtain the desired statistical accuracy.

10.2.2 For case (1), the scanner should find one edge of the region containing tracks and systematically cover the total area. A proven method (6) is to align the grid carefully so that the vertical lines are parallel to the y motion—a track or surface blemish should move on a grid line as the stage is moved along the y -axis. Do not count tracks touching or crossing the left and top grid lines, count those touching or crossing the right and bottom grid lines. When all the tracks in a given field are counted from left to right and from top to bottom as in reading, a track or blemish crossing or touching the top line is moved in the y direction until it is in the corresponding position on the bottom line. After the tracks moved into the field are counted as before, repeat the process until all the tracks in the given y swath have been counted. If tracks on the right edge of the region containing the tracks have been counted, move a track or surface blemish on the left line to the corresponding position on the right grid line, and count all of the tracks in a new y swath. Repeat this procedure over the entire area containing tracks; count all tracks. If track densities are sufficiently small, tracks may be counted as they cross a horizontal grid line as the SSTR is moved continuously in the y direction, instead of counting tracks field by field.

10.2.3 In case (2), the procedure is the same, except that a region removed from the edges of the track distribution is selected for counting. The area scanned is determined by observing the initial and final readings of the calibrated dial for the y -axis, and multiplying the difference by the width of the grid as measured by a stage micrometer. This may be repeated for more scanning swaths which need not be adjacent. This case offers the advantage to the scanner of selecting the best counting region if surface blemishes mar certain regions of the SSTR.

10.2.4 Count tracks with a tally counter; the scanner should be free to work the fine focus control while tracks are being counted so that tracks will be kept in sharp focus.

10.2.5 When scanners are first trained, they should *not* be told what to count. Rather, they should be asked to examine regions of the SSTR that do not contain tracks, so that they teach themselves to distinguish surface blemishes from fission tracks. In this way, careful scanners generally converge quickly to good agreement. If difficulties persist, different scanners may be asked to count tracks in the same field in order to remove small discrepancies. By using this procedure, observer biases are generally minimized and objectivity is established.

10.2.6 It is important that the SSTR surface be clean when scanned. Accomplish this by putting a cover glass over the surface of a clean SSTR ready for counting. If this is not feasible the SSTR should be cleaned, if necessary, before the tracks are counted.

11. Automated Track Counting

11.1 Introduction:

11.1.1 A major inconvenience of detection methods using tracks is the necessity for manual, visual measurement of tracks, a task that requires care, patience, and dedication. This drawback is especially significant for precision measurements, where inherent statistical limitations require the observation of large numbers of tracks, making the task time consuming and expensive. As a consequence, worldwide expertise in precision applications of SSTR methods is quite limited. A more detailed discussion of these requirements can be found in a critical review of the SSTR method (17).

11.1.2 Elimination of the human element is highly desirable for precise track measurements, since it allows the observation of larger numbers of tracks and permits the introduction of more quantitative standards of track identification and background subtraction. Such standards would obviate problems of personal bias in manual track measurements, which can otherwise compromise experimental accuracy. In order to attain high accuracy, such biases must constantly be guarded against in manual track scanning. Therefore, a considerable interest has existed, and continues to exist, in the automation of this scanning task. A perhaps tacit, but certainly reasonable assumption is that any such automated system must provide at least comparable accuracy to manual scanning techniques. Only under such a condition can the high accuracy goals of current SSTR applications be maintained.

11.2 Background:

11.2.1 Since the late 1960s, considerable effort has been expended by many groups in attempts to automate track scanning. Spark scanning methods have been developed (18-21) but have not been widely used due to limitations in accuracy (10-20 %) and track density (less than $10^3/\text{cm}^3$). More sophisticated systems employed an optical microscope under computer control (22-30). The availability of inexpensive minicomputers and microprocessors has afforded considerable progress in automated scanning capability (31-33). Of equal significance has been the development high-quality video camera image analysis systems. In addition to scarcely compromising microscopic resolution and contrast, modern CCD camera systems provide fast digital signals that afford dramatic improvements for automated pattern recognition. In view of this rapid evolution, it is best to consult the most recent literature for details on these highly-specialized techniques.

11.3 Automated Track Counting System:

11.3.1 Equipment and Calibration:

11.3.1.1 A good research-quality microscope is required, equipped with a motor-driven stage that can be controlled by a computer and can be repositioned with an accuracy of $\pm 1 \mu\text{m}$.

11.3.1.2 A computer input corresponding to the visual image from the microscope must be obtained. One method is to view the microscope image with a video camera and digitize the video image for input to a suitable image analysis computer.

11.3.1.3 A computer with sufficient speed and capacity to carry out the necessary steps for identification and correlation of track data is required.

11.3.2 Automated Track Counting Procedure:

11.3.2.1 A consistent and verifiable procedure (software or hardware, or both) must be developed for the identification and counting of tracks. This procedure may include gray level discrimination, image-enhancement, pattern recognition or other procedures that aid in track identification, or combination thereof.

11.3.2.2 Following optimization of the automated track counter parameters, counting of a series of track standards is required to verify the operation of the scanner within the desired accuracy. Whenever the scanning parameters are changed, recalibration with standards using the new parameters is required.

11.3.2.3 It is important that the SSTR surface be clean when scanned. Accomplish this by putting a cover glass over the surface of a clean SSTR ready for counting. For automated scanning, the quality of the SSTR can be particularly important. Care should be taken to ensure that the SSTR surface is as free as possible of cracks, scratches, dust, or other sources of visual interference.

11.4 High Precision Applications:

11.4.1 *Low and Medium Track Density Analysis*—Analysis of SSTR with low track densities can be done by counting tracks taking each contiguous area as one track. Corrections for pile-up are small and may be made by a variety of methods. It is also necessary to correct for background arising from imperfections in the track recorder, which the automated system may identify as tracks. Other methods normally applied for high track densities can also be used for low track densities, if the background can be handled accurately.

11.4.2 High Track Density Analysis:

11.4.2.1 At extremely high track densities, overlap of tracks can become so great that individual tracks can no longer be distinguished. An analysis of track density uncertainty as a function of track density appears in reference 34. The uncertainty attained in track density measurements will likely be a different function of track density for different automated scanning systems. In recent efforts (34), track density uncertainties less than 2 % were found to be generally unattainable for track densities greater than 8×10^5 tracks/cm². The high track density limit will also depend on the degree of uniformity of the fissionable deposits, and