



Designation: **E1336 – 11** **E1336 – 11 (Reapproved 2017)**

## Standard Test Method for Obtaining Colorimetric Data From a Visual Display Unit by Spectroradiometry<sup>1</sup>

This standard is issued under the fixed designation E1336; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

### INTRODUCTION

The fundamental procedure for characterizing the color and luminance of a visual display unit (VDU) is to obtain the spectroradiometric data under specified measurement conditions, and from these data to compute CIE chromaticity coordinates and absolute luminance values based on the 1931 CIE Standard Observer. The considerations involved and the procedures to be used to obtain precision colorimetric data for this purpose are contained in this test method. The values and procedures for computing CIE chromaticity coordinates are contained in Practice **E308**. The procedures for obtaining spectroradiometric data are contained in Test Method **E1341**. This test method includes some modifications to the procedures given in Practice **E308** that are necessary for computing the absolute luminance values of VDUs. This procedure is intended to be generally applicable to any VDU device, including but not limited to cathode ray tubes (CRT), liquid crystal displays (LCD), and electroluminescent displays (ELD).

### 1. Scope

- 1.1 This test method prescribes the instrumental measurements required for characterizing the color and brightness of VDUs.
- 1.2 This test method is specific in scope rather than general as to type of instrument and object.
- 1.3 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.
- 1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*
- 1.5 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

### 2. Referenced Documents

- 2.1 *ASTM Standards:*<sup>2</sup>
  - E284 Terminology of Appearance**
  - E308 Practice for Computing the Colors of Objects by Using the CIE System**
  - E1341 Practice for Obtaining Spectroradiometric Data from Radiant Sources for Colorimetry**
- 2.2 *CIE Publications:*
  - Publication CIE No. 18 Principles of Light Measurements**<sup>3</sup>
  - Publication CIE No. 15.2 Colorimetry, 2nd ed., 1986**<sup>3</sup>
  - Publication CIE No. 63 Spectroradiometric Measurement of Light Sources, 1984**<sup>3</sup>

<sup>1</sup> This test method is under the jurisdiction of ASTM Committee **E12** on Color and Appearance and is the direct responsibility of Subcommittee **E12.06** on Display, Imaging and Imaging Colorimetry.

Current edition approved Nov. 1, 2011/May 1, 2017. Published November 2011/May 2017. Originally approved in 1991. Last previous edition approved in 2003/2011 as **E1336 – 96 (2003)**:**E1336 – 11**. DOI: 10.1520/E1336-11/10.1520/E1336-11R17.

<sup>2</sup> For referenced ASTM standards, visit the ASTM website, [www.astm.org](http://www.astm.org), or contact ASTM Customer Service at [service@astm.org](mailto:service@astm.org). For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

<sup>3</sup> Available from U.S. National Committee of the CIE (International Commission on Illumination), c/o Thomas M. Lemons, TLA Lighting Consultants, Inc., 7 Pond St., Salem, MA 01970, <http://www.cie-usnc.org>, <http://www.cie.co.at> or <http://www.techstreet.com>.

### 2.3 IEC Publications:

Publication No. 441 Photometric and Colorimetric Methods of Measurement of the Light Emitted by a Cathode-Ray Tube Screen, 1974<sup>4</sup>

## 3. Terminology

3.1 The definitions of appearance terms in Terminology E284 are applicable to this test method.

## 4. Summary of Test Method

4.1 Procedures are given for obtaining spectroradiometric data and for the calculation of CIE tristimulus values and other color coordinates to describe the colors of VDUs. Modifications to the standard calculation procedures of Practice E308 are described.

## 5. Significance and Use

5.1 The most fundamental method for obtaining CIE tristimulus values or other color coordinates for describing the colors of visual display units (VDUs) is by the use of spectroradiometric data. (See CIE No. 18 and 63.) These data are used by summation together with numerical values representing the 1931 CIE Standard Observer and normalized to  $K_m$ , the maximum spectral luminous efficacy function.

5.2 The special requirements for characterizing VDUs possessing narrow or discontinuous spectra are presented and discussed. Modifications to the requirements of Practice E308 are given to correct for the unusual nature of narrow or discontinuous sources.

## 6. Requirements When Using Spectroradiometry

6.1 When describing the measurement of VDUs by spectroradiometry, the following must be specified:

6.1.1 The radiometric quantity determined, such as the irradiance ( $W/m^2$ ) or radiance ( $W/m^2\text{-sr}$ ), or the photometric quantity determined, such as illuminance ( $lm/m^2$ ) or luminance ( $lm/m^2\text{-sr}$  or  $cd/m^2$ ). The use of older, less descriptive names or units such as phot, nit, stilb is not recommended.

6.1.2 The geometry of the measurement conditions, including, whether a diffuser was used and the material from which it was constructed, the distances from the VDU, the size of the area to be measured on the VDU, the uniformity of the VDU across the area to be measured, the microstructure of the VDU picture elements, and the presence of any special intermediate optical devices such as integrating spheres.

6.1.3 The spectral parameters, including the spectral region, wavelength measurement interval, and spectral bandwidth. These must be specified since the various VDU technologies may demand more or less stringent requirements.

6.1.4 The type of standard used to calibrate the system, a standard lamp, a calibrated source, or a calibrated detector, and the source of the calibration.

6.1.5 The physical and temporal characteristics of the VDU including, refresh or field rate, convergence and purity adjustments (if the manufacturer allows such), luminance level, and any spectral line character in the emission from the VDU. The integration time of the detector system should be noted in relation to the refresh or field rate of the VDU. (See IEC No. 441.)

## 7. Apparatus

7.1 The basic instrument requirement is a spectroradiometric system designed for the measurement of spectral radiance or irradiance of light sources. See Practice E1341 for details on each of the parts of a spectroradiometer and how to calibrate and use the instrument.

### 7.2 Calibration Sources:

7.2.1 The standard calibration source for spectroradiometry is a tungsten-filament lamp operated at a specified current. It is preferable to have more than one standard lamp to permit cross-checks and to allow calibration at a range of luminance levels.

7.2.2 Monochromatic emission sources such as a low-pressure mercury arc lamp or tunable laser should also be available for use in calibrating the wavelength scale.

7.2.3 The electrical supplies for the calibration sources should be of the constant current type. The supply should be linear and not a switching supply. Current regulation should be maintained to better than 0.1 %. At this level the radiant flux from the calibration source is at least an order of magnitude more stable than the flux from a VDU.

7.2.4 There should be a standard for length measurements available (such as a high quality metric rule) since absolute irradiance calibration must be performed at an exact distance from the filament of the calibration lamp.

7.3 *Receiving Optics*—To maximize the light throughput the number of optical surfaces between the source of light should be kept to a minimum. In extended diffuse sources (such as VDUs) only a set of limiting apertures will be needed. In some instances, it may be desirable to image the VDU with an intermediate focusing lens or mirror assembly. Care should be taken to use a magnification that will adequately fill the entrance slit when viewing both the calibration and test source.

<sup>4</sup> Available from International Electrotechnical Commission (IEC), 3 rue de Varembé, Case postale 131, CH-1211, Geneva 20, Switzerland, <http://www.iec.ch>.