

Designation: E170 - 17

Standard Terminology Relating to Radiation Measurements and Dosimetry¹

This standard is issued under the fixed designation E170; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ε) indicates an editorial change since the last revision or reapproval.

INTRODUCTION

This terminology generally covers terms that apply to radiation measurements and dosimetry associated with energy deposition and radiation effects, or damage, in materials caused by interactions by high-energy radiation fields. The common radiation fields considered are X-rays, gamma rays, electrons, alpha particles, neutrons, and mixtures of these fields. This treatment is not intended to be exhaustive but reflects special and common terms used in technology and applications of interest to Committee E10, as for example, in areas of radiation effects on components of nuclear power reactors, radiation hardness testing of electronics, and radiation processing of materials.

This terminology uses recommended definitions and concepts of quantities, with units, for radiation measurements as contained in the International Commission on Radiation Units and Measurements (ICRU) Report 85a on "Fundamental Quantities and Units for Ionizing Radiation," October 2011.² Those terms that are defined essentially according to the terminology of ICRU Report 85a will be followed by ICRU in parentheses. It should also be noted that the units for quantities used are the latest adopted according to the International System of Units (SI) which are contained in Appendix X1 as taken from a table in ICRU Report 85a.² This terminology also uses recommended definitions of two JCGM documents,³ namely "International vocabulary of metrology" (VIM, 2012, unless indicated otherwise) and "Guide to the expression of uncertainty in measurement" (GUM, 2008). Those terms that are defined essentially according to the terminology of these documents will be followed by either VIM or GUM in parentheses.

A term is boldfaced when it is defined in this standard. For some terms, text in italics is used just before the definition to limit its field of application, for example, see activity.

1.1 ASTM Standards:⁴

¹ This terminology is under the jurisdiction of ASTM Committee E10 on Nuclear Technology and Applications and is the direct responsibility of Subcommittee E10.93 on Editorial

1. Referenced Documents catalog/standards/sist/07d8c498-c4 E380 Practice for Use of the International System of Units (SI) (the Modernized Metric System) (Withdrawn 1997)⁵ E722 Practice for Characterizing Neutron Fluence Spectra in

Terms of an Equivalent Monoenergetic Neutron Fluence for Radiation-Hardness Testing of Electronics

E910 Test Method for Application and Analysis of Helium Accumulation Fluence Monitors for Reactor Vessel Surveillance, E706 (IIIC)

1.2 Joint Committee for Guides in Metrology (JCGM) Reports:³

JCGM 100:2008, GUM 1995, with minor corrections, Evaluation of measurement data – Guide to the expression of uncertainty in measurement

JCGM 200:2012, VIM International vocabulary of metrology – Basic and general concepts and associated terms

Current edition approved June 1, 2017. Published June 2017. Originally approved in 1963. Last previous edition approved in 2016 as E170 - 16a. DOI: 10.1520/E0170-17.

² ICRU Report 60 has been superseded by ICRU Report 85a on "Fundamental Quantities and Units for Ionizing Radiation," October 2011. Both of these documents are available from International Commission on Radiation Units and Measurements (ICRU), 7910 Woodmont Ave., Suite 800, Bethesda, MD 20814.

³ Document produced by Working Groups of the Joint Committee for Guides in Metrology (JCGM). Available free of charge at BIPM website (http://

⁴ For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For Annual Book of ASTM Standards volume information, refer to the standard's Document Summary page on the ASTM website.

⁵ The last approved version of this historical standard is referenced on www.astm.org.

1.3 ICRU Documents:²

ICRU 60 Fundamental Quantities and Units for Ionizing Radiation, December 30, 1998

ICRU 85a Fundamental Quantities and Units for Ionizing Radiation, October, 2011

1.4 NIST Document:⁶

NIST Technical Note 1297 Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results, 1994

1.5 ISO Standard:⁷

ISO 10012 Measurement management systems – Requirements for measurement processes and measuring equipment

2. Terminology

absorbed dose (*D*)—quotient of $d\bar{\varepsilon}$ by dm, where $d\bar{\varepsilon}$ is the mean incremental energy imparted by ionizing radiation to matter of incremental mass dm. (ICRU), thus

$$D = d\bar{\varepsilon}/dm \tag{1}$$

DISCUSSION—The SI unit of absorbed dose is the gray (Gy), where 1 gray is equivalent to the absorption of 1 joule per kilogram of the specified material (1 Gy = 1 J/kg). The unit rad (1 rad = 100 erg/g = 0.01 Gy) is still widely used in the nuclear community; however, its continued use is not encouraged. For a photon source under conditions of charged particle equilibrium, the absorbed dose, D, may be expressed as follows:

$$D = \Phi \cdot E \cdot \mu_{\rm en}/\rho, \qquad (2)$$

where:

 Φ = fluence (m⁻²),

E = energy of the ionizing radiation (J), and

 μ_{en}/ρ = mass energy absorption coefficient (m²/kg).

If bremsstrahlung production within the specified material is negligible, the mass energy absorption coefficient (μ_{en}/ρ) is equal to the mass energy transfer coefficient (μ_{tr}/ρ) , and absorbed dose is equal to kerma if, in addition, charged particle equilibrium exists.

absorbed dose rate (*D*)—quotient of d*D* by d*t* where d*D* is the increment of absorbed dose in the time interval d*t* (ICRU), thus

$$\dot{D} = dD/dt \tag{3}$$

SI unit: $Gy \cdot s^{-1}$.

Discussion—The absorbed-dose rate is often specified as an average value over a longer time interval, for example, in units of $Gy \cdot min^{-1}$ or $Gy \cdot h^{-1}$.

accuracy—closeness of agreement between a measured **quantity value** and a true quantity value of a **measurand** (VIM).

DISCUSSION—

- (1) The concept "accuracy" is not a quantity and is not given a numerical quantity value. A measurement is said to be more accurate when it offers a smaller measurement error.
- (2) The term "accuracy" should not be used for measurement trueness and the term "**precision**" should not be used for "accuracy," which, however, is related to both these concepts.

(3) "Accuracy" is sometimes understood as closeness of agreement between measured quantity values that are being attributed to the measurand.

activation cross section—cross section for a specific direct or compound nuclear interaction in which the product nucleus is radioactive.

DISCUSSION—Fission and spallation processes produce a statistical ensemble of outgoing nuclear channels, but they are not considered to be activation reactions.

activity (A)—of an amount of radionuclide in a particular energy state at a given time, quotient of -dN by dt, where dN is the mean change in the number of nuclei in that energy state due to spontaneous nuclear transformations in the time interval dt (ICRU), thus

$$A = -dN/dt (4)$$

Unit: s⁻¹

The special name for the unit of activity is the becquerel (Bq), where

$$1 Bq = 1 s^{-1}$$
 (5

Discussion—The former special unit of activity was the curie (Ci), where

1 Ci =
$$3.7 \times 10^{10} \text{ s}^{-1}$$
 (exactly). (6)

The negative sign in Eq 4 is an indication that the activity is decreasing with time. The "particular energy state" is the ground state of the nuclide unless otherwise specified. The activity of an amount of radionuclide in a particular energy state is equal to the product of the decay constant for that state and the number of nuclei in that state (that is, $A = N\lambda$). (See **decay constant**.)

aleatory uncertainty—uncertainty representing random uncertainty contributors where there is little possibility of reducing this uncertainty contributor by consideration of a more controlled scenario.

DISCUSSION-

- (1) One paradigm decomposes uncertainty into epistemic and aleatory components. This division of uncertainty categories is very dependent upon what question is being posed in a given application. Aleatory uncertainties can be transformed into epistemic uncertainties depending upon the application. The uncertainties underlying a quantity may be classified as aleatory or epistemic according to the goals of the process.
- (2) Aleatory uncertainty, also referred to as variability, stochastic uncertainty or irreducible uncertainty, is used to describe inherent variation associated with a quantity or phenomenon of interest. The determination of material properties or operating conditions of a physical system typically leads to aleatory uncertainties; additional experimental characterization might provide more conclusive description of the variability but cannot eliminate it completely. Aleatory uncertainty is normally characterized using probabilistic approaches.

analysis bandwidth—spectral band used in an instrument, such as a densitometer, for a measurement.

analysis wavelength—wavelength used in a spectrophotometric instrument for the measurement of optical absorbance or reflectance.

annihilation radiation—gamma radiation produced by the annihilation of a positron and an electron.

⁶ Available from National Institute of Standards and Technology (NIST), 100 Bureau Dr., Stop 1070, Gaithersburg, MD 20899-1070, USA, http://www.nist.gov

⁷ Available from International Organization for Standardization (ISO), ISO Central Secretariat, BIBC II, Chemin de Blandonnet 8, CP 401, 1214 Vernier, Geneva, Switzerland, http://www.iso.org.

Discussion—For particles at rest, two photons are produced, each having an energy corresponding to the rest mass of an electron (511 keV).

backscatter peak—peak in the observed photon spectrum resulting from large-angle (>110°) Compton scattering of gamma rays from materials near the detector.

Discussion—This peak is normally below about 0.25 MeV. Also, it will not have the same shape as the full-energy peaks (being wider and skewed toward lower energy).

benchmark neutron field—well-characterized irradiation environment which provides a fluence or fluence rate of neutrons suitable for the validation or calibration of experimental techniques and methods as well as for validation of cross sections and other nuclear data, where following classification for reactor dosimetry has been made:⁸

controlled neutron field—neutron field physically well-defined, and with some spectrum definition, employed for a restricted set of validation experiments.

reference neutron field—permanent and reproducible neutron field less well characterized than a standard field but accepted as a measurement reference by a community of users.

standard neutron field—permanent and reproducible neutron field that is characterized to state-of-the-art accuracy in terms of neutron fluence rate and energy spectra, and their associated spatial and angular distributions, where important field quantities need to be verified by interlaboratory measurements.

Discussion—A type of neutron field is considered to be a "standard" over a specified energy range and there is only one type of "standard neutron field" for a given energy range. Currently, the ²⁵²Cf spontaneous fission field is a "standard neutron field" from 0.5 MeV to 8 MeV. The deuterium-tritium (DT) accelerator field is considered to be the "standard neutron field" from 13.5 to 15 MeV. The thermal Maxwellian and epithermal 1/E slowing-down field are also considered to be "standard neutron fields."

bremsstrahlung—broad-spectrum electromagnetic radiation emitted when an energetic charged particle is influenced by a strong electric field, such as the Coulomb field of an atomic nucleus.

Discussion—In radiation processing, bremsstrahlung photons are generated by the deceleration or deflection of energetic electrons in a target material. When an electron passes close to an atomic nucleus, the strong Coulomb field causes the electron to deviate from its original motion. This interaction results in a loss of kinetic energy by the electron with the emission of electromagnetic radiation; the photon energy distribution extends up to the maximum kinetic energy of the incident electron. This bremsstrahlung spectrum depends on the electron energy, the composition and thickness of the target, and the angle of emission with respect to the incident electron.

buildup factor—for radiation passing through a medium, ratio of the total value of a specified radiation quantity (such as absorbed dose) at any point in that medium to the contribution to that quantity from the incident uncollided radiation reaching that point.

cadmium ratio—ratio of the neutron reaction rate measured with a given bare neutron detector to the neutron reaction rate measured with an identical neutron detector enclosed by a particular cadmium cover and exposed in the same neutron field at the same or an equivalent spatial location.

Discussion—In practice, meaningful experimental values can be obtained in an isotropic neutron field by using a cadmium filter approximately 1 mm thick.

calibrated instrument—instrument that has been through a **calibration** process at established time intervals.

Discussion—Measurements carried out by this instrument have metrological traceability to the reference standard if calibration is properly carried out.

calibration—set of operations that establish, under specified conditions, the relationship between values of quantities indicated by a measuring instrument or measuring system, or values represented by a material measure or a reference material, and the corresponding values realized by standards (VIM: 1993).

DISCUSSION-

- (1) Calibration conditions include environmental and irradiation conditions present during calibration.
- (2) These standards should have metrological traceability to a national or international standard.
- (3) To be reliable, calibration should be carried out at regular time intervals frequency may depend on the final use of the data. Often, the frequency is specified by regulatory authorities.

calibration source or field—see electron standard field, gamma-ray standard field, and X-ray standard field.

calorimeter—instrument capable of making measurements of energy deposition (or **absorbed dose**) in a material through measurement of change in its temperature and knowledge of the characteristics of the material and the details of its construction.

Discussion—Calorimeter is generally designated as a **primary-standard dosimeter**.

certified reference material (**CRM**)—reference material, accompanied by documentation issued by an authoritative body and providing one or more specified property values with associated uncertainties and traceabilities, using valid procedures (VIM).

Discussion—"Certified reference material" should be differentiated from "Standard Reference Material" which is a National Institute of Standards and Technology (NIST) trademarked nomenclature.

charged particle equilibrium—condition that exists in an incremental volume within a material under irradiation if the kinetic energies and number of charged particles (of each type) entering that volume are equal to those leaving that volume.

Discussion—When electrons are the predominant charged particle, the term "electron equilibrium" is often used to describe charged particle equilibrium. See also the discussions attached to the definitions of **kerma** and **absorbed dose**.

coincidence sum peak—for gamma spectroscopy, peak in the observed photon spectrum produced at an energy corresponding to the sum of the energies of two or more gamma-or x-rays from a single nuclear event when the emitted

⁸ The following three definitions are derived from: *Neutron Cross Sections for Reactor Dosimetry*, International Atomic Energy Agency, Laboratory Activities, Vienna, Vol 1, 1978, p. 62 and Vlasov, M., IAEA Program on Benchmark Neutron Fields Applications for Reactor Dosimetry, Report INDC(SEC)-54/L+Dos, IAEA, Vienna, 1976.

photons interact with the detector within the resolving time of the measurement system.

combined standard uncertainty—standard uncertainty that is obtained using the individual standard uncertainties associated with the input quantities in a measurement model (VIM).

Discussion—In case of correlations of input quantities contributing to the resulting uncertainty, covariances must also be taken into account when calculating the combined standard uncertainty.

Compton edge (E_c)—maximum energy value of electrons of the Compton scattering continuum, which is given by:

$$E_{c} = E_{\gamma} - \frac{E_{\gamma}}{1 + \frac{2E_{\gamma}}{0.511}} \tag{7}$$

Discussion—This value corresponds to 180° scattering of the primary photon of energy E_{γ} (MeV). For a 1 MeV photon, the Compton edge is about 0.8 MeV.

Compton scattering—elastic scattering of a photon by an atomic electron, under the condition of conservation of momentum, that is, the vector sum of the momenta of the outgoing electron and photon is equal to the momentum of the incident photon.

Discussion—The scattered photon energy, E'_{ν} (in MeV), is given by

$$E'_{\gamma} = \frac{E_{\gamma}}{1 + \frac{E_{\gamma}(1 - \cos \theta)}{0.511}}$$
 (8)

where E_{γ} is the incident photon energy in MeV and θ is the angle between the direction of the incident and scattered photon. The electron energy, $E_{\rm e}$, is equal to $E_{\gamma} - E_{\gamma}$.

continuum—for gamma spectroscopy, smooth distribution of energy deposited in a gamma detector arising from partial energy absorption from **Compton scattering** or other processes (for example, **bremsstrahlung**). See **Compton scattering**.

coverage factor (*k*)—number larger than one by which a **combined standard uncertainty** is multiplied to obtain an **expanded uncertainty** (VIM).

Discussion—Coverage factor is typically in the range of 2 to 3.

cross section (σ)—of a target entity, for a particular interaction produced by incident charged or uncharged particles of a given type and energy, quotient of $N_{\rm int}$ by Φ , where $N_{\rm int}$ is the mean number of such interactions per target entity subjected to the **fluence** Φ (adapted from ICRU), thus

$$\sigma = N_{in}/\Phi \tag{9}$$

Unit: m²

Discussion—The special unit of cross section is the barn, b, where

$$1 b = 10^{-28} m^2 = 10^{-24} cm^2$$
 (10)

cumulative fission yield—total number of atoms of a specific nuclide produced directly by a fission event and via radioactive decay of the precursors.

DISCUSSION—This definition is from INDC(NDS)-0534. The fission yield (either independent or cumulative) varies with the energy of the

neutron that initiates the fission event. The fission yields for a spontaneous fission event are distinct from those for a thermal or fast neutron-induced fission.

decay constant (λ)—of a radionuclide in a particular energy state, quotient of -dN/N by dt, where dN/N is the mean fractional change in the number of nuclei in that energy state due to spontaneous nuclear transformations in the time interval dt (ICRU), thus

$$\lambda = \frac{-dN/N}{dt} \tag{11}$$

Unit: s⁻¹

DISCUSSION—The quantity (ln 2)/ λ is commonly called the half-life, T_{V_2} , of the radionuclide, that is, the time taken for the activity of an amount of radionuclide to become half its initial value.

depth-dose distribution—variation of absorbed dose with depth from the incident surface of a material exposed to a given radiation.

displacement dose (D_d) —quotient of $d\bar{\epsilon}_d$ by dm, where $d\bar{\epsilon}_d$ is that part of the mean incremental energy which produces atomic displacements (that is, excluding the energy that produces ionization and excitation of electrons) imparted by radiation to matter of incremental mass dm, thus

$$D_d = d\bar{\varepsilon}_d / dm \tag{12}$$

Unit: J ⋅ kg⁻¹

Discussion—A more common unit is **displacements per atom (dpa)** (see definition).

displacements per atom (dpa)—mean number of times each atom of a solid is displaced from its lattice site during its exposure to radiation.

Discussion—This quantity is calculated from the displacement dose using a dislocation efficiency model such as Kinchin-Pease¹⁰ or Norget-Robinson-Torrens (NRT) model.¹¹

displacement threshold energy (E_d)—minimum kinetic energy imparted to a lattice atom to permanently displace it from its initial lattice site.

Discussion—This energy refers to the energy required to create the initial Frenkel pair, that is, a vacancy-interstitial defect pair, and is independent of subsequent defect interaction or thermal recombination effects. This energy can have an angle-dependence and, in polyatomic lattices, can be different for different types of lattice atoms. Displacement threshold energies in typical solids are on the order of 10-50 eV.

dosimeter—device that, when irradiated, exhibits a quantifiable change that can be related to a dosimetric quantity using appropriate measurement instrument(s) and procedures.

Discussion—As discussed in ICRU-85a, dosimetric quantities provide a physical measure to correlate with actual or potential effects. They are products of radiometric quantities and interaction coefficients. In calculations, the values of these quantities and coefficients must be known, while measurements might not require this information. Dosimetric quantities include air **kerma**, **exposure** and **absorbed dose to a specified material**.

⁹ Handbook of Nuclear Data for Safeguards: Database Extensions, August 2008.

¹⁰ Kinchin, G. H., and Pease, R. S., "The Displacement of Atoms in Solids by Radiation," *Reports on Progress in Physics*, Vol 18, 1955, pp. 1-51.

¹¹ Norgett, M. J., Robinson, M. T., and Torrens, I. M., "A Proposed Method of Calculating Displacement Dose Rates," *Nuclear Engineering and Design*, Vol 33, 1975, pp. 50-54.

dosimetry system—interrelated elements used for determining a dosimetric quantity, including **dosimeters**, measurement instruments and their associated reference standards, and procedures for their use.

DISCUSSION—As discussed in ICRU-85a, dosimetric quantities provide a physical measure to correlate with actual or potential effects. They are products of radiometric quantities and interaction coefficients. In calculations, the values of these quantities and coefficients must be known, while measurements might not require this information. Dosimetric quantities include air **kerma**, **exposure** and **absorbed dose to a specified material**.

- effective cadmium cut-off energy (E_{Cd})—energy at which a specified thickness of cadmium results in the same reaction rate in a $1/\nu$ detector as a perfect filter which has the following properties in a neutron field with a 1/E energy dependence of the neutron fluence spectrum:
- (1) for all energies below $E_{\rm Cd}$, no neutrons are present after the filter, and
- (2) for all energies above $E_{\rm Cd}$, neutron reactions after the filter occur at the same rate as if the cadmium were not present.

Discussion— $E_{\rm Cd}$ varies with cadmium thickness used as the filter and geometry, angular distribution of incident neutrons, and ambient temperature. The definition is applicable for detectors whose cross sections do no depart significantly from a $1/\nu$ dependence in the region of the cut-off energy, and also for neutron fields whose neutron fluence spectrum does not depart significantly from a 1/E energy dependence in region of the cut-off energy.

efficiency—see total efficiency and full-energy peak efficiency.

electron equilibrium—charged-particle equilibrium for electrons.

electron standard field—electron field whose particle energy and direction, spatial uniformity, temporal profile, and **fluence rate** uniformity are well established and reproducible.

energy calibration—process of establishing the relationship between photon or particle energy and channel number in the spectrometer.

Discussion—The energy calibration may be as simple as building a table of two or more energy-channel pairs or as complex as using a least squares algorithm to establish a function describing the energy versus channel relationship.

epistemic uncertainty—uncertainty component solely due to a lack of knowledge.

DISCUSSION-

- (1) One paradigm decomposes uncertainty into epistemic and aleatory components. This division of uncertainty categories is very dependent upon what question is being posed in a given application. Epistemic uncertainties can be transformed into aleatory uncertainties depending upon the application. The uncertainties underlying a quantity may be classified as aleatory or epistemic according to the goals of the process.
- (2) The epistemic component is also called the reducible uncertainty and can arise from assumptions introduced in the derivation of the mathematical model used or simplifications related to the correlation or dependence between physical processes. This epistemic uncertainty has the possibility of being reduced if one can gather more data or refine modeling assumptions. Epistemic uncertainty is not well characterized

by probabilistic approaches because it might be difficult to infer any statistical information due to the nominal lack of knowledge.

epithermal neutrons—general classification of neutrons with energies above those of thermal neutrons; or frequently, neutrons with energies in the resonance range, between the thermal limit and some upper limit, such as 0.1 MeV (see **thermal neutrons**).

DISCUSSION—The term "epithermal neutrons" is generally used in thermal neutron systems when two groups of neutrons are considered. The term is not used to describe high energy neutrons in other types of systems such as fast or fusion reactors.

equivalent fission fluence—fluence of fission spectrum neutrons that would give a detector or material response for a particular reaction equal to that in a given neutron field.

equivalent 2200 m/s fluence ($\Phi_{\rm w}$)—measure of the effective thermal neutron fluence made with an ideal 1/v detector and using the 2200 m/s cross section, thus

$$\Phi_{w} = \int_{0}^{t_{h}} n(t) v_{0} dt \tag{13}$$

where:

n(t) = neutron density, at time t after the beginning of the exposure of the detector,

 $v_0 = 2200$ m/s, and

 t_i = duration of the exposure of the detector.

Discussion—The equivalent 2200 m/s fluence is often referred to as the Westcott convention fluence, or simply the Westcott fluence. ¹² The symbol nv_0t is sometimes used. All neutrons are included in n(t) (not just thermal neutrons). The equivalent 2200 m/s is especially useful when cadmium is not being used.

Reactions =
$$\int_{0}^{\infty} \int_{0}^{\infty} n(E,t) v \sigma(E) E dt = \int_{0}^{\infty} \int_{0}^{\infty} n(E,t) v \frac{\sigma_{0} v_{0}}{v} dE dt$$

$$= \int_{0}^{\infty} \frac{1}{n(t)} \sigma_{0} v_{0} dt = \Phi_{w} \sigma_{0}$$
(14)

 $\sigma(E)$ may be expressed as $\frac{\sigma_0 v_0}{v}$ for a 1/v cross section. Φ_w may not be a measured value (that is, made with a 1/v detector). It may be a calculated quantity.

equivalent monoenergetic neutron fluence $(\Phi_{eq}(E_o))$ —
measure of an incident energy fluence spectrum, $\Phi(E)$, in
terms of the fluence of monoenergetic neutrons at a specific
energy, E_o , that produces the same displacement kerma, K_D ,
in a specific material (for example, silicon) as $\Phi(E)$.

Discussion—In applying this definition, total kerma is divided into two parts, ionization and displacement kerma (see Practice E722).

escape or pair production peak—peak in a gamma ray spectrum resulting from the pair production process within the detector, subsequent annihilation of the positron produced, and escape from the detector of one or both of the annihilation photons (see pair production and annihilation radiation).

¹² Westcott, C. H., Walker, W. H., and Alexander, T. K., "Effective Cross Sections and Cadmium Ratios for the Neutron Spectra of Thermal Reactors," Proceedings of the International Conference on Peaceful Uses of Atomic Energy, United Nations, Vol 16, 1958, p. 70.

single escape peak—gamma ray spectrum peak corresponding to escape of one of the annihilation photons from the active volume of the detector, where the peak energy is equal to the original gamma ray energy minus 511 keV.

double escape peak—gamma ray spectrum peak corresponding to escape of both of the annihilation photons from the active volume of the detector, where the peak energy is equal to the original gamma ray energy minus 1.022 MeV.

- **expanded uncertainty**—quantity defining an interval about the result of a measurement that may be expected to encompass a large fraction of the distribution of values that could reasonably be attributed to the **measurand** (GUM).
- (1) The fractions may be viewed as the coverage probability or level of confidence of the interval.
- (2) To associate a specific level of confidence with the interval defined by the expanded uncertainty requires explicit or implicit assumptions regarding the probability distribution characterized by the measurement result and its **combined standard uncertainty**. The level of confidence that may be attributed to this interval can be known only to the extent to which such assumptions may be justified.
- (3) Expanded uncertainty is also referred to as overall uncertainty.

exposure (X)—quotient of dq by dm, where dq is the absolute value of the mean total charge of the ions of one sign produced when all electrons and positrons liberated or created by photons incident on a mass dm of dry air are completely stopped in dry air (ICRU), thus

$$X = \mathrm{d}q/\mathrm{d}m \tag{15}$$

Unit: $C \cdot kg^{-1}$

Discussion—Formerly, the special unit of exposure was the röntgen (R), where

https://stan 1 R =
$$2.58 \times 10^{-4} \text{ C} \cdot \text{kg}^{-1} \text{ (exactly)}$$
 (16)

exposure rate (\dot{X}) —quotient of dX by dt, where dX is the increment of exposure in the time interval, dt (ICRU), thus

$$\dot{X} = dX/dt \tag{17}$$

Unit: $C \cdot kg^{-1} s^{-1}$

fast neutrons—neutrons of energy exceeding some threshold that must be specified (typically 0.1 or 1 MeV).

DISCUSSION—This term is often associated with those neutrons predominantly responsible for displacement damage of materials in neutron radiation fields.

fission chamber—ionization chamber containing one or more surfaces coated with fissionable material.

fluence (Φ)—quotient of dN by da, where dN is the number of particles incident on a sphere of cross-sectional area da (ICRU), thus

$$\Phi = dN/da \tag{18}$$

Unit: m⁻²

DISCUSSION—In order to distinguish this quantity from the energy fluence, this term is sometimes referred to as "particle fluence." The fluence may also be expressed as the time integral of the fluence rate.

fluence rate (ϕ)—quotient of $d\Phi$ by dt, where $d\Phi$ is the increment of fluence in the time interval dt (ICRU), thus

$$\varphi = \frac{\mathrm{d}\Phi}{\mathrm{d}t} = \frac{\mathrm{d}^2 N}{\mathrm{d}a\,\mathrm{d}t} \tag{19}$$

Unit: $m^{-2} \cdot s^{-1}$

DISCUSSION—In order to distinguish this quantity from the energy fluence rate, this term is sometimes referred to as "particle fluence rate." The term **flux density** may be used but the term fluence rate conforms to the adoption of a uniform set of terms and units as prescribed by ICRU and SI units. Historically, the term *neutron flux* has been understood to mean neutron flux density (fluence rate). This term still is widely used in the nuclear community.

Fricke dosimetry system—consists of a liquid chemical dosimeter (composed of ferrous sulfate or ferrous ammonium sulfate in aqueous sulfuric acid solution), a spectrophotometer (to measure optical absorbance) and its associated reference standards, and procedures for its use.

DISCUSSION-

- (1) It is considered to be a reference-standard dosimetry system.
- (2) Sodium chloride is usually added to dosimetric solution to minimize the effects of organic impurities.

full-energy peak—for gamma spectroscopy, peak in an energy spectrum recorded by a photon detector that occurs when the full energy of an incident photon is absorbed by the detector.

DISCUSSION—This is sometimes referred to as the photopeak.

full-energy peak efficiency— for gamma spectroscopy, ratio of the net count rate in the full-energy peak to the emission rate of the photons from a sample giving rise to the peak.

Discussion—The value is dependent on the source-detector-shield geometry and the photon energy. This is sometimes referred to as the photopeak efficiency.

gamma-ray standard field—well-characterized gamma-ray field that is well established and reproducible at a designated location and it used for validation or calibration of experimental techniques and methods.

Discussion—An example is the gamma-ray field produce by ⁶⁰Co.

G value—see radiation chemical yield.

half-life—see decay constant.

helium accumulation fluence monitor (HAFM)—passive neutron dosimeter where the neutron fluence is obtained by dividing the measured concentration of the reaction product helium by the **spectrum-averaged cross section**. (See Test Method E910).

independent fission yield—number of atoms of a specific nuclide produced directly by a fission event (not via radioactive decay of the precursors).

DISCUSSION—This definition is from INDC (NDS)-0534. The fission yield (either independent or cumulative) varies with the energy of the neutron that initiates the fission event. The fission yields for a spontaneous fission event are distinct from those for a thermal or fast neutron-induced fission.

influence quantity—quantity that, in a direct measurement, does not affect the quantity that is actually measured, but affects the relation between the indication and the measurement result (VIM).