![SIST logo]

SLOVENSKI STANDARD
SIST EN 27029:1999

01-november-1999

Akustika - Slušni prag v odvisnosti od starosti in spola otološke normalne osebe (ISO 7029:1984)

Acoustics - Threshold of hearing by air conduction as a function of age and sex for otologically normal persons (ISO 7029:1984)

Akustik - Luftleitungshörschwelle in Abhängigkeit von Alter und Geschlecht otologisch normaler Personen (ISO 7029:1984)

Acoustique - Seuil normal d'audition par conduction aérienne en fonction de l'âge et du sexe pour les personnes otologiquement normales (ISO 7029:1984)

Ta slovenski standard je istoveten z: **EN 27029:1991**

<u>**ICS:**</u>

13.140	Vpliv hrupa na ljudi	Noise with respect to human beings
17.140.99	Drugi standardi v zvezi z akustiko	Other standards related to acoustics

SIST EN 27029:1999 **en**

iTeh STANDARD PREVIEW
(standards.iteh.ai)

REPUBLIKA SLOVENIJA
MINISTRSTVO ZA ZNANOST IN TEHNOLOGIJO
Urad RS za standardizacijo in meroslovje
LJUBLJANA
SIST..EN. 27079............... -11- 1999
PREVZET PO METODI RAZGLASITVE

EUROPEAN STANDARD

NORME EUROPEENNE

EUROPÄISCHE NORM

EN 27029:1991

October 1991

UDC 534.75:612.85

Descriptors : Acoustics, audiometry, audibility, auditory threshold, humans

English version

Acoustics - Threshold of hearing by air conduction
as a function of age and sex otologically persons
(Identical with ISO 7029:1984)

Acoustique - Seuil normal d'audition
par conduction aérienne en function de
l'âge et du sexe pour les personnes
otologiquement normales (Identique à
l'ISO 7029:1984)

Akustik - Luftleitungshörschwelle in
Abhängigkeit von Alter und Geschlecht
otologisch normaler Personen (Identisch
mit ISO 7029:1984)

This European Standard was approved by CEN on 1991-10-07 and is identical to
the ISO standard as referred to.
CEN members are bound to comply with the CEN/CENELEC Internal Regulations
which stipulate the conditions for giving this European Standard the
status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national
standards may be obtained on application to the Central Secretariat or to any
CEN member.

This European Standard exists in three official versions (English, French,
German). A version in any other language made by translation under the
responsibility of a CEN member into its own language and notified to the
Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Denmark,
Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg,
Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

CEN

European Committee for Standardization
Comité Européen de Normalisation
Europäisches Komitee für Normung

Central Secretariat: rue de Stassart 36, B-1050 Brussels

Ref. No. EN 27029:1991 E

Page 2
EN 27029:1991

Foreword

This European Standard has been taken over by CEN/TC 211 "Acoustics" from the work of the International Organization for Standardization (ISO).

This document has been submitted to the formal vote and has been approved.

Natonal Standards identical to this European Standard shall be published at the latest by 92-04-09 and conflicting national standards shall be withdrawn at the latest 92-04-09.

In accordance with the CEN/CENELEC Internal Regulations, the following countries are bound to implement this European Standard : Austria, Belgium, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland, United Kingdom.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Endorsement notice

The text of the International Standard ISO 7029:1984 has been approved by CEN as a European Standard without any modification.

International Standard ISO 7029

Acoustics — Threshold of hearing by air conduction as a function of age and sex for otologically normal persons

Acoustique — Seuil normal d'audition par conduction aérienne en fonction de l'âge et du sexe pour les personnes otologiquement normales

First edition — 1984-12-15

iTeh STANDARD PREVIEW
(standards.iteh.ai)

UDC 534.75 : 612.85

Ref. No. ISO 7029-1984 (E)

Descriptors : acoustics, audiometry, humans, audibility, auditory threshold.

ISO 7029-1984 (E)

Price based on 8 pages

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work.

Draft International Standards adopted by the technical committees are circulated to the member bodies for approval before their acceptance as International Standards by the ISO Council. They are approved in accordance with ISO procedures requiring at least 75 % approval by the member bodies voting.

International Standard ISO 7029 was prepared by Technical Committee ISO/TC 43, *Acoustics*.

INTERNATIONAL STANDARD

ISO 7029-1984 (E)

Acoustics — Threshold of hearing by air conduction as a function of age and sex for otologically normal persons

0 Introduction

It is well known that the sensitivity of human hearing to pure tones falls progressively with age and that the loss of hearing is more rapid for high-frequency tones than for low-frequency tones. Moreover, the magnitude of this effect varies considerably between individuals.

In the case of young persons, data have been standardized (see ISO 389) specifying the modal values of the threshold of hearing for the otologically normal (as defined in ISO 389); these data form the basis for calibrating pure tone audiometers. These data do not discriminate between male and female persons and they do not specify the technique of audiometry to which they apply. It is known that these distinctions should, in principle, be made in order to obtain the most accurate value for audiometric zero. In this International Standard the distinction between males and females has been made since the difference is found to be significantly large in the case of older age groups. When an individual, who is more than 18 years old, is being tested with an audiometer, part of any observed hearing loss will most probably be associated with the person's age, and it is, therefore, important to be aware of this when estimating the amount of hearing loss attributable to other causes under investigation. Furthermore, there is a need for generally acceptable data to be standardized for this purpose.

Numerous data on the elevation of hearing threshold levels increasing with age exist in various publications, but there are certain numerical differences between them which may be attributed to the use of different criteria of selection for test populations, different audiometric techniques, etc. However, a thorough examination of the data has enabled a representative set of values to be established. This International Standard is based on these values which refer to screened populations of otologically normal persons as defined herein.

1 Scope and field of application

This International Standard specifies, for the range of audiometric frequencies from 125 to 8 000 Hz and for groups of otologically normal persons of a given age within the age limits of 18 to 70 years inclusive :

a) the expected value of the median hearing threshold shift relative to a group of persons 18 years of age;

b) the expected statistical distribution above and below the median value (see note 2 to 4.1).

The data in this International Standard are applicable as descriptive statistics of the hearing levels of populations of various ages. An example of this application is as a baseline of comparison for estimating the amount of hearing loss caused in a population due to a specific agent, for example, noise, and in this application the data of this International Standard are included as "Data Base A" in ISO 1999. The data may also be used to compare an individual's hearing with the normal distribution of the hearing threshold levels for the person's age group. In the audiological diagnosis of an individual person, it is, however, not possible to determine precisely which changes of the hearing threshold level are attributable to an accumulation of detrimental effects upon the hearing increasing with age and which changes have been caused by other factors such as, for example, noise.

2 References

ISO 389, *Acoustics — Standard reference zero for the calibration of pure tone audiometers.*

ISO 1999, *Acoustics — Determination of occupational noise exposure and estimation of noise-induced hearing impairment.* [1]

ISO 6189, *Acoustics — Pure tone air conduction threshold audiometry for hearing conservation purposes.*

ISO 8253, *Acoustics — Pure tone audiometric test methods.* [2]

IEC Publication 645, *Audiometers.*

---

1) At present at the stage of draft. (Revision of ISO 1999-1975.)

2) At present at the stage of draft.

ISO 7029-1984 (E)

3 Definitions

For the purpose of this International Standard, the following definitions apply.

3.1 otologically normal person : A person in a normal state of health who at the time of testing is free from excess wax in the ear canals, is without known ear pathology and who has no history of undue exposure to noise.

3.2 hearing threshold level (of a given ear of a person at a given frequency) : The threshold of hearing as determined in a stated manner by means of a pure tone air conduction audiometer, expressed as the hearing level in decibels.

NOTES

1 Specifications for audiometers are given in IEC Publication 645.

2 For the calibration of audiometers, see ISO 389.

3 For appropriate test conditions see, for example, ISO 6189 and ISO 8253.

3.3 hearing level : The sound pressure level of a tone of given frequency which is generated by the earphone of a pure tone audiometer in an acoustic coupler of specified type, when expressed in decibels relative to the reference equivalent threshold sound pressure level (audiometric zero) for that frequency and for the pattern of earphone and acoustic coupler in question.

NOTE — Values of reference equivalent threshold sound pressure level for various patterns of earphone are specified in ISO 389.

4 Specification

4.1 Median values

The median value of hearing threshold deviation of otologically normal persons aged Y years shall be given by the following equation :

$$H_{md,Y} = \alpha(Y - 18\ \text{years})^2 + H_{md,18} \qquad \dots \text{(1)}$$

where $H_{md,18}$ is the median value of hearing threshold level of otologically normal persons of the same sex aged 18 years.

NOTES

1 For practical purposes, $H_{md,18}$ may be taken as zero, corresponding to the hearing threshold for young persons, as specified in ISO 389.

2 The median values specified in this International Standard are not the modal values used in ISO 389.

Values of the coefficient α for males and females shall be those given in table 1.

The range of Y for which formula (1) is valid is from 18 to 70 years inclusive.

Table 1 — Values of the coefficient α used to determine the median value of hearing threshold deviation for an otologically normal population of given age

Frequency in Hz	Value of α, in dB/year²	
	Males	Females
125	0,003 0	0,003 0
250	0,003 0	0,003 0
500	0,003 5	0,003 5
1 000	0,004 0	0,004 0
1 500	0,005 5	0,005 0
2 000	0,007 0	0,006 0
3 000	0,011 5	0,007 5
4 000	0,016 0	0,009 0
6 000	0,018 0	0,012 0
8 000	0,022 0	0,015 0

4.2 Statistical distribution

4.2.1 The statistical distribution of hearing threshold levels of otologically normal persons of a given age shall be approximated, for the purpose of this International Standard, by the halves of two normal (Gaussian) distributions. One half lies above the median value, $H_{md,Y}$, and has the larger dispersion characterized by the parameter s_u; the other half lies below the median and has the smaller dispersion characterized by the parameter s_l.

Values of the parameters s_u and s_l, in decibels, are given by the following equations :

$$s_u = b_u + 0,445\ H_{md,Y} \qquad \dots \text{(2)}$$

$$s_l = b_l + 0,356\ H_{md,Y} \qquad \dots \text{(3)}$$

where b_u and b_l have the values given in table 2.

NOTE — The parameters s_u and s_l are defined so that they correspond to the standard deviation of the full normal distributions of which the upper and lower halves respectively comprise the actual distribution of hearing threshold levels.

Table 2 — Values of the parameters b_u and b_l used to determine respectively the upper and lower parts of the statistical distribution of hearing threshold levels, centred on the median value, for an otologically normal population of given age

Frequency in Hz	Value of b_u, in dB		Value of b_l, in dB	
	Males	Females	Males	Females
125	7,23	6,67	5,78	5,34
250	6,67	6,12	5,34	4,89
500	6,12	6,12	4,89	4,89
1 000	6,12	6,12	4,89	4,89
1 500	6,67	6,67	5,34	5,34
2 000	7,23	6,67	5,78	5,34
3 000	7,78	7,23	6,23	5,78
4 000	8,34	7,78	6,67	6,23
6 000	9,45	8,90	7,56	7,12
8 000	10,56	10,56	8,45	8,45

To calculate s_u, refer to table 2 for the appropriate frequency and sex to determine b_u and then apply equation (2) to determine s_u. Similarly, determine b_l from table 2 and then apply equation (3) to determine s_l.

4.2.2 To determine the hearing threshold level which can be expected to be exceeded by a given fraction, Q, of an otologically normal population of given age, proceed as described in 4.2.2.1 or 4.2.2.2.

NOTE — Values are given, for selected values of the parameters, in annex B.

4.2.2.1 For a fraction Q of the population such that $0,05 < Q < 0,5$, the value is given by the following equation :

$$H_{Q,Y} = H_{md,Y} + k \times s_u \qquad \ldots \ (4)$$

where k is a function of Q as specified in 4.2.2.3.

4.2.2.2 For a fraction Q of the population such that $0,5 < Q < 0,95$, the value is given by the following equation :

$$H_{Q,Y} = H_{md,Y} - k \times s_l \qquad \ldots \ (5)$$

where k is a function of Q as specified in 4.2.2.3.

4.2.2.3 Values of the multiplier k, corresponding to the normal (Gaussian) distribution, are given in table 3 for values of the fraction Q at intervals of 0,01 (1 %) from 0,05 to 0,95 (5 % to 95 %).

Table 3 — Values of the multiplier k

Q		k	Q		k
0,05	0,95	1,645	0,26	0,74	0,643
0,06	0,94	1,555	0,27	0,73	0,613
0,07	0,93	1,476	0,28	0,72	0,583
0,08	0,92	1,405	0,29	0,71	0,553
0,09	0,91	1,341	0,30	0,70	0,524
0,10	0,90	1,282	0,31	0,69	0,496
0,11	0,89	1,227	0,32	0,68	0,468
0,12	0,88	1,175	0,33	0,67	0,440
0,13	0,87	1,126	0,34	0,66	0,413
0,14	0,86	1,080	0,35	0,65	0,385
0,15	0,85	1,036	0,36	0,64	0,359
0,16	0,84	0,995	0,37	0,63	0,332
0,17	0,83	0,954	0,38	0,62	0,306
0,18	0,82	0,915	0,39	0,61	0,279
0,19	0,81	0,878	0,40	0,60	0,253
0,20	0,80	0,842	0,41	0,59	0,228
0,21	0,79	0,806	0,42	0,58	0,202
0,22	0,78	0,772	0,43	0,57	0,176
0,23	0,77	0,739	0,44	0,56	0,151
0,24	0,76	0,706	0,45	0,55	0,126
0,25	0,75	0,675	0,46	0,54	0,100
			0,47	0,53	0,075
			0,48	0,52	0,050
			0,49	0,51	0,025
			0,50		0,000

NOTE — Due to uncertainties in the experimental data on which this International Standard is based, tails of the statistical distributions for $0 < Q < 0,05$ and for $0,95 < Q < 1$ are unreliable and are not tabulated.