



## Standard Test Method for Preparing Quantitative Pole Figures<sup>1</sup>

This standard is issued under the fixed designation E81; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

### 1. Scope

- 1.1 This test method covers the use of the X-ray diffractometer to prepare quantitative pole figures.
- 1.2 The test method consists of several experimental procedures. Some of the procedures **(1-5)**<sup>2</sup> permit preparation of a complete pole figure. Others must be used in combination to produce a complete pole figure.
- 1.3 Pole figures **(6)** and inverse pole figures **(7-10)** are two dimensional averages of the three-dimensional crystallite orientation distribution. Pole figures may be used to construct either inverse pole figures **(11-13)** or the crystallite orientation distribution **(14-21)**. Development of series expansions of the crystallite orientation distribution from reflection pole figures **(22, 23)** makes it possible to obtain a series expansion of a complete pole figure from several incomplete pole figures. Pole figures or inverse pole figures derived by such methods shall be termed calculated. These techniques will not be described herein.
- 1.4 Provided the orientation is homogeneous through the thickness of the sheet, certain procedures **(1-3)** may be used to obtain a complete pole figure.
- 1.5 Provided the orientation has mirror symmetry with respect to planes perpendicular to the rolling, transverse, and normal directions, certain procedures **(4, 5, 24)** may be used to obtain a complete pole figure.
- 1.6 The test method emphasizes the Schulz reflection technique **(25)**. Other techniques **(3, 4, 5, 24)** may be considered variants of the Schulz technique and are cited as options, but not described herein.
- 1.7 The test method also includes a description of the transmission technique of Decker, et al **(26)**, which may be used in conjunction with the Schulz reflection technique to obtain a complete pole figure.
- 1.8 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*
- 1.9 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

### 2. Summary of Test Method

2.1 The test method consists of characterizing the distribution of orientations of selected lattice planes with respect to sample-fixed coordinates **(6)**. The distribution will usually be obtained by measurement of the intensity of X rays diffracted by the sample. In such measurements the detector and associated limiting slits are fixed at twice the appropriate Bragg angle, and the diffracted intensity is recorded as the orientation of the sample is changed **(1-6, 25, 26, 27)**. After the measured data have been corrected, as necessary, for background, defocusing, and absorption, and normalized to have an average value of unity, the results may be plotted in stereographic or equal-area projection.

2.2 The geometry of the Schulz **(25)** reflection method is illustrated in Fig. 1. Goniometers employing this geometry are commercially available. The source of X rays is indicated by *L*. Slit S1 limits divergence of the incident beam in the plane of projection. Slit S2 limits divergence perpendicular to the plane of projection. The sample, indicated by crosshatching, may be tilted about the axis *FF'*, which is perpendicular to the diffractometer axis and lies in the plane of the sample. The tilt angle was denoted

<sup>1</sup> This test method is under the jurisdiction of ASTM Committee E04 on Metallography and is the direct responsibility of Subcommittee E04.11 on X-Ray and Electron Metallography.

Current edition approved Oct. 1, 2011; June 1, 2017. Published December 2011; June 2017. Originally approved in 1949. Last previous edition approved in 2007 as E81 – 96 (2007); (2011). DOI: 10.1520/E0081-96R11.10.1520/E0081-96R17.

<sup>2</sup> The boldface numbers in parentheses refer to the list of references at the end of this test method.




FIG. 1 Geometry of Reflection Method.

$\phi$  by Schulz (25). The sample position shown in Fig. 1 corresponds to  $\phi = 0$  deg, for which approximate parafocusing conditions exist at the detector slit, S3. With the application of a defocusing correction, this method is useful over a range of colatitude  $\phi$  from 0 deg to approximately 75 deg.

2.2.1 Tilting the sample about  $FF'$ , so as to reduce the distance between  $L$  and points in the sample surface above the plane of projection, causes X rays diffracted from these points to be displaced to the left of the center of S3, while X rays diffracted from points in the sample below the plane of projection are displaced to the right of the center of S3. The displacement is equal to  $2D \tan \phi \cos \theta$ , where  $D$  is the distance above or below the plane of projection. The integrated, or total, diffracted intensity is influenced only slightly by tilting the sample (28). Insofar as possible, the detector slit shall be of sufficient width to include the defocused line profile corresponding to the maximum sample tilt for which measurements are to be made. Because of interferences from neighboring diffraction peaks and physical limitations on sample size and detector slit width, it is necessary to limit vertical divergence of the incident beam. A widely used pole figure goniometer with a focal spot to the center of the sample distance of 172 mm employs a 0.5-mm slit located 30 mm from the center of the sample for this purpose. Measured intensities may be corrected for defocusing by comparison with intensities diffracted by a randomly oriented specimen of similar material, or byemploying the theoretically calculated corrections (28).

2.3 The geometry of the transmission technique of Decker, et al (26) is shown in Fig. 2. In contrast to the reflection method, X rays diffracted from different points in the sample diverge, making the resolution of adjacent peaks more difficult. The ratio of the diffracted intensity at  $\alpha = -5, -10, \dots, -70$  deg, to the diffracted intensity at  $\alpha = 0$  deg, calculated in accordance with the expression given by Decker, et al (26) for linear absorption thickness product,  $\mu t$ , = 1.0, 1.4, ..., 3.0, and, for  $\theta = 5, 10, \dots, 25$  deg is given in Table 1. These data may be used as a guide to determine the useful range of  $\alpha$  for a given  $\mu t$  and  $\theta$ . If, for example,  $I_{\alpha} / I_0$  is restricted to values  $\geq 0.5$ , one arrives at the series of curves shown in Fig. 3.

### 3. Significance and Use

3.1 Pole figures are two-dimensional graphic representations, on polar coordinate paper, of the average distribution of crystallite orientations in three dimensions. Data for constructing pole figures are obtained with X-ray diffractometers, using reflection and transmission techniques.

3.2 Several alternative procedures may be used. Some produce complete pole figures. Others yield partial pole figures, which may be combined to produce a complete figure.




FIG. 2 Geometry of Transmission Method.

TABLE 1 ( $I_{\alpha}/I_0$ )  $\times$  1000

$\theta$	$-\alpha$														
	$\mu t$	5	10	15	20	25	30	35	40	45	50	55	60	65	70
5	1.0	992	984	976	966	954	939	918	890	851	796	703	617	480	313
	1.4	991	978	962	941	915	882	840	786	719	636	533	412	277	146
	1.8	989	972	948	917	878	828	768	695	608	508	395	276	162	070
	2.2	988	966	935	893	842	778	702	614	515	406	294	186	095	034
	2.6	986	960	922	871	807	731	643	544	436	326	219	126	057	017
	3.0	985	954	909	849	775	687	589	481	370	261	164	086	034	009
10	1.0	984	969	952	934	912	887	855	815	762	694	603	486	344	191
	1.4	983	962	938	908	873	831	779	716	640	548	440	320	198	094
	1.8	981	956	924	884	836	779	710	630	538	435	325	215	119	049
	2.2	980	950	911	861	801	730	649	556	455	348	242	147	074	027
	2.6	978	944	898	839	768	686	593	492	385	280	183	103	047	016
	3.0	977	938	885	817	737	644	543	436	328	226	139	073	030	009
15	1.0	976	952	927	900	868	832	789	735	668	583	477	349	209	085
	1.4	975	946	912	874	829	776	714	640	553	453	342	227	123	046
	1.8	973	939	898	850	792	725	648	560	462	358	252	155	078	027
	2.2	972	933	885	826	758	678	590	492	389	286	190	110	052	017
	2.6	970	927	872	804	725	636	538	435	331	232	146	080	036	011
	3.0	968	921	859	783	695	597	493	386	283	190	115	060	025	007
20	1.0	968	935	901	863	822	774	718	649	566	465	345	214	093	000
	1.4	966	928	885	836	781	717	643	557	460	354	243	140	058	000
	1.8	964	921	870	811	743	666	579	484	381	278	180	099	039	000
	2.2	963	915	857	788	709	621	525	424	321	224	139	074	028	000
	2.6	961	909	843	766	678	582	479	375	274	185	111	057	020	000
	3.0	960	903	831	746	650	547	440	335	238	155	090	044	015	000
25	1.0	959	917	872	824	771	710	639	555	455	339	214	096	000	000
	1.4	957	909	856	796	728	651	565	468	362	253	151	065	000	000
	1.8	955	902	840	770	690	602	505	402	298	200	115	048	000	000
	2.2	953	895	826	746	657	560	456	352	253	164	092	038	000	000
	2.6	952	889	812	724	627	523	417	314	219	139	076	031	000	000
	3.0	950	883	800	705	601	493	384	283	194	121	065	025	000	000




FIG. 3  $\alpha$  versus  $\mu t$  for  $I_{\alpha}/I_0 = 0.5$ ,  $\theta = 5, 10, \dots, 25$  deg.

#### 4. Apparatus

4.1 *Source of X Rays*—A beam of characteristic X rays of substantially constant intensity is required. Characteristic Kalpha radiation of chromium, iron, cobalt, nickel, copper, molybdenum, and silver have all been used successfully, depending on the chemical composition of the specimen. Insofar as possible, the radiation selected shall provide sufficient angular dispersion to permit the resolution of peaks to be measured, and shall not produce excessive fluorescence in the sample. Linear absorption coefficients (29) for selected elements are given in Table 2. Lower energy radiation (Cr, Fe, Co, Ni, Cu) is generally preferred for reflection pole figure measurements as it provides greater angular dispersion. Higher energy radiation (Mo, Ag) is generally preferred for transmission measurements.

4.2 *Slits*—Suitable slits shall be provided to limit horizontal (in the plane of projection of Figs. 1 and 2) and vertical (perpendicular to the plane of projection of Figs. 1 and 2) divergence of the incident beam. Horizontal divergences of 1 to 3 deg for reflection and 0.5 deg for transmission are typical. Vertical divergences of 0.2 deg for reflection and 1 deg for transmission are typical. Insofar as possible, the receiving slit shall be of sufficient width to include the diffracted peak. Receiving slits corresponding to 1 deg 2-theta are typical.

4.3 *Specimen Holder—Reflection Method:*

**TABLE 2 Linear Absorption Coefficient  $\mu$  ( $\text{cm}^{-1}$ ) for Selected Wavelengths and Elements**

		K-alpha Radiation						
		Ag	Mo	Cu	Ni	Co	Fe	Cr
$\lambda$		0.5608	0.7107	1.5418	1.6591	1.7902	1.9373	2.2909
Absorber								
6	C	0.90	1.41	10.4	12.8	15.9	20.0	32.6
12	Mg	3.69	7.15	67.2	83.0	104	130	211
13	Al	7.15	13.9	131	162	202	253	410
22	Ti	55.4	109	936	1134	1386	1696	2570
24	Cr	114	224	1869	2258	2739	3329	574
25	Mn	131	257	2115	2545	3072	424	539
26	Fe	155	303	2424	2912	416	523	850
27	Co	194	378	2786	436	544	684	1112
28	Ni	214	415	407	503	627	789	1282
29	Cu	236	455	472	585	729	920	1482
30	Zn	205	395	430	532	663	834	1348
40	Zr	380	103	930	1144	1404	1742	2724
42	Mo	661	188	1652	2020	2479	3060	4723
47	Ag	137	271	2287	2769	3367	4102	6147
48	Cd	121	238	1998	2413	2924	3564	5302
50	Sn	116	227	1869	2256	2723	3292	4833
74	W	1023	1912	3320	4014	4883	5944	8839
79	Au	1215	2215	4006	4815	5817	7030	10250
82	Pb	768	1361	2631	3153	3788	4559	6566

4.3.1 The specimen holder for the reflection method shall preferably employ the Schulz reflection geometry illustrated in Fig. 1 and described in 2.2. It is desirable that the specimen holder be equipped with a means for oscillating the sample in the plane of its surface without changing the orientation of the sample. It is also desirable that the magnitude of the oscillation be variable. The specimen holder shall preferably be provided with automatic means for changing colatitude and longitude of the sample.

4.3.2 Alternative reflection geometries include those of Bakarian (1), Field and Marchant (27), and Jetter and Borie (2). The method of Bakarian requires machining a number of cylindrical specimens whose axes are perpendicular to the sheet normal direction. Each specimen provides intensity data along one parallel of longitude. The method of Jetter and Borie entails the preparation of a spherical specimen. In the methods of Bakarian and of Jetter and Borie, the sample shall, insofar as possible, be prepared from homogeneous material. These methods have the advantage that intensity data need not be corrected for absorption or defocusing. They do not permit oscillation of the sample. Equipment is not currently commercially available for these methods.

4.3.3 The method of Field and Marchant (27) requires an absorption correction. If this method is used in conjunction with the transmission method of Decker, et al (26), it is necessary to use either different orders of reflection or different radiations in order to obtain a complete pole figure.

4.4 *Specimen Holder—Transmission Method*—If the transmission method is used, the specimen holder shall employ the geometry of Decker, et al (26), shown in Fig. 2 and described in 2.3. It is desirable that the specimen holder be equipped with a means for oscillating the sample in the plane of its surface without changing the orientation of the sample. The specimen holder shall preferably be provided with automatic means for changing colatitude and longitude of the sample.

4.5 *Detector*—The detector shall preferably be of an energy-dispersive type, for example, a solid state, proportional, or scintillation counter, and used in conjunction with a pulse height selector circuit to discriminate against X rays whose energies differ markedly from that of the characteristic K-alpha radiation being used. Reduction of the characteristic K-beta radiation requires the use of a monochromator or appropriate beta filter. Pd, Zr, Ni, Co, Fe, Mn, and V are appropriate beta filters for Ag, Mo, Cu, Ni, Co, Fe, and Cr, respectively.

## 5. Test Specimens

5.1 For the reflection method, the sample shall be of sufficient thickness that loss of intensity due to transmission through the sample may be ignored. If a maximum loss of 1 % the incident beam is acceptable, the specimen must have a linear absorption thickness product equal to or greater than  $2.3 \sin \theta$ . For an iron sample with molybdenum K-alpha radiation, this requires that  $\mu t$  be greater than 0.4, 0.6, and 0.7 for the (110), (200), and (211) reflections, respectively.

5.1.1 Surface preparation is particularly important in the reflection method. Calculations due to Borie (30), who assumed a sawtooth surface of spacing  $a$  on a material with linear absorption coefficient  $\mu$ , indicate that the product  $\mu a$  should be less than 0.5 if significant intensity losses are to be avoided. For an iron sample with cobalt K-alpha radiation,  $\mu = 416 \text{ cm}^{-1}$ , corresponding to  $a \leq 12 \mu\text{m}$ .

5.2 For the transmission method, maximum intensity is obtained for a linear absorption thickness product equal to  $\cos \theta$ . For an iron sample with molybdenum K-alpha, this corresponds to  $\mu t$  equal to 0.98, 0.97, and 0.95 for the (110), (200), and (211) reflections, respectively. Thus, a suitable transmission sample can also be used for reflection measurements.